

LAB CHRONICLE

OrderID:	Q2202	OrderDate:	6/4/2025 11:59:00 AM
Client:	PARSONS Engineering of New York, Inc.	Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05
Contact:	Stephen Liberatore	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2202-01	MW-9-20250603	Water	VOCMS Group1	8260-Low	06/03/25		06/10/25	06/03/25
Q2202-02	MW-11-20250603	Water	VOCMS Group1	8260-Low	06/03/25		06/10/25	06/03/25
Q2202-03	MW-12-20250603	Water	VOCMS Group1	8260-Low	06/03/25		06/10/25	06/03/25
Q2202-03DL	MW-12-20250603DL	Water	VOCMS Group1	8260-Low	06/03/25		06/11/25	06/03/25
Q2202-04	MW-13D-20250603	Water	VOCMS Group1	8260-Low	06/03/25		06/11/25	06/03/25
Q2202-05	MW-1A-20250603	Water	VOCMS Group1	8260-Low	06/03/25		06/11/25	06/03/25
Q2202-06	TB-20250603	Water	VOCMS Group1	8260-Low	06/03/25		06/11/25	06/03/25

Hit Summary Sheet
SW-846

SDG No.: Q2202

Client: PARSONS Engineering of New York, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: MW-9-20250603								
Q2202-01	MW-9-20250603	Water	Acetone	2.20	J	1.50	5.00	ug/L
Q2202-01	MW-9-20250603	Water	2-Butanone	1.30	J	0.98	5.00	ug/L
Q2202-01	MW-9-20250603	Water	Chloroform	7.10		0.25	1.00	ug/L
Q2202-01	MW-9-20250603	Water	Trichloroethene	0.32	J	0.090	1.00	ug/L
Q2202-01	MW-9-20250603	Water	Tetrachloroethene	0.39	J	0.23	1.00	ug/L
			Total Voc :			11.3		
Q2202-01	MW-9-20250603	Water	sec-Butylbenzene	* 0.24	J	0.13	1.00	ug/L
			Total Tics :			0.24		
			Total Concentration:			11.6		
Client ID: MW-11-20250603								
Q2202-02	MW-11-20250603	Water	Acetone	1.90	J	1.50	5.00	ug/L
Q2202-02	MW-11-20250603	Water	Methyl tert-butyl Ether	0.64	J	0.16	1.00	ug/L
Q2202-02	MW-11-20250603	Water	Chloroform	0.51	J	0.25	1.00	ug/L
Q2202-02	MW-11-20250603	Water	Tetrachloroethene	1.90		0.23	1.00	ug/L
			Total Voc :			4.95		
			Total Concentration:			4.95		
Client ID: MW-12-20250603								
Q2202-03	MW-12-20250603	Water	Acetone	3.30	J	1.50	5.00	ug/L
Q2202-03	MW-12-20250603	Water	Carbon Disulfide	19.5		0.21	1.00	ug/L
Q2202-03	MW-12-20250603	Water	Methyl tert-butyl Ether	1.20		0.16	1.00	ug/L
Q2202-03	MW-12-20250603	Water	Methylcyclohexane	18.5		0.16	1.00	ug/L
Q2202-03	MW-12-20250603	Water	Benzene	4.10		0.15	1.00	ug/L
Q2202-03	MW-12-20250603	Water	Trichloroethene	0.56	J	0.090	1.00	ug/L
Q2202-03	MW-12-20250603	Water	Toluene	1.90		0.14	1.00	ug/L
Q2202-03	MW-12-20250603	Water	Ethyl Benzene	240	E	0.13	1.00	ug/L
Q2202-03	MW-12-20250603	Water	m/p-Xylenes	13.1		0.24	2.00	ug/L
Q2202-03	MW-12-20250603	Water	o-Xylene	13.5		0.12	1.00	ug/L
Q2202-03	MW-12-20250603	Water	Isopropylbenzene	65.7		0.12	1.00	ug/L
			Total Voc :			381		
Q2202-03	MW-12-20250603	Water	Benzene, 1,2,4,5-tetramethyl-	* 15.1	J	0	0	ug/L
Q2202-03	MW-12-20250603	Water	Azulene	* 250	J	0	0	ug/L
Q2202-03	MW-12-20250603	Water	Indane	* 130	J	0	0	ug/L
Q2202-03	MW-12-20250603	Water	Benzene, 1-ethyl-2-methyl-	* 35.4	J	0	0	ug/L
Q2202-03	MW-12-20250603	Water	Indan, 1-methyl-	* 54.8	J	0	0	ug/L
Q2202-03	MW-12-20250603	Water	1H-Indene, 2,3-dihydro-4-meth	* 35.4	J	0	0	ug/L
Q2202-03	MW-12-20250603	Water	Benzene, 1-ethyl-2,4-dimethyl-	* 45.1	J	0	0	ug/L

Hit Summary Sheet SW-846

SDG No.: Q2202

Client: PARSONS Engineering of New York, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q2202-03	MW-12-20250603	Water	2-Methylindene	* 29.2	J	0	0	ug/L
Q2202-03	MW-12-20250603	Water	Sulfur dioxide	* 39.3	J	0	0	ug/L
Q2202-03	MW-12-20250603	Water	n-propylbenzene	* 73.5	J	0.13	1.00	ug/L
Q2202-03	MW-12-20250603	Water	1,3,5-Trimethylbenzene	* 26.1	J	0.15	1.00	ug/L
Q2202-03	MW-12-20250603	Water	1,2,4-Trimethylbenzene	* 240	J	0.14	1.00	ug/L
Q2202-03	MW-12-20250603	Water	sec-Butylbenzene	* 3.20	J	0.13	1.00	ug/L
Q2202-03	MW-12-20250603	Water	p-Isopropyltoluene	* 6.20	J	0.13	1.00	ug/L
Q2202-03	MW-12-20250603	Water	n-Butylbenzene	* 4.40	J	0.15	1.00	ug/L
Total Tics :				988				
Total Concentration:				1370				
Client ID:	MW-12-20250603DL							
Q2202-03DL	MW-12-20250603E	Water	Carbon Disulfide	28.4	D	1.10	5.00	ug/L
Q2202-03DL	MW-12-20250603E	Water	Methylcyclohexane	17.3	D	0.80	5.00	ug/L
Q2202-03DL	MW-12-20250603E	Water	Benzene	3.90	JD	0.75	5.00	ug/L
Q2202-03DL	MW-12-20250603E	Water	Toluene	2.00	JD	0.70	5.00	ug/L
Q2202-03DL	MW-12-20250603E	Water	Ethyl Benzene	220	D	0.65	5.00	ug/L
Q2202-03DL	MW-12-20250603E	Water	m/p-Xylenes	13.1	D	1.20	10.0	ug/L
Q2202-03DL	MW-12-20250603E	Water	o-Xylene	12.5	D	0.60	5.00	ug/L
Q2202-03DL	MW-12-20250603E	Water	Isopropylbenzene	58.5	D	0.60	5.00	ug/L
Total Voc :				356				
Total Concentration:				356				
Client ID:	MW-13D-20250603							
Q2202-04	MW-13D-20250603	Water	Acetone	1.50	J	1.50	5.00	ug/L
Q2202-04	MW-13D-20250603	Water	Chloroform	2.00		0.25	1.00	ug/L
Q2202-04	MW-13D-20250603	Water	Tetrachloroethene	2.80		0.23	1.00	ug/L
Total Voc :				6.30				
Q2202-04	MW-13D-20250603	Water	Naphthalene	* 0.26	J	0.20	1.00	ug/L
Total Tics :				0.26				
Total Concentration:				6.56				
Client ID:	MW-1A-20250603							
Q2202-05	MW-1A-20250603	Water	Acetone	4.40	J	1.50	5.00	ug/L
Q2202-05	MW-1A-20250603	Water	2-Butanone	1.20	J	0.98	5.00	ug/L
Q2202-05	MW-1A-20250603	Water	Chloroform	1.50		0.25	1.00	ug/L
Q2202-05	MW-1A-20250603	Water	Benzene	2.20		0.15	1.00	ug/L
Q2202-05	MW-1A-20250603	Water	4-Methyl-2-Pentanone	0.85	J	0.68	5.00	ug/L
Total Voc :				10.2				
Q2202-05	MW-1A-20250603	Water	Naphthalene	* 0.34	J	0.20	1.00	ug/L
Total Tics :				0.34				

Hit Summary Sheet
SW-846

SDG No.: Q2202

Client: PARSONS Engineering of New York, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Total Concentration:				10.5				



SAMPLE DATA

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-9-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086934.D	1		06/10/25 17:09	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	2.20	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	1.30	J	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	7.10		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.32	J	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.			Date Collected:	06/03/25
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05			Date Received:	06/03/25
Client Sample ID:	MW-9-20250603			SDG No.:	Q2202
Lab Sample ID:	Q2202-01			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086934.D	1		06/10/25 17:09	VN061025

[illegible]

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-11-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086935.D	1		06/10/25 17:31	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.90	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.64	J	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.51	J	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-11-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086935.D	1		06/10/25 17:31	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	1.90		0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	49.7		74 - 125	99%	SPK: 50
1868-53-7	Dibromofluoromethane	50.2		75 - 124	100%	SPK: 50
2037-26-5	Toluene-d8	52.1		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.9		77 - 121	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	320000	8.235			
540-36-3	1,4-Difluorobenzene	603000	9.106			
3114-55-4	Chlorobenzene-d5	539000	11.864			
3855-82-1	1,4-Dichlorobenzene-d4	260000	13.794			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-11-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086935.D	1		06/10/25 17:31	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-12-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-03		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086936.D	1		06/10/25 17:52	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	3.30	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	19.5		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	1.20		0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	18.5		0.16	1.00	ug/L
71-43-2	Benzene	4.10		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.56	J	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	1.90		0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/03/25
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/03/25
Client Sample ID:	MW-12-20250603	SDG No.:	Q2202
Lab Sample ID:	Q2202-03	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086936.D	1		06/10/25 17:52	VN061025

[illegible]

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.			Date Collected:	06/03/25
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05			Date Received:	06/03/25
Client Sample ID:	MW-12-20250603			SDG No.:	Q2202
Lab Sample ID:	Q2202-03			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086936.D	1		06/10/25 17:52	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
007446-09-5	Sulfur dioxide	39.3	J		2.33	ug/L
103-65-1	n-propylbenzene	73.5	J		13.0	ug/L
108-67-8	1,3,5-Trimethylbenzene	26.1	J		13.2	ug/L
000611-14-3	Benzene, 1-ethyl-2-methyl-	35.4	J		13.3	ug/L
95-63-6	1,2,4-Trimethylbenzene	240	J		13.5	ug/L
135-98-8	sec-Butylbenzene	3.20	J		13.6	ug/L
99-87-6	p-Isopropyltoluene	6.20	J		13.7	ug/L
000496-11-7	Indane	130	J		14.0	ug/L
104-51-8	n-Butylbenzene	4.40	J		14.1	ug/L
000767-58-8	Indan, 1-methyl-	54.8	J		14.4	ug/L
000095-93-2	Benzene, 1,2,4,5-tetramethyl-	15.1	J		14.7	ug/L
000824-22-6	1H-Indene, 2,3-dihydro-4-methyl-	35.4	J		14.9	ug/L
000874-41-9	Benzene, 1-ethyl-2,4-dimethyl-	45.1	J		15.0	ug/L
002177-47-1	2-Methylindene	29.2	J		15.2	ug/L
000275-51-4	Azulene	250	J		15.6	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-12-20250603DL		SDG No.:	Q2202	
Lab Sample ID:	Q2202-03DL		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086947.D	5		06/11/25 14:44	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	UD	1.10	5.00	ug/L
74-87-3	Chloromethane	1.60	UD	1.60	5.00	ug/L
75-01-4	Vinyl Chloride	1.30	UD	1.30	5.00	ug/L
74-83-9	Bromomethane	7.20	UD	7.20	25.0	ug/L
75-00-3	Chloroethane	2.40	UD	2.40	5.00	ug/L
75-69-4	Trichlorofluoromethane	1.70	UD	1.70	5.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	1.30	UD	1.30	5.00	ug/L
75-35-4	1,1-Dichloroethene	1.20	UD	1.20	5.00	ug/L
67-64-1	Acetone	7.60	UD	7.60	25.0	ug/L
75-15-0	Carbon Disulfide	28.4	D	1.10	5.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.80	UD	0.80	5.00	ug/L
79-20-9	Methyl Acetate	1.40	UD	1.40	5.00	ug/L
75-09-2	Methylene Chloride	1.40	UD	1.40	5.00	ug/L
156-60-5	trans-1,2-Dichloroethene	1.20	UD	1.20	5.00	ug/L
75-34-3	1,1-Dichloroethane	1.20	UD	1.20	5.00	ug/L
110-82-7	Cyclohexane	7.30	UD	7.30	25.0	ug/L
78-93-3	2-Butanone	4.90	UD	4.90	25.0	ug/L
56-23-5	Carbon Tetrachloride	1.30	UD	1.30	5.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.95	UD	0.95	5.00	ug/L
74-97-5	Bromochloromethane	1.10	UD	1.10	5.00	ug/L
67-66-3	Chloroform	1.30	UD	1.30	5.00	ug/L
71-55-6	1,1,1-Trichloroethane	1.00	UD	1.00	5.00	ug/L
108-87-2	Methylcyclohexane	17.3	D	0.80	5.00	ug/L
71-43-2	Benzene	3.90	JD	0.75	5.00	ug/L
107-06-2	1,2-Dichloroethane	1.10	UD	1.10	5.00	ug/L
79-01-6	Trichloroethene	0.47	UD	0.47	5.00	ug/L
78-87-5	1,2-Dichloropropane	1.00	UD	1.00	5.00	ug/L
75-27-4	Bromodichloromethane	1.10	UD	1.10	5.00	ug/L
108-10-1	4-Methyl-2-Pentanone	3.40	UD	3.40	25.0	ug/L
108-88-3	Toluene	2.00	JD	0.70	5.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-12-20250603DL		SDG No.:	Q2202	
Lab Sample ID:	Q2202-03DL		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086947.D	5		06/11/25 14:44	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.85	UD	0.85	5.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.80	UD	0.80	5.00	ug/L
79-00-5	1,1,2-Trichloroethane	1.10	UD	1.10	5.00	ug/L
591-78-6	2-Hexanone	4.50	UD	4.50	25.0	ug/L
124-48-1	Dibromochloromethane	0.90	UD	0.90	5.00	ug/L
106-93-4	1,2-Dibromoethane	0.75	UD	0.75	5.00	ug/L
127-18-4	Tetrachloroethene	1.20	UD	1.20	5.00	ug/L
108-90-7	Chlorobenzene	0.60	UD	0.60	5.00	ug/L
100-41-4	Ethyl Benzene	220	D	0.65	5.00	ug/L
179601-23-1	m/p-Xylenes	13.1	D	1.20	10.0	ug/L
95-47-6	o-Xylene	12.5	D	0.60	5.00	ug/L
100-42-5	Styrene	0.75	UD	0.75	5.00	ug/L
75-25-2	Bromoform	0.95	UD	0.95	5.00	ug/L
98-82-8	Isopropylbenzene	58.5	D	0.60	5.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	1.30	UD	1.30	5.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.80	UD	0.80	5.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.95	UD	0.95	5.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.80	UD	0.80	5.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	2.70	UD	2.70	5.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	1.00	UD	1.00	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	1.00	UD	1.00	5.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	50.3		74 - 125	101%	SPK: 50
1868-53-7	Dibromofluoromethane	50.1		75 - 124	100%	SPK: 50
2037-26-5	Toluene-d8	52.2		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.3		77 - 121	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	328000	8.236			
540-36-3	1,4-Difluorobenzene	626000	9.106			
3114-55-4	Chlorobenzene-d5	565000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	268000	13.788			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-13D-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-04		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086952.D	1		06/11/25 16:35	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	2.00		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.			Date Collected:	06/03/25
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05			Date Received:	06/03/25
Client Sample ID:	MW-13D-20250603			SDG No.:	Q2202
Lab Sample ID:	Q2202-04			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086952.D	1		06/11/25 16:35	VN061125

[illegible]

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/03/25
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/03/25
Client Sample ID:	MW-13D-20250603	SDG No.:	Q2202
Lab Sample ID:	Q2202-04	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086952.D	1		06/11/25 16:35	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
91-20-3	Naphthalene	0.26	J		15.6	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-1A-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086951.D	1		06/11/25 16:12	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	4.40	J	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	1.20	J	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	1.50		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	2.20		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.85	J	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.			Date Collected:	06/03/25
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05			Date Received:	06/03/25
Client Sample ID:	MW-1A-20250603			SDG No.:	Q2202
Lab Sample ID:	Q2202-05			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID :	0.25	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086951.D	1		06/11/25 16:12	VN061125

[illegible]

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	MW-1A-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-05		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086951.D	1		06/11/25 16:12	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
91-20-3	Naphthalene	0.34	J		15.6	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	TB-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086949.D	1		06/11/25 15:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.	Date Collected:	06/03/25
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05	Date Received:	06/03/25
Client Sample ID:	TB-20250603	SDG No.:	Q2202
Lab Sample ID:	Q2202-06	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086949.D	1		06/11/25 15:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.9		74 - 125	98%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3		75 - 124	99%	SPK: 50
2037-26-5	Toluene-d8	51.8		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.7		77 - 121	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	324000	8.23			
540-36-3	1,4-Difluorobenzene	611000	9.106			
3114-55-4	Chlorobenzene-d5	539000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	263000	13.794			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	06/03/25	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	06/03/25	
Client Sample ID:	TB-20250603		SDG No.:	Q2202	
Lab Sample ID:	Q2202-06		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086949.D	1		06/11/25 15:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q2202**

Client: **PARSONS Engineering of New York, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2202-01	MW-9-20250603	1,2-Dichloroethane-d4	50	49.2	98	74	125
		Dibromofluoromethane	50	49.4	99	75	124
		Toluene-d8	50	52.2	104	86	113
		4-Bromofluorobenzene	50	50.1	100	77	121
Q2202-02	MW-11-20250603	1,2-Dichloroethane-d4	50	49.7	99	74	125
		Dibromofluoromethane	50	50.2	100	75	124
		Toluene-d8	50	52.1	104	86	113
		4-Bromofluorobenzene	50	49.9	100	77	121
Q2202-03	MW-12-20250603	1,2-Dichloroethane-d4	50	49.2	98	74	125
		Dibromofluoromethane	50	49.6	99	75	124
		Toluene-d8	50	52.6	105	86	113
		4-Bromofluorobenzene	50	50.2	100	77	121
Q2202-03DL	MW-12-20250603DL	1,2-Dichloroethane-d4	50	50.3	101	74	125
		Dibromofluoromethane	50	50.1	100	75	124
		Toluene-d8	50	52.2	104	86	113
		4-Bromofluorobenzene	50	51.3	103	77	121
Q2202-04	MW-13D-20250603	1,2-Dichloroethane-d4	50	48.4	97	74	125
		Dibromofluoromethane	50	49.0	98	75	124
		Toluene-d8	50	51.3	103	86	113
		4-Bromofluorobenzene	50	48.9	98	77	121
Q2202-05	MW-1A-20250603	1,2-Dichloroethane-d4	50	50.7	101	74	125
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	52.0	104	86	113
		4-Bromofluorobenzene	50	49.8	100	77	121
Q2202-06	TB-20250603	1,2-Dichloroethane-d4	50	48.9	98	74	125
		Dibromofluoromethane	50	49.4	99	75	124
		Toluene-d8	50	51.8	104	86	113
		4-Bromofluorobenzene	50	49.7	99	77	121
VN0610WBL01	VN0610WBL01	1,2-Dichloroethane-d4	50	46.7	93	74	125
		Dibromofluoromethane	50	49.0	98	75	124
		Toluene-d8	50	51.5	103	86	113
		4-Bromofluorobenzene	50	49.0	98	77	121
VN0610WBS01	VN0610WBS01	1,2-Dichloroethane-d4	50	45.4	91	74	125
		Dibromofluoromethane	50	49.5	99	75	124
		Toluene-d8	50	48.6	97	86	113
		4-Bromofluorobenzene	50	48.3	97	77	121
VN0610WBSD0	VN0610WBSD01	1,2-Dichloroethane-d4	50	47.0	94	74	125
		Dibromofluoromethane	50	49.8	100	75	124
		Toluene-d8	50	48.1	96	86	113
		4-Bromofluorobenzene	50	48.6	97	77	121
VN0611WBL01	VN0611WBL01	1,2-Dichloroethane-d4	50	48.3	97	74	125
		Dibromofluoromethane	50	49.3	99	75	124
		Toluene-d8	50	51.7	103	86	113
		4-Bromofluorobenzene	50	49.4	99	77	121
VN0611WBS02	VN0611WBS02	1,2-Dichloroethane-d4	50	47.1	94	74	125
		Dibromofluoromethane	50	51.1	102	75	124
		Toluene-d8	50	48.7	97	86	113
		4-Bromofluorobenzene	50	50.0	100	77	121

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2202

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VN086916.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0610WBS01	Dichlorodifluoromethane	20	19.0	ug/L	95			69	116	
	Chloromethane	20	15.9	ug/L	79			65	116	
	Vinyl chloride	20	18.0	ug/L	90			65	117	
	Bromomethane	20	19.7	ug/L	99			58	125	
	Chloroethane	20	18.4	ug/L	92			56	128	
	Trichlorofluoromethane	20	18.3	ug/L	92			73	115	
	1,1,2-Trichlorotrifluoroethane	20	18.2	ug/L	91			80	112	
	1,1-Dichloroethene	20	18.7	ug/L	94			74	110	
	Acetone	100	80.3	ug/L	80			60	125	
	Carbon disulfide	20	17.1	ug/L	86			64	112	
	Methyl tert-butyl Ether	20	18.7	ug/L	94			78	114	
	Methyl Acetate	20	17.1	ug/L	86			67	125	
	Methylene Chloride	20	17.4	ug/L	87			72	114	
	trans-1,2-Dichloroethene	20	17.9	ug/L	90			75	108	
	1,1-Dichloroethane	20	18.2	ug/L	91			78	112	
	Cyclohexane	20	16.5	ug/L	83			75	110	
	2-Butanone	100	84.0	ug/L	84			65	122	
	Carbon Tetrachloride	20	18.8	ug/L	94			77	113	
	cis-1,2-Dichloroethene	20	18.7	ug/L	94			77	110	
	Bromochloromethane	20	16.6	ug/L	83			70	124	
	Chloroform	20	18.2	ug/L	91			79	113	
	1,1,1-Trichloroethane	20	18.1	ug/L	91			80	108	
	Methylcyclohexane	20	15.9	ug/L	79			72	115	
	Benzene	20	18.5	ug/L	93			82	109	
	1,2-Dichloroethane	20	18.7	ug/L	94			80	115	
	Trichloroethene	20	19.3	ug/L	97			77	113	
	1,2-Dichloropropane	20	18.6	ug/L	93			83	111	
	Bromodichloromethane	20	18.8	ug/L	94			83	110	
	4-Methyl-2-Pentanone	100	91.5	ug/L	92			74	118	
	Toluene	20	19.0	ug/L	95			82	110	
	t-1,3-Dichloropropene	20	19.6	ug/L	98			79	110	
	cis-1,3-Dichloropropene	20	19.3	ug/L	97			82	110	
	1,1,2-Trichloroethane	20	19.5	ug/L	98			83	112	
	2-Hexanone	100	87.4	ug/L	87			73	117	
	Dibromochloromethane	20	19.5	ug/L	98			82	110	
	1,2-Dibromoethane	20	18.9	ug/L	95			81	110	
	Tetrachloroethene	20	17.9	ug/L	90			67	123	
	Chlorobenzene	20	18.9	ug/L	95			82	109	
	Ethyl Benzene	20	18.3	ug/L	92			83	109	
	m/p-Xylenes	40	37.7	ug/L	94			82	110	
	o-Xylene	20	18.9	ug/L	95			83	109	
	Styrene	20	19.0	ug/L	95			80	111	
	Bromoform	20	20.1	ug/L	101			79	109	
	Isopropylbenzene	20	18.1	ug/L	91			83	112	
	1,1,2,2-Tetrachloroethane	20	19.4	ug/L	97			76	118	
	1,3-Dichlorobenzene	20	18.8	ug/L	94			82	108	
	1,4-Dichlorobenzene	20	18.7	ug/L	94			82	107	
	1,2-Dichlorobenzene	20	18.8	ug/L	94			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2202

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VN086916.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0610WBS01	1,2-Dibromo-3-Chloropropane	20	18.6	ug/L	93			68	112	
	1,2,4-Trichlorobenzene	20	17.0	ug/L	85			75	113	
	1,2,3-Trichlorobenzene	20	15.7	ug/L	79			76	114	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2202

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VN086917.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0610WBSD01	Dichlorodifluoromethane	20	19.0	ug/L	95	0		69	116	20
	Chloromethane	20	16.1	ug/L	81	3		65	116	20
	Vinyl chloride	20	18.2	ug/L	91	1		65	117	20
	Bromomethane	20	19.2	ug/L	96	3		58	125	20
	Chloroethane	20	18.9	ug/L	95	3		56	128	20
	Trichlorofluoromethane	20	18.7	ug/L	94	2		73	115	20
	1,1,2-Trichlorotrifluoroethane	20	18.4	ug/L	92	1		80	112	20
	1,1-Dichloroethene	20	18.8	ug/L	94	0		74	110	20
	Acetone	100	84.2	ug/L	84	5		60	125	20
	Carbon disulfide	20	17.4	ug/L	87	1		64	112	20
	Methyl tert-butyl Ether	20	19.7	ug/L	99	5		78	114	20
	Methyl Acetate	20	18.6	ug/L	93	8		67	125	20
	Methylene Chloride	20	18.6	ug/L	93	7		72	114	20
	trans-1,2-Dichloroethene	20	18.3	ug/L	92	2		75	108	20
	1,1-Dichloroethane	20	18.9	ug/L	95	4		78	112	20
	Cyclohexane	20	16.9	ug/L	85	2		75	110	20
	2-Butanone	100	90.8	ug/L	91	8		65	122	20
	Carbon Tetrachloride	20	19.2	ug/L	96	2		77	113	20
	cis-1,2-Dichloroethene	20	19.2	ug/L	96	2		77	110	20
	Bromochloromethane	20	17.7	ug/L	89	7		70	124	20
	Chloroform	20	18.7	ug/L	94	3		79	113	20
	1,1,1-Trichloroethane	20	18.4	ug/L	92	1		80	108	20
	Methylcyclohexane	20	16.3	ug/L	81	3		72	115	20
	Benzene	20	19.0	ug/L	95	2		82	109	20
	1,2-Dichloroethane	20	19.9	ug/L	100	6		80	115	20
	Trichloroethene	20	19.7	ug/L	99	2		77	113	20
	1,2-Dichloropropane	20	19.5	ug/L	98	5		83	111	20
	Bromodichloromethane	20	20.0	ug/L	100	6		83	110	20
	4-Methyl-2-Pentanone	100	98.3	ug/L	98	6		74	118	20
	Toluene	20	19.4	ug/L	97	2		82	110	20
	t-1,3-Dichloropropene	20	20.6	ug/L	103	5		79	110	20
	cis-1,3-Dichloropropene	20	20.3	ug/L	102	5		82	110	20
	1,1,2-Trichloroethane	20	20.2	ug/L	101	3		83	112	20
	2-Hexanone	100	94.9	ug/L	95	9		73	117	20
	Dibromochloromethane	20	20.9	ug/L	104	6		82	110	20
	1,2-Dibromoethane	20	20.3	ug/L	102	7		81	110	20
	Tetrachloroethene	20	18.4	ug/L	92	2		67	123	20
	Chlorobenzene	20	19.8	ug/L	99	4		82	109	20
	Ethyl Benzene	20	18.8	ug/L	94	2		83	109	20
	m/p-Xylenes	40	38.0	ug/L	95	1		82	110	20
	o-Xylene	20	19.7	ug/L	99	4		83	109	20
	Styrene	20	19.8	ug/L	99	4		80	111	20
	Bromoform	20	21.0	ug/L	105	4		79	109	20
	Isopropylbenzene	20	18.6	ug/L	93	2		83	112	20
	1,1,2,2-Tetrachloroethane	20	20.9	ug/L	104	7		76	118	20
	1,3-Dichlorobenzene	20	19.5	ug/L	98	4		82	108	20
	1,4-Dichlorobenzene	20	19.3	ug/L	97	3		82	107	20
	1,2-Dichlorobenzene	20	19.8	ug/L	99	5		82	109	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2202

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VN086917.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0610WBSD01	1,2-Dibromo-3-Chloropropane	20	19.3	ug/L	97	4		68	112	20
	1,2,4-Trichlorobenzene	20	17.2	ug/L	86	1		75	113	20
	1,2,3-Trichlorobenzene	20	16.6	ug/L	83	5		76	114	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2202

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VN086944.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0611WBS02	Dichlorodifluoromethane	20	19.7	ug/L	99			69	116	
	Chloromethane	20	16.1	ug/L	81			65	116	
	Vinyl chloride	20	18.9	ug/L	95			65	117	
	Bromomethane	20	16.6	ug/L	83			58	125	
	Chloroethane	20	19.4	ug/L	97			56	128	
	Trichlorofluoromethane	20	19.2	ug/L	96			73	115	
	1,1,2-Trichlorotrifluoroethane	20	19.6	ug/L	98			80	112	
	1,1-Dichloroethene	20	19.2	ug/L	96			74	110	
	Acetone	100	99.0	ug/L	99			60	125	
	Carbon disulfide	20	17.8	ug/L	89			64	112	
	Methyl tert-butyl Ether	20	19.6	ug/L	98			78	114	
	Methyl Acetate	20	19.0	ug/L	95			67	125	
	Methylene Chloride	20	19.0	ug/L	95			72	114	
	trans-1,2-Dichloroethene	20	18.5	ug/L	93			75	108	
	1,1-Dichloroethane	20	19.4	ug/L	97			78	112	
	Cyclohexane	20	17.6	ug/L	88			75	110	
	2-Butanone	100	90.4	ug/L	90			65	122	
	Carbon Tetrachloride	20	19.2	ug/L	96			77	113	
	cis-1,2-Dichloroethene	20	19.5	ug/L	98			77	110	
	Bromochloromethane	20	19.9	ug/L	100			70	124	
	Chloroform	20	19.5	ug/L	98			79	113	
	1,1,1-Trichloroethane	20	19.0	ug/L	95			80	108	
	Methylcyclohexane	20	16.9	ug/L	85			72	115	
	Benzene	20	19.3	ug/L	97			82	109	
	1,2-Dichloroethane	20	19.5	ug/L	98			80	115	
	Trichloroethene	20	19.9	ug/L	100			77	113	
	1,2-Dichloropropane	20	19.6	ug/L	98			83	111	
	Bromodichloromethane	20	19.6	ug/L	98			83	110	
	4-Methyl-2-Pentanone	100	98.0	ug/L	98			74	118	
	Toluene	20	19.5	ug/L	98			82	110	
	t-1,3-Dichloropropene	20	19.9	ug/L	100			79	110	
	cis-1,3-Dichloropropene	20	20.1	ug/L	101			82	110	
	1,1,2-Trichloroethane	20	20.4	ug/L	102			83	112	
	2-Hexanone	100	89.2	ug/L	89			73	117	
	Dibromochloromethane	20	20.5	ug/L	103			82	110	
	1,2-Dibromoethane	20	19.7	ug/L	99			81	110	
	Tetrachloroethene	20	19.2	ug/L	96			67	123	
	Chlorobenzene	20	20.0	ug/L	100			82	109	
	Ethyl Benzene	20	19.6	ug/L	98			83	109	
	m/p-Xylenes	40	39.6	ug/L	99			82	110	
	o-Xylene	20	20.2	ug/L	101			83	109	
	Styrene	20	20.1	ug/L	101			80	111	
	Bromoform	20	21.0	ug/L	105			79	109	
	Isopropylbenzene	20	19.3	ug/L	97			83	112	
	1,1,2,2-Tetrachloroethane	20	21.0	ug/L	105			76	118	
	1,3-Dichlorobenzene	20	20.3	ug/L	102			82	108	
	1,4-Dichlorobenzene	20	20.4	ug/L	102			82	107	
	1,2-Dichlorobenzene	20	20.0	ug/L	100			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2202

Client: PARSONS Engineering of New York, Inc.

Analytical Method: SW8260-Low

Datafile : VN086944.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VN0611WBS02	1,2-Dibromo-3-Chloropropane	20	19.2	ug/L	96			68	112	
	1,2,4-Trichlorobenzene	20	18.4	ug/L	92			75	113	
	1,2,3-Trichlorobenzene	20	17.8	ug/L	89			76	114	

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0610WBL01

Lab Name: CHEMTECH

Contract: PARS02

Lab Code: CHEM Case No.: Q2202

SAS No.: Q2202 SDG NO.: Q2202

Lab File ID: VN086915.D

Lab Sample ID: VN0610WBL01

Date Analyzed: 06/10/2025

Time Analyzed: 10:09

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0610WBS01	VN0610WBS01	VN086916.D	06/10/2025
VN0610WBSD01	VN0610WBSD01	VN086917.D	06/10/2025
MW-9-20250603	Q2202-01	VN086934.D	06/10/2025
MW-11-20250603	Q2202-02	VN086935.D	06/10/2025
MW-12-20250603	Q2202-03	VN086936.D	06/10/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0611WBL01

Lab Name: CHEMTECH

Contract: PARS02

Lab Code: CHEM Case No.: Q2202

SAS No.: Q2202 SDG NO.: Q2202

Lab File ID: VN086942.D

Lab Sample ID: VN0611WBL01

Date Analyzed: 06/11/2025

Time Analyzed: 12:28

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0611WBS02	VN0611WBS02	VN086944.D	06/11/2025
MW-12-20250603DL	Q2202-03DL	VN086947.D	06/11/2025
TB-20250603	Q2202-06	VN086949.D	06/11/2025
MW-1A-20250603	Q2202-05	VN086951.D	06/11/2025
MW-13D-20250603	Q2202-04	VN086952.D	06/11/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202
Lab File ID: VN086861.D BFB Injection Date: 06/06/2025
Instrument ID: MSVOA_N BFB Injection Time: 07:59
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.3
75	30.0 - 60.0% of mass 95	48.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	66.6
175	5.0 - 9.0% of mass 174	4.7 (7.1) 1
176	95.0 - 101.0% of mass 174	65.3 (98.1) 1
177	5.0 - 9.0% of mass 176	4.4 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VN086862.D	06/06/2025	12:44
VSTDIC005	VSTDIC005	VN086863.D	06/06/2025	13:17
VSTDIC020	VSTDIC020	VN086864.D	06/06/2025	13:40
VSTDIC050	VSTDIC050	VN086865.D	06/06/2025	14:03
VSTDIC100	VSTDIC100	VN086866.D	06/06/2025	14:26
VSTDIC150	VSTDIC150	VN086867.D	06/06/2025	14:49

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202
Lab File ID: VN086912.D BFB Injection Date: 06/10/2025
Instrument ID: MSVOA_N BFB Injection Time: 08:39
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.1
75	30.0 - 60.0% of mass 95	47.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7
173	Less than 2.0% of mass 174	0.6 (0.8) 1
174	50.0 - 100.0% of mass 95	73.1
175	5.0 - 9.0% of mass 174	5.3 (7.2) 1
176	95.0 - 101.0% of mass 174	71.3 (97.5) 1
177	5.0 - 9.0% of mass 176	4.9 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN086913.D	06/10/2025	09:13
VN0610WBL01	VN0610WBL01	VN086915.D	06/10/2025	10:09
VN0610WBS01	VN0610WBS01	VN086916.D	06/10/2025	10:30
VN0610WBSD01	VN0610WBSD01	VN086917.D	06/10/2025	11:04
MW-9-20250603	Q2202-01	VN086934.D	06/10/2025	17:09
MW-11-20250603	Q2202-02	VN086935.D	06/10/2025	17:31
MW-12-20250603	Q2202-03	VN086936.D	06/10/2025	17:52

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202
Lab File ID: VN086939.D BFB Injection Date: 06/11/2025
Instrument ID: MSVOA_N BFB Injection Time: 10:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.7
75	30.0 - 60.0% of mass 95	46.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.1
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	71.1
175	5.0 - 9.0% of mass 174	5.4 (7.5) 1
176	95.0 - 101.0% of mass 174	68.8 (96.8) 1
177	5.0 - 9.0% of mass 176	4.8 (7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN086940.D	06/11/2025	11:32
VN0611WBL01	VN0611WBL01	VN086942.D	06/11/2025	12:28
VN0611WBS02	VN0611WBS02	VN086944.D	06/11/2025	13:27
MW-12-20250603DL	Q2202-03DL	VN086947.D	06/11/2025	14:44
TB-20250603	Q2202-06	VN086949.D	06/11/2025	15:28
MW-1A-20250603	Q2202-05	VN086951.D	06/11/2025	16:12
MW-13D-20250603	Q2202-04	VN086952.D	06/11/2025	16:35

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202
Lab File ID: VN086913.D Date Analyzed: 06/10/2025
Instrument ID: MSVOA_N Time Analyzed: 09:13
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	239452	8.23	422119	9.11	362285	11.87
UPPER LIMIT	478904	8.73	844238	9.606	724570	12.365
LOWER LIMIT	119726	7.73	211060	8.606	181143	11.365
EPA SAMPLE NO.						
MW-9-20250603	312249	8.24	586192	9.11	514503	11.87
MW-11-20250603	320004	8.24	603215	9.11	538656	11.86
MW-12-20250603	311388	8.23	590517	9.11	527900	11.87
VN0610WBL01	241251	8.23	440901	9.11	386425	11.87
VN0610WBS01	249843	8.23	443174	9.11	389297	11.87
VN0610WBSD01	231054	8.23	413772	9.11	362257	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202
 Lab File ID: VN086913.D Date Analyzed: 06/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	170987	13.788				
UPPER LIMIT	341974	14.288				
LOWER LIMIT	85493.5	13.288				
EPA SAMPLE NO.						
MW-9-20250603	245448	13.79				
MW-11-20250603	259517	13.79				
MW-12-20250603	243200	13.79				
VN0610WBL01	183380	13.79				
VN0610WBS01	189620	13.79				
VN0610WBSD01	176612	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202
Lab File ID: VN086940.D Date Analyzed: 06/11/2025
Instrument ID: MSVOA_N Time Analyzed: 11:32
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	205098	8.23	357643	9.11	312173	11.87
UPPER LIMIT	410196	8.73	715286	9.606	624346	12.365
LOWER LIMIT	102549	7.73	178822	8.606	156087	11.365
EPA SAMPLE NO.						
MW-12-20250603DL	327908	8.24	625927	9.11	564664	11.87
MW-13D-20250603	281008	8.24	529071	9.11	457419	11.87
MW-1A-20250603	321797	8.23	621476	9.11	553688	11.87
TB-20250603	324235	8.23	611383	9.11	538730	11.87
VN0611WBL01	328120	8.23	618651	9.11	543433	11.87
VN0611WBS02	206799	8.24	370524	9.11	321188	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202
 Lab File ID: VN086940.D Date Analyzed: 06/11/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:32
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	152136	13.788				
UPPER LIMIT	304272	14.288				
LOWER LIMIT	76068	13.288				
EPA SAMPLE NO.						
MW-12-20250603DL	268318	13.79				
MW-13D-20250603	218595	13.79				
MW-1A-20250603	263952	13.79				
TB-20250603	263158	13.79				
VN0611WBL01	257257	13.79				
VN0611WBS02	157403	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0610WBL01		SDG No.:	Q2202
Lab Sample ID:	VN0610WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086915.D	1		06/10/25 10:09	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0610WBL01		SDG No.:	Q2202
Lab Sample ID:	VN0610WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086915.D	1		06/10/25 10:09	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.7		74 - 125	93%	SPK: 50
1868-53-7	Dibromofluoromethane	49.0		75 - 124	98%	SPK: 50
2037-26-5	Toluene-d8	51.5		86 - 113	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.0		77 - 121	98%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	241000	8.23			
540-36-3	1,4-Difluorobenzene	441000	9.106			
3114-55-4	Chlorobenzene-d5	386000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	183000	13.788			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0610WBL01		SDG No.:	Q2202
Lab Sample ID:	VN0610WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086915.D	1		06/10/25 10:09	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0611WBL01		SDG No.:	Q2202
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0611WBL01		SDG No.:	Q2202
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	48.3		74 - 125	97%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3		75 - 124	99%	SPK: 50
2037-26-5	Toluene-d8	51.7		86 - 113	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.4		77 - 121	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	328000	8.23			
540-36-3	1,4-Difluorobenzene	619000	9.106			
3114-55-4	Chlorobenzene-d5	543000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	257000	13.788			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0611WBL01		SDG No.:	Q2202
Lab Sample ID:	VN0611WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086942.D	1		06/11/25 12:28	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0610WBS01		SDG No.:	Q2202
Lab Sample ID:	VN0610WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086916.D	1		06/10/25 10:30	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	19.0		0.22	1.00	ug/L
74-87-3	Chloromethane	15.9		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	18.0		0.26	1.00	ug/L
74-83-9	Bromomethane	19.7		1.40	5.00	ug/L
75-00-3	Chloroethane	18.4		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.3		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.2		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.7		0.23	1.00	ug/L
67-64-1	Acetone	80.3		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	17.1		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	18.7		0.16	1.00	ug/L
79-20-9	Methyl Acetate	17.1		0.27	1.00	ug/L
75-09-2	Methylene Chloride	17.4		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	17.9		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.2		0.23	1.00	ug/L
110-82-7	Cyclohexane	16.5		1.50	5.00	ug/L
78-93-3	2-Butanone	84.0		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.8		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.7		0.19	1.00	ug/L
74-97-5	Bromochloromethane	16.6		0.22	1.00	ug/L
67-66-3	Chloroform	18.2		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.1		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	15.9		0.16	1.00	ug/L
71-43-2	Benzene	18.5		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.7		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.3		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	18.6		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	18.8		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	91.5		0.68	5.00	ug/L
108-88-3	Toluene	19.0		0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0610WBS01		SDG No.:	Q2202
Lab Sample ID:	VN0610WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086916.D	1		06/10/25 10:30	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.6		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.3		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	19.5		0.21	1.00	ug/L
591-78-6	2-Hexanone	87.4		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	19.5		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	18.9		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	17.9		0.23	1.00	ug/L
108-90-7	Chlorobenzene	18.9		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	18.3		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	37.7		0.24	2.00	ug/L
95-47-6	o-Xylene	18.9		0.12	1.00	ug/L
100-42-5	Styrene	19.0		0.15	1.00	ug/L
75-25-2	Bromoform	20.1		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	18.1		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.4		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.8		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.7		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	18.8		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	18.6		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.0		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	15.7		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	45.4		74 - 125	91%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		75 - 124	99%	SPK: 50
2037-26-5	Toluene-d8	48.6		86 - 113	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.3		77 - 121	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	250000	8.23			
540-36-3	1,4-Difluorobenzene	443000	9.106			
3114-55-4	Chlorobenzene-d5	389000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	190000	13.788			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0610WBS01		SDG No.:	Q2202
Lab Sample ID:	VN0610WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086916.D	1		06/10/25 10:30	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0611WBS02		SDG No.:	Q2202
Lab Sample ID:	VN0611WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	19.7		0.22	1.00	ug/L
74-87-3	Chloromethane	16.1		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	18.9		0.26	1.00	ug/L
74-83-9	Bromomethane	16.6		1.40	5.00	ug/L
75-00-3	Chloroethane	19.4		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	19.2		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.6		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.2		0.23	1.00	ug/L
67-64-1	Acetone	99.0		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	17.8		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.6		0.16	1.00	ug/L
79-20-9	Methyl Acetate	19.0		0.27	1.00	ug/L
75-09-2	Methylene Chloride	19.0		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.5		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.4		0.23	1.00	ug/L
110-82-7	Cyclohexane	17.6		1.50	5.00	ug/L
78-93-3	2-Butanone	90.4		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.2		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.5		0.19	1.00	ug/L
74-97-5	Bromochloromethane	19.9		0.22	1.00	ug/L
67-66-3	Chloroform	19.5		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.0		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	16.9		0.16	1.00	ug/L
71-43-2	Benzene	19.3		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.5		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.9		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.6		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	19.6		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	98.0		0.68	5.00	ug/L
108-88-3	Toluene	19.5		0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0611WBS02		SDG No.:	Q2202
Lab Sample ID:	VN0611WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	19.9		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.1		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.4		0.21	1.00	ug/L
591-78-6	2-Hexanone	89.2		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	20.5		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.7		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	19.2		0.23	1.00	ug/L
108-90-7	Chlorobenzene	20.0		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	19.6		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	39.6		0.24	2.00	ug/L
95-47-6	o-Xylene	20.2		0.12	1.00	ug/L
100-42-5	Styrene	20.1		0.15	1.00	ug/L
75-25-2	Bromoform	21.0		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	19.3		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.0		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	20.3		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.4		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	20.0		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.2		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.4		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	17.8		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	47.1		74 - 125	94%	SPK: 50
1868-53-7	Dibromofluoromethane	51.1		75 - 124	102%	SPK: 50
2037-26-5	Toluene-d8	48.7		86 - 113	97%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.0		77 - 121	100%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	207000	8.235			
540-36-3	1,4-Difluorobenzene	371000	9.106			
3114-55-4	Chlorobenzene-d5	321000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	157000	13.788			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0611WBS02		SDG No.:	Q2202
Lab Sample ID:	VN0611WBS02		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086944.D	1		06/11/25 13:27	VN061125

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0610WBSD01		SDG No.:	Q2202
Lab Sample ID:	VN0610WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086917.D	1		06/10/25 11:04	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	19.0		0.22	1.00	ug/L
74-87-3	Chloromethane	16.1		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	18.2		0.26	1.00	ug/L
74-83-9	Bromomethane	19.2		1.40	5.00	ug/L
75-00-3	Chloroethane	18.9		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.7		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.4		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.8		0.23	1.00	ug/L
67-64-1	Acetone	84.2		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	17.4		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.7		0.16	1.00	ug/L
79-20-9	Methyl Acetate	18.6		0.27	1.00	ug/L
75-09-2	Methylene Chloride	18.6		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.3		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	18.9		0.23	1.00	ug/L
110-82-7	Cyclohexane	16.9		1.50	5.00	ug/L
78-93-3	2-Butanone	90.8		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.2		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.2		0.19	1.00	ug/L
74-97-5	Bromochloromethane	17.7		0.22	1.00	ug/L
67-66-3	Chloroform	18.7		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	18.4		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	16.3		0.16	1.00	ug/L
71-43-2	Benzene	19.0		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.9		0.22	1.00	ug/L
79-01-6	Trichloroethene	19.7		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.5		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	20.0		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	98.3		0.68	5.00	ug/L
108-88-3	Toluene	19.4		0.14	1.00	ug/L

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0610WBSD01		SDG No.:	Q2202
Lab Sample ID:	VN0610WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086917.D	1		06/10/25 11:04	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.6		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.3		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.2		0.21	1.00	ug/L
591-78-6	2-Hexanone	94.9		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	20.9		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.3		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	18.4		0.23	1.00	ug/L
108-90-7	Chlorobenzene	19.8		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	18.8		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	38.0		0.24	2.00	ug/L
95-47-6	o-Xylene	19.7		0.12	1.00	ug/L
100-42-5	Styrene	19.8		0.15	1.00	ug/L
75-25-2	Bromoform	21.0		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	18.6		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.9		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.5		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.3		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.8		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.3		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.2		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	16.6		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	46.9		74 - 125	94%	SPK: 50
1868-53-7	Dibromofluoromethane	49.8		75 - 124	100%	SPK: 50
2037-26-5	Toluene-d8	48.1		86 - 113	96%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.6		77 - 121	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	231000	8.23			
540-36-3	1,4-Difluorobenzene	414000	9.106			
3114-55-4	Chlorobenzene-d5	362000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	177000	13.788			

Report of Analysis

Client:	PARSONS Engineering of New York, Inc.		Date Collected:	
Project:	Con Ed Non MGP – Atlantic Ave 453957.600024.05		Date Received:	
Client Sample ID:	VN0610WBSD01		SDG No.:	Q2202
Lab Sample ID:	VN0610WBSD01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN086917.D	1		06/10/25 11:04	VN061025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG No.: Q2202

Instrument ID: MSVOA_N Calibration Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Calibration Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:								
RRF001 = VN086862.D			RRF005 = VN086863.D			RRF020 = VN086864.D		
RRF050 = VN086865.D			RRF100 = VN086866.D			RRF150 = VN086867.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.467	0.444	0.539	0.501	0.535	0.506	0.499	7.5
Chloromethane	0.762	0.654	0.645	0.597	0.617	0.587	0.644	9.9
Vinyl Chloride	0.670	0.670	0.684	0.640	0.673	0.648	0.664	2.5
Bromomethane		0.375	0.380	0.357	0.379	0.368	0.372	2.6
Chloroethane	0.460	0.444	0.442	0.408	0.418	0.402	0.429	5.4
Trichlorofluoromethane	0.882	0.903	0.904	0.834	0.858	0.825	0.868	3.9
1,1,2-Trichlorotrifluoroethane	0.554	0.567	0.563	0.519	0.546	0.520	0.545	3.8
1,1-Dichloroethene	0.573	0.593	0.563	0.533	0.550	0.527	0.557	4.4
Acetone	0.426	0.366	0.366	0.322	0.334	0.316	0.355	11.5
Carbon Disulfide	1.718	1.622	1.542	1.426	1.496	1.433	1.539	7.4
Methyl tert-butyl Ether	2.120	2.038	2.051	1.933	2.021	1.926	2.015	3.7
Methyl Acetate	1.035	1.049	1.078	0.986	1.049	1.011	1.035	3.1
Methylene Chloride	0.822	0.688	0.643	0.605	0.629	0.601	0.665	12.5
trans-1,2-Dichloroethene	0.700	0.674	0.621	0.567	0.591	0.561	0.619	9.3
1,1-Dichloroethane	1.192	1.153	1.156	1.063	1.110	1.043	1.120	5.2
Cyclohexane		1.303	1.116	1.004	1.030	0.976	1.086	12.2
2-Butanone	0.604	0.598	0.604	0.551	0.573	0.533	0.577	5.2
Carbon Tetrachloride	0.453	0.449	0.434	0.409	0.435	0.421	0.433	3.9
cis-1,2-Dichloroethene	0.786	0.766	0.762	0.699	0.729	0.701	0.740	4.9
Bromochloromethane	0.579	0.564	0.616	0.466	0.517	0.560	0.550	9.5
Chloroform	1.235	1.152	1.145	1.061	1.085	1.030	1.118	6.7
1,1,1-Trichloroethane	1.029	0.995	0.969	0.895	0.925	0.893	0.951	5.9
Methylcyclohexane	0.633	0.645	0.588	0.570	0.603	0.589	0.605	4.8
Benzene	1.588	1.501	1.444	1.345	1.414	1.371	1.444	6.2
1,2-Dichloroethane	0.473	0.456	0.444	0.411	0.430	0.413	0.438	5.6
Trichloroethene	0.359	0.360	0.341	0.327	0.340	0.328	0.342	4.2
1,2-Dichloropropane	0.366	0.369	0.354	0.332	0.352	0.335	0.351	4.4
Bromodichloromethane	0.510	0.484	0.480	0.457	0.483	0.465	0.480	3.8
4-Methyl-2-Pentanone	0.505	0.549	0.576	0.538	0.562	0.528	0.543	4.6
Toluene	0.918	0.914	0.885	0.835	0.883	0.859	0.882	3.6

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>PARS02</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q2202</u>
Instrument ID:	<u>MSVOA_N</u>	SAS No.:	<u>Q2202</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>Q2202</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>06/06/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>12:44</u>
			<u>14:49</u>

LAB FILE ID:	RRF001 = VN086862.D	RRF005 = VN086863.D	RRF020 = VN086864.D	RRF050 = VN086865.D	RRF100 = VN086866.D	RRF150 = VN086867.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.571	0.526	0.524	0.522	0.548	0.530	0.537	3.6
cis-1,3-Dichloropropene	0.609	0.577	0.564	0.551	0.584	0.561	0.574	3.6
1,1,2-Trichloroethane	0.359	0.355	0.342	0.322	0.335	0.323	0.340	4.7
2-Hexanone	0.312	0.282	0.363	0.368	0.397	0.376	0.350	12.4
Dibromochloromethane	0.361	0.351	0.358	0.340	0.363	0.349	0.354	2.5
1,2-Dibromoethane	0.353	0.358	0.354	0.329	0.354	0.341	0.348	3.2
Tetrachloroethene	0.355	0.331	0.312	0.294	0.313	0.293	0.316	7.5
Chlorobenzene	1.233	1.135	1.107	1.023	1.089	1.030	1.103	7
Ethyl Benzene	1.989	1.975	1.907	1.796	1.913	1.816	1.900	4.2
m/p-Xylenes	0.730	0.751	0.741	0.701	0.736	0.703	0.727	2.8
o-Xylene	0.714	0.699	0.702	0.674	0.711	0.678	0.696	2.4
Styrene	1.177	1.193	1.226	1.162	1.229	1.164	1.192	2.5
Bromoform	0.217	0.266	0.276	0.265	0.286	0.267	0.263	9.1
Isopropylbenzene	3.864	3.749	3.649	3.426	3.621	3.546	3.643	4.2
1,1,2,2-Tetrachloroethane	1.292	1.299	1.273	1.178	1.205	1.157	1.234	5
1,3-Dichlorobenzene	1.763	1.709	1.657	1.554	1.612	1.566	1.644	5
1,4-Dichlorobenzene	1.820	1.786	1.657	1.572	1.642	1.576	1.676	6.3
1,2-Dichlorobenzene	1.675	1.651	1.596	1.500	1.557	1.496	1.579	4.8
1,2-Dibromo-3-Chloropropane	0.339	0.317	0.290	0.272	0.283	0.270	0.295	9.3
1,2,4-Trichlorobenzene	1.042	1.037	0.991	0.969	1.016	0.994	1.008	2.8
1,2,3-Trichlorobenzene	1.073	1.009	0.993	0.950	1.000	0.987	1.002	4
1,2-Dichloroethane-d4		0.732	0.707	0.500	0.656	0.751	0.669	15.1
Dibromofluoromethane		0.303	0.310	0.219	0.298	0.351	0.296	16.2
Toluene-d8		1.245	1.203	0.861	1.178	1.377	1.173	16.2
4-Bromofluorobenzene		0.441	0.446	0.325	0.446	0.521	0.436	16.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG No.: Q2202

Instrument ID: MSVOA_N Calibration Date/Time: 06/10/2025 09:13

Lab File ID: VN086913.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.499	0.502		0.6	20
Chloromethane	0.644	0.552	0.1	-14.29	20
Vinyl Chloride	0.664	0.653		-1.66	20
Bromomethane	0.372	0.358		-3.76	20
Chloroethane	0.429	0.421		-1.87	20
Trichlorofluoromethane	0.868	0.848		-2.3	20
1,1,2-Trichlorotrifluoroethane	0.545	0.530		-2.75	20
1,1-Dichloroethene	0.557	0.548		-1.62	20
Acetone	0.355	0.348		-1.97	20
Carbon Disulfide	1.539	1.428		-7.21	20
Methyl tert-butyl Ether	2.015	1.999		-0.79	20
Methyl Acetate	1.035	0.941		-9.08	20
Methylene Chloride	0.665	0.617		-7.22	20
trans-1,2-Dichloroethene	0.619	0.587		-5.17	20
1,1-Dichloroethane	1.120	1.079	0.1	-3.66	20
Cyclohexane	1.086	0.907		-16.48	20
2-Butanone	0.577	0.493		-14.56	20
Carbon Tetrachloride	0.433	0.429		-0.92	20
cis-1,2-Dichloroethene	0.740	0.721		-2.57	20
Bromochloromethane	0.550	0.473		-14	20
Chloroform	1.118	1.053		-5.81	20
1,1,1-Trichloroethane	0.951	0.894		-5.99	20
Methylcyclohexane	0.605	0.523		-13.55	20
Benzene	1.444	1.400		-3.05	20
1,2-Dichloroethane	0.438	0.426		-2.74	20
Trichloroethene	0.342	0.345		0.88	20
1,2-Dichloropropane	0.351	0.344		-1.99	20
Bromodichloromethane	0.480	0.479		-0.21	20
4-Methyl-2-Pentanone	0.543	0.513		-5.53	20
Toluene	0.882	0.871		-1.25	20
t-1,3-Dichloropropene	0.537	0.556		3.54	20
cis-1,3-Dichloropropene	0.574	0.592		3.14	20
1,1,2-Trichloroethane	0.340	0.338		-0.59	20
2-Hexanone	0.350	0.339		-3.14	20
Dibromochloromethane	0.354	0.363		2.54	20
1,2-Dibromoethane	0.348	0.350		0.57	20
Tetrachloroethene	0.316	0.308		-2.53	20
Chlorobenzene	1.103	1.117	0.3	1.27	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG No.: Q2202

Instrument ID: MSVOA_N Calibration Date/Time: 06/10/2025 09:13

Lab File ID: VN086913.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.900	1.844		-2.95	20
m/p-Xylenes	0.727	0.727		0	20
o-Xylene	0.696	0.714		2.59	20
Styrene	1.192	1.232		3.36	20
Bromoform	0.263	0.287	0.1	9.13	20
Isopropylbenzene	3.643	3.637		-0.19	20
1,1,2,2-Tetrachloroethane	1.234	1.268	0.3	2.76	20
1,3-Dichlorobenzene	1.644	1.678		2.07	20
1,4-Dichlorobenzene	1.676	1.702		1.55	20
1,2-Dichlorobenzene	1.579	1.605		1.65	20
1,2-Dibromo-3-Chloropropane	0.295	0.277		-6.1	20
1,2,4-Trichlorobenzene	1.008	0.918		-8.93	20
1,2,3-Trichlorobenzene	1.002	0.859		-14.27	20
1,2-Dichloroethane-d4	0.669	0.613		-8.37	20
Dibromofluoromethane	0.296	0.301		1.69	20
Toluene-d8	1.173	1.153		-1.71	20
4-Bromofluorobenzene	0.436	0.420		-3.67	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG No.: Q2202

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 11:32

Lab File ID: VN086940.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.499	0.538		7.82	20
Chloromethane	0.644	0.555	0.1	-13.82	20
Vinyl Chloride	0.664	0.701		5.57	20
Bromomethane	0.372	0.312		-16.13	20
Chloroethane	0.429	0.457		6.53	20
Trichlorofluoromethane	0.868	0.930		7.14	20
1,1,2-Trichlorotrifluoroethane	0.545	0.582		6.79	20
1,1-Dichloroethene	0.557	0.593		6.46	20
Acetone	0.355	0.384		8.17	20
Carbon Disulfide	1.539	1.505		-2.21	20
Methyl tert-butyl Ether	2.015	2.172		7.79	20
Methyl Acetate	1.035	1.085		4.83	20
Methylene Chloride	0.665	0.673		1.2	20
trans-1,2-Dichloroethene	0.619	0.633		2.26	20
1,1-Dichloroethane	1.120	1.178	0.1	5.18	20
Cyclohexane	1.086	1.017		-6.35	20
2-Butanone	0.577	0.568		-1.56	20
Carbon Tetrachloride	0.433	0.472		9.01	20
cis-1,2-Dichloroethene	0.740	0.793		7.16	20
Bromochloromethane	0.550	0.578		5.09	20
Chloroform	1.118	1.176		5.19	20
1,1,1-Trichloroethane	0.951	0.976		2.63	20
Methylcyclohexane	0.605	0.576		-4.79	20
Benzene	1.444	1.554		7.55	20
1,2-Dichloroethane	0.438	0.472		7.76	20
Trichloroethene	0.342	0.382		11.7	20
1,2-Dichloropropane	0.351	0.384		9.4	20
Bromodichloromethane	0.480	0.532		10.83	20
4-Methyl-2-Pentanone	0.543	0.590		8.66	20
Toluene	0.882	0.978		10.88	20
t-1,3-Dichloropropene	0.537	0.609		13.41	20
cis-1,3-Dichloropropene	0.574	0.661		15.16	20
1,1,2-Trichloroethane	0.340	0.383		12.65	20
2-Hexanone	0.350	0.395		12.86	20
Dibromochloromethane	0.354	0.414		16.95	20
1,2-Dibromoethane	0.348	0.385		10.63	20
Tetrachloroethene	0.316	0.337		6.65	20
Chlorobenzene	1.103	1.222	0.3	10.79	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: PARS02

Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG No.: Q2202

Instrument ID: MSVOA_N Calibration Date/Time: 06/11/2025 11:32

Lab File ID: VN086940.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:44 14:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.900	2.059		8.37	20
m/p-Xylenes	0.727	0.802		10.32	20
o-Xylene	0.696	0.782		12.36	20
Styrene	1.192	1.358		13.93	20
Bromoform	0.263	0.319	0.1	21.29	20
Isopropylbenzene	3.643	3.898		7	20
1,1,2,2-Tetrachloroethane	1.234	1.402	0.3	13.61	20
1,3-Dichlorobenzene	1.644	1.843		12.1	20
1,4-Dichlorobenzene	1.676	1.887		12.59	20
1,2-Dichlorobenzene	1.579	1.761		11.53	20
1,2-Dibromo-3-Chloropropane	0.295	0.311		5.42	20
1,2,4-Trichlorobenzene	1.008	1.052		4.36	20
1,2,3-Trichlorobenzene	1.002	0.995		-0.7	20
1,2-Dichloroethane-d4	0.669	0.622		-7.03	20
Dibromofluoromethane	0.296	0.309		4.39	20
Toluene-d8	1.173	1.181		0.68	20
4-Bromofluorobenzene	0.436	0.440		0.92	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.