

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202
Lab File ID: VN086913.D Date Analyzed: 06/10/2025
Instrument ID: MSVOA_N Time Analyzed: 09:13
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	239452	8.23	422119	9.11	362285	11.87
UPPER LIMIT	478904	8.73	844238	9.606	724570	12.365
LOWER LIMIT	119726	7.73	211060	8.606	181143	11.365
EPA SAMPLE NO.						
MW-9-20250603	312249	8.24	586192	9.11	514503	11.87
MW-11-20250603	320004	8.24	603215	9.11	538656	11.86
MW-12-20250603	311388	8.23	590517	9.11	527900	11.87
VN0610WBL01	241251	8.23	440901	9.11	386425	11.87
VN0610WBS01	249843	8.23	443174	9.11	389297	11.87
VN0610WBSD01	231054	8.23	413772	9.11	362257	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.