

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PARS02
 Lab Code: CHEM Case No.: Q2202 SAS No.: Q2202 SDG NO.: Q2202
 Lab File ID: VN086913.D Date Analyzed: 06/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:13
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	170987	13.788				
UPPER LIMIT	341974	14.288				
LOWER LIMIT	85493.5	13.288				
EPA SAMPLE NO.						
MW-9-20250603	245448	13.79				
MW-11-20250603	259517	13.79				
MW-12-20250603	243200	13.79				
VN0610WBL01	183380	13.79				
VN0610WBS01	189620	13.79				
VN0610WBSD01	176612	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.