



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Tully Environmental, Inc

Project Name: Transfer Station-SPDES

Project # N/A

Order ID # Q2203

Test Name: VOC-BTEX

A. Number of Samples and Date of Receipt:

2 Water samples were received on 06/04/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, BOD5, Metals Group 10, Oil and Grease, TSS and VOC-BTEX. This data package contains results for VOC-BTEX.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI. The analysis of VOC-BTEX was based on method 624.1.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Samples 001-WILLETS-PT-BLVD(JUNE), 002-35TH-AVE(JUNE) were diluted due to high concentrations.

E. Additional Comments:

“As per method, MS/MSD is required to be performed with the sample analysis.

However, Lab did not receive sufficient volume to perform the MS/MSD therefore

MS/MSD were not performed for this project. However, Lab has performed LCS/LCSD instead.”



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____