

Initial and Continuing Calibration Verification

Client: Tully Environmental, Inc

SDG No.: Q2203

Project: Transfer Station-SPDES

RunNo.: LB136039

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV1 Ammonia as N	mg/L	1	1	100	90-110	06/06/2025
Sample ID: CCV2 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV3 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV4 Ammonia as N	mg/L	1.1	1	110	90-110	06/06/2025
Sample ID: CCV5 Ammonia as N	mg/L	1	1	100	90-110	06/06/2025