

## Duplicate Sample Summary

|                                         |                                           |
|-----------------------------------------|-------------------------------------------|
| <b>Client:</b> Tully Environmental, Inc | <b>SDG No.:</b> Q2203                     |
| <b>Project:</b> Transfer Station-SPDES  | <b>Sample ID:</b> Q2191-01                |
| <b>Client ID:</b> EFFLUENTMSD           | <b>Percent Solids for Spike Sample:</b> 0 |

| Analyte        | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|----------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Oil and Grease | mg/L  | +/-18               | 90.8             |                    | 90.9                |                    | 1                  | 0.11       |      | 06/09/2025       |

## Duplicate Sample Summary

|                                         |                                           |
|-----------------------------------------|-------------------------------------------|
| <b>Client:</b> Tully Environmental, Inc | <b>SDG No.:</b> Q2203                     |
| <b>Project:</b> Transfer Station-SPDES  | <b>Sample ID:</b> Q2197-05                |
| <b>Client ID:</b> DSN003DUP             | <b>Percent Solids for Spike Sample:</b> 0 |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| BOD5    | mg/L  | +/-20               | 8.84             |                    | 8.64                |                    | 1                  | 2.29       |      | 06/04/2025       |
| TSS     | mg/L  | +/-5                | 7.90             |                    | 8.00                |                    | 1                  | 1.26       |      | 06/05/2025       |

## Duplicate Sample Summary

|                                                |                                           |
|------------------------------------------------|-------------------------------------------|
| <b>Client:</b> Tully Environmental, Inc        | <b>SDG No.:</b> Q2203                     |
| <b>Project:</b> Transfer Station-SPDES         | <b>Sample ID:</b> Q2203-01                |
| <b>Client ID:</b> 001-WILLETS-PT-BLVD(JUNE)DUP | <b>Percent Solids for Spike Sample:</b> 0 |

| Analyte      | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|--------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Ammonia as N | mg/L  | +/-20               | 5.50             | OR                 | 5.40                | OR                 | 1                  | 2          |      | 06/06/2025       |
| Ammonia as N | mg/L  | +/-20               | 5.30             | D                  | 5.30                | D                  | 5                  | 0          |      | 06/06/2025       |

### Duplicate Sample Summary

|                   |                              |                                         |          |
|-------------------|------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | Tully Environmental, Inc     | <b>SDG No.:</b>                         | Q2203    |
| <b>Project:</b>   | Transfer Station-SPDES       | <b>Sample ID:</b>                       | Q2203-01 |
| <b>Client ID:</b> | 001-WILLETS-PT-BLVD(JUNE)MSD | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte      | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|--------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Ammonia as N | mg/L  | +/-20               | 6.70             | OR                 | 6.50                | OR                 | 1                  | 3          |      | 06/06/2025       |

### Duplicate Sample Summary

|                   |                              |                                         |          |
|-------------------|------------------------------|-----------------------------------------|----------|
| <b>Client:</b>    | Tully Environmental, Inc     | <b>SDG No.:</b>                         | Q2203    |
| <b>Project:</b>   | Transfer Station-SPDES       | <b>Sample ID:</b>                       | Q2243-02 |
| <b>Client ID:</b> | WATER-TREATMENT-DISCHARGEMSD | <b>Percent Solids for Spike Sample:</b> | 0        |

| Analyte        | Units | Acceptance<br>Limit | Sample<br>Result | Conc.<br>Qualifier | Duplicate<br>Result | Conc.<br>Qualifier | Dilution<br>Factor | RPD/<br>AD | Qual | Analysis<br>Date |
|----------------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| Oil and Grease | mg/L  | +/-18               | 23.9             |                    | 24.3                |                    | 1                  | 1.66       |      | 06/09/2025       |