

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136035

Review By	Janvi	Review On	6/6/2025 2:54:51 PM
Supervise By	MOHAN	Supervise On	6/6/2025 7:09:42 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/05/25 15:11		Jaswal	OK
2	S1	S1	CAL2	06/05/25 15:16		Jaswal	OK
3	S2	S2	CAL3	06/05/25 15:20		Jaswal	OK
4	S3	S3	CAL4	06/05/25 15:24		Jaswal	OK
5	S4	S4	CAL5	06/05/25 15:28		Jaswal	OK
6	S5	S5	CAL6	06/05/25 15:32		Jaswal	OK
7	ICV01	ICV01	ICV	06/05/25 15:55	ICV fail for Ba,Be,Tl (200.7) (95-105)	Jaswal	OK
8	LLICV01	LLICV01	LLICV	06/05/25 16:09		Jaswal	OK
9	ICB01	ICB01	ICB	06/05/25 16:14		Jaswal	OK
10	CRI01	CRI01	CRDL	06/05/25 16:19		Jaswal	OK
11	ICSA01	ICSA01	ICSA	06/05/25 16:23		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/05/25 16:32		Jaswal	OK
13	ICSADL	ICSADL	ICSA	06/05/25 16:40		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	06/05/25 16:45		Jaswal	OK
15	CCV01	CCV01	CCV	06/05/25 16:50		Jaswal	OK
16	CCB01	CCB01	CCB	06/05/25 16:54		Jaswal	OK
17	Q2205-01	001-WILLETS-PT-BL	SAM	06/05/25 16:59		Jaswal	OK
18	Q2205-02	002-35TH-AVE(MAY)	SAM	06/05/25 17:03		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136035

Review By	Janvi	Review On	6/6/2025 2:54:51 PM
Supervise By	MOHAN	Supervise On	6/6/2025 7:09:42 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

19	Q2172-01	TP06-MHQ	SAM	06/05/25 17:08		Jaswal	OK
20	Q2172-01DUP	TP06-MHQDUP	DUP	06/05/25 17:12		Jaswal	OK
21	Q2172-01L	TP06-MHQL	SD	06/05/25 17:16		Jaswal	OK
22	Q2172-01A	TP06-MHQA	PS	06/05/25 17:29		Jaswal	OK
23	Q2203-01	001-WILLETS-PT-BL	SAM	06/05/25 17:33		Jaswal	OK
24	Q2172-01MS	TP06-MHQMS	MS	06/05/25 17:37		Jaswal	OK
25	Q2172-01MSD	TP06-MHQMSD	MSD	06/05/25 17:41		Jaswal	OK
26	CCV02	CCV02	CCV	06/05/25 17:46		Jaswal	OK
27	CCB02	CCB02	CCB	06/05/25 17:50		Jaswal	OK
28	Q2203-04	002-35TH-AVE(JUNE	SAM	06/05/25 17:54		Jaswal	OK
29	Q2199-01	ETGI-343	SAM	06/05/25 17:59		Jaswal	OK
30	Q2199-03	VNJ-231	SAM	06/05/25 18:03		Jaswal	OK
31	Q2199-05	72-11978	SAM	06/05/25 18:07		Jaswal	OK
32	Q2148-02	COMP	SAM	06/05/25 19:05		Jaswal	OK
33	Q2148-02DUP	COMPDUP	DUP	06/05/25 19:09		Jaswal	OK
34	Q2148-02L	COMPL	SD	06/05/25 19:14		Jaswal	OK
35	Q2148-02MS	COMPMS	MS	06/05/25 19:18		Jaswal	OK
36	Q2148-02MSD	COMPMSD	MSD	06/05/25 19:22		Jaswal	OK
37	CCV03	CCV03	CCV	06/05/25 19:43		Jaswal	OK
38	CCB03	CCB03	CCB	06/05/25 19:47		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136035

Review By	Janvi	Review On	6/6/2025 2:54:51 PM
Supervise By	MOHAN	Supervise On	6/6/2025 7:09:42 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

39	Q2148-02A	COMPA	PS	06/05/25 19:52		Jaswal	OK
40	Q2194-01	COMP-12	SAM	06/05/25 19:56		Jaswal	OK
41	Q2194-03	COMP-13	SAM	06/05/25 20:00		Jaswal	OK
42	Q2195-01	OK-01-060325	SAM	06/05/25 20:04		Jaswal	OK
43	Q2195-01DUP	OK-01-060325DUP	DUP	06/05/25 20:08		Jaswal	OK
44	Q2195-01L	OK-01-060325L	SD	06/05/25 20:12		Jaswal	OK
45	Q2195-01MS	OK-01-060325MS	MS	06/05/25 20:16		Jaswal	OK
46	Q2195-01MSD	OK-01-060325MSD	MSD	06/05/25 20:20		Jaswal	OK
47	Q2195-01A	OK-01-060325A	PS	06/05/25 20:24		Jaswal	OK
48	CCV04	CCV04	CCV	06/05/25 20:28		Jaswal	OK
49	CCB04	CCB04	CCB	06/05/25 20:32		Jaswal	OK
50	Q2198-01	B-202-SB02	SAM	06/05/25 20:37		Jaswal	OK
51	Q2198-03	B-207-SB02	SAM	06/05/25 20:41		Jaswal	OK
52	PB168279BL	PB168279BL	MB	06/05/25 20:45	LCS fail for Na	Jaswal	Not Ok
53	PB168279BS	PB168279BS	LCS	06/05/25 20:49	LCS fail for Na	Jaswal	Not Ok
54	PB168291BL	PB168291BL	MB	06/05/25 20:53	LCS fail for Fe,Zn	Jaswal	Not Ok
55	PB168291BS	PB168291BS	LCS	06/05/25 20:58	LCS fail for Fe,Zn	Jaswal	Not Ok
56	Q2214-01	RBR251425	SAM	06/05/25 21:02		Jaswal	OK
57	Q2219-01	72-11995	SAM	06/05/25 21:06		Jaswal	OK
58	Q2224-01	EO-01-06042025	SAM	06/05/25 21:10		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136035

Review By	Janvi	Review On	6/6/2025 2:54:51 PM
Supervise By	MOHAN	Supervise On	6/6/2025 7:09:42 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

59	CCV05	CCV05	CCV	06/05/25 23:51		Jaswal	OK
60	CCB05	CCB05	CCB	06/05/25 23:56		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136052

Review By	Janvi	Review On	6/9/2025 3:49:01 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:57:27 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	06/06/25 12:50		Jaswal	OK
2	S1	S1	CAL2	06/06/25 12:54		Jaswal	OK
3	S2	S2	CAL3	06/06/25 12:59		Jaswal	OK
4	S3	S3	CAL4	06/06/25 13:03		Jaswal	OK
5	S4	S4	CAL5	06/06/25 13:07		Jaswal	OK
6	S5	S5	CAL6	06/06/25 13:12		Jaswal	OK
7	ICV01	ICV01	ICV	06/06/25 13:16		Jaswal	OK
8	LLICV01	LLICV01	LLICV	06/06/25 13:29		Jaswal	OK
9	ICB01	ICB01	ICB	06/06/25 13:33		Jaswal	OK
10	CRI01	CRI01	CRDL	06/06/25 13:37		Jaswal	OK
11	ICSA01	ICSA01	ICSA	06/06/25 13:42		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	06/06/25 13:46		Jaswal	OK
13	ICSADL	ICSADL	ICSA	06/06/25 13:50		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	06/06/25 13:54		Jaswal	OK
15	CCV01	CCV01	CCV	06/06/25 14:05		Jaswal	OK
16	CCB01	CCB01	CCB	06/06/25 14:09		Jaswal	OK
17	PB168279BL	PB168279BL	MB	06/06/25 14:13		Jaswal	OK
18	PB168279BS	PB168279BS	LCS	06/06/25 14:18		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136052

Review By	Janvi	Review On	6/9/2025 3:49:01 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:57:27 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

19	PB168284BL	PB168284BL	MB	06/06/25 14:22		Jaswal	OK
20	PB168284BS	PB168284BS	LCS	06/06/25 14:26		Jaswal	OK
21	PB168302BL	PB168302BL	MB	06/06/25 14:30		Jaswal	OK
22	PB168302BS	PB168302BS	LCS	06/06/25 14:35		Jaswal	OK
23	PB168307BL	PB168307BL	MB	06/06/25 14:39		Jaswal	OK
24	PB168307BS	PB168307BS	LCS	06/06/25 14:43		Jaswal	OK
25	Q2237-02	TW-WTS-10	SAM	06/06/25 14:47		Jaswal	OK
26	PB168291BL	PB168291BL	MB	06/06/25 14:52		Jaswal	OK
27	CCV02	CCV02	CCV	06/06/25 14:56		Jaswal	OK
28	CCB02	CCB02	CCB	06/06/25 15:00		Jaswal	OK
29	PB168291BS	PB168291BS	LCS	06/06/25 15:06		Jaswal	OK
30	PB168271TB	PB168271TB	MB	06/06/25 15:10		Jaswal	OK
31	Q2192-01	SB-1	SAM	06/06/25 15:15		Jaswal	OK
32	Q2194-02	COMP-12	SAM	06/06/25 15:19	Need to Confirm TCLP Result For Ba	Jaswal	OK
33	Q2194-04	COMP-13	SAM	06/06/25 15:24	Need to Confirm TCLP Result For Ba	Jaswal	OK
34	Q2198-02	B-202-SB02	SAM	06/06/25 15:28		Jaswal	OK
35	Q2198-04	B-207-SB02	SAM	06/06/25 15:33		Jaswal	OK
36	Q2206-04	TP-1	SAM	06/06/25 15:37		Jaswal	OK
37	Q2207-09	BU-703-COMP-01	SAM	06/06/25 15:41	Need to Confirm TCLP Result For Ba,Zn	Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136052

Review By	Janvi	Review On	6/9/2025 3:49:01 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:57:27 PM

STD. NAME	STD REF.#
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868
ICV Standard	MP85872
CCV Standard	MP85875
ICSA Standard	MP85873,MP85874
CRI Standard	MP85897
LCS Standard	
Chk Standard	MP85876,MP85877

38	Q2207-18	BU-703-COMP-02	SAM	06/06/25 15:46	Need to Confirm TCLP Result For Ba,Zn	Jaswal	OK
39	CCV03	CCV03	CCV	06/06/25 15:50		Jaswal	OK
40	CCB03	CCB03	CCB	06/06/25 15:57		Jaswal	OK
41	Q2207-27	BU-703-COMP-03	SAM	06/06/25 16:01		Jaswal	OK
42	Q2207-36	BU-703-COMP-04	SAM	06/06/25 16:10	Need to Confirm TCLP Result For Ba	Jaswal	OK
43	Q2207-45	BU-703-COMP-05	SAM	06/06/25 16:15		Jaswal	OK
44	Q2208-09	BU-703-COMP-06	SAM	06/06/25 16:19	Need to Confirm TCLP Result For Ba	Jaswal	OK
45	Q2208-18	BU-703-COMP-07	SAM	06/06/25 16:24	Need to Confirm TCLP Result For Zn	Jaswal	OK
46	Q2208-27	BU-703-COMP-08	SAM	06/06/25 16:28	Need to Confirm TCLP Result For Ba	Jaswal	OK
47	Q2208-36	BU-703-COMP-09	SAM	06/06/25 16:33	Not Use	Jaswal	Not Ok
48	Q2206-01	TP-1	SAM	06/06/25 16:37		Jaswal	OK
49	Q2207-01	BU-703-COMP-01	SAM	06/06/25 16:41		Jaswal	OK
50	CCV04	CCV04	CCV	06/06/25 16:46		Jaswal	OK
51	CCB04	CCB04	CCB	06/06/25 16:51		Jaswal	OK
52	Q2207-10	BU-703-COMP-02	SAM	06/06/25 16:55		Jaswal	OK
53	Q2207-19	BU-703-COMP-03	SAM	06/06/25 16:59		Jaswal	OK
54	Q2207-28	BU-703-COMP-04	SAM	06/06/25 17:03		Jaswal	OK
55	Q2207-37	BU-703-COMP-05	SAM	06/06/25 17:11		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136052

Review By	Janvi	Review On	6/9/2025 3:49:01 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:57:27 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

56	Q2208-01	BU-703-COMP-06	SAM	06/06/25 17:15		Jaswal	OK
57	Q2208-10	BU-703-COMP-07	SAM	06/06/25 17:19		Jaswal	OK
58	Q2208-19	BU-703-COMP-08	SAM	06/06/25 17:23		Jaswal	OK
59	Q2208-28	BU-703-COMP-09	SAM	06/06/25 17:28		Jaswal	OK
60	Q2218-01	72-11934	SAM	06/06/25 17:32		Jaswal	OK
61	Q2218-03DL	3309DL	SAM	06/06/25 17:36	Test not on log in page	Jaswal	Not Ok
62	CCV05	CCV05	CCV	06/06/25 17:40		Jaswal	OK
63	CCB05	CCB05	CCB	06/06/25 17:46		Jaswal	OK
64	Q2223-01	HR-01-06042025	SAM	06/06/25 17:51	CCV06 Fail For Many parameters	Jaswal	Not Ok
65	Q2223-03	HR-04-06042025	SAM	06/06/25 17:55	CCV06 Fail For Many parameters	Jaswal	Not Ok
66	Q2225-01	SU-03-06042025	SAM	06/06/25 17:59	CCV06 Fail For Many parameters	Jaswal	Not Ok
67	Q2226-01	TP06-MHI-WC	SAM	06/06/25 18:03	CCV06 Fail For Many parameters	Jaswal	Not Ok
68	Q2227-01	TP07-MHH-WC	SAM	06/06/25 18:07	CCV06 Fail For Many parameters	Jaswal	Not Ok
69	Q2228-01	TP08-MHI-WC	SAM	06/06/25 19:31	CCV06 Fail For Many parameters	Jaswal	Not Ok
70	Q2208-36DUP	BU-703-COMP-09DU	DUP	06/06/25 22:54	CCV06 Fail For Many parameters	Jaswal	Not Ok
71	Q2208-36L	BU-703-COMP-09L	SD	06/07/25 01:22	CCV06 Fail For Many parameters	Jaswal	Not Ok
72	Q2208-36MS	BU-703-COMP-09MS	MS	06/07/25 02:54	CCV06 Fail For Many parameters	Jaswal	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB136052

Review By	Janvi	Review On	6/9/2025 3:49:01 PM
Supervise By	jaswal	Supervise On	6/10/2025 4:57:27 PM
STD. NAME	STD REF.#		
ICAL Standard	MP85867,MP85897,MP85871,MP85870,MP85869,MP85868		
ICV Standard	MP85872		
CCV Standard	MP85875		
ICSA Standard	MP85873,MP85874		
CRI Standard	MP85897		
LCS Standard			
Chk Standard	MP85876,MP85877		

73	CCV06	CCV06	CCV	06/07/25 03:03		Jaswal	OK
74	CCB06	CCB06	CCB	06/07/25 03:07		Jaswal	OK
75	Q2208-36MSD	BU-703-COMP-09MS	MSD	06/07/25 03:11	CCV06,07 Fail For Many parameters	Jaswal	Not Ok
76	Q2208-36A	BU-703-COMP-09A	PS	06/07/25 03:15	CCV06,07 Fail For Many parameters	Jaswal	Not Ok
77	Q2226-01DUP	TP06-MHI-WCDUP	DUP	06/07/25 03:20	CCV06,07 Fail For Many parameters	Jaswal	Not Ok
78	Q2226-01L	TP06-MHI-WCL	SD	06/07/25 03:24	CCV06,07 Fail For Many parameters	Jaswal	Not Ok
79	Q2226-01MS	TP06-MHI-WCMS	MS	06/07/25 03:28	CCV06,07 Fail For Many parameters	Jaswal	Not Ok
80	Q2226-01MSD	TP06-MHI-WCMSD	MSD	06/07/25 03:32	CCV06,07 Fail For Many parameters	Jaswal	Not Ok
81	Q2226-01A	TP06-MHI-WCA	PS	06/07/25 03:36	CCV06,07 Fail For Many parameters	Jaswal	Not Ok
82	CCV07	CCV07	CCV	06/07/25 03:40		Jaswal	OK
83	CCB07	CCB07	CCB	06/07/25 03:44		Jaswal	OK