

**Hit Summary Sheet**  
**SW-846**

|                 |                          |                    |                        |
|-----------------|--------------------------|--------------------|------------------------|
| <b>SDG No.:</b> | Q2203                    | <b>Order ID:</b>   | Q2203                  |
| <b>Client:</b>  | Tully Environmental, Inc | <b>Project ID:</b> | Transfer Station-SPDES |

| Sample ID                                    | Client ID                | Matrix | Parameter | Concentration | C | MDL  | RDL  | Units |
|--|--------------------------|--------|-----------|---------------|---|------|------|-------|
| <b>Client ID : 001-WILLETS-PT-BLVD(JUNE)</b> |                          |        |           |               |   |      |      |       |
| Q2203-01                                     | 001-WILLETS-PT-BLVD(JUNE | Water  | Copper    | 70.8          |   | 1.89 | 10.0 | ug/L  |
| Q2203-01                                     | 001-WILLETS-PT-BLVD(JUNE | Water  | Iron      | 19100         |   | 18.0 | 50.0 | ug/L  |
| Q2203-01                                     | 001-WILLETS-PT-BLVD(JUNE | Water  | Lead      | 69.0          |   | 1.21 | 6.00 | ug/L  |
| <b>Client ID : 002-35TH-AVE(JUNE)</b>        |                          |        |           |               |   |      |      |       |
| Q2203-04                                     | 002-35TH-AVE(JUNE)       | Water  | Copper    | 84.0          |   | 1.89 | 10.0 | ug/L  |
| Q2203-04                                     | 002-35TH-AVE(JUNE)       | Water  | Iron      | 18100         |   | 18.0 | 50.0 | ug/L  |
| Q2203-04                                     | 002-35TH-AVE(JUNE)       | Water  | Lead      | 72.6          |   | 1.21 | 6.00 | ug/L  |



# SAMPLE DATA

A

B

C

D

## Report of Analysis

|                   |                           |                 |          |
|-------------------|---------------------------|-----------------|----------|
| Client:           | Tully Environmental, Inc  | Date Collected: | 06/03/25 |
| Project:          | Transfer Station-SPDES    | Date Received:  | 06/04/25 |
| Client Sample ID: | 001-WILLETS-PT-BLVD(JUNE) | SDG No.:        | Q2203    |
| Lab Sample ID:    | Q2203-01                  | Matrix:         | Water    |
| Level (low/med):  | low                       | % Solid:        | 0        |

| Cas       | Parameter | Conc. | Qua. | DF | MDL  | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met.  | Prep Met. |
|-----------|-----------|-------|------|----|------|------------|-------|----------------|----------------|-----------|-----------|
| 7440-50-8 | Copper    | 70.8  |      | 1  | 1.89 | 10.0       | ug/L  | 06/04/25 15:15 | 06/05/25 17:33 | EPA 200.7 |           |
| 7439-89-6 | Iron      | 19100 |      | 1  | 18.0 | 50.0       | ug/L  | 06/04/25 15:15 | 06/05/25 17:33 | EPA 200.7 |           |
| 7439-92-1 | Lead      | 69.0  |      | 1  | 1.21 | 6.00       | ug/L  | 06/04/25 15:15 | 06/05/25 17:33 | EPA 200.7 |           |

|               |                 |                 |        |            |
|---------------|-----------------|-----------------|--------|------------|
| Color Before: | light Brown     | Clarity Before: | Cloudy | Texture:   |
| Color After:  | Colorless       | Clarity After:  | Clear  | Artifacts: |
| Comments:     | Metals Group 10 |                 |        |            |

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits

## Report of Analysis

|                   |                          |                 |          |
|-------------------|--------------------------|-----------------|----------|
| Client:           | Tully Environmental, Inc | Date Collected: | 06/03/25 |
| Project:          | Transfer Station-SPDES   | Date Received:  | 06/04/25 |
| Client Sample ID: | 002-35TH-AVE(JUNE)       | SDG No.:        | Q2203    |
| Lab Sample ID:    | Q2203-04                 | Matrix:         | Water    |
| Level (low/med):  | low                      | % Solid:        | 0        |

| Cas       | Parameter | Conc. | Qua. | DF | MDL  | LOQ / CRQL | Units | Prep Date      | Date Ana.      | Ana Met.  | Prep Met. |
|-----------|-----------|-------|------|----|------|------------|-------|----------------|----------------|-----------|-----------|
| 7440-50-8 | Copper    | 84.0  |      | 1  | 1.89 | 10.0       | ug/L  | 06/04/25 15:15 | 06/05/25 17:54 | EPA 200.7 |           |
| 7439-89-6 | Iron      | 18100 |      | 1  | 18.0 | 50.0       | ug/L  | 06/04/25 15:15 | 06/05/25 17:54 | EPA 200.7 |           |
| 7439-92-1 | Lead      | 72.6  |      | 1  | 1.21 | 6.00       | ug/L  | 06/04/25 15:15 | 06/05/25 17:54 | EPA 200.7 |           |

|               |                 |                 |        |            |
|---------------|-----------------|-----------------|--------|------------|
| Color Before: | light Brown     | Clarity Before: | Cloudy | Texture:   |
| Color After:  | Colorless       | Clarity After:  | Clear  | Artifacts: |
| Comments:     | Metals Group 10 |                 |        |            |

U = Not Detected  
 LOQ = Limit of Quantitation  
 MDL = Method Detection Limit  
 LOD = Limit of Detection  
 D = Dilution  
 Q = indicates LCS control criteria did not meet requirements

J = Estimated Value  
 B = Analyte Found in Associated Method Blank  
 \* = indicates the duplicate analysis is not within control limits.  
 E = Indicates the reported value is estimated because of the presence of interference.  
 OR = Over Range  
 N =Spiked sample recovery not within control limits

### LAB CHRONICLE

|                 |                          |                   |                        |
|-----------------|--------------------------|-------------------|------------------------|
| <b>OrderID:</b> | Q2203                    | <b>OrderDate:</b> | 6/4/2025 12:05:00 PM   |
| <b>Client:</b>  | Tully Environmental, Inc | <b>Project:</b>   | Transfer Station-SPDES |
| <b>Contact:</b> | Dean Devoe               | <b>Location:</b>  | N31,VOA Ref. #3 Water  |

| LabID    | ClientID                      | Matrix | Test            | Method | Sample Date | Prep Date | Anal Date | Received |
|----------|-------------------------------|--------|-----------------|--------|-------------|-----------|-----------|----------|
| Q2203-01 | 001-WILLETS-PT-BLV<br>D(JUNE) | Water  |                 |        | 06/03/25    |           |           | 06/04/25 |
|          |                               |        | Metals Group 10 | 200.7  |             | 06/04/25  | 06/05/25  |          |
| Q2203-04 | 002-35TH-AVE(JUNE)            | Water  |                 |        | 06/03/25    |           |           | 06/04/25 |
|          |                               |        | Metals Group 10 | 200.7  |             | 06/04/25  | 06/05/25  |          |