

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2203
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2203 **SAS No.:** Q2203
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	527	510	103	95 - 105	P	06/05/2025	15:55	LB136035
	Iron	10100	10000	101	95 - 105	P	06/05/2025	15:55	LB136035
	Lead	985	1000	98	95 - 105	P	06/05/2025	15:55	LB136035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2203
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2203 **SAS No.:** Q2203
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.5	20.0	107	80 - 120	P	06/05/2025	16:09	LB136035
	Iron	114	100	114	80 - 120	P	06/05/2025	16:09	LB136035
	Lead	14.0	12.0	117	80 - 120	P	06/05/2025	16:09	LB136035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc SDG No.: Q2203
Contract: TULL01 Lab Code: CHEM Case No.: Q2203 SAS No.: Q2203
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1260	1250	101	90 - 110	P	06/05/2025	16:50	LB136035
	Iron	4890	5000	98	90 - 110	P	06/05/2025	16:50	LB136035
	Lead	4930	5000	99	90 - 110	P	06/05/2025	16:50	LB136035
CCV02	Copper	1260	1250	101	90 - 110	P	06/05/2025	17:46	LB136035
	Iron	5140	5000	103	90 - 110	P	06/05/2025	17:46	LB136035
	Lead	4870	5000	97	90 - 110	P	06/05/2025	17:46	LB136035
CCV03	Copper	1230	1250	98	90 - 110	P	06/05/2025	19:43	LB136035
	Iron	5110	5000	102	90 - 110	P	06/05/2025	19:43	LB136035
	Lead	4870	5000	97	90 - 110	P	06/05/2025	19:43	LB136035
CCV04	Copper	1270	1250	102	90 - 110	P	06/05/2025	20:28	LB136035
	Iron	5230	5000	105	90 - 110	P	06/05/2025	20:28	LB136035
	Lead	4990	5000	100	90 - 110	P	06/05/2025	20:28	LB136035
CCV05	Copper	1250	1250	100	90 - 110	P	06/05/2025	23:51	LB136035
	Iron	5320	5000	106	90 - 110	P	06/05/2025	23:51	LB136035
	Lead	4970	5000	100	90 - 110	P	06/05/2025	23:51	LB136035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2203
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2203 **SAS No.:** Q2203
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	527	510	103	95 - 105	P	06/06/2025	13:16	LB136052
	Iron	10300	10000	103	95 - 105	P	06/06/2025	13:16	LB136052
	Lead	980	1000	98	95 - 105	P	06/06/2025	13:16	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2203
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2203 **SAS No.:** Q2203
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.0	20.0	105	80 - 120	P	06/06/2025	13:29	LB136052
	Iron	106	100	106	80 - 120	P	06/06/2025	13:29	LB136052
	Lead	13.9	12.0	116	80 - 120	P	06/06/2025	13:29	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2203
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2203 **SAS No.:** Q2203
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	06/06/2025	14:05	LB136052
	Iron	4810	5000	96	90 - 110	P	06/06/2025	14:05	LB136052
	Lead	4910	5000	98	90 - 110	P	06/06/2025	14:05	LB136052
CCV02	Copper	1280	1250	102	90 - 110	P	06/06/2025	14:56	LB136052
	Iron	4920	5000	98	90 - 110	P	06/06/2025	14:56	LB136052
	Lead	4960	5000	99	90 - 110	P	06/06/2025	14:56	LB136052
CCV03	Copper	1280	1250	103	90 - 110	P	06/06/2025	15:50	LB136052
	Iron	4640	5000	93	90 - 110	P	06/06/2025	15:50	LB136052
	Lead	5000	5000	100	90 - 110	P	06/06/2025	15:50	LB136052
CCV04	Copper	1240	1250	99	90 - 110	P	06/06/2025	16:46	LB136052
	Iron	4980	5000	100	90 - 110	P	06/06/2025	16:46	LB136052
	Lead	4890	5000	98	90 - 110	P	06/06/2025	16:46	LB136052
CCV05	Copper	1250	1250	100	90 - 110	P	06/06/2025	17:40	LB136052
	Iron	4770	5000	95	90 - 110	P	06/06/2025	17:40	LB136052
	Lead	4810	5000	96	90 - 110	P	06/06/2025	17:40	LB136052
CCV06	Copper	1170	1250	94	90 - 110	P	06/07/2025	03:03	LB136052
	Iron	4160	5000	83	90 - 110	P	06/07/2025	03:03	LB136052
	Lead	4580	5000	92	90 - 110	P	06/07/2025	03:03	LB136052
CCV07	Copper	1180	1250	94	90 - 110	P	06/07/2025	03:40	LB136052
	Iron	4870	5000	98	90 - 110	P	06/07/2025	03:40	LB136052
	Lead	4630	5000	92	90 - 110	P	06/07/2025	03:40	LB136052