



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q2203
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2203 **SAS No.:** Q2203
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	21.2	20.0	106	65 - 135	P	06/05/2025	16:19	LB136035
	Iron	111	100	111	65 - 135	P	06/05/2025	16:19	LB136035
	Lead	13.0	12.0	108	65 - 135	P	06/05/2025	16:19	LB136035
CRI01	Copper	20.8	20.0	104	65 - 135	P	06/06/2025	13:37	LB136052
	Iron	112	100	112	65 - 135	P	06/06/2025	13:37	LB136052
	Lead	13.8	12.0	115	65 - 135	P	06/06/2025	13:37	LB136052