



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2203

Contract: TULL01

Lab Code: CHEM

Case No.: Q2203

SAS No.: Q2203

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	06/05/2025	16:14	LB136035
	Iron	100	+/-100	U	100	P	06/05/2025	16:14	LB136035
	Lead	12.0	+/-12.0	U	12.0	P	06/05/2025	16:14	LB136035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Environmental, Inc</u>		SDG No.: <u>Q2203</u>							
Contract: <u>TULL01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2203</u>	SAS No.: <u>Q2203</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	06/05/2025	16:54	LB136035
	Iron	100	+/-100	U	100	P	06/05/2025	16:54	LB136035
	Lead	12.0	+/-12.0	U	12.0	P	06/05/2025	16:54	LB136035
CCB02	Copper	20.0	+/-20.0	U	20.0	P	06/05/2025	17:50	LB136035
	Iron	100	+/-100	U	100	P	06/05/2025	17:50	LB136035
	Lead	12.0	+/-12.0	U	12.0	P	06/05/2025	17:50	LB136035
CCB03	Copper	20.0	+/-20.0	U	20.0	P	06/05/2025	19:47	LB136035
	Iron	100	+/-100	U	100	P	06/05/2025	19:47	LB136035
	Lead	12.0	+/-12.0	U	12.0	P	06/05/2025	19:47	LB136035
CCB04	Copper	20.0	+/-20.0	U	20.0	P	06/05/2025	20:32	LB136035
	Iron	100	+/-100	U	100	P	06/05/2025	20:32	LB136035
	Lead	12.0	+/-12.0	U	12.0	P	06/05/2025	20:32	LB136035
CCB05	Copper	20.0	+/-20.0	U	20.0	P	06/05/2025	23:56	LB136035
	Iron	100	+/-100	U	100	P	06/05/2025	23:56	LB136035
	Lead	12.0	+/-12.0	U	12.0	P	06/05/2025	23:56	LB136035

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2203

Contract: TULL01

Lab Code: CHEM

Case No.: Q2203

SAS No.: Q2203

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	20.0	+/-20.0	U	20.0	P	06/06/2025	13:33	LB136052
	Iron	100	+/-100	U	100	P	06/06/2025	13:33	LB136052
	Lead	12.0	+/-12.0	U	12.0	P	06/06/2025	13:33	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Environmental, Inc</u>		SDG No.: <u>Q2203</u>							
Contract: <u>TULL01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2203</u>	SAS No.: <u>Q2203</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	20.0	+/-20.0	U	20.0	P	06/06/2025	14:09	LB136052
	Iron	100	+/-100	U	100	P	06/06/2025	14:09	LB136052
	Lead	12.0	+/-12.0	U	12.0	P	06/06/2025	14:09	LB136052
CCB02	Copper	20.0	+/-20.0	U	20.0	P	06/06/2025	15:00	LB136052
	Iron	100	+/-100	U	100	P	06/06/2025	15:00	LB136052
	Lead	12.0	+/-12.0	U	12.0	P	06/06/2025	15:00	LB136052
CCB03	Copper	20.0	+/-20.0	U	20.0	P	06/06/2025	15:57	LB136052
	Iron	100	+/-100	U	100	P	06/06/2025	15:57	LB136052
	Lead	12.0	+/-12.0	U	12.0	P	06/06/2025	15:57	LB136052
CCB04	Copper	20.0	+/-20.0	U	20.0	P	06/06/2025	16:51	LB136052
	Iron	100	+/-100	U	100	P	06/06/2025	16:51	LB136052
	Lead	12.0	+/-12.0	U	12.0	P	06/06/2025	16:51	LB136052
CCB05	Copper	20.0	+/-20.0	U	20.0	P	06/06/2025	17:46	LB136052
	Iron	100	+/-100	U	100	P	06/06/2025	17:46	LB136052
	Lead	12.0	+/-12.0	U	12.0	P	06/06/2025	17:46	LB136052
CCB06	Copper	20.0	+/-20.0	U	20.0	P	06/07/2025	03:07	LB136052
	Iron	100	+/-100	U	100	P	06/07/2025	03:07	LB136052
	Lead	12.0	+/-12.0	U	12.0	P	06/07/2025	03:07	LB136052
CCB07	Copper	20.0	+/-20.0	U	20.0	P	06/07/2025	03:44	LB136052
	Iron	100	+/-100	U	100	P	06/07/2025	03:44	LB136052
	Lead	12.0	+/-12.0	U	12.0	P	06/07/2025	03:44	LB136052