

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q2203
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2203 **SAS No.:** Q2203
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	19.3	2.0	965	-18	22	06/05/2025	16:23	LB136035
	Iron	103000	101000	102	85600	116500	06/05/2025	16:23	LB136035
	Lead	-0.76			-12	12	06/05/2025	16:23	LB136035
ICSAB01	Copper	517	511	101	434	588	06/05/2025	16:32	LB136035
	Iron	99900	99300	101	84400	114500	06/05/2025	16:32	LB136035
	Lead	48.0	49.0	98	37	61	06/05/2025	16:32	LB136035
ICSA	Copper	55.0	2.0	2750	-18	22	06/05/2025	16:40	LB136035
	Iron	96000	101000	95	85600	116500	06/05/2025	16:40	LB136035
	Lead	-8.58			-12	12	06/05/2025	16:40	LB136035
ICSAB	Copper	524	511	102	434	588	06/05/2025	16:45	LB136035
	Iron	105000	99300	106	84400	114500	06/05/2025	16:45	LB136035
	Lead	64.2	49.0	131	37	61	06/05/2025	16:45	LB136035
ICSA01	Copper	-6.16	2.0	308	-18	22	06/06/2025	13:42	LB136052
	Iron	90700	101000	90	85600	116500	06/06/2025	13:42	LB136052
	Lead	2.11			-12	12	06/06/2025	13:42	LB136052
ICSAB01	Copper	482	511	94	434	588	06/06/2025	13:46	LB136052
	Iron	104000	99300	105	84400	114500	06/06/2025	13:46	LB136052
	Lead	47.1	49.0	96	37	61	06/06/2025	13:46	LB136052
ICSA	Copper	-6.80	2.0	340	-18	22	06/06/2025	13:50	LB136052
	Iron	95000	101000	94	85600	116500	06/06/2025	13:50	LB136052
	Lead	17.5			-12	12	06/06/2025	13:50	LB136052
ICSAB	Copper	487	511	95	434	588	06/06/2025	13:54	LB136052
	Iron	88600	99300	89	84400	114500	06/06/2025	13:54	LB136052
	Lead	62.6	49.0	128	37	61	06/06/2025	13:54	LB136052