

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2203</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2203</u> sas no.: <u>Q2203</u>
matrix: <u>Water</u>	sample id: <u>Q2148-02</u>	client id: <u>COMPMS</u>
Percent Solids for Sample: NA	Spiked ID: Q2148-02MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	266		99.1		150	112		P
Iron	ug/L	75 - 125	4990		3180		1500	121		P
Lead	ug/L	75 - 125	536		9.01		500	105		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q2203</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q2203</u> sas no.: <u>Q2203</u>
matrix: <u>Water</u>	sample id: <u>Q2148-02</u>	client id: <u>COMPMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q2148-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	257		99.1		150	105		P
Iron	ug/L	75 - 125	5050		3180		1500	125		P
Lead	ug/L	75 - 125	513		9.01		500	101		P