

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc **SDG No.:** Q2203
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2203 **SAS No.:** Q2203

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168291BS							
Copper	ug/L	150	161		107	85 - 115	P
Iron	ug/L	1500	1500		100	85 - 115	P
Lead	ug/L	500	491		98	85 - 115	P