

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q2203 SAS No.: Q2203 SDG No.: Q2203

Instrument ID: MSVOA_X Calibration Date/Time: 06/10/2025 20:40

Lab File ID: VX046615.D Init. Calib. Date(s): 06/10/2025 06/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 20:19 21:44

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.387	1.500		8.15	
Toluene	1.626	1.749		7.57	
Ethyl Benzene	1.856	1.964		5.82	
m/p-Xylenes	0.687	0.734		6.84	
o-Xylene	0.661	0.703		6.35	
1,2-Dichloroethane-d4	2.428	2.442		0.58	
Toluene-d8	1.340	1.354		1.04	
4-Bromofluorobenzene	0.506	0.506		0	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.