

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q2203 SAS No.: Q2203 SDG No.: Q2203

Instrument ID: MSVOA_X Calibration Date/Time: 06/11/2025 11:12

Lab File ID: VX046628.D Init. Calib. Date(s): 06/10/2025 06/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 20:19 21:44

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Benzene	1.387	1.393	0.5	0.43	
Toluene	1.626	1.647	0.4	1.29	
Ethyl Benzene	1.856	1.841	0.1	-0.81	
m/p-Xylenes	0.687	0.689	0.3	0.29	
o-Xylene	0.661	0.658	0.3	-0.45	
1,2-Dichloroethane-d4	2.428	2.484	0.01	2.31	
Toluene-d8	1.340	1.361	0.01	1.57	
4-Bromofluorobenzene	0.506	0.516	0.2	1.98	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.