


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061025\
Data File : VX046613.D
Acq On : 10 Jun 2025 19:57
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Title : SW846 8260
Last Update : Fri Jun 06 16:56:12 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	7615	PASS
75	95	30	60	54.4	20195	PASS
95	95	100	100	100.0	37107	PASS
96	95	5	9	6.9	2556	PASS
173	174	0.00	2	0.7	178	PASS
174	95	50	100	68.0	25240	PASS
175	174	5	9	7.8	1961	PASS
176	174	95	101	95.5	24109	PASS
177	176	5	9	6.8	1649	PASS