

Surrogate Summary

SDG No.: Q2203

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2203-01	001-WILLETS-PT-BLVD(JUNE)	1,2-Dichloroethane-d4	30	30.1	100	91	110
		Toluene-d8	30	30.4	101	91	112
		4-Bromofluorobenzene	30	29.0	97	63	112
Q2203-01DL	001-WILLETS-PT-BLVD(JUNE)DL	1,2-Dichloroethane-d4	30	32.3	108	91	110
		Toluene-d8	30	29.6	99	91	112
		4-Bromofluorobenzene	30	30.0	100	63	112
Q2203-04	002-35TH-AVE(JUNE)	1,2-Dichloroethane-d4	30	30.9	103	91	110
		Toluene-d8	30	30.0	100	91	112
		4-Bromofluorobenzene	30	28.5	95	63	112
Q2203-04DL	002-35TH-AVE(JUNE)DL	1,2-Dichloroethane-d4	30	30.8	103	91	110
		Toluene-d8	30	29.7	99	91	112
		4-Bromofluorobenzene	30	29.0	97	63	112
VX0610WBL02	VX0610WBL02	1,2-Dichloroethane-d4	30	31.4	105	91	110
		Toluene-d8	30	30.1	100	91	112
		4-Bromofluorobenzene	30	29.7	99	63	112
VX0610WBS02	VX0610WBS02	1,2-Dichloroethane-d4	30	30.2	101	91	110
		Toluene-d8	30	30.7	102	91	112
		4-Bromofluorobenzene	30	30.7	102	63	112
VX0610WBSD0	VX0610WBSD02	1,2-Dichloroethane-d4	30	30.2	101	91	110
		Toluene-d8	30	30.2	101	91	112
		4-Bromofluorobenzene	30	30.4	101	63	112
VX0611WBL01	VX0611WBL01	1,2-Dichloroethane-d4	30	30.5	102	91	110
		Toluene-d8	30	29.8	99	91	112
		4-Bromofluorobenzene	30	30.3	101	63	112
VX0611WBS01	VX0611WBS01	1,2-Dichloroethane-d4	30	29.9	100	91	110
		Toluene-d8	30	29.9	100	91	112
		4-Bromofluorobenzene	30	30.1	100	63	112