

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2203

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VX046621.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0610WBS02	Benzene	20	18.1	ug/L	91			65	135	
	Toluene	20	18.7	ug/L	94			70	130	
	Ethyl Benzene	20	18.5	ug/L	93			60	140	
	m/p-Xylenes	40	37.8	ug/L	95			87	111	
	o-Xylene	20	18.6	ug/L	93			87	111	