

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2203

Client: Tully Environmental, Inc

Analytical Method: SW624.1

Datafile : VX046622.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0610WBSD02	Benzene	20	19.8	ug/L	99	8		65	135	20
	Toluene	20	20.3	ug/L	102	8		70	130	20
	Ethyl Benzene	20	19.9	ug/L	100	7		60	140	20
	m/p-Xylenes	40	40.9	ug/L	102	7		87	111	20
	o-Xylene	20	20.2	ug/L	101	8		87	111	20