

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: Q2203 SAS No.: Q2203 SDG NO.: Q2203
 Lab File ID: VX046615.D Date Analyzed: 06/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 20:40
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	19846	4.92	104620	6.77	96769	10.06
UPPER LIMIT	39692	5.416	209240	7.269	193538	10.555
LOWER LIMIT	9923	4.416	52310	6.269	48384.5	9.555
EPA SAMPLE NO.						
001-WILLETS-PT-BLVD (JUNE)	18793	4.91	96292	6.77	85225	10.06
002-35TH-AVE (JUNE)	17480	4.92	92940	6.77	83680	10.06
VX0610WBL02	18694	4.92	99149	6.77	91863	10.06
VX0610WBS02	19222	4.92	103988	6.77	93147	10.06
VX0610WBSD02	18375	4.92	97353	6.77	88537	10.06

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.