

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
 Lab Code: CHEM Case No.: Q2203 SAS No.: Q2203 SDG NO.: Q2203
 Lab File ID: VX046628.D Date Analyzed: 06/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:12
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
001-WILLETTS-PT-BLVD (JUNE) DL	0	0.00				
002-35TH-AVE (JUNE) DL	0	0.00				
VX0611WBL01	0	0.00				
VX0611WBS01	0	0.00				

IS4 =

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.