



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

## **CASE NARRATIVE**

**Tully Environmental, Inc**

**Project Name: Transfer Station-SPDES**

**Project # N/A**

**Order ID # Q2205**

**Test Name: Metals Group 12**

### **A. Number of Samples and Date of Receipt:**

2 Water samples were received on 06/04/2025.

### **B. Parameters:**

According to the Chain of Custody document, the following analyses were requested: Metals Group 12 and TSS. This data package contains results for Metals Group 12.

### **C. Analytical Techniques:**

The analysis and digestion of Metals Group 12 was based on method 200.7.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

### **E. Additional Comments:**

In analytical sequence LB136052, The % recovery was outside of acceptance limit for Iron of CCV06 but no any sample associated under this CCV.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_