



Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	10100	10000	101	95 - 105	P	06/05/2025	15:55	LB136035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2205
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2205 **SAS No.:** Q2205
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	114	100	114	80 - 120	P	06/05/2025	16:09	LB136035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2205
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2205 **SAS No.:** Q2205
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4890	5000	98	90 - 110	P	06/05/2025	16:50	LB136035
CCV02	Iron	5140	5000	103	90 - 110	P	06/05/2025	17:46	LB136035
CCV03	Iron	5110	5000	102	90 - 110	P	06/05/2025	19:43	LB136035
CCV04	Iron	5230	5000	105	90 - 110	P	06/05/2025	20:28	LB136035
CCV05	Iron	5320	5000	106	90 - 110	P	06/05/2025	23:51	LB136035

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2205
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2205 **SAS No.:** Q2205
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	10300	10000	103	95 - 105	P	06/06/2025	13:16	LB136052



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2205
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2205 **SAS No.:** Q2205
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	106	100	106	80 - 120	P	06/06/2025	13:29	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q2205
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2205 **SAS No.:** Q2205
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	4810	5000	96	90 - 110	P	06/06/2025	14:05	LB136052
CCV02	Iron	4920	5000	98	90 - 110	P	06/06/2025	14:56	LB136052
CCV03	Iron	4640	5000	93	90 - 110	P	06/06/2025	15:50	LB136052
CCV04	Iron	4980	5000	100	90 - 110	P	06/06/2025	16:46	LB136052
CCV05	Iron	4770	5000	95	90 - 110	P	06/06/2025	17:40	LB136052
CCV06	Iron	4160	5000	83	90 - 110	P	06/07/2025	03:03	LB136052
CCV07	Iron	4870	5000	98	90 - 110	P	06/07/2025	03:40	LB136052