

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q2205
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q2205 **SAS No.:** Q2205
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	103000	101000	102	85600	116500	06/05/2025	16:23	LB136035
ICSAB01	Iron	99900	99300	101	84400	114500	06/05/2025	16:32	LB136035
ICSA	Iron	96000	101000	95	85600	116500	06/05/2025	16:40	LB136035
ICSAB	Iron	105000	99300	106	84400	114500	06/05/2025	16:45	LB136035
ICSA01	Iron	90700	101000	90	85600	116500	06/06/2025	13:42	LB136052
ICSAB01	Iron	104000	99300	105	84400	114500	06/06/2025	13:46	LB136052
ICSA	Iron	95000	101000	94	85600	116500	06/06/2025	13:50	LB136052
ICSAB	Iron	88600	99300	89	84400	114500	06/06/2025	13:54	LB136052