

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q2205
contract: TULL01 **lab code:** CHEM **case no.:** Q2205 **sas no.:** Q2205
matrix: Water **sample id:** Q2148-02 **client id:** COMPMS
Percent Solids for Sample: NA **Spiked ID:** Q2148-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	4990		3180		1500	121		P

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q2205
contract: TULL01 **lab code:** CHEM **case no.:** Q2205 **sas no.:** Q2205
matrix: Water **sample id:** Q2148-02 **client id:** COMPMSD
Percent Solids for Sample: NA **Spiked ID:** Q2148-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	75 - 125	5050		3180		1500	125		P