

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q2209	OrderDate:	6/4/2025 1:52:00 PM
Client:	G Environmental	Project:	Power
Contact:	Gary Landis	Location:	N31,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2209-01	P01W	Water			06/04/25			06/04/25
			SVOCMS Group2	8270E		06/06/25	06/09/25	
			SVOC-SIMGroup1	8270-Modified		06/06/25	06/10/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q2209
Client: G Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID :				0.000				
Total Svoc :					0.00			
Total Concentration:					0.00			



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2209

Client: G Environmental

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168336BL	PB168336BL	2-Methylnaphthalene-d10	0.4	0.36	91		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.40	99		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.37	92		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.40	101		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.42	105		30 (54)	130 (175)
PB168336BS	PB168336BS	2-Methylnaphthalene-d10	0.4	0.36	90		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.30	76		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.36	90		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.38	95		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.38	95		30 (54)	130 (175)
Q2209-01	P01W	2-Methylnaphthalene-d10	0.4	0.38	94		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.47	117		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.39	97		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.43	106		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.45	113		30 (54)	130 (175)
Q2250-02MS	MW-11A-13.5-060525MS	2-Methylnaphthalene-d10	0.4	0.30	75		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.37	92		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.32	79		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.34	86		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.47	118		30 (54)	130 (175)
Q2250-03MSD	MW-11A-13.5-060525MSD	2-Methylnaphthalene-d10	0.4	0.30	75		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.37	91		30 (54)	150 (157)
		Nitrobenzene-d5	0.4	0.32	79		30 (27)	130 (154)
		2-Fluorobiphenyl	0.4	0.35	86		30 (30)	130 (155)
		Terphenyl-d14	0.4	0.45	113		30 (54)	130 (175)

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2209

Client: G Environmental

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2250-02MS		Client Sample ID: MW-11A-13.5-060525MS		DataFile: BN037192.D							
Benzo(a)anthracene	0.42	0	0.45	ug/L	107				70 (51)	130 (146)	
Benzo(b)fluoranthene	0.42	0	0.40	ug/L	95				70 (60)	130 (123)	
Benzo(k)fluoranthene	0.42	0	0.38	ug/L	90				70 (61)	130 (122)	
Benzo(a)pyrene	0.42	0	0.40	ug/L	95				70 (54)	130 (129)	
Benzo(g,h,i)perylene	0.42	0	0.39	ug/L	93				70 (44)	130 (132)	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2209

Client: G Environmental

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2250-03MSD		Client Sample ID: MW-11A-13.5-060525MSD		DataFile: BN037193.D							
Benzo(a)anthracene	0.4	0	0.44	ug/L	110	3			70 (51)	130 (146)	20 (20)
Benzo(b)fluoranthene	0.4	0	0.38	ug/L	95	0			70 (60)	130 (123)	20 (20)
Benzo(k)fluoranthene	0.4	0	0.39	ug/L	98	9			70 (61)	130 (122)	20 (20)
Benzo(a)pyrene	0.4	0	0.38	ug/L	95	0			70 (54)	130 (129)	20 (20)
Benzo(g,h,i)perylene	0.4	0	0.39	ug/L	98	5			70 (44)	130 (132)	20 (20)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2209

Client: G Environmental

Analytical Method: 8270-Modified DataFile: BN037201.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168336BS	Benzo(a)anthracene	0.4	0.36	ug/L	90				70 (54)	130 (130)	
	Benzo(b)fluoranthene	0.4	0.35	ug/L	88				70 (65)	130 (121)	
	Benzo(k)fluoranthene	0.4	0.36	ug/L	90				70 (72)	130 (119)	
	Benzo(a)pyrene	0.4	0.39	ug/L	98				70 (68)	130 (120)	
	Benzo(g,h,i)perylene	0.4	0.42	ug/L	105				70 (76)	130 (117)	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168336BL

Lab Name: CHEMTECH

Contract: GENV01

Lab Code: CHEM Case No.: Q2209

SAS No.: Q2209 SDG NO.: Q2209

Lab File ID: BN037190.D

Lab Sample ID: PB168336BL

Instrument ID: BNA_N

Date Extracted: 06/06/2025

Matrix: (soil/water) Water

Date Analyzed: 06/09/2025

Level: (low/med) LOW

Time Analyzed: 11:30

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168336BS	PB168336BS	BN037201.D	06/09/2025
MW-11A-13.5-060525MS	Q2250-02MS	BN037192.D	06/09/2025
MW-11A-13.5-060525MSD	Q2250-03MSD	BN037193.D	06/09/2025
P01W	Q2209-01	BN037211.D	06/10/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: GENV01Lab Code: CHEMSAS No.: Q2209SDG NO.: Q2209Lab File ID: BN037142.DDFTPP Injection Date: 06/03/2025Instrument ID: BNA_NDFTPP Injection Time: 10:21

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	69.8
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	58.7
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	53.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	24.4
365	Greater than 1% of mass 198	4.5
441	Present, but less than mass 443	10.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	12.1 (19.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037143.D	06/03/2025	11:39
SSTDICC0.2	SSTDICC0.2	BN037144.D	06/03/2025	12:15
SSTDICCC0.4	SSTDICCC0.4	BN037145.D	06/03/2025	12:51
SSTDICC0.8	SSTDICC0.8	BN037146.D	06/03/2025	13:26
SSTDICC1.6	SSTDICC1.6	BN037147.D	06/03/2025	14:02
SSTDICC3.2	SSTDICC3.2	BN037148.D	06/03/2025	14:38
SSTDICC5.0	SSTDICC5.0	BN037149.D	06/03/2025	15:14



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: GENV01Lab Code: CHEMSAS No.: Q2209 SDG NO.: Q2209Lab File ID: BN037188.DDFTPP Injection Date: 06/09/2025Instrument ID: BNA_NDFTPP Injection Time: 10:15

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	74
68	Less than 2.0% of mass 69	0.4 (0.7) 1
69	Mass 69 relative abundance	59.6
70	Less than 2.0% of mass 69	0.4 (0.6) 1
127	10.0 - 80.0% of mass 198	53
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	24.1
365	Greater than 1% of mass 198	4.4
441	Present, but less than mass 443	8.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.5 (18.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037189.D	06/09/2025	10:54
PB168336BL	PB168336BL	BN037190.D	06/09/2025	11:30
MW-11A-13.5-060525MS	Q2250-02MS	BN037192.D	06/09/2025	14:33
MW-11A-13.5-060525MSD	Q2250-03MSD	BN037193.D	06/09/2025	15:47
PB168336BS	PB168336BS	BN037201.D	06/09/2025	20:40



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: GENV01Lab Code: CHEMSAS No.: Q2209 SDG NO.: Q2209Lab File ID: BN037203.DDFTPP Injection Date: 06/09/2025Instrument ID: BNA_NDFTPP Injection Time: 22:32

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	79.8
68	Less than 2.0% of mass 69	1 (1.5) 1
69	Mass 69 relative abundance	65.5
70	Less than 2.0% of mass 69	0.4 (0.6) 1
127	10.0 - 80.0% of mass 198	56.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	24.4
365	Greater than 1% of mass 198	4.3
441	Present, but less than mass 443	8.8
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.6 (20.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037204.D	06/09/2025	23:11
P01W	Q2209-01	BN037211.D	06/10/2025	03:25

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG NO.: Q2209

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/09/2025

Lab File ID: BN037189.D Time Analyzed: 10:54

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2093	7.589	5342	10.36	2894	14.24
UPPER LIMIT	4186	8.089	10684	10.862	5788	14.735
LOWER LIMIT	1046.5	7.089	2671	9.862	1447	13.735
EPA SAMPLE NO.						
01 PB168336BL	1816	7.59	4227	10.37	2101	14.25
02 MW-11A-13.5-060525MS	2144	7.59	5670	10.36	2991	14.23
03 MW-11A-13.5-060525MSD	2169	7.59	5646	10.36	2926	14.24
04 PB168336BS	2227	7.59	5466	10.36	2607	14.23

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG NO.: Q2209

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/09/2025

Lab File ID: BN037189.D Time Analyzed: 10:54

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	5308	16.984	3516	21.18	3185	23.377
UPPER LIMIT	10616	17.484	7032	21.68	6370	23.877
LOWER LIMIT	2654	16.484	1758	20.68	1592.5	22.877
EPA SAMPLE NO.						
01 PB168336BL	3500	17.00	2446	21.19	2291	23.39
02 MW-11A-13.5-060525MS	5389	16.98	3448	21.19	3177	23.38
03 MW-11A-13.5-060525MSD	5139	16.98	3419	21.18	3336	23.37
04 PB168336BS	4253	16.98	2468	21.19	2373	23.38

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG NO.: Q2209

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/09/2025

Lab File ID: BN037204.D Time Analyzed: 23:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1688	7.589	4447	10.37	2457	14.23
UPPER LIMIT	3376	8.089	8894	10.872	4914	14.734
LOWER LIMIT	844	7.089	2223.5	9.872	1228.5	13.734
EPA SAMPLE NO.						
01 P01W	1897	7.59	4879	10.36	2497	14.24

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG NO.: Q2209

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/09/2025

Lab File ID: BN037204.D Time Analyzed: 23:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	4471	16.984	2829	21.18	2684	23.377
UPPER LIMIT	8942	17.484	5658	21.68	5368	23.877
LOWER LIMIT	2235.5	16.484	1414.5	20.68	1342	22.877
EPA SAMPLE NO.						
01 P01W	4818	16.98	4180	21.18	4091	23.37

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	G Environmental	Date Collected:	06/04/25
Project:	Power	Date Received:	06/04/25
Client Sample ID:	P01W	SDG No.:	Q2209
Lab Sample ID:	Q2209-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037211.D	1	06/06/25 11:54	06/10/25 03:25	PB168336

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
56-55-3	Benzo(a)anthracene	0.040	U	0.040	0.10	ug/L
205-99-2	Benzo(b)fluoranthene	0.040	U	0.040	0.10	ug/L
207-08-9	Benzo(k)fluoranthene	0.050	U	0.050	0.10	ug/L
50-32-8	Benzo(a)pyrene	0.040	U	0.040	0.10	ug/L
191-24-2	Benzo(g,h,i)perylene	0.040	U	0.040	0.10	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.38		30 (20) - 150 (139)	94%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.47		30 (54) - 150 (157)	117%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		30 (27) - 130 (154)	97%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.43		30 (30) - 130 (155)	106%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		30 (54) - 130 (175)	113%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	1900	7.59			
1146-65-2	Naphthalene-d8	4880	10.362			
15067-26-2	Acenaphthene-d10	2500	14.235			
1517-22-2	Phenanthrene-d10	4820	16.984			
1719-03-5	Chrysene-d12	4180	21.18			
1520-96-3	Perylene-d12	4090	23.374			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037211.D
 Acq On : 10 Jun 2025 03:25
 Operator : RC/JU
 Sample : Q2209-01
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 P01W

Quant Time: Jun 10 04:06:20 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.590	152	1897	0.400	ng	0.00
7) Naphthalene-d8	10.362	136	4879	0.400	ng	#-0.01
13) Acenaphthene-d10	14.235	164	2497	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	4818	0.400	ng	0.00
29) Chrysene-d12	21.180	240	4180	0.400	ng	# 0.00
35) Perylene-d12	23.374	264	4091	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.185	112	1194	0.255	ng	0.00
5) Phenol-d6	6.766	99	907	0.160	ng	0.00
8) Nitrobenzene-d5	8.739	82	2010	0.390	ng	0.00
11) 2-Methylnaphthalene-d10	11.966	152	2570	0.378	ng	0.00
14) 2,4,6-Tribromophenol	15.730	330	436	0.434	ng	-0.01
15) 2-Fluorobiphenyl	12.858	172	4536	0.426	ng	0.00
27) Fluoranthene-d10	19.022	212	5720	0.467	ng	0.00
31) Terphenyl-d14	19.630	244	4444	0.452	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.108	149	869	0.091	ng	# 94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037211.D
Acq On : 10 Jun 2025 03:25
Operator : RC/JU
Sample : Q2209-01
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
P01W

Quant Time: Jun 10 04:06:20 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:52:03 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.590 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25




Instrument :
BNA_N
ClientSampleId :
P01W


Tgt Ion:152 Resp: 1897
Ion Ratio Lower Upper
152 100
150 148.5 123.2 184.8
115 69.1 56.6 85.0



#4
2-Fluorophenol
Concen: 0.255 ng
RT: 5.185 min Scan# 319
Delta R.T. 0.000 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25

Tgt Ion:112 Resp: 1194
Ion Ratio Lower Upper
112 100
64 68.8 56.3 84.5
63 44.4 36.2 54.4









#5
Phenol-d6
Concen: 0.160 ng
RT: 6.766 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25




Instrument :
BNA_N
ClientSampleId :
P01W


Tgt Ion: 99 Resp: 907
Ion Ratio Lower Upper
99 100
42 68.5 31.3 46.9#
71 57.8 38.2 57.2#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.362 min Scan# 944
Delta R.T. -0.011 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25

Tgt Ion: 136 Resp: 4879
Ion Ratio Lower Upper
136 100
137 12.9 9.7 14.5
54 16.0 9.7 14.5#
68 8.8 5.4 8.2#









#8
Nitrobenzene-d5
Concen: 0.390 ng
RT: 8.739 min Scan# 792
Delta R.T. 0.000 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25




Instrument :
BNA_N
ClientSampleId :
P01W


Tgt Ion: 82 Resp: 2010
Ion Ratio Lower Upper
82 100
128 35.4 26.9 40.3
54 68.4 61.4 92.2



#11
2-Methylnaphthalene-d10
Concen: 0.378 ng
RT: 11.966 min Scan# 1142
Delta R.T. -0.005 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25

Tgt Ion: 152 Resp: 2570
Ion Ratio Lower Upper
152 100
151 22.0 17.1 25.7









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.235 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25




Instrument :
BNA_N
ClientSampleId :
P01W


Tgt Ion:164 Resp: 2497
Ion Ratio Lower Upper
164 100
162 105.9 85.5 128.3
160 54.2 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 0.434 ng
RT: 15.730 min Scan# 1599
Delta R.T. -0.012 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25

Tgt Ion:330 Resp: 436
Ion Ratio Lower Upper
330 100
332 98.6 77.1 115.7
141 68.6 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 0.426 ng
RT: 12.858 min Scan# 11
Delta R.T. -0.005 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25




Instrument :
BNA_N
ClientSampleId :
P01W


Tgt Ion:172 Resp: 4536
Ion Ratio Lower Upper
172 100
171 36.0 29.6 44.4
170 25.3 20.3 30.5



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.984 min Scan# 1700
Delta R.T. 0.000 min
Lab File: BN037211.D
Acq: 10 Jun 2025 03:25

Tgt Ion:188 Resp: 4818
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.7 11.3 16.9





#27

Fluoranthene-d10

Concen: 0.467 ng

RT: 19.022 min Scan# 1933

Delta R.T. -0.005 min

Lab File: BN037211.D

Acq: 10 Jun 2025 03:25

Instrument :

BNA_N

ClientSampleId :

P01W





Tgt Ion:212 Resp: 5720

Ion Ratio Lower Upper

212 100

106 13.8 10.6 15.8

104 7.9 6.6 9.8



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.180 min Scan# 2276

Delta R.T. -0.009 min

Lab File: BN037211.D

Acq: 10 Jun 2025 03:25




Tgt Ion:240 Resp: 4180


Ion Ratio Lower Upper

240 100

120 18.5 9.0 13.4

236 30.4 23.0 34.4









#31
 Terphenyl-d14
 Concen: 0.452 ng
 RT: 19.630 min Scan# 2065
 Delta R.T. -0.005 min
 Lab File: BN037211.D
 Acq: 10 Jun 2025 03:25




Instrument :
 BNA_N
 ClientSampleId :
 P01W


Tgt Ion:244 Resp: 4444
 Ion Ratio Lower Upper
 244 100
 212 12.0 10.0 15.0
 122 17.1 13.2 19.8



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.091 ng
 RT: 21.108 min Scan# 2268
 Delta R.T. -0.009 min
 Lab File: BN037211.D
 Acq: 10 Jun 2025 03:25

Tgt Ion:149 Resp: 869
 Ion Ratio Lower Upper
 149 100
 167 29.1 21.0 31.4
 279 6.3 2.9 4.3#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.374 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN037211.D


Acq: 10 Jun 2025 03:25

Instrument :

BNA_N

ClientSampleId :

P01W





Tgt Ion:264 Resp: 4091

Ion Ratio Lower Upper

264 100

260 26.8 22.1 33.1

265 66.6 55.8 83.8





CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG No.: Q2209
 Instrument ID: BNA_N Calibration Date(s): 06/03/2025 06/03/2025
 Calibration Time(s): 11:39 15:14

LAB FILE ID:		RRF0.1 = BN037143.D		RRF0.2 = BN037144.D		RRF0.4 = BN037145.D		
		RRF0.8 = BN037146.D		RRF1.6 = BN037147.D		RRF3.2 = BN037148.D		
COMPOUND	RRF0.1	RRF0.2	RRF0.4	RRF0.8	RRF1.6	RRF3.2	RRF	% RSD
2-Methylnaphthalene-d10	0.520	0.515	0.562	0.536	0.598	0.577	0.557	6.0
Fluoranthene-d10	0.969	0.937	0.975	0.956	1.092	1.071	1.016	7.2
2-Fluorophenol	1.027	1.017	0.940	0.945	1.036	0.984	0.989	3.9
Phenol-d6	1.156	1.144	1.127	1.126	1.293	1.261	1.199	6.4
Nitrobenzene-d5	0.393	0.383	0.421	0.407	0.455	0.450	0.422	6.9
2-Fluorobiphenyl	1.722	1.691	1.626	1.654	1.814	1.706	1.705	3.5
2,4,6-Tribromophenol	0.124	0.147	0.146	0.157	0.185	0.182	0.161	15.0
Terphenyl-d14	0.964	0.909	0.896	0.941	1.006	0.952	0.942	4.0
Benzo(a)anthracene	1.369	1.367	1.291	1.404	1.582	1.553	1.448	8.1
Benzo(b)fluoranthene	1.529	1.520	1.421	1.575	1.763	1.713	1.615	8.6
Benzo(k)fluoranthene	1.576	1.565	1.461	1.612	1.777	1.743	1.648	7.8
Benzo(a)pyrene	1.310	1.287	1.219	1.294	1.451	1.426	1.352	7.3
Benzo(g,h,i)perylene	1.368	1.450	1.351	1.372	1.477	1.424	1.410	3.3

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN060325.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jun 04 01:52:03 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037143.D 0.2 =BN037144.D 0.4 =BN037145.D 0.8 =BN037146.D 1.6 =BN037147.D 3.2 =BN037148.D 5.0 =BN037149.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.598	0.657	0.510	0.506	0.526	0.477	0.458	0.533	13.16
3)	n-Nitrosodimet...	1.098	1.031	1.061	1.067	1.163	1.061	1.012	1.071	4.60
4) S	2-Fluorophenol	1.027	1.017	0.940	0.945	1.036	0.984	0.975	0.989	3.91
5) S	Phenol-d6	1.156	1.144	1.127	1.126	1.293	1.261	1.285	1.199	6.42
6)	bis(2-Chloroet...	1.138	1.139	1.128	1.089	1.223	1.146	1.146	1.144	3.51
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.393	0.383	0.421	0.407	0.455	0.450	0.446	0.422	6.86
9)	Naphthalene	1.183	1.125	1.119	1.111	1.215	1.165	1.160	1.154	3.31
10)	Hexachlorobuta...	0.253	0.249	0.261	0.247	0.266	0.246	0.238	0.251	3.81
11) SURR	2-Methylnaphth...	0.520	0.515	0.562	0.536	0.598	0.577	0.588	0.557	5.97
12)	2-Methylnaphth...	0.704	0.680	0.691	0.719	0.809	0.783	0.793	0.740	7.22
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.124	0.147	0.146	0.157	0.185	0.182	0.186	0.161	15.03
15) S	2-Fluorobiphenyl	1.722	1.691	1.626	1.654	1.814	1.706	1.725	1.705	3.52
16)	Acenaphthylene	1.946	1.905	1.768	1.871	2.112	2.050	2.075	1.961	6.32
17)	Acenaphthene	1.290	1.253	1.159	1.212	1.370	1.309	1.320	1.273	5.59
18)	Fluorene	1.701	1.577	1.518	1.611	1.823	1.736	1.752	1.674	6.48
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.039	0.050	0.067	0.090	0.102	0.114	0.077		38.58
21)	4-Bromophenyl-...	0.256	0.253	0.244	0.254	0.281	0.276	0.271	0.262	5.32
22)	Hexachlorobenzene	0.289	0.284	0.269	0.279	0.301	0.284	0.274	0.283	3.72
23)	Atrazine	0.194	0.200	0.187	0.209	0.241	0.238	0.247	0.216	11.42
24)	Pentachlorophenol	0.086	0.086	0.092	0.107	0.140	0.153	0.165	0.124	26.72
25)	Phenanthrene	1.285	1.242	1.193	1.248	1.386	1.357	1.361	1.296	5.64
26)	Anthracene	1.098	1.099	1.036	1.143	1.294	1.290	1.317	1.183	9.71
27) SURR	Fluoranthene-d10	0.969	0.937	0.975	0.956	1.092	1.071	1.114	1.016	7.22
28)	Fluoranthene	1.339	1.294	1.277	1.365	1.579	1.563	1.605	1.432	10.09
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	2.051	1.974	1.827	1.928	2.048	1.955	1.885	1.953	4.20
31) S	Terphenyl-d14	0.964	0.909	0.896	0.941	1.006	0.952	0.923	0.942	3.96
32)	Benzo(a)anthra...	1.369	1.367	1.291	1.404	1.582	1.553	1.570	1.448	8.15
33)	Chrysene	1.755	1.636	1.473	1.582	1.698	1.584	1.556	1.612	5.81
34)	Bis(2-ethylhex...	1.032	0.859	0.774	0.858	0.956	0.914	1.002	0.914	9.90
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\


Method File : 8270-SIM-BN060325.M

36)	Indeno(1,2,3-c...	1.443	1.605	1.501	1.526	1.695	1.673	1.697	1.591	6.44
37)	Benzo(b)fluora...	1.529	1.520	1.421	1.575	1.763	1.713	1.781	1.615	8.58
38)	Benzo(k)fluora...	1.576	1.565	1.461	1.612	1.777	1.743	1.805	1.648	7.79
39) C	Benzo(a)pyrene	1.310	1.287	1.219	1.294	1.451	1.426	1.481	1.352	7.32
40)	Dibenzo(a,h)an...	1.074	1.167	1.160	1.196	1.333	1.332	1.328	1.227	8.48
41)	Benzo(g,h,i)pe...	1.368	1.450	1.351	1.372	1.477	1.424	1.425	1.410	3.33

(#) = Out of Range

Pentachlorophenol

Response Ratio



$$\text{Response} = 1.694\text{e-}001 * \text{Amt} - 9.118\text{e-}002$$


Coef of Det (r^2) = 0.997707 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA N\Methods\8270-SIM-BN060325.M

Calibration Table Last Updated: Wed Jun 04 01:52:03 2025

4,6-Dinitro-2-methylphenol

Response Ratio



$$\text{Response} = 1.172\text{e-}001 * \text{Amt} - 8.037\text{e-}002$$

Coef of Det (r^2) = 0.995947 Curve Fit: Linear

Method Name: Z:\svoasrv\HPCHEM1\BNA N\Methods\8270-SIM-BN060325.M

Calibration Table Last Updated: Wed Jun 04 01:52:03 2025

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037143.D
 Acq On : 03 Jun 2025 11:39
 Operator : RC/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Jun 04 01:41:16 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:40:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.590	152	2014	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	4942	0.400	ng	0.00
13) Acenaphthene-d10	14.235	164	2688	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	5431	0.400	ng	0.00
29) Chrysene-d12	21.189	240	3598	0.400	ng	# 0.00
35) Perylene-d12	23.380	264	3208	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	517	0.104	ng	0.00
5) Phenol-d6	6.766	99	582	0.096	ng	0.00
8) Nitrobenzene-d5	8.739	82	486	0.093	ng	0.00
11) 2-Methylnaphthalene-d10	11.966	152	643	0.093	ng	0.00
14) 2,4,6-Tribromophenol	15.743	330	83	0.077	ng	0.00
15) 2-Fluorobiphenyl	12.858	172	1157	0.101	ng	0.00
27) Fluoranthene-d10	19.026	212	1316	0.095	ng	0.00
31) Terphenyl-d14	19.635	244	867	0.102	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.119	88	301	0.112	ng	97
3) n-Nitrosodimethylamine	3.444	42	553	0.103	ng	95
6) bis(2-Chloroethyl)ether	7.019	93	573	0.099	ng	95
9) Naphthalene	10.415	128	1462	0.103	ng	# 93
10) Hexachlorobutadiene	10.703	225	312	0.100	ng	# 98
12) 2-Methylnaphthalene	12.042	142	870	0.095	ng	97
16) Acenaphthylene	13.957	152	1308	0.099	ng	100
17) Acenaphthene	14.299	154	867	0.101	ng	99
18) Fluorene	15.293	166	1143	0.102	ng	99
21) 4-Bromophenyl-phenylether	16.190	248	347	0.097	ng	96
22) Hexachlorobenzene	16.301	284	392	0.102	ng	99
23) Atrazine	16.463	200	264	0.090	ng	# 91
25) Phenanthrene	17.034	178	1745	0.099	ng	98
26) Anthracene	17.120	178	1491	0.093	ng	98
28) Fluoranthene	19.054	202	1818	0.094	ng	99
30) Pyrene	19.421	202	1845	0.105	ng	99
32) Benzo(a)anthracene	21.171	228	1231	0.095	ng	93
33) Chrysene	21.225	228	1579	0.109	ng	97
34) Bis(2-ethylhexyl)phtha...	21.117	149	928	0.113	ng	97
36) Indeno(1,2,3-cd)pyrene	25.576	276	1157	0.091	ng	97
37) Benzo(b)fluoranthene	22.725	252	1226	0.095	ng	# 65
38) Benzo(k)fluoranthene	22.766	252	1264	0.096	ng	# 64
39) Benzo(a)pyrene	23.284	252	1051	0.097	ng	# 52
40) Dibenzo(a,h)anthracene	25.594	278	861	0.087	ng	# 56
41) Benzo(g,h,i)perylene	26.248	276	1097	0.097	ng	# 82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037143.D
Acq On : 03 Jun 2025 11:39
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Jun 04 01:41:16 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:40:18 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.590 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2014
Ion Ratio Lower Upper
152 100
150 148.3 123.2 184.8
115 66.9 56.6 85.0



#2
1,4-Dioxane
Concen: 0.112 ng
RT: 3.119 min Scan# 33
Delta R.T. -0.007 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion: 88 Resp: 301
Ion Ratio Lower Upper
88 100
43 58.8 43.5 65.3
58 85.4 67.7 101.5









#3
n-Nitrosodimethylamine
Concen: 0.103 ng
RT: 3.444 min Scan# 71
Delta R.T. 0.007 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 553
Ion Ratio Lower Upper
42 100
74 70.2 53.0 79.4
44 6.9 5.9 8.9



#4
2-Fluorophenol
Concen: 0.104 ng
RT: 5.184 min Scan# 319
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion: 112 Resp: 517
Ion Ratio Lower Upper
112 100
64 70.8 56.3 84.5
63 39.7 36.2 54.4









#5
Phenol-d6
Concen: 0.096 ng
RT: 6.766 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	Ratio	Lower	Upper
99	100		
42	37.3	31.3	46.9
71	49.7	38.2	57.2



#6
bis(2-Chloroethyl)ether
Concen: 0.099 ng
RT: 7.019 min Scan# 573
Delta R.T. -0.007 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion	Ratio	Lower	Upper
93	100		
63	91.4	68.6	103.0
95	31.6	24.3	36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	4942
Ion	Ratio	Lower	Upper
136	100		
137	12.3	9.7	14.5
54	10.3	9.7	14.5
68	6.7	5.4	8.2



#8
Nitrobenzene-d5
Concen: 0.093 ng
RT: 8.739 min Scan# 792
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion:	82	Resp:	486
Ion	Ratio	Lower	Upper
82	100		
128	45.6	26.9	40.3#
54	79.2	61.4	92.2









#9
Naphthalene
Concen: 0.103 ng
RT: 10.415 min Scan# 949
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 1462
Ion Ratio Lower Upper
128 100
129 15.3 9.8 14.8#
127 18.4 12.3 18.5



#10
Hexachlorobutadiene
Concen: 0.100 ng
RT: 10.703 min Scan# 976
Delta R.T. -0.011 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion:225 Resp: 312
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.093 ng
RT: 11.966 min Scan# 1142
Delta R.T. -0.005 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 643
Ion Ratio Lower Upper
152 100
151 23.0 17.1 25.7



#12
2-Methylnaphthalene
Concen: 0.095 ng
RT: 12.042 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion:142 Resp: 870
Ion Ratio Lower Upper
142 100
141 91.6 74.6 111.8
115 55.6 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.235 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 2688
Ion Ratio Lower Upper
164 100
162 108.7 85.5 128.3
160 57.9 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 0.077 ng
RT: 15.743 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion:330 Resp: 83
Ion Ratio Lower Upper
330 100
332 97.6 77.1 115.7
141 56.6 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 0.101 ng
RT: 12.858 min Scan# 1319
Delta R.T. -0.005 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion	Ratio	Lower	Upper
172	100		
171	40.6	29.6	44.4
170	28.7	20.3	30.5



#16
Acenaphthylene
Concen: 0.099 ng
RT: 13.957 min Scan# 1436
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.3	16.3	24.5
153	13.1	10.6	15.8





#17

Acenaphthene

Concen: 0.101 ng

RT: 14.299 min Scan# 1468

Delta R.T. 0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:154 Resp: 867

Ion Ratio Lower Upper

154 100

153 115.6 93.8 140.8

152 64.1 50.5 75.7



#18

Fluorene

Concen: 0.102 ng

RT: 15.293 min Scan# 1561

Delta R.T. 0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39




Tgt Ion:166 Resp: 1143


Ion Ratio Lower Upper

166 100

165 102.8 81.1 121.7

167 13.3 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 1700

Delta R.T. 0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:188 Resp: 5431

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.2 11.3 16.9



#21

4-Bromophenyl-phenylether

Concen: 0.097 ng

RT: 16.190 min Scan# 1636

Delta R.T. 0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39




Tgt Ion:248 Resp: 347


Ion Ratio Lower Upper

248 100

250 97.2 76.1 114.1

141 69.0 60.1 90.1





#22

Hexachlorobenzene

Concen: 0.102 ng

RT: 16.301 min Scan# 1645

Delta R.T. -0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:284 Resp: 392

Ion Ratio Lower Upper

284 100

142 54.1 44.0 66.0

249 37.5 29.7 44.5



#23

Atrazine

Concen: 0.090 ng

RT: 16.463 min Scan# 1658

Delta R.T. 0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39




Tgt Ion:200 Resp: 264


Ion Ratio Lower Upper

200 100

173 43.7 28.1 42.1#

215 52.9 39.3 58.9









#25
Phenanthrene
Concen: 0.099 ng
RT: 17.034 min Scan# 1704
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	Ratio	Lower	Upper
178	100		
176	21.3	15.7	23.5
179	15.6	12.3	18.5



#26
Anthracene
Concen: 0.093 ng
RT: 17.120 min Scan# 1711
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.0	15.2	22.8
179	15.8	12.9	19.3





#27

Fluoranthene-d10

Concen: 0.095 ng

RT: 19.026 min Scan# 1934

Delta R.T. 0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:212 Resp: 1316

Ion Ratio Lower Upper

212 100

106 13.6 10.6 15.8

104 8.4 6.6 9.8



#28

Fluoranthene

Concen: 0.094 ng

RT: 19.054 min Scan# 1940

Delta R.T. 0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39




Tgt Ion:202 Resp: 1818


Ion Ratio Lower Upper

202 100

101 11.5 8.7 13.1

203 17.4 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.189 min Scan# 2118
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:240 Resp: 3598
Ion Ratio Lower Upper
240 100
120 14.4 9.0 13.4#
236 29.8 23.0 34.4



#30
Pyrene
Concen: 0.105 ng
RT: 19.421 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion:202 Resp: 1845
Ion Ratio Lower Upper
202 100
200 21.6 17.0 25.6
203 17.5 14.2 21.4









#31
Terphenyl-d14
Concen: 0.102 ng
RT: 19.635 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	Ratio	Lower	Upper
244	100		
212	17.9	10.0	15.0#
122	22.3	13.2	19.8#



#32
Benzo(a)anthracene
Concen: 0.095 ng
RT: 21.171 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.9	22.6	33.8
229	23.2	16.2	24.2





#33

Chrysene

Concen: 0.109 ng

RT: 21.225 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:228 Resp: 1579

Ion Ratio Lower Upper

228 100

226 33.4 25.2 37.8

229 22.1 16.8 25.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.113 ng

RT: 21.117 min Scan# 2269

Delta R.T. 0.000 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39




Tgt Ion:149 Resp: 928


Ion Ratio Lower Upper

149 100

167 27.6 21.0 31.4

279 4.0 2.9 4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.380 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:264 Resp: 3208
Ion Ratio Lower Upper
264 100
260 27.5 22.1 33.1
265 74.8 55.8 83.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.091 ng
RT: 25.576 min Scan# 3497
Delta R.T. 0.003 min
Lab File: BN037143.D
Acq: 03 Jun 2025 11:39

Tgt Ion:276 Resp: 1157
Ion Ratio Lower Upper
276 100
138 23.5 21.0 31.6
277 24.3 19.4 29.2





#37

Benzo(b)fluoranthene

Concen: 0.095 ng

RT: 22.725 min Scan# 2521

Delta R.T. 0.003 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:252 Resp: 1226

Ion Ratio Lower Upper

252 100

253 44.9 22.3 33.5#

125 33.0 13.2 19.8#



#38

Benzo(k)fluoranthene

Concen: 0.096 ng

RT: 22.766 min Scan# 2536

Delta R.T. 0.003 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39




Tgt Ion:252 Resp: 1264


Ion Ratio Lower Upper

252 100

253 45.6 22.2 33.4#

125 33.9 13.2 19.8#





#39

Benzo(a)pyrene

Concen: 0.097 ng

RT: 23.284 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:252 Resp: 1051

Ion Ratio Lower Upper

252 100

253 54.8 25.0 37.4#

125 46.9 17.0 25.6#



#40

Dibenzo(a,h)anthracene

Concen: 0.087 ng

RT: 25.594 min Scan# 3503

Delta R.T. 0.006 min

Lab File: BN037143.D

Acq: 03 Jun 2025 11:39




Tgt Ion:278 Resp: 861


Ion Ratio Lower Upper

278 100

139 35.4 17.6 26.4#

279 63.2 26.0 39.0#





#41

Benzo(g,h,i)perylene

Concen: 0.097 ng

RT: 26.248 min Scan# 31

Delta R.T. 0.009 min

Lab File: BN037143.D


Acq: 03 Jun 2025 11:39

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:276 Resp: 1097

Ion Ratio Lower Upper

276 100

277 35.2 20.9 31.3#

138 35.4 20.8 31.2#



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037144.D
 Acq On : 03 Jun 2025 12:15
 Operator : RC/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Jun 04 01:41:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:40:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	2031	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	5053	0.400	ng	0.00
13) Acenaphthene-d10	14.234	164	2624	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	4950	0.400	ng	0.00
29) Chrysene-d12	21.188	240	3306	0.400	ng	0.00
35) Perylene-d12	23.374	264	3099	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	1033	0.206	ng	0.00
5) Phenol-d6	6.773	99	1162	0.191	ng	0.00
8) Nitrobenzene-d5	8.739	82	967	0.181	ng	0.00
11) 2-Methylnaphthalene-d10	11.970	152	1302	0.185	ng	0.00
14) 2,4,6-Tribromophenol	15.742	330	193	0.183	ng	0.00
15) 2-Fluorobiphenyl	12.863	172	2219	0.198	ng	0.00
27) Fluoranthene-d10	19.026	212	2318	0.184	ng	0.00
31) Terphenyl-d14	19.635	244	1502	0.193	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.126	88	667	0.246	ng	86
3) n-Nitrosodimethylamine	3.444	42	1047	0.193	ng	95
6) bis(2-Chloroethyl)ether	7.026	93	1157	0.199	ng	97
9) Naphthalene	10.415	128	2843	0.195	ng	97
10) Hexachlorobutadiene	10.714	225	630	0.198	ng	# 95
12) 2-Methylnaphthalene	12.046	142	1717	0.184	ng	98
16) Acenaphthylene	13.956	152	2500	0.194	ng	99
17) Acenaphthene	14.299	154	1644	0.197	ng	99
18) Fluorene	15.293	166	2069	0.188	ng	99
20) 4,6-Dinitro-2-methylph...	15.400	198	97	0.341	ng	# 54
21) 4-Bromophenyl-phenylether	16.189	248	626	0.193	ng	97
22) Hexachlorobenzene	16.301	284	703	0.201	ng	98
23) Atrazine	16.462	200	494	0.184	ng	94
24) Pentachlorophenol	16.673	266	214	0.317	ng	97
25) Phenanthrene	17.033	178	3073	0.192	ng	99
26) Anthracene	17.120	178	2721	0.186	ng	98
28) Fluoranthene	19.054	202	3203	0.181	ng	99
30) Pyrene	19.421	202	3263	0.202	ng	100
32) Benzo(a)anthracene	21.171	228	2259	0.189	ng	95
33) Chrysene	21.215	228	2704	0.203	ng	97
34) Bis(2-ethylhexyl)phtha...	21.108	149	1420	0.188	ng	98
36) Indeno(1,2,3-cd)pyrene	25.570	276	2487	0.202	ng	100
37) Benzo(b)fluoranthene	22.719	252	2355	0.188	ng	# 87
38) Benzo(k)fluoranthene	22.763	252	2425	0.190	ng	# 86
39) Benzo(a)pyrene	23.281	252	1994	0.190	ng	# 81
40) Dibenzo(a,h)anthracene	25.587	278	1809	0.190	ng	# 85
41) Benzo(g,h,i)perylene	26.239	276	2247	0.206	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037144.D
Acq On : 03 Jun 2025 12:15
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Quant Time: Jun 04 01:41:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:40:18 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.589 min Scan# 61
Delta R.T. -0.001 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2







Tgt Ion:152 Resp: 2031
Ion Ratio Lower Upper
152 100
150 154.2 123.2 184.8
115 71.5 56.6 85.0



#2
1,4-Dioxane
Concen: 0.246 ng
RT: 3.126 min Scan# 34
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion: 88 Resp: 667
Ion Ratio Lower Upper
88 100
43 46.2 43.5 65.3
58 70.9 67.7 101.5









#3
n-Nitrosodimethylamine
Concen: 0.193 ng
RT: 3.444 min Scan# 71
Delta R.T. 0.007 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 1047
Ion Ratio Lower Upper
42 100
74 70.1 53.0 79.4
44 7.7 5.9 8.9



#4
2-Fluorophenol
Concen: 0.206 ng
RT: 5.184 min Scan# 319
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion: 112 Resp: 1033
Ion Ratio Lower Upper
112 100
64 69.4 56.3 84.5
63 44.3 36.2 54.4









#5
Phenol-d6
Concen: 0.191 ng
RT: 6.773 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
99	100		
42	38.3	31.3	46.9
71	47.2	38.2	57.2



#6
bis(2-Chloroethyl)ether
Concen: 0.199 ng
RT: 7.026 min Scan# 574
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion	Ratio	Lower	Upper
93	100		
63	82.2	68.6	103.0
95	30.9	24.3	36.5





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.372 min Scan# 945

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2

Tgt Ion:136 Resp: 5053





Ion Ratio Lower Upper

136 100

137 12.1 9.7 14.5

54 11.9 9.7 14.5

68 6.9 5.4 8.2



#8

Nitrobenzene-d5

Concen: 0.181 ng

RT: 8.739 min Scan# 792

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15




Tgt Ion: 82 Resp: 967


Ion Ratio Lower Upper

82 100

128 38.5 26.9 40.3

54 81.9 61.4 92.2





#9
Naphthalene

Concen: 0.195 ng

RT: 10.415 min Scan# 949

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:128 Resp: 2843

Ion	Ratio	Lower	Upper
-----	-------	-------	-------

128	100		
-----	-----	--	--

129	13.5	9.8	14.8
-----	------	-----	------

127	16.6	12.3	18.5
-----	------	------	------



#10

Hexachlorobutadiene

Concen: 0.198 ng

RT: 10.714 min Scan# 977

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15




Tgt Ion:225 Resp: 630


Ion	Ratio	Lower	Upper
-----	-------	-------	-------

225	100		
-----	-----	--	--

223	0.0	0.0	0.0
-----	-----	-----	-----

227	67.0	50.3	75.5
-----	------	------	------









#11
2-Methylnaphthalene-d10
Concen: 0.185 ng
RT: 11.970 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 1302
Ion Ratio Lower Upper
152 100
151 22.8 17.1 25.7



#12
2-Methylnaphthalene
Concen: 0.184 ng
RT: 12.046 min Scan# 1158
Delta R.T. 0.005 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion:142 Resp: 1717
Ion Ratio Lower Upper
142 100
141 92.1 74.6 111.8
115 53.4 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.234 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:164 Resp: 2624
Ion Ratio Lower Upper
164 100
162 107.7 85.5 128.3
160 57.0 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 0.183 ng
RT: 15.742 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion:330 Resp: 193
Ion Ratio Lower Upper
330 100
332 99.5 77.1 115.7
141 52.3 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 0.198 ng
RT: 12.863 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	172	Resp:	2219
Ion	Ratio	Lower	Upper
172	100		
171	37.4	29.6	44.4
170	27.2	20.3	30.5



#16
Acenaphthylene
Concen: 0.194 ng
RT: 13.956 min Scan# 1436
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion:	152	Resp:	2500
Ion	Ratio	Lower	Upper
152	100		
151	20.4	16.3	24.5
153	12.6	10.6	15.8









#17
Acenaphthene
Concen: 0.197 ng
RT: 14.299 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:154 Resp: 1644
Ion Ratio Lower Upper
154 100
153 117.0 93.8 140.8
152 65.3 50.5 75.7



#18
Fluorene
Concen: 0.188 ng
RT: 15.293 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion:166 Resp: 2069
Ion Ratio Lower Upper
166 100
165 102.4 81.1 121.7
167 13.8 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:188 Resp: 4950

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.1 11.3 16.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.341 ng

RT: 15.400 min Scan# 1571

Delta R.T. 0.011 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15




Tgt Ion:198 Resp: 97


Ion Ratio Lower Upper

198 100

51 226.6 125.2 187.8#

105 96.2 57.1 85.7#









#21
4-Bromophenyl-phenylether
Concen: 0.193 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
248	100		
250	97.7	76.1	114.1
141	78.0	60.1	90.1



#22
Hexachlorobenzene
Concen: 0.201 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion	Ratio	Lower	Upper
284	100		
142	57.2	44.0	66.0
249	36.7	29.7	44.5









#23
Atrazine
Concen: 0.184 ng
RT: 16.462 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:200 Resp: 494
Ion Ratio Lower Upper
200 100
173 41.3 28.1 42.1
215 50.7 39.3 58.9



#24
Pentachlorophenol
Concen: 0.317 ng
RT: 16.673 min Scan# 1675
Delta R.T. 0.012 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion:266 Resp: 214
Ion Ratio Lower Upper
266 100
264 57.9 49.3 73.9
268 62.1 49.0 73.4









#25
Phenanthrene
Concen: 0.192 ng
RT: 17.033 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:178 Resp: 3073
Ion Ratio Lower Upper
178 100
176 20.0 15.7 23.5
179 16.0 12.3 18.5



#26
Anthracene
Concen: 0.186 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion:178 Resp: 2721
Ion Ratio Lower Upper
178 100
176 18.9 15.2 22.8
179 14.5 12.9 19.3





#27

Fluoranthene-d10

Concen: 0.184 ng

RT: 19.026 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:212 Resp: 2318

Ion Ratio Lower Upper

212 100

106 13.2 10.6 15.8

104 8.7 6.6 9.8



#28

Fluoranthene

Concen: 0.181 ng

RT: 19.054 min Scan# 1940

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15




Tgt Ion:202 Resp: 3203


Ion Ratio Lower Upper

202 100

101 10.2 8.7 13.1

203 17.0 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.188 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:240 Resp: 3306

Ion Ratio Lower Upper

240 100

120 12.5 9.0 13.4

236 29.5 23.0 34.4



#30

Pyrene

Concen: 0.202 ng

RT: 19.421 min Scan# 2019

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15




Tgt Ion:202 Resp: 3263


Ion Ratio Lower Upper

202 100

200 21.5 17.0 25.6

203 17.7 14.2 21.4









#31
Terphenyl-d14
Concen: 0.193 ng
RT: 19.635 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
244	100		
212	14.8	10.0	15.0
122	19.2	13.2	19.8



#32
Benzo(a)anthracene
Concen: 0.189 ng
RT: 21.171 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BN037144.D
Acq: 03 Jun 2025 12:15

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.1	22.6	33.8
229	22.5	16.2	24.2





#33

Chrysene

Concen: 0.203 ng

RT: 21.215 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:228 Resp: 2704

Ion Ratio Lower Upper

228 100

226 33.4 25.2 37.8

229 21.6 16.8 25.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.188 ng

RT: 21.108 min Scan# 2268

Delta R.T. -0.009 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15




Tgt Ion:149 Resp: 1420


Ion Ratio Lower Upper

149 100

167 27.5 21.0 31.4

279 4.1 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.374 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:264 Resp: 3099

Ion Ratio Lower Upper

264 100

260 27.7 22.1 33.1

265 71.8 55.8 83.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.202 ng

RT: 25.570 min Scan# 3495

Delta R.T. -0.003 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15




Tgt Ion:276 Resp: 2487


Ion Ratio Lower Upper

276 100

138 26.3 21.0 31.6

277 24.6 19.4 29.2





#37

Benzo(b)fluoranthene

Concen: 0.188 ng

RT: 22.719 min Scan# 2520

Delta R.T. -0.003 min

Lab File: BN037144.D


Acq: 03 Jun 2025 12:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 2355

Ion Ratio Lower Upper

252 100

253 34.4 22.3 33.5#

125 22.8 13.2 19.8#



#38

Benzo(k)fluoranthene

Concen: 0.190 ng

RT: 22.763 min Scan# 2535

Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15




Tgt Ion:252 Resp: 2425


Ion Ratio Lower Upper

252 100

253 35.0 22.2 33.4#

125 22.2 13.2 19.8#





#39

Benzo(a)pyrene

Concen: 0.190 ng

RT: 23.281 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037144.D


Acq: 03 Jun 2025 12:15

Instrument :




BNA_N

ClientSampleId :

SSTDICC0.2



Tgt Ion:	252	Resp:	1994
Ion	Ratio	Lower	Upper
252	100		
253	39.9	25.0	37.4#
125	32.6	17.0	25.6#



#40

Dibenzo(a,h)anthracene

Concen: 0.190 ng




RT: 25.587 min Scan# 3501


Delta R.T. -0.000 min

Lab File: BN037144.D

Acq: 03 Jun 2025 12:15

Tgt Ion:	278	Resp:	1809
Ion	Ratio	Lower	Upper
278	100		
139	25.9	17.6	26.4
279	43.9	26.0	39.0#





#41

Benzo(g,h,i)perylene

Concen: 0.206 ng

RT: 26.239 min Scan# 31

Delta R.T. -0.000 min

Lab File: BN037144.D


Acq: 03 Jun 2025 12:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:276 Resp: 2247

Ion Ratio Lower Upper

276 100

277 29.6 20.9 31.3

138 32.2 20.8 31.2#



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037145.D
 Acq On : 03 Jun 2025 12:51
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Jun 04 01:42:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:40:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.590	152	2099	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	5221	0.400	ng	0.00
13) Acenaphthene-d10	14.235	164	2886	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	5446	0.400	ng	0.00
29) Chrysene-d12	21.189	240	3769	0.400	ng	0.00
35) Perylene-d12	23.377	264	3386	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	1973	0.380	ng	0.00
5) Phenol-d6	6.773	99	2365	0.376	ng	0.00
8) Nitrobenzene-d5	8.739	82	2197	0.399	ng	0.00
11) 2-Methylnaphthalene-d10	11.971	152	2936	0.404	ng	0.00
14) 2,4,6-Tribromophenol	15.743	330	421	0.362	ng	0.00
15) 2-Fluorobiphenyl	12.863	172	4694	0.381	ng	0.00
27) Fluoranthene-d10	19.026	212	5309	0.384	ng	0.00
31) Terphenyl-d14	19.635	244	3376	0.381	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.126	88	1070	0.382	ng	100
3) n-Nitrosodimethylamine	3.437	42	2228	0.397	ng	100
6) bis(2-Chloroethyl)ether	7.026	93	2367	0.394	ng	100
9) Naphthalene	10.415	128	5840	0.388	ng	100
10) Hexachlorobutadiene	10.714	225	1363	0.415	ng	# 100
12) 2-Methylnaphthalene	12.042	142	3609	0.374	ng	100
16) Acenaphthylene	13.957	152	5103	0.361	ng	100
17) Acenaphthene	14.299	154	3344	0.364	ng	100
18) Fluorene	15.293	166	4380	0.363	ng	100
20) 4,6-Dinitro-2-methylph...	15.389	198	272	0.445	ng	100
21) 4-Bromophenyl-phenylether	16.190	248	1329	0.372	ng	100
22) Hexachlorobenzene	16.301	284	1463	0.380	ng	100
23) Atrazine	16.463	200	1018	0.345	ng	100
24) Pentachlorophenol	16.661	266	502	0.433	ng	100
25) Phenanthrene	17.034	178	6496	0.368	ng	100
26) Anthracene	17.120	178	5644	0.351	ng	100
28) Fluoranthene	19.054	202	6952	0.357	ng	100
30) Pyrene	19.421	202	6885	0.374	ng	100
32) Benzo(a)anthracene	21.171	228	4866	0.357	ng	100
33) Chrysene	21.225	228	5553	0.366	ng	100
34) Bis(2-ethylhexyl)phtha...	21.117	149	2918	0.339	ng	100
36) Indeno(1,2,3-cd)pyrene	25.573	276	5083	0.377	ng	100
37) Benzo(b)fluoranthene	22.722	252	4811	0.352	ng	100
38) Benzo(k)fluoranthene	22.763	252	4946	0.354	ng	100
39) Benzo(a)pyrene	23.281	252	4127	0.360	ng	100
40) Dibenzo(a,h)anthracene	25.588	278	3928	0.378	ng	100
41) Benzo(g,h,i)perylene	26.240	276	4576	0.383	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037145.D
Acq On : 03 Jun 2025 12:51
Operator : RC/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Jun 04 01:42:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:40:18 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.590 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 2099
Ion Ratio Lower Upper
152 100
150 154.0 123.2 184.8
115 70.8 56.6 85.0



#2
1,4-Dioxane
Concen: 0.382 ng
RT: 3.126 min Scan# 34
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion: 88 Resp: 1070
Ion Ratio Lower Upper
88 100
43 54.3 43.5 65.3
58 84.5 67.7 101.5









#3
n-Nitrosodimethylamine
Concen: 0.397 ng
RT: 3.437 min Scan# 71
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 2228
Ion Ratio Lower Upper
42 100
74 66.2 53.0 79.4
44 7.4 5.9 8.9



#4
2-Fluorophenol
Concen: 0.380 ng
RT: 5.184 min Scan# 319
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion: 112 Resp: 1973
Ion Ratio Lower Upper
112 100
64 70.4 56.3 84.5
63 45.3 36.2 54.4









#5
Phenol-d6
Concen: 0.376 ng
RT: 6.773 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	99	Resp:	2365
Ion	Ratio	Lower	Upper
99	100		
42	39.1	31.3	46.9
71	47.7	38.2	57.2



#6
bis(2-Chloroethyl)ether
Concen: 0.394 ng
RT: 7.026 min Scan# 574
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion:	93	Resp:	2367
Ion	Ratio	Lower	Upper
93	100		
63	85.5	68.6	103.0
95	30.3	24.3	36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	5221
Ion	Ratio	Lower	Upper
136	100		
137	12.1	9.7	14.5
54	12.1	9.7	14.5
68	6.8	5.4	8.2



#8
Nitrobenzene-d5
Concen: 0.399 ng
RT: 8.739 min Scan# 792
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion:	82	Resp:	2197
Ion	Ratio	Lower	Upper
82	100		
128	33.6	26.9	40.3
54	76.8	61.4	92.2









#9
Naphthalene
Concen: 0.388 ng
RT: 10.415 min Scan# 949
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:128 Resp: 5840
Ion Ratio Lower Upper
128 100
129 12.3 9.8 14.8
127 15.4 12.3 18.5



#10
Hexachlorobutadiene
Concen: 0.415 ng
RT: 10.714 min Scan# 977
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion:225 Resp: 1363
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.9 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.404 ng
RT: 11.971 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 2936
Ion Ratio Lower Upper
152 100
151 21.4 17.1 25.7



#12
2-Methylnaphthalene
Concen: 0.374 ng
RT: 12.042 min Scan# 1157
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion:142 Resp: 3609
Ion Ratio Lower Upper
142 100
141 93.2 74.6 111.8
115 51.2 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.235 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 2886
Ion Ratio Lower Upper
164 100
162 106.9 85.5 128.3
160 55.8 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 0.362 ng
RT: 15.743 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion:330 Resp: 421
Ion Ratio Lower Upper
330 100
332 96.7 77.1 115.7
141 58.0 46.4 69.6





#15

2-Fluorobiphenyl

Concen: 0.381 ng

RT: 12.863 min Scan# 1319

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:172 Resp: 4694

Ion Ratio Lower Upper

172 100

171 37.0 29.6 44.4

170 25.4 20.3 30.5



#16

Acenaphthylene

Concen: 0.361 ng

RT: 13.957 min Scan# 1436

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51




Tgt Ion:152 Resp: 5103


Ion Ratio Lower Upper

152 100

151 20.4 16.3 24.5

153 13.2 10.6 15.8









#17
Acenaphthene
Concen: 0.364 ng
RT: 14.299 min Scan# 1468
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 3344
Ion Ratio Lower Upper
154 100
153 117.3 93.8 140.8
152 63.1 50.5 75.7



#18
Fluorene
Concen: 0.363 ng
RT: 15.293 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion:166 Resp: 4380
Ion Ratio Lower Upper
166 100
165 101.4 81.1 121.7
167 13.5 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 11

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:188 Resp: 5446

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.1 11.3 16.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.445 ng

RT: 15.389 min Scan# 1570

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51




Tgt Ion:198 Resp: 272


Ion Ratio Lower Upper

198 100

51 156.5 125.2 187.8

105 71.4 57.1 85.7









#21
4-Bromophenyl-phenylether
Concen: 0.372 ng
RT: 16.190 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.1	76.1	114.1
141	75.1	60.1	90.1



#22
Hexachlorobenzene
Concen: 0.380 ng
RT: 16.301 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion	Ratio	Lower	Upper
284	100		
142	55.0	44.0	66.0
249	37.1	29.7	44.5









#23
Atrazine
Concen: 0.345 ng
RT: 16.463 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
200	100		
173	35.1	28.1	42.1
215	49.1	39.3	58.9



#24
Pentachlorophenol
Concen: 0.433 ng
RT: 16.661 min Scan# 1674
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion	Ratio	Lower	Upper
266	100		
264	61.8	49.3	73.9
268	61.4	49.0	73.4









#25
Phenanthrene
Concen: 0.368 ng
RT: 17.034 min Scan# 1704
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 6496
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.4 12.3 18.5



#26
Anthracene
Concen: 0.351 ng
RT: 17.120 min Scan# 1711
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion:178 Resp: 5644
Ion Ratio Lower Upper
178 100
176 19.0 15.2 22.8
179 16.1 12.9 19.3





#27

Fluoranthene-d10

Concen: 0.384 ng

RT: 19.026 min Scan# 1934

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:212 Resp: 5309

Ion Ratio Lower Upper

212 100

106 13.2 10.6 15.8

104 8.2 6.6 9.8



#28

Fluoranthene

Concen: 0.357 ng

RT: 19.054 min Scan# 1940

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51




Tgt Ion:202 Resp: 6952


Ion Ratio Lower Upper

202 100

101 10.9 8.7 13.1

203 16.9 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.189 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:240 Resp: 3769

Ion Ratio Lower Upper

240 100

120 11.2 9.0 13.4

236 28.7 23.0 34.4



#30

Pyrene

Concen: 0.374 ng

RT: 19.421 min Scan# 2019

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51




Tgt Ion:202 Resp: 6885


Ion Ratio Lower Upper

202 100

200 21.3 17.0 25.6

203 17.8 14.2 21.4









#31
Terphenyl-d14
Concen: 0.381 ng
RT: 19.635 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	12.5	10.0	15.0
122	16.5	13.2	19.8



#32
Benzo(a)anthracene
Concen: 0.357 ng
RT: 21.171 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BN037145.D
Acq: 03 Jun 2025 12:51

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.2	22.6	33.8
229	20.2	16.2	24.2





#33

Chrysene

Concen: 0.366 ng

RT: 21.225 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:228 Resp: 5553

Ion Ratio Lower Upper

228 100

226 31.5 25.2 37.8

229 21.0 16.8 25.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.339 ng

RT: 21.117 min Scan# 2269

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51




Tgt Ion:149 Resp: 2918


Ion Ratio Lower Upper

149 100

167 26.2 21.0 31.4

279 3.6 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.377 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:264 Resp: 3386

Ion Ratio Lower Upper

264 100

260 27.6 22.1 33.1

265 69.8 55.8 83.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.377 ng

RT: 25.573 min Scan# 3496

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51




Tgt Ion:276 Resp: 5083


Ion Ratio Lower Upper

276 100

138 26.3 21.0 31.6

277 24.3 19.4 29.2





#37

Benzo(b)fluoranthene

Concen: 0.352 ng

RT: 22.722 min Scan# 2521

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 4811

Ion Ratio Lower Upper

252 100

253 27.9 22.3 33.5

125 16.5 13.2 19.8



#38

Benzo(k)fluoranthene

Concen: 0.354 ng

RT: 22.763 min Scan# 2535

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51




Tgt Ion:252 Resp: 4946


Ion Ratio Lower Upper

252 100

253 27.8 22.2 33.4

125 16.5 13.2 19.8





#39

Benzo(a)pyrene

Concen: 0.360 ng

RT: 23.281 min Scan# 2712

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 4127

Ion Ratio Lower Upper

252 100

253 31.2 25.0 37.4

125 21.3 17.0 25.6



#40

Dibenzo(a,h)anthracene

Concen: 0.378 ng

RT: 25.588 min Scan# 3501

Delta R.T. 0.000 min

Lab File: BN037145.D

Acq: 03 Jun 2025 12:51




Tgt Ion:278 Resp: 3928


Ion Ratio Lower Upper

278 100

139 22.0 17.6 26.4

279 32.5 26.0 39.0





#41

Benzo(g,h,i)perylene

Concen: 0.383 ng

RT: 26.240 min Scan# 31

Delta R.T. 0.000 min

Lab File: BN037145.D


Acq: 03 Jun 2025 12:51

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:276 Resp: 4576

Ion Ratio Lower Upper

276 100

277 26.1 20.9 31.3

138 26.0 20.8 31.2



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037146.D
 Acq On : 03 Jun 2025 13:26
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Jun 04 01:42:31 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:40:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.597	152	1776	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	4485	0.400	ng	0.00
13) Acenaphthene-d10	14.234	164	2508	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	4816	0.400	ng	0.00
29) Chrysene-d12	21.189	240	3381	0.400	ng	0.00
35) Perylene-d12	23.380	264	2909	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.192	112	3356	0.764	ng	0.00
5) Phenol-d6	6.773	99	4000	0.751	ng	0.00
8) Nitrobenzene-d5	8.739	82	3647	0.771	ng	0.00
11) 2-Methylnaphthalene-d10	11.971	152	4809	0.770	ng	0.00
14) 2,4,6-Tribromophenol	15.743	330	789	0.781	ng	0.00
15) 2-Fluorobiphenyl	12.863	172	8296	0.776	ng	0.00
27) Fluoranthene-d10	19.026	212	9209	0.753	ng	0.00
31) Terphenyl-d14	19.635	244	6364	0.800	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.119	88	1799	0.760	ng	97
3) n-Nitrosodimethylamine	3.437	42	3791	0.798	ng	# 97
6) bis(2-Chloroethyl)ether	7.026	93	3867	0.761	ng	98
9) Naphthalene	10.415	128	9964	0.770	ng	99
10) Hexachlorobutadiene	10.714	225	2212	0.785	ng	# 99
12) 2-Methylnaphthalene	12.041	142	6445	0.777	ng	99
16) Acenaphthylene	13.957	152	9385	0.763	ng	99
17) Acenaphthene	14.299	154	6080	0.762	ng	100
18) Fluorene	15.293	166	8080	0.770	ng	99
20) 4,6-Dinitro-2-methylph...	15.389	198	643	0.730	ng	# 56
21) 4-Bromophenyl-phenylether	16.189	248	2446	0.775	ng	97
22) Hexachlorobenzene	16.301	284	2690	0.790	ng	98
23) Atrazine	16.462	200	2011	0.772	ng	92
24) Pentachlorophenol	16.649	266	1033	0.722	ng	96
25) Phenanthrene	17.033	178	12021	0.770	ng	100
26) Anthracene	17.120	178	11005	0.773	ng	99
28) Fluoranthene	19.054	202	13144	0.763	ng	100
30) Pyrene	19.421	202	13035	0.790	ng	99
32) Benzo(a)anthracene	21.171	228	9494	0.776	ng	100
33) Chrysene	21.224	228	10700	0.785	ng	98
34) Bis(2-ethylhexyl)phtha...	21.117	149	5801	0.751	ng	100
36) Indeno(1,2,3-cd)pyrene	25.573	276	8881	0.767	ng	99
37) Benzo(b)fluoranthene	22.722	252	9165	0.780	ng	95
38) Benzo(k)fluoranthene	22.763	252	9379	0.782	ng	93
39) Benzo(a)pyrene	23.281	252	7528	0.765	ng	# 91
40) Dibenzo(a,h)anthracene	25.588	278	6960	0.780	ng	93
41) Benzo(g,h,i)perylene	26.240	276	7980	0.778	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037146.D
Acq On : 03 Jun 2025 13:26
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Quant Time: Jun 04 01:42:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:40:18 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.597 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8







Tgt Ion:152 Resp: 1776
Ion Ratio Lower Upper
152 100
150 150.3 123.2 184.8
115 68.8 56.6 85.0



#2
1,4-Dioxane
Concen: 0.760 ng
RT: 3.119 min Scan# 33
Delta R.T. -0.007 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion: 88 Resp: 1799
Ion Ratio Lower Upper
88 100
43 52.9 43.5 65.3
58 87.4 67.7 101.5









#3
n-Nitrosodimethylamine
Concen: 0.798 ng
RT: 3.437 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 3791
Ion Ratio Lower Upper
42 100
74 64.4 53.0 79.4
44 4.8 5.9 8.9#



#4
2-Fluorophenol
Concen: 0.764 ng
RT: 5.192 min Scan# 320
Delta R.T. 0.007 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion: 112 Resp: 3356
Ion Ratio Lower Upper
112 100
64 69.7 56.3 84.5
63 44.0 36.2 54.4









#5
Phenol-d6
Concen: 0.751 ng
RT: 6.773 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
99	100		
42	40.1	31.3	46.9
71	47.5	38.2	57.2



#6
bis(2-Chloroethyl)ether
Concen: 0.761 ng
RT: 7.026 min Scan# 574
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion	Ratio	Lower	Upper
93	100		
63	87.7	68.6	103.0
95	31.2	24.3	36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	4485
Ion	Ratio	Lower	Upper
136	100		
137	12.5	9.7	14.5
54	11.7	9.7	14.5
68	6.8	5.4	8.2



#8
Nitrobenzene-d5
Concen: 0.771 ng
RT: 8.739 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion:	82	Resp:	3647
Ion	Ratio	Lower	Upper
82	100		
128	33.5	26.9	40.3
54	73.4	61.4	92.2





#9

Naphthalene

Concen: 0.770 ng

RT: 10.415 min Scan# 949

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:128 Resp: 9964

Ion Ratio Lower Upper

128 100

129 11.7 9.8 14.8

127 15.1 12.3 18.5



#10

Hexachlorobutadiene

Concen: 0.785 ng

RT: 10.714 min Scan# 977

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26




Tgt Ion:225 Resp: 2212


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.9 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.770 ng
RT: 11.971 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 4809
Ion Ratio Lower Upper
152 100
151 21.8 17.1 25.7



#12
2-Methylnaphthalene
Concen: 0.777 ng
RT: 12.041 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion:142 Resp: 6445
Ion Ratio Lower Upper
142 100
141 92.8 74.6 111.8
115 50.3 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.234 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 2508
Ion Ratio Lower Upper
164 100
162 108.2 85.5 128.3
160 56.8 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 0.781 ng
RT: 15.743 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion:330 Resp: 789
Ion Ratio Lower Upper
330 100
332 97.5 77.1 115.7
141 58.3 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 0.776 ng
RT: 12.863 min Scan# 1319
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.6	29.6	44.4
170	24.7	20.3	30.5



#16
Acenaphthylene
Concen: 0.763 ng
RT: 13.957 min Scan# 1436
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.2	16.3	24.5
153	12.9	10.6	15.8









#17
Acenaphthene
Concen: 0.762 ng
RT: 14.299 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:154 Resp: 6080
Ion Ratio Lower Upper
154 100
153 117.4 93.8 140.8
152 62.7 50.5 75.7



#18
Fluorene
Concen: 0.770 ng
RT: 15.293 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion:166 Resp: 8080
Ion Ratio Lower Upper
166 100
165 100.5 81.1 121.7
167 13.3 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:188 Resp: 4816

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.4 11.3 16.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.730 ng

RT: 15.389 min Scan# 1570

Delta R.T. 0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26




Tgt Ion:198 Resp: 643


Ion Ratio Lower Upper

198 100

51 89.6 125.2 187.8#

105 49.2 57.1 85.7#









#21
4-Bromophenyl-phenylether
Concen: 0.775 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
248	100		
250	97.2	76.1	114.1
141	70.7	60.1	90.1



#22
Hexachlorobenzene
Concen: 0.790 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion	Ratio	Lower	Upper
284	100		
142	53.3	44.0	66.0
249	36.2	29.7	44.5









#23
Atrazine
Concen: 0.772 ng
RT: 16.462 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
200	100		
173	28.8	28.1	42.1
215	45.0	39.3	58.9



#24
Pentachlorophenol
Concen: 0.722 ng
RT: 16.649 min Scan# 1673
Delta R.T. -0.012 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion	Ratio	Lower	Upper
266	100		
264	64.1	49.3	73.9
268	64.2	49.0	73.4









#25
Phenanthrene
Concen: 0.770 ng
RT: 17.033 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 12021
Ion Ratio Lower Upper
178 100
176 20.0 15.7 23.5
179 15.4 12.3 18.5



#26
Anthracene
Concen: 0.773 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion:178 Resp: 11005
Ion Ratio Lower Upper
178 100
176 19.0 15.2 22.8
179 15.2 12.9 19.3





#27

Fluoranthene-d10

Concen: 0.753 ng

RT: 19.026 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:212 Resp: 9209

Ion Ratio Lower Upper

212 100

106 13.3 10.6 15.8

104 8.1 6.6 9.8



#28

Fluoranthene

Concen: 0.763 ng

RT: 19.054 min Scan# 1940

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26




Tgt Ion:202 Resp: 13144


Ion Ratio Lower Upper

202 100

101 10.6 8.7 13.1

203 17.0 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.189 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 3381

Ion Ratio Lower Upper

240 100

120 11.3 9.0 13.4

236 29.1 23.0 34.4



#30

Pyrene

Concen: 0.790 ng

RT: 19.421 min Scan# 2019

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26




Tgt Ion:202 Resp: 13035


Ion Ratio Lower Upper

202 100

200 21.6 17.0 25.6

203 17.6 14.2 21.4









#31
Terphenyl-d14
Concen: 0.800 ng
RT: 19.635 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.5	10.0	15.0
122	15.5	13.2	19.8



#32
Benzo(a)anthracene
Concen: 0.776 ng
RT: 21.171 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.8	22.6	33.8
229	20.2	16.2	24.2





#33

Chrysene

Concen: 0.785 ng

RT: 21.224 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:228 Resp: 10700

Ion Ratio Lower Upper

228 100

226 30.9 25.2 37.8

229 19.7 16.8 25.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.751 ng

RT: 21.117 min Scan# 2269

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26




Tgt Ion:149 Resp: 5801


Ion Ratio Lower Upper

149 100

167 26.3 21.0 31.4

279 3.3 2.9 4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.380 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:264 Resp: 2909
Ion Ratio Lower Upper
264 100
260 27.3 22.1 33.1
265 76.0 55.8 83.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.767 ng
RT: 25.573 min Scan# 3496
Delta R.T. -0.000 min
Lab File: BN037146.D
Acq: 03 Jun 2025 13:26

Tgt Ion:276 Resp: 8881
Ion Ratio Lower Upper
276 100
138 27.1 21.0 31.6
277 24.5 19.4 29.2





#37

Benzo(b)fluoranthene

Concen: 0.780 ng

RT: 22.722 min Scan# 2521

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 9165

Ion Ratio Lower Upper

252 100

253 25.0 22.3 33.5

125 14.3 13.2 19.8



#38

Benzo(k)fluoranthene

Concen: 0.782 ng

RT: 22.763 min Scan# 2535

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26




Tgt Ion:252 Resp: 9379


Ion Ratio Lower Upper

252 100

253 24.5 22.2 33.4

125 13.5 13.2 19.8





#39

Benzo(a)pyrene

Concen: 0.765 ng

RT: 23.281 min Scan# 2712

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 7528

Ion Ratio Lower Upper

252 100

253 26.6 25.0 37.4

125 16.4 17.0 25.6#



#40

Dibenzo(a,h)anthracene

Concen: 0.780 ng

RT: 25.588 min Scan# 3501

Delta R.T. -0.000 min

Lab File: BN037146.D

Acq: 03 Jun 2025 13:26




Tgt Ion:278 Resp: 6960


Ion Ratio Lower Upper

278 100

139 19.5 17.6 26.4

279 27.7 26.0 39.0





#41

Benzo(g,h,i)perylene

Concen: 0.778 ng

RT: 26.240 min Scan# 31

Delta R.T. -0.000 min

Lab File: BN037146.D


Acq: 03 Jun 2025 13:26

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:276 Resp: 7980

Ion Ratio Lower Upper

276 100

277 24.8 20.9 31.3

138 26.1 20.8 31.2



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037147.D
 Acq On : 03 Jun 2025 14:02
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Jun 04 01:42:52 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:40:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	1779	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	4585	0.400	ng	0.00
13) Acenaphthene-d10	14.235	164	2603	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	5057	0.400	ng	0.00
29) Chrysene-d12	21.189	240	3873	0.400	ng	0.00
35) Perylene-d12	23.377	264	3320	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	7371	1.676	ng	0.00
5) Phenol-d6	6.766	99	9203	1.726	ng	0.00
8) Nitrobenzene-d5	8.739	82	8338	1.724	ng	0.00
11) 2-Methylnaphthalene-d10	11.966	152	10971	1.719	ng	0.00
14) 2,4,6-Tribromophenol	15.730	330	1930	1.842	ng	-0.01
15) 2-Fluorobiphenyl	12.858	172	18891	1.702	ng	0.00
27) Fluoranthene-d10	19.026	212	22095	1.720	ng	0.00
31) Terphenyl-d14	19.630	244	15589	1.710	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.119	88	3745	1.579	ng	97
3) n-Nitrosodimethylamine	3.429	42	8276	1.738	ng	# 97
6) bis(2-Chloroethyl)ether	7.026	93	8706	1.711	ng	98
9) Naphthalene	10.415	128	22289	1.685	ng	97
10) Hexachlorobutadiene	10.714	225	4883	1.694	ng	# 99
12) 2-Methylnaphthalene	12.042	142	14834	1.750	ng	97
16) Acenaphthylene	13.957	152	21986	1.723	ng	100
17) Acenaphthene	14.299	154	14266	1.722	ng	99
18) Fluorene	15.293	166	18976	1.742	ng	100
20) 4,6-Dinitro-2-methylph...	15.378	198	1819	1.502	ng	# 41
21) 4-Bromophenyl-phenylether	16.189	248	5692	1.717	ng	90
22) Hexachlorobenzene	16.301	284	6093	1.703	ng	99
23) Atrazine	16.463	200	4865	1.778	ng	# 88
24) Pentachlorophenol	16.649	266	2828	1.535	ng	96
25) Phenanthrene	17.021	178	28034	1.711	ng	100
26) Anthracene	17.120	178	26183	1.751	ng	99
28) Fluoranthene	19.054	202	31942	1.765	ng	100
30) Pyrene	19.416	202	31728	1.678	ng	100
32) Benzo(a)anthracene	21.171	228	24513	1.748	ng	99
33) Chrysene	21.224	228	26306	1.685	ng	98
34) Bis(2-ethylhexyl)phtha...	21.117	149	14807	1.674	ng	100
36) Indeno(1,2,3-cd)pyrene	25.573	276	22504	1.704	ng	98
37) Benzo(b)fluoranthene	22.722	252	23418	1.747	ng	# 90
38) Benzo(k)fluoranthene	22.763	252	23602	1.725	ng	# 90
39) Benzo(a)pyrene	23.281	252	19265	1.716	ng	# 86
40) Dibenzo(a,h)anthracene	25.588	278	17700	1.738	ng	# 90
41) Benzo(g,h,i)perylene	26.243	276	19619	1.677	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037147.D
Acq On : 03 Jun 2025 14:02
Operator : RC/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Quant Time: Jun 04 01:42:52 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:40:18 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.589 min Scan# 61
Delta R.T. -0.001 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6







Tgt Ion:152 Resp: 1779
Ion Ratio Lower Upper
152 100
150 160.2 123.2 184.8
115 71.5 56.6 85.0



#2
1,4-Dioxane
Concen: 1.579 ng
RT: 3.119 min Scan# 33
Delta R.T. -0.007 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion: 88 Resp: 3745
Ion Ratio Lower Upper
88 100
43 53.6 43.5 65.3
58 88.6 67.7 101.5









#3
n-Nitrosodimethylamine
Concen: 1.738 ng
RT: 3.429 min Scan# 70
Delta R.T. -0.007 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 8276
Ion Ratio Lower Upper
42 100
74 64.1 53.0 79.4
44 5.2 5.9 8.9#



#4
2-Fluorophenol
Concen: 1.676 ng
RT: 5.184 min Scan# 319
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion: 112 Resp: 7371
Ion Ratio Lower Upper
112 100
64 69.5 56.3 84.5
63 44.8 36.2 54.4









#5
Phenol-d6
Concen: 1.726 ng
RT: 6.766 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 9203
Ion Ratio Lower Upper
99 100
42 39.7 31.3 46.9
71 47.7 38.2 57.2



#6
bis(2-Chloroethyl)ether
Concen: 1.711 ng
RT: 7.026 min Scan# 574
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion: 93 Resp: 8706
Ion Ratio Lower Upper
93 100
63 88.1 68.6 103.0
95 31.3 24.3 36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	4585
Ion	Ratio	Lower	Upper
136	100		
137	12.5	9.7	14.5
54	11.3	9.7	14.5
68	6.6	5.4	8.2



#8
Nitrobenzene-d5
Concen: 1.724 ng
RT: 8.739 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion:	82	Resp:	8338
Ion	Ratio	Lower	Upper
82	100		
128	34.0	26.9	40.3
54	70.4	61.4	92.2









#9
Naphthalene
Concen: 1.685 ng
RT: 10.415 min Scan# 949
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 22289
Ion Ratio Lower Upper
128 100
129 11.3 9.8 14.8
127 14.1 12.3 18.5



#10
Hexachlorobutadiene
Concen: 1.694 ng
RT: 10.714 min Scan# 977
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion:225 Resp: 4883
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.4 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 1.719 ng
RT: 11.966 min Scan# 1142
Delta R.T. -0.005 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 10971
Ion Ratio Lower Upper
152 100
151 21.7 17.1 25.7



#12
2-Methylnaphthalene
Concen: 1.750 ng
RT: 12.042 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion:142 Resp: 14834
Ion Ratio Lower Upper
142 100
141 91.4 74.6 111.8
115 48.3 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.235 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 2603
Ion Ratio Lower Upper
164 100
162 112.4 85.5 128.3
160 57.8 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 1.842 ng
RT: 15.730 min Scan# 1599
Delta R.T. -0.012 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion:330 Resp: 1930
Ion Ratio Lower Upper
330 100
332 97.3 77.1 115.7
141 62.6 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 1.702 ng
RT: 12.858 min Scan# 1318
Delta R.T. -0.005 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 18891
Ion Ratio Lower Upper
172 100
171 36.2 29.6 44.4
170 24.5 20.3 30.5



#16
Acenaphthylene
Concen: 1.723 ng
RT: 13.957 min Scan# 1436
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion:152 Resp: 21986
Ion Ratio Lower Upper
152 100
151 20.3 16.3 24.5
153 13.0 10.6 15.8









#17
Acenaphthene
Concen: 1.722 ng
RT: 14.299 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 14266
Ion Ratio Lower Upper
154 100
153 117.1 93.8 140.8
152 61.2 50.5 75.7



#18
Fluorene
Concen: 1.742 ng
RT: 15.293 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion:166 Resp: 18976
Ion Ratio Lower Upper
166 100
165 101.0 81.1 121.7
167 13.2 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN037147.D

Acq: 03 Jun 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 5057

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 13.6 11.3 16.9



#20

4,6-Dinitro-2-methylphenol

Concen: 1.502 ng

RT: 15.378 min Scan# 1569

Delta R.T. -0.011 min

Lab File: BN037147.D

Acq: 03 Jun 2025 14:02




Tgt Ion:198 Resp: 1819


Ion Ratio Lower Upper

198 100

51 65.5 125.2 187.8#

105 42.0 57.1 85.7#









#21
4-Bromophenyl-phenylether
Concen: 1.717 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
248	100		
250	99.4	76.1	114.1
141	61.1	60.1	90.1



#22
Hexachlorobenzene
Concen: 1.703 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion	Ratio	Lower	Upper
284	100		
142	54.4	44.0	66.0
249	35.9	29.7	44.5









#23
Atrazine
Concen: 1.778 ng
RT: 16.463 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 4865
Ion Ratio Lower Upper
200 100
173 25.0 28.1 42.1#
215 44.1 39.3 58.9



#24
Pentachlorophenol
Concen: 1.535 ng
RT: 16.649 min Scan# 1673
Delta R.T. -0.012 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion:266 Resp: 2828
Ion Ratio Lower Upper
266 100
264 63.4 49.3 73.9
268 64.9 49.0 73.4









#25
Phenanthrene
Concen: 1.711 ng
RT: 17.021 min Scan# 1703
Delta R.T. -0.012 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 28034
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.2 12.3 18.5



#26
Anthracene
Concen: 1.751 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion:178 Resp: 26183
Ion Ratio Lower Upper
178 100
176 19.0 15.2 22.8
179 15.2 12.9 19.3





#27

Fluoranthene-d10

Concen: 1.720 ng

RT: 19.026 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BN037147.D

Acq: 03 Jun 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:212 Resp: 22095

Ion Ratio Lower Upper

212 100

106 13.5 10.6 15.8

104 8.2 6.6 9.8



#28

Fluoranthene

Concen: 1.765 ng

RT: 19.054 min Scan# 1940

Delta R.T. -0.000 min

Lab File: BN037147.D

Acq: 03 Jun 2025 14:02




Tgt Ion:202 Resp: 31942


Ion Ratio Lower Upper

202 100

101 10.9 8.7 13.1

203 16.9 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.189 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037147.D

Acq: 03 Jun 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 3873

Ion Ratio Lower Upper

240 100

120 10.5 9.0 13.4

236 29.2 23.0 34.4



#30

Pyrene

Concen: 1.678 ng

RT: 19.416 min Scan# 2018

Delta R.T. -0.005 min

Lab File: BN037147.D

Acq: 03 Jun 2025 14:02




Tgt Ion:202 Resp: 31728


Ion Ratio Lower Upper

202 100

200 21.5 17.0 25.6

203 17.6 14.2 21.4









#31
Terphenyl-d14
Concen: 1.710 ng
RT: 19.630 min Scan# 2064
Delta R.T. -0.005 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 15589
Ion Ratio Lower Upper
244 100
212 11.0 10.0 15.0
122 16.5 13.2 19.8



#32
Benzo(a)anthracene
Concen: 1.748 ng
RT: 21.171 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BN037147.D
Acq: 03 Jun 2025 14:02

Tgt Ion:228 Resp: 24513
Ion Ratio Lower Upper
228 100
226 28.0 22.6 33.8
229 19.6 16.2 24.2





#33

Chrysene

Concen: 1.685 ng

RT: 21.224 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037147.D

Acq: 03 Jun 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 26306

Ion Ratio Lower Upper

228 100

226 30.4 25.2 37.8

229 19.9 16.8 25.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.674 ng

RT: 21.117 min Scan# 2269

Delta R.T. -0.000 min

Lab File: BN037147.D

Acq: 03 Jun 2025 14:02




Tgt Ion:149 Resp: 14807


Ion Ratio Lower Upper

149 100

167 26.1 21.0 31.4

279 3.0 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.377 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037147.D


Acq: 03 Jun 2025 14:02

Instrument :



BNA_N

ClientSampleId :

SSTDICC1.6



Tgt Ion:	264	Resp:	3320
Ion	Ratio	Lower	Upper
264	100		
260	28.1	22.1	33.1
265	72.6	55.8	83.8



#36

Indeno(1,2,3-cd)pyrene


Concen: 1.704 ng

RT: 25.573 min Scan# 3496



Delta R.T. -0.000 min


Lab File: BN037147.D

Acq: 03 Jun 2025 14:02



Tgt Ion:	276	Resp:	22504
Ion	Ratio	Lower	Upper
276	100		
138	27.8	21.0	31.6
277	25.1	19.4	29.2





#37

Benzo(b)fluoranthene

Concen: 1.747 ng

RT: 22.722 min Scan# 2521

Delta R.T. -0.000 min

Lab File: BN037147.D


Acq: 03 Jun 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 23418

Ion Ratio Lower Upper

252 100

253 22.8 22.3 33.5

125 12.3 13.2 19.8#



#38

Benzo(k)fluoranthene


Concen: 1.725 ng

RT: 22.763 min Scan# 2535

Delta R.T. -0.000 min

Lab File: BN037147.D

Acq: 03 Jun 2025 14:02




Tgt Ion:252 Resp: 23602


Ion Ratio Lower Upper

252 100

253 22.8 22.2 33.4

125 12.3 13.2 19.8#





#39

Benzo(a)pyrene

Concen: 1.716 ng

RT: 23.281 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037147.D


Acq: 03 Jun 2025 14:02

Instrument :




BNA_N

ClientSampleId :

SSTDICC1.6



Tgt Ion:	252	Resp:	19265
Ion	Ratio	Lower	Upper
252	100		
253	24.1	25.0	37.4#
125	14.4	17.0	25.6#



#40

Dibenzo(a,h)anthracene


Concen: 1.738 ng

RT: 25.588 min Scan# 3501



Delta R.T. -0.000 min


Lab File: BN037147.D

Acq: 03 Jun 2025 14:02



Tgt Ion:	278	Resp:	17700
Ion	Ratio	Lower	Upper
278	100		
139	19.9	17.6	26.4
279	24.9	26.0	39.0#





#41

Benzo(g,h,i)perylene

Concen: 1.677 ng

RT: 26.243 min Scan# 31

Delta R.T. 0.003 min

Lab File: BN037147.D


Acq: 03 Jun 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:276 Resp: 19619

Ion Ratio Lower Upper

276 100

277 24.0 20.9 31.3

138 24.7 20.8 31.2



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037148.D
 Acq On : 03 Jun 2025 14:38
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Jun 04 01:43:12 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:40:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	2082	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	5272	0.400	ng	0.00
13) Acenaphthene-d10	14.234	164	3042	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	5757	0.400	ng	0.00
29) Chrysene-d12	21.180	240	4564	0.400	ng	# 0.00
35) Perylene-d12	23.377	264	3760	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	16397	3.185	ng	0.00
5) Phenol-d6	6.766	99	21011	3.367	ng	0.00
8) Nitrobenzene-d5	8.739	82	18977	3.412	ng	0.00
11) 2-Methylnaphthalene-d10	11.965	152	24344	3.317	ng	0.00
14) 2,4,6-Tribromophenol	15.730	330	4423	3.611	ng	-0.01
15) 2-Fluorobiphenyl	12.858	172	41512	3.201	ng	0.00
27) Fluoranthene-d10	19.026	212	49341	3.373	ng	0.00
31) Terphenyl-d14	19.630	244	34773	3.237	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.119	88	7944	2.862	ng	97
3) n-Nitrosodimethylamine	3.429	42	17675	3.172	ng	# 98
6) bis(2-Chloroethyl)ether	7.026	93	19085	3.205	ng	96
9) Naphthalene	10.415	128	49155	3.232	ng	97
10) Hexachlorobutadiene	10.714	225	10373	3.131	ng	# 99
12) 2-Methylnaphthalene	12.041	142	33014	3.386	ng	96
16) Acenaphthylene	13.956	152	49883	3.345	ng	99
17) Acenaphthene	14.299	154	31857	3.290	ng	98
18) Fluorene	15.293	166	42258	3.320	ng	99
20) 4,6-Dinitro-2-methylph...	15.378	198	4695	3.058	ng	# 35
21) 4-Bromophenyl-phenylether	16.189	248	12710	3.369	ng	# 89
22) Hexachlorobenzene	16.301	284	13101	3.217	ng	98
23) Atrazine	16.462	200	10979	3.524	ng	# 87
24) Pentachlorophenol	16.649	266	7055	3.108	ng	98
25) Phenanthrene	17.033	178	62481	3.350	ng	100
26) Anthracene	17.120	178	59394	3.490	ng	99
28) Fluoranthene	19.054	202	72004	3.495	ng	100
30) Pyrene	19.416	202	71397	3.205	ng	100
32) Benzo(a)anthracene	21.171	228	56691	3.431	ng	98
33) Chrysene	21.215	228	57833	3.144	ng	98
34) Bis(2-ethylhexyl)phtha...	21.117	149	33369	3.201	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.573	276	50327	3.364	ng	97
37) Benzo(b)fluoranthene	22.722	252	51531	3.395	ng	# 89
38) Benzo(k)fluoranthene	22.763	252	52430	3.384	ng	# 89
39) Benzo(a)pyrene	23.281	252	42893	3.374	ng	# 84
40) Dibenzo(a,h)anthracene	25.587	278	40079	3.474	ng	# 89
41) Benzo(g,h,i)perylene	26.239	276	42842	3.233	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037148.D
Acq On : 03 Jun 2025 14:38
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Quant Time: Jun 04 01:43:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:40:18 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.589 min Scan# 61
Delta R.T. -0.001 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:152 Resp: 2082
Ion Ratio Lower Upper
152 100
150 154.4 123.2 184.8
115 70.5 56.6 85.0



#2
1,4-Dioxane
Concen: 2.862 ng
RT: 3.119 min Scan# 33
Delta R.T. -0.007 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion: 88 Resp: 7944
Ion Ratio Lower Upper
88 100
43 53.4 43.5 65.3
58 88.6 67.7 101.5









#3
n-Nitrosodimethylamine
Concen: 3.172 ng
RT: 3.429 min Scan# 70
Delta R.T. -0.007 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 17675
Ion Ratio Lower Upper
42 100
74 65.5 53.0 79.4
44 4.8 5.9 8.9#



#4
2-Fluorophenol
Concen: 3.185 ng
RT: 5.184 min Scan# 319
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion: 112 Resp: 16397
Ion Ratio Lower Upper
112 100
64 70.6 56.3 84.5
63 45.1 36.2 54.4









#5
Phenol-d6
Concen: 3.367 ng
RT: 6.766 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 21011
Ion Ratio Lower Upper
99 100
42 39.5 31.3 46.9
71 47.7 38.2 57.2



#6
bis(2-Chloroethyl)ether
Concen: 3.205 ng
RT: 7.026 min Scan# 574
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion: 93 Resp: 19085
Ion Ratio Lower Upper
93 100
63 89.3 68.6 103.0
95 31.8 24.3 36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:136 Resp: 5272
Ion Ratio Lower Upper
136 100
137 11.8 9.7 14.5
54 11.0 9.7 14.5
68 6.2 5.4 8.2



#8
Nitrobenzene-d5
Concen: 3.412 ng
RT: 8.739 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion: 82 Resp: 18977
Ion Ratio Lower Upper
82 100
128 34.3 26.9 40.3
54 69.2 61.4 92.2









#9
Naphthalene
Concen: 3.232 ng
RT: 10.415 min Scan# 949
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion	128	129	127
Resp:	49155		
Ion Ratio	100	10.9	14.0
Lower		9.8	12.3
Upper		14.8	18.5



#10
Hexachlorobutadiene
Concen: 3.131 ng
RT: 10.714 min Scan# 977
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion	225	223	227
Resp:	10373		
Ion Ratio	100	0.0	63.4
Lower		0.0	50.3
Upper		0.0	75.5









#11
2-Methylnaphthalene-d10
Concen: 3.317 ng
RT: 11.965 min Scan# 1142
Delta R.T. -0.005 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 24344
Ion Ratio Lower Upper
152 100
151 21.5 17.1 25.7



#12
2-Methylnaphthalene
Concen: 3.386 ng
RT: 12.041 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion:142 Resp: 33014
Ion Ratio Lower Upper
142 100
141 90.1 74.6 111.8
115 47.6 41.0 61.4





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.234 min Scan# 1462

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:164 Resp: 3042

Ion Ratio Lower Upper

164 100

162 108.4 85.5 128.3

160 55.8 44.6 67.0



#14

2,4,6-Tribromophenol

Concen: 3.611 ng

RT: 15.730 min Scan# 1599

Delta R.T. -0.013 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38




Tgt Ion:330 Resp: 4423


Ion Ratio Lower Upper

330 100

332 96.7 77.1 115.7

141 65.0 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 3.201 ng
RT: 12.858 min Scan# 1318
Delta R.T. -0.005 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 41512
Ion Ratio Lower Upper
172 100
171 36.4 29.6 44.4
170 24.5 20.3 30.5



#16
Acenaphthylene
Concen: 3.345 ng
RT: 13.956 min Scan# 1436
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion:152 Resp: 49883
Ion Ratio Lower Upper
152 100
151 20.3 16.3 24.5
153 12.9 10.6 15.8





#17

Acenaphthene

Concen: 3.290 ng

RT: 14.299 min Scan# 1468

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:154 Resp: 31857

Ion Ratio Lower Upper

154 100

153 116.6 93.8 140.8

152 59.9 50.5 75.7



#18

Fluorene

Concen: 3.320 ng

RT: 15.293 min Scan# 1561

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38




Tgt Ion:166 Resp: 42258


Ion Ratio Lower Upper

166 100

165 99.8 81.1 121.7

167 13.3 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 5757

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 13.1 11.3 16.9



#20

4,6-Dinitro-2-methylphenol

Concen: 3.058 ng

RT: 15.378 min Scan# 1569

Delta R.T. -0.011 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38




Tgt Ion:198 Resp: 4695


Ion Ratio Lower Upper

198 100

51 56.8 125.2 187.8#

105 38.9 57.1 85.7#









#21
4-Bromophenyl-phenylether
Concen: 3.369 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 12710
Ion Ratio Lower Upper
248 100
250 97.3 76.1 114.1
141 56.2 60.1 90.1#



#22
Hexachlorobenzene
Concen: 3.217 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion:284 Resp: 13101
Ion Ratio Lower Upper
284 100
142 53.4 44.0 66.0
249 36.5 29.7 44.5









#23
Atrazine
Concen: 3.524 ng
RT: 16.462 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 10979
Ion Ratio Lower Upper
200 100
173 24.8 28.1 42.1#
215 42.7 39.3 58.9



#24
Pentachlorophenol
Concen: 3.108 ng
RT: 16.649 min Scan# 1673
Delta R.T. -0.012 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion:266 Resp: 7055
Ion Ratio Lower Upper
266 100
264 63.3 49.3 73.9
268 63.1 49.0 73.4









#25
Phenanthrene
Concen: 3.350 ng
RT: 17.033 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 62481
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.1 12.3 18.5



#26
Anthracene
Concen: 3.490 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion:178 Resp: 59394
Ion Ratio Lower Upper
178 100
176 19.1 15.2 22.8
179 15.2 12.9 19.3





#27

Fluoranthene-d10

Concen: 3.373 ng

RT: 19.026 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:212 Resp: 49341

Ion Ratio Lower Upper

212 100

106 13.9 10.6 15.8

104 8.4 6.6 9.8



#28

Fluoranthene

Concen: 3.495 ng

RT: 19.054 min Scan# 1940

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38




Tgt Ion:202 Resp: 72004


Ion Ratio Lower Upper

202 100

101 10.9 8.7 13.1

203 17.2 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.180 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:240 Resp: 4564

Ion Ratio Lower Upper

240 100

120 18.4 9.0 13.4#

236 30.1 23.0 34.4



#30

Pyrene

Concen: 3.205 ng

RT: 19.416 min Scan# 2018

Delta R.T. -0.005 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38




Tgt Ion:202 Resp: 71397


Ion Ratio Lower Upper

202 100

200 21.5 17.0 25.6


203 17.7 14.2 21.4







#31
Terphenyl-d14
Concen: 3.237 ng
RT: 19.630 min Scan# 2064
Delta R.T. -0.005 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:244 Resp: 34773
Ion Ratio Lower Upper
244 100
212 10.8 10.0 15.0
122 16.5 13.2 19.8



#32
Benzo(a)anthracene
Concen: 3.431 ng
RT: 21.171 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BN037148.D
Acq: 03 Jun 2025 14:38

Tgt Ion:228 Resp: 56691
Ion Ratio Lower Upper
228 100
226 27.0 22.6 33.8
229 19.7 16.2 24.2





#33

Chrysene

Concen: 3.144 ng

RT: 21.215 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 57833

Ion Ratio Lower Upper

228 100

226 32.0 25.2 37.8

229 19.1 16.8 25.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 3.201 ng

RT: 21.117 min Scan# 2269

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38




Tgt Ion:149 Resp: 33369


Ion Ratio Lower Upper

149 100

167 25.9 21.0 31.4

279 2.8 2.9 4.3#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.377 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037148.D


Acq: 03 Jun 2025 14:38

Instrument :

BNA_N



ClientSampleId :

SSTDICC3.2



Tgt Ion:264 Resp: 3760

Ion	Ratio	Lower	Upper
264	100		
260	27.7	22.1	33.1
265	67.7	55.8	83.8



#36

Indeno(1,2,3-cd)pyrene


Concen: 3.364 ng

RT: 25.573 min Scan# 3496

Delta R.T. -0.000 min



Lab File: BN037148.D


Acq: 03 Jun 2025 14:38



Tgt Ion:276 Resp: 50327

Ion	Ratio	Lower	Upper
276	100		
138	29.0	21.0	31.6
277	24.9	19.4	29.2





#37

Benzo(b)fluoranthene

Concen: 3.395 ng

RT: 22.722 min Scan# 2521

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 51531

Ion Ratio Lower Upper

252 100

253 22.3 22.3 33.5

125 12.0 13.2 19.8#



#38

Benzo(k)fluoranthene

Concen: 3.384 ng

RT: 22.763 min Scan# 2535

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38




Tgt Ion:252 Resp: 52430


Ion Ratio Lower Upper

252 100

253 22.0 22.2 33.4#

125 11.8 13.2 19.8#





#39

Benzo(a)pyrene

Concen: 3.374 ng

RT: 23.281 min Scan# 2712

Delta R.T. -0.000 min

Lab File: BN037148.D


Acq: 03 Jun 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 42893

Ion Ratio Lower Upper

252 100

253 22.8 25.0 37.4#

125 13.2 17.0 25.6#



#40

Dibenzo(a,h)anthracene

Concen: 3.474 ng

RT: 25.587 min Scan# 3501

Delta R.T. -0.000 min

Lab File: BN037148.D

Acq: 03 Jun 2025 14:38




Tgt Ion:278 Resp: 40079


Ion Ratio Lower Upper

278 100

139 19.6 17.6 26.4

279 23.9 26.0 39.0#





#41

Benzo(g,h,i)perylene

Concen: 3.233 ng

RT: 26.239 min Scan# 31

Delta R.T. -0.000 min

Lab File: BN037148.D


Acq: 03 Jun 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:276 Resp: 42842

Ion Ratio Lower Upper

276 100

277 23.6 20.9 31.3

138 24.9 20.8 31.2



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037149.D
 Acq On : 03 Jun 2025 15:14
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Jun 04 01:43:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:40:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	2274	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	5863	0.400	ng	0.00
13) Acenaphthene-d10	14.235	164	3467	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	6764	0.400	ng	0.00
29) Chrysene-d12	21.180	240	5784	0.400	ng	# 0.00
35) Perylene-d12	23.374	264	4620	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	27710	4.928	ng	0.00
5) Phenol-d6	6.766	99	36531	5.359	ng	0.00
8) Nitrobenzene-d5	8.739	82	32690	5.285	ng	0.00
11) 2-Methylnaphthalene-d10	11.966	152	43118	5.283	ng	0.00
14) 2,4,6-Tribromophenol	15.730	330	8081	5.789	ng	-0.01
15) 2-Fluorobiphenyl	12.858	172	74736	5.056	ng	0.00
27) Fluoranthene-d10	19.026	212	94210	5.482	ng	0.00
31) Terphenyl-d14	19.630	244	66718	4.900	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.119	88	13025	4.297	ng	97
3) n-Nitrosodimethylamine	3.422	42	28752	4.724	ng	# 97
6) bis(2-Chloroethyl)ether	7.026	93	32562	5.006	ng	97
9) Naphthalene	10.415	128	84983	5.024	ng	96
10) Hexachlorobutadiene	10.714	225	17448	4.735	ng	# 99
12) 2-Methylnaphthalene	12.042	142	58095	5.358	ng	96
16) Acenaphthylene	13.957	152	89938	5.291	ng	99
17) Acenaphthene	14.299	154	57215	5.184	ng	97
18) Fluorene	15.293	166	75931	5.233	ng	98
20) 4,6-Dinitro-2-methylph...	15.378	198	9612	5.124	ng	# 31
21) 4-Bromophenyl-phenylether	16.189	248	22933	5.173	ng	# 87
22) Hexachlorobenzene	16.301	284	23190	4.847	ng	98
23) Atrazine	16.463	200	20848	5.696	ng	# 87
24) Pentachlorophenol	16.649	266	13952	5.085	ng	98
25) Phenanthrene	17.021	178	115105	5.252	ng	99
26) Anthracene	17.120	178	111362	5.569	ng	99
28) Fluoranthene	19.054	202	135672	5.604	ng	99
30) Pyrene	19.416	202	136281	4.827	ng	100
32) Benzo(a)anthracene	21.171	228	113532	5.423	ng	99
33) Chrysene	21.216	228	112475	4.825	ng	98
34) Bis(2-ethylhexyl)phtha...	21.117	149	72476	5.486	ng	# 99
36) Indeno(1,2,3-cd)pyrene	25.573	276	97988	5.331	ng	96
37) Benzo(b)fluoranthene	22.720	252	102873	5.516	ng	# 89
38) Benzo(k)fluoranthene	22.763	252	104252	5.475	ng	# 89
39) Benzo(a)pyrene	23.278	252	85510	5.474	ng	# 83
40) Dibenzo(a,h)anthracene	25.585	278	76694	5.411	ng	# 89
41) Benzo(g,h,i)perylene	26.243	276	82277	5.054	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037149.D
Acq On : 03 Jun 2025 15:14
Operator : RC/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Jun 04 01:43:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:40:18 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.589 min Scan# 61
Delta R.T. -0.001 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 2274
Ion Ratio Lower Upper
152 100
150 156.4 123.2 184.8
115 69.7 56.6 85.0



#2
1,4-Dioxane
Concen: 4.297 ng
RT: 3.119 min Scan# 33
Delta R.T. -0.007 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion: 88 Resp: 13025
Ion Ratio Lower Upper
88 100
43 52.8 43.5 65.3
58 88.1 67.7 101.5









#3
n-Nitrosodimethylamine
Concen: 4.724 ng
RT: 3.422 min Scan# 71
Delta R.T. -0.015 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 28752
Ion Ratio Lower Upper
42 100
74 68.0 53.0 79.4
44 5.0 5.9 8.9#



#4
2-Fluorophenol
Concen: 4.928 ng
RT: 5.184 min Scan# 319
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion: 112 Resp: 27710
Ion Ratio Lower Upper
112 100
64 70.9 56.3 84.5
63 44.5 36.2 54.4









#5
Phenol-d6
Concen: 5.359 ng
RT: 6.766 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 36531
Ion Ratio Lower Upper
99 100
42 38.3 31.3 46.9
71 46.9 38.2 57.2



#6
bis(2-Chloroethyl)ether
Concen: 5.006 ng
RT: 7.026 min Scan# 574
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion: 93 Resp: 32562
Ion Ratio Lower Upper
93 100
63 89.2 68.6 103.0
95 31.7 24.3 36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion	136	Resp	5863
Ion	Ratio	Lower	Upper
136	100		
137	12.2	9.7	14.5
54	10.7	9.7	14.5
68	6.1	5.4	8.2



#8
Nitrobenzene-d5
Concen: 5.285 ng
RT: 8.739 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion	82	Resp	32690
Ion	Ratio	Lower	Upper
82	100		
128	34.7	26.9	40.3
54	68.6	61.4	92.2









#9
Naphthalene
Concen: 5.024 ng
RT: 10.415 min Scan# 949
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion	128	129	127
Resp:	84983		
Ion Ratio	100	10.9	13.6
Lower		9.8	12.3
Upper		14.8	18.5



#10
Hexachlorobutadiene
Concen: 4.735 ng
RT: 10.714 min Scan# 977
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion	225	223	227
Resp:	17448		
Ion Ratio	100	0.0	63.6
Lower		0.0	50.3
Upper		0.0	75.5









#11
2-Methylnaphthalene-d10
Concen: 5.283 ng
RT: 11.966 min Scan# 1142
Delta R.T. -0.005 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 43118
Ion Ratio Lower Upper
152 100
151 21.3 17.1 25.7



#12
2-Methylnaphthalene
Concen: 5.358 ng
RT: 12.042 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion:142 Resp: 58095
Ion Ratio Lower Upper
142 100
141 90.7 74.6 111.8
115 47.2 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.235 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 3467
Ion Ratio Lower Upper
164 100
162 108.9 85.5 128.3
160 55.0 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 5.789 ng
RT: 15.730 min Scan# 1599
Delta R.T. -0.012 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion:330 Resp: 8081
Ion Ratio Lower Upper
330 100
332 96.5 77.1 115.7
141 66.9 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 5.056 ng
RT: 12.858 min Scan# 1318
Delta R.T. -0.005 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 74736
Ion Ratio Lower Upper
172 100
171 36.0 29.6 44.4
170 24.4 20.3 30.5



#16
Acenaphthylene
Concen: 5.291 ng
RT: 13.957 min Scan# 1436
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion:152 Resp: 89938
Ion Ratio Lower Upper
152 100
151 20.1 16.3 24.5
153 12.9 10.6 15.8









#17
Acenaphthene
Concen: 5.184 ng
RT: 14.299 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 57215
Ion Ratio Lower Upper
154 100
153 116.0 93.8 140.8
152 58.8 50.5 75.7



#18
Fluorene
Concen: 5.233 ng
RT: 15.293 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion:166 Resp: 75931
Ion Ratio Lower Upper
166 100
165 99.5 81.1 121.7
167 13.4 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:188 Resp: 6764

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 13.2 11.3 16.9



#20

4,6-Dinitro-2-methylphenol

Concen: 5.124 ng

RT: 15.378 min Scan# 1569

Delta R.T. -0.011 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14




Tgt Ion:198 Resp: 9612


Ion Ratio Lower Upper

198 100

51 51.5 125.2 187.8#

105 35.5 57.1 85.7#









#21
4-Bromophenyl-phenylether
Concen: 5.173 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:248 Resp: 22933
Ion Ratio Lower Upper
248 100
250 99.0 76.1 114.1
141 54.8 60.1 90.1#



#22
Hexachlorobenzene
Concen: 4.847 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion:284 Resp: 23190
Ion Ratio Lower Upper
284 100
142 53.5 44.0 66.0
249 36.6 29.7 44.5









#23
Atrazine
Concen: 5.696 ng
RT: 16.463 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:200 Resp: 20848
Ion Ratio Lower Upper
200 100
173 25.1 28.1 42.1#
215 42.7 39.3 58.9



#24
Pentachlorophenol
Concen: 5.085 ng
RT: 16.649 min Scan# 1673
Delta R.T. -0.012 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion:266 Resp: 13952
Ion Ratio Lower Upper
266 100
264 63.2 49.3 73.9
268 62.7 49.0 73.4









#25
Phenanthrene
Concen: 5.252 ng
RT: 17.021 min Scan# 1703
Delta R.T. -0.012 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 115105
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.1 12.3 18.5



#26
Anthracene
Concen: 5.569 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion:178 Resp: 111362
Ion Ratio Lower Upper
178 100
176 19.1 15.2 22.8
179 15.1 12.9 19.3





#27

Fluoranthene-d10

Concen: 5.482 ng

RT: 19.026 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:212 Resp: 94210

Ion Ratio Lower Upper

212 100

106 14.3 10.6 15.8

104 8.6 6.6 9.8



#28

Fluoranthene

Concen: 5.604 ng

RT: 19.054 min Scan# 1940

Delta R.T. -0.000 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14




Tgt Ion:202 Resp: 135672


Ion Ratio Lower Upper

202 100

101 11.4 8.7 13.1

203 17.2 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.180 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:240 Resp: 5784

Ion Ratio Lower Upper

240 100

120 18.0 9.0 13.4#

236 30.0 23.0 34.4



#30

Pyrene

Concen: 4.827 ng

RT: 19.416 min Scan# 2018

Delta R.T. -0.005 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14




Tgt Ion:202 Resp: 136281


Ion Ratio Lower Upper

202 100

200 21.5 17.0 25.6

203 17.6 14.2 21.4









#31
Terphenyl-d14
Concen: 4.900 ng
RT: 19.630 min Scan# 2064
Delta R.T. -0.005 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:244 Resp: 66718
Ion Ratio Lower Upper
244 100
212 10.7 10.0 15.0
122 16.5 13.2 19.8



#32
Benzo(a)anthracene
Concen: 5.423 ng
RT: 21.171 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BN037149.D
Acq: 03 Jun 2025 15:14

Tgt Ion:228 Resp: 113532
Ion Ratio Lower Upper
228 100
226 27.3 22.6 33.8
229 19.7 16.2 24.2





#33

Chrysene

Concen: 4.825 ng

RT: 21.216 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:228 Resp: 112475

Ion Ratio Lower Upper

228 100

226 31.8 25.2 37.8

229 19.1 16.8 25.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 5.486 ng

RT: 21.117 min Scan# 2269

Delta R.T. -0.000 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14




Tgt Ion:149 Resp: 72476


Ion Ratio Lower Upper

149 100

167 25.8 21.0 31.4

279 2.6 2.9 4.3#





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.374 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:264 Resp: 4620

Ion Ratio Lower Upper

264 100

260 26.9 22.1 33.1

265 62.7 55.8 83.8



#36

Indeno(1,2,3-cd)pyrene

Concen: 5.331 ng

RT: 25.573 min Scan# 3496

Delta R.T. -0.000 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14




Tgt Ion:276 Resp: 97988


Ion Ratio Lower Upper

276 100

138 29.9 21.0 31.6

277 25.0 19.4 29.2





#37

Benzo(b)fluoranthene

Concen: 5.516 ng

RT: 22.720 min Scan# 2520

Delta R.T. -0.003 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:252 Resp: 102873

Ion Ratio Lower Upper

252 100

253 22.0 22.3 33.5#

125 12.1 13.2 19.8#



#38

Benzo(k)fluoranthene

Concen: 5.475 ng

RT: 22.763 min Scan# 2535

Delta R.T. -0.000 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14




Tgt Ion:252 Resp: 104252


Ion Ratio Lower Upper

252 100

253 21.8 22.2 33.4#

125 12.0 13.2 19.8#





#39

Benzo(a)pyrene

Concen: 5.474 ng

RT: 23.278 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN037149.D


Acq: 03 Jun 2025 15:14

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:252 Resp: 85510

Ion Ratio Lower Upper

252 100

253 21.9 25.0 37.4#

125 13.3 17.0 25.6#



#40

Dibenzo(a,h)anthracene

Concen: 5.411 ng

RT: 25.585 min Scan# 3500

Delta R.T. -0.003 min

Lab File: BN037149.D

Acq: 03 Jun 2025 15:14




Tgt Ion:278 Resp: 76694


Ion Ratio Lower Upper

278 100

139 19.8 17.6 26.4

279 23.7 26.0 39.0#





#41

Benzo(g,h,i)perylene

Concen: 5.054 ng

RT: 26.243 min Scan# 31

Delta R.T. 0.003 min

Lab File: BN037149.D


Acq: 03 Jun 2025 15:14

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:276 Resp: 82277

Ion Ratio Lower Upper

276 100

277 23.7 20.9 31.3

138 26.0 20.8 31.2



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037150.D
 Acq On : 03 Jun 2025 15:53
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060325

Quant Time: Jun 04 01:54:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.597	152	2254	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	5892	0.400	ng	0.00
13) Acenaphthene-d10	14.245	164	3320	0.400	ng	0.01
19) Phenanthrene-d10	16.984	188	6332	0.400	ng	0.00
29) Chrysene-d12	21.189	240	4292	0.400	ng	0.00
35) Perylene-d12	23.377	264	3564	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	2014	0.361	ng	0.00
5) Phenol-d6	6.773	99	2419	0.358	ng	0.00
8) Nitrobenzene-d5	8.739	82	2473	0.398	ng	0.00
11) 2-Methylnaphthalene-d10	11.971	152	3266	0.398	ng	0.00
14) 2,4,6-Tribromophenol	15.743	330	440	0.329	ng	0.00
15) 2-Fluorobiphenyl	12.858	172	5696	0.402	ng	0.00
27) Fluoranthene-d10	19.026	212	6037	0.375	ng	0.00
31) Terphenyl-d14	19.635	244	4196	0.415	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.119	88	1242	0.413	ng	98
3) n-Nitrosodimethylamine	3.437	42	2355	0.390	ng	100
6) bis(2-Chloroethyl)ether	7.026	93	2715	0.421	ng	100
9) Naphthalene	10.426	128	6699	0.394	ng	98
10) Hexachlorobutadiene	10.714	225	1489	0.402	ng	# 99
12) 2-Methylnaphthalene	12.047	142	3839	0.352	ng	96
16) Acenaphthylene	13.957	152	6575	0.404	ng	100
17) Acenaphthene	14.299	154	3940	0.373	ng	100
18) Fluorene	15.293	166	5182	0.373	ng	99
20) 4,6-Dinitro-2-methylph...	15.389	198	348	0.462	ng	84
21) 4-Bromophenyl-phenylether	16.189	248	1544	0.372	ng	98
22) Hexachlorobenzene	16.301	284	1743	0.389	ng	99
23) Atrazine	16.463	200	1282	0.374	ng	97
24) Pentachlorophenol	16.661	266	591	0.436	ng	97
25) Phenanthrene	17.033	178	7869	0.384	ng	100
26) Anthracene	17.120	178	7105	0.380	ng	99
28) Fluoranthene	19.054	202	8186	0.361	ng	99
30) Pyrene	19.421	202	8148	0.389	ng	99
32) Benzo(a)anthracene	21.171	228	5949	0.383	ng	99
33) Chrysene	21.224	228	6661	0.385	ng	99
34) Bis(2-ethylhexyl)phtha...	21.117	149	3635	0.371	ng	99
36) Indeno(1,2,3-cd)pyrene	25.573	276	5461	0.385	ng	100
37) Benzo(b)fluoranthene	22.720	252	5432	0.378	ng	100
38) Benzo(k)fluoranthene	22.763	252	5859	0.399	ng	100
39) Benzo(a)pyrene	23.281	252	4925	0.409	ng	99
40) Dibenzo(a,h)anthracene	25.588	278	4227	0.387	ng	100
41) Benzo(g,h,i)perylene	26.248	276	4620	0.368	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037150.D
Acq On : 03 Jun 2025 15:53
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Quant Time: Jun 04 01:54:08 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:52:03 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.597 min Scan# 61
Delta R.T. 0.007 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Instrument :
BNA_N
ClientSampleId :
ICVBN060325







Tgt Ion:152 Resp: 2254
Ion Ratio Lower Upper
152 100
150 149.2 123.2 184.8
115 65.5 56.6 85.0



#2
1,4-Dioxane
Concen: 0.413 ng
RT: 3.119 min Scan# 33
Delta R.T. -0.007 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion: 88 Resp: 1242
Ion Ratio Lower Upper
88 100
43 50.8 43.5 65.3
58 84.5 67.7 101.5









#3
n-Nitrosodimethylamine
Concen: 0.390 ng
RT: 3.437 min Scan# 77
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion: 42 Resp: 2355
Ion Ratio Lower Upper
42 100
74 66.5 53.0 79.4
44 6.9 5.9 8.9



#4
2-Fluorophenol
Concen: 0.361 ng
RT: 5.184 min Scan# 319
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion: 112 Resp: 2014
Ion Ratio Lower Upper
112 100
64 72.0 56.3 84.5
63 42.3 36.2 54.4









#5
Phenol-d6
Concen: 0.358 ng
RT: 6.773 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion	Ratio	Lower	Upper
99	100		
42	38.5	31.3	46.9
71	45.7	38.2	57.2



#6
bis(2-Chloroethyl)ether
Concen: 0.421 ng
RT: 7.026 min Scan# 574
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion	Ratio	Lower	Upper
93	100		
63	85.3	68.6	103.0
95	30.5	24.3	36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion:	136	Resp:	5892
Ion	Ratio	Lower	Upper
136	100		
137	11.9	9.7	14.5
54	11.7	9.7	14.5
68	7.0	5.4	8.2



#8
Nitrobenzene-d5
Concen: 0.398 ng
RT: 8.739 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion:	82	Resp:	2473
Ion	Ratio	Lower	Upper
82	100		
128	33.6	26.9	40.3
54	71.8	61.4	92.2









#9
Naphthalene
Concen: 0.394 ng
RT: 10.426 min Scan# 949
Delta R.T. 0.011 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion:128 Resp: 6699
Ion Ratio Lower Upper
128 100
129 12.0 9.8 14.8
127 14.0 12.3 18.5



#10
Hexachlorobutadiene
Concen: 0.402 ng
RT: 10.714 min Scan# 977
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion:225 Resp: 1489
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.398 ng
RT: 11.971 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion:152 Resp: 3266
Ion Ratio Lower Upper
152 100
151 21.7 17.1 25.7



#12
2-Methylnaphthalene
Concen: 0.352 ng
RT: 12.047 min Scan# 1158
Delta R.T. 0.005 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion:142 Resp: 3839
Ion Ratio Lower Upper
142 100
141 90.3 74.6 111.8
115 48.0 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.245 min Scan# 1463
Delta R.T. 0.011 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion:164 Resp: 3320
Ion Ratio Lower Upper
164 100
162 103.3 85.5 128.3
160 52.8 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 0.329 ng
RT: 15.743 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion:330 Resp: 440
Ion Ratio Lower Upper
330 100
332 96.1 77.1 115.7
141 55.9 46.4 69.6





#15

2-Fluorobiphenyl

Concen: 0.402 ng

RT: 12.858 min Scan# 1318

Delta R.T. -0.005 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53

Instrument :

BNA_N

ClientSampleId :

ICVBN060325





Tgt Ion:172 Resp: 5696

Ion Ratio Lower Upper

172 100

171 37.1 29.6 44.4

170 25.1 20.3 30.5



#16

Acenaphthylene

Concen: 0.404 ng

RT: 13.957 min Scan# 1436

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53




Tgt Ion:152 Resp: 6575


Ion Ratio Lower Upper

152 100

151 20.3 16.3 24.5

153 13.1 10.6 15.8





#17

Acenaphthene

Concen: 0.373 ng

RT: 14.299 min Scan# 1468

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53

Instrument :

BNA_N

ClientSampleId :

ICVBN060325





Tgt Ion:154 Resp: 3940

Ion Ratio Lower Upper

154 100

153 118.0 93.8 140.8

152 63.2 50.5 75.7



#18

Fluorene

Concen: 0.373 ng

RT: 15.293 min Scan# 1561

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53




Tgt Ion:166 Resp: 5182


Ion Ratio Lower Upper

166 100

165 100.6 81.1 121.7

167 13.2 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53

Instrument :

BNA_N

ClientSampleId :

ICVBN060325





Tgt Ion:188 Resp: 6332

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 15.8 11.3 16.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.462 ng

RT: 15.389 min Scan# 1570

Delta R.T. 0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53




Tgt Ion:198 Resp: 348


Ion Ratio Lower Upper

198 100

51 132.7 125.2 187.8

105 62.8 57.1 85.7









#21
4-Bromophenyl-phenylether
Concen: 0.372 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion	Ratio	Lower	Upper
248	100		
250	96.0	76.1	114.1
141	77.6	60.1	90.1



#22
Hexachlorobenzene
Concen: 0.389 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion	Ratio	Lower	Upper
284	100		
142	53.8	44.0	66.0
249	37.4	29.7	44.5









#23
Atrazine
Concen: 0.374 ng
RT: 16.463 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion	Ratio	Lower	Upper
200	100		
173	35.5	28.1	42.1
215	46.5	39.3	58.9



#24
Pentachlorophenol
Concen: 0.436 ng
RT: 16.661 min Scan# 1674
Delta R.T. 0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion	Ratio	Lower	Upper
266	100		
264	60.2	49.3	73.9
268	64.0	49.0	73.4









#25
Phenanthrene
Concen: 0.384 ng
RT: 17.033 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion:178 Resp: 7869
Ion Ratio Lower Upper
178 100
176 19.5 15.7 23.5
179 15.7 12.3 18.5



#26
Anthracene
Concen: 0.380 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion:178 Resp: 7105
Ion Ratio Lower Upper
178 100
176 19.0 15.2 22.8
179 15.5 12.9 19.3





#27

Fluoranthene-d10

Concen: 0.375 ng

RT: 19.026 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53

Instrument :

BNA_N

ClientSampleId :

ICVBN060325





Tgt Ion:212 Resp: 6037

Ion Ratio Lower Upper

212 100

106 13.7 10.6 15.8

104 8.2 6.6 9.8



#28

Fluoranthene

Concen: 0.361 ng

RT: 19.054 min Scan# 1940

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53




Tgt Ion:202 Resp: 8186


Ion Ratio Lower Upper

202 100

101 11.4 8.7 13.1

203 16.7 13.5 20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.189 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53

Instrument :

BNA_N

ClientSampleId :

ICVBN060325





Tgt Ion:240 Resp: 4292

Ion Ratio Lower Upper

240 100

120 11.1 9.0 13.4

236 28.5 23.0 34.4



#30

Pyrene

Concen: 0.389 ng

RT: 19.421 min Scan# 2019

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53




Tgt Ion:202 Resp: 8148


Ion Ratio Lower Upper

202 100

200 21.0 17.0 25.6

203 17.3 14.2 21.4









#31
Terphenyl-d14
Concen: 0.415 ng
RT: 19.635 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53




Instrument :
BNA_N
ClientSampleId :
ICVBN060325


Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.6	10.0	15.0
122	16.3	13.2	19.8



#32
Benzo(a)anthracene
Concen: 0.383 ng
RT: 21.171 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BN037150.D
Acq: 03 Jun 2025 15:53

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.6	22.6	33.8
229	20.5	16.2	24.2





#33

Chrysene

Concen: 0.385 ng

RT: 21.224 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53

Instrument :

BNA_N

ClientSampleId :

ICVBN060325





Tgt Ion:228 Resp: 6661

Ion Ratio Lower Upper

228 100

226 30.8 25.2 37.8

229 20.4 16.8 25.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.371 ng

RT: 21.117 min Scan# 2269

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53




Tgt Ion:149 Resp: 3635


Ion Ratio Lower Upper

149 100

167 25.6 21.0 31.4

279 3.4 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.377 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037150.D


Acq: 03 Jun 2025 15:53

Instrument :



BNA_N

ClientSampleId :

ICVBN060325



Tgt Ion:	264	Resp:	3564
Ion	Ratio	Lower	Upper
264	100		
260	27.5	22.1	33.1
265	74.5	55.8	83.8



#36

Indeno(1,2,3-cd)pyrene


Concen: 0.385 ng

RT: 25.573 min Scan# 3496



Delta R.T. -0.000 min


Lab File: BN037150.D

Acq: 03 Jun 2025 15:53



Tgt Ion:	276	Resp:	5461
Ion	Ratio	Lower	Upper
276	100		
138	26.2	21.0	31.6
277	24.3	19.4	29.2





#37

Benzo(b)fluoranthene

Concen: 0.378 ng

RT: 22.720 min Scan# 2520

Delta R.T. -0.003 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53

Instrument :

BNA_N

ClientSampleId :

ICVBN060325





Tgt Ion:252 Resp: 5432

Ion Ratio Lower Upper

252 100

253 27.8 22.3 33.5

125 16.7 13.2 19.8



#38

Benzo(k)fluoranthene

Concen: 0.399 ng

RT: 22.763 min Scan# 2535

Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53




Tgt Ion:252 Resp: 5859


Ion Ratio Lower Upper

252 100

253 27.6 22.2 33.4

125 16.7 13.2 19.8





#39

Benzo(a)pyrene

Concen: 0.409 ng

RT: 23.281 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037150.D


Acq: 03 Jun 2025 15:53

Instrument :



BNA_N

ClientSampleId :

ICVBN060325



Tgt Ion:	252	Resp:	4925
Ion Ratio	Lower	Upper	
252	100		
253	30.9	25.0	37.4
125	20.8	17.0	25.6



#40

Dibenzo(a,h)anthracene

Concen: 0.387 ng




RT: 25.588 min Scan# 3501


Delta R.T. -0.000 min

Lab File: BN037150.D

Acq: 03 Jun 2025 15:53

Tgt Ion:	278	Resp:	4227
Ion Ratio	Lower	Upper	
278	100		
139	21.7	17.6	26.4
279	32.3	26.0	39.0





#41

Benzo(g,h,i)perylene

Concen: 0.368 ng

RT: 26.248 min Scan# 31

Delta R.T. 0.009 min

Lab File: BN037150.D


Acq: 03 Jun 2025 15:53

Instrument :

BNA_N

ClientSampleId :

ICVBN060325





Tgt Ion:276 Resp: 4620

Ion Ratio Lower Upper

276 100

277 27.0 20.9 31.3

138 26.8 20.8 31.2



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037150.D
 Acq On : 03 Jun 2025 15:53
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICBN060325

Quant Time: Jun 04 01:54:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	107	0.00
2	1,4-Dioxane	0.533	0.551	-3.4	116	0.00
3	n-Nitrosodimethylamine	1.071	1.045	2.4	106	0.00
4 S	2-Fluorophenol	0.989	0.894	9.6	102	0.00
5 S	Phenol-d6	1.199	1.073	10.5	102	0.00
6	bis(2-Chloroethyl)ether	1.144	1.205	-5.3	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	113	0.00
8 S	Nitrobenzene-d5	0.422	0.420	0.5	113	0.00
9	Naphthalene	1.154	1.137	1.5	115	0.01
10	Hexachlorobutadiene	0.251	0.253	-0.8	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.557	0.554	0.5	111	0.00
12	2-Methylnaphthalene	0.740	0.652	11.9	106	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.01
14 S	2,4,6-Tribromophenol	0.161	0.133	17.4	105	0.00
15 S	2-Fluorobiphenyl	1.705	1.716	-0.6	121	0.00
16	Acenaphthylene	1.961	1.980	-1.0	129	0.00
17	Acenaphthene	1.273	1.187	6.8	118	0.00
18	Fluorene	1.674	1.561	6.8	118	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4,6-Dinitro-2-methylphenol	0.077	0.055	28.6#	128	0.00
21	4-Bromophenyl-phenylether	0.262	0.244	6.9	116	0.00
22	Hexachlorobenzene	0.283	0.275	2.8	119	0.00
23	Atrazine	0.216	0.202	6.5	126	0.00
24	Pentachlorophenol	0.124	0.093	25.0	118	0.00
25	Phenanthrene	1.296	1.243	4.1	121	0.00
26	Anthracene	1.183	1.122	5.2	126	0.00
27 SURR	Fluoranthene-d10	1.016	0.953	6.2	114	0.00
28	Fluoranthene	1.432	1.293	9.7	118	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	114	0.00
30	Pyrene	1.953	1.898	2.8	118	0.00
31 S	Terphenyl-d14	0.942	0.978	-3.8	124	0.00
32	Benzo(a)anthracene	1.448	1.386	4.3	122	0.00
33	Chrysene	1.612	1.552	3.7	120	0.00
34	Bis(2-ethylhexyl)phthalate	0.914	0.847	7.3	125	0.00
35 I	Perylene-d12	1.000	1.000	0.0	105	0.00
36	Indeno(1,2,3-cd)pyrene	1.591	1.532	3.7	107	0.00
37	Benzo(b)fluoranthene	1.615	1.524	5.6	113	0.00
38	Benzo(k)fluoranthene	1.648	1.644	0.2	118	0.00
39 C	Benzo(a)pyrene	1.352	1.382	-2.2	119	0.00
40	Dibenzo(a,h)anthracene	1.227	1.186	3.3	108	0.00
41	Benzo(g,h,i)perylene	1.410	1.296	8.1	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN060325

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
 Data File : BN037150.D
 Acq On : 03 Jun 2025 15:53
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN060325

Quant Time: Jun 04 01:54:08 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	107	0.00
2	1,4-Dioxane	0.400	0.413	-3.2	116	0.00
3	n-Nitrosodimethylamine	0.400	0.390	2.5	106	0.00
4 S	2-Fluorophenol	0.400	0.361	9.8	102	0.00
5 S	Phenol-d6	0.400	0.358	10.5	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.421	-5.2	115	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	113	0.00
8 S	Nitrobenzene-d5	0.400	0.398	0.5	113	0.00
9	Naphthalene	0.400	0.394	1.5	115	0.01
10	Hexachlorobutadiene	0.400	0.402	-0.5	109	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	111	0.00
12	2-Methylnaphthalene	0.400	0.352	12.0	106	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	115	0.01
14 S	2,4,6-Tribromophenol	0.400	0.329	17.8	105	0.00
15 S	2-Fluorobiphenyl	0.400	0.402	-0.5	121	0.00
16	Acenaphthylene	0.400	0.404	-1.0	129	0.00
17	Acenaphthene	0.400	0.373	6.8	118	0.00
18	Fluorene	0.400	0.373	6.8	118	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	116	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.462	-15.5	128	0.00
21	4-Bromophenyl-phenylether	0.400	0.372	7.0	116	0.00
22	Hexachlorobenzene	0.400	0.389	2.8	119	0.00
23	Atrazine	0.400	0.374	6.5	126	0.00
24	Pentachlorophenol	0.400	0.436	-9.0	118	0.00
25	Phenanthrene	0.400	0.384	4.0	121	0.00
26	Anthracene	0.400	0.380	5.0	126	0.00
27 SURR	Fluoranthene-d10	0.400	0.375	6.3	114	0.00
28	Fluoranthene	0.400	0.361	9.8	118	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	114	0.00
30	Pyrene	0.400	0.389	2.8	118	0.00
31 S	Terphenyl-d14	0.400	0.415	-3.7	124	0.00
32	Benzo(a)anthracene	0.400	0.383	4.3	122	0.00
33	Chrysene	0.400	0.385	3.8	120	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.371	7.3	125	0.00
35 I	Perylene-d12	0.400	0.400	0.0	105	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.385	3.8	107	0.00
37	Benzo(b)fluoranthene	0.400	0.378	5.5	113	0.00
38	Benzo(k)fluoranthene	0.400	0.399	0.3	118	0.00
39 C	Benzo(a)pyrene	0.400	0.409	-2.2	119	0.00
40	Dibenzo(a,h)anthracene	0.400	0.387	3.3	108	0.00
41	Benzo(g,h,i)perylene	0.400	0.368	8.0	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN060325



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG No.: Q2209

Instrument ID: BNA_N Calibration Date/Time: 06/09/2025 10:54

Lab File ID: BN037189.D Init. Calib. Date(s): 06/03/2025 06/03/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:39 15:14

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.557	0.558		0.2	20.0
Fluoranthene-d10	1.016	0.960		-5.5	20.0
2-Fluorophenol	0.989	0.924		-6.6	20.0
Phenol-d6	1.199	1.124		-6.3	20.0
Nitrobenzene-d5	0.422	0.422		0.0	20.0
2-Fluorobiphenyl	1.705	1.691		-0.8	20.0
2,4,6-Tribromophenol	0.161	0.142		-11.8	20.0
Terphenyl-d14	0.942	0.909		-3.5	20.0
Benzo(a)anthracene	1.448	1.282		-11.5	20.0
Benzo(b)fluoranthene	1.615	1.446		-10.5	20.0
Benzo(k)fluoranthene	1.648	1.440		-12.6	20.0
Benzo(a)pyrene	1.352	1.193		-11.8	20.0
Benzo(g,h,i)perylene	1.410	1.306		-7.4	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037189.D
 Acq On : 09 Jun 2025 10:54
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Jun 09 11:21:12 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	2093	0.400	ng	0.00
7) Naphthalene-d8	10.362	136	5342	0.400	ng	#-0.01
13) Acenaphthene-d10	14.235	164	2894	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	5308	0.400	ng	0.00
29) Chrysene-d12	21.180	240	3516	0.400	ng	# 0.00
35) Perylene-d12	23.377	264	3185	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	1933	0.373	ng	0.00
5) Phenol-d6	6.766	99	2353	0.375	ng	0.00
8) Nitrobenzene-d5	8.739	82	2254	0.400	ng	0.00
11) 2-Methylnaphthalene-d10	11.966	152	2980	0.401	ng	0.00
14) 2,4,6-Tribromophenol	15.730	330	410	0.352	ng	-0.01
15) 2-Fluorobiphenyl	12.858	172	4893	0.397	ng	0.00
27) Fluoranthene-d10	19.026	212	5096	0.378	ng	0.00
31) Terphenyl-d14	19.630	244	3196	0.386	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.119	88	1086	0.389	ng	96
3) n-Nitrosodimethylamine	3.437	42	2219	0.396	ng	99
6) bis(2-Chloroethyl)ether	7.019	93	2354	0.393	ng	98
9) Naphthalene	10.415	128	6000	0.389	ng	99
10) Hexachlorobutadiene	10.703	225	1350	0.402	ng	# 100
12) 2-Methylnaphthalene	12.042	142	3683	0.373	ng	97
16) Acenaphthylene	13.957	152	5056	0.356	ng	99
17) Acenaphthene	14.299	154	3370	0.366	ng	100
18) Fluorene	15.293	166	4348	0.359	ng	99
20) 4,6-Dinitro-2-methylph...	15.378	198	331	0.487	ng	94
21) 4-Bromophenyl-phenylether	16.189	248	1282	0.369	ng	# 88
22) Hexachlorobenzene	16.301	284	1488	0.396	ng	98
23) Atrazine	16.463	200	1020	0.355	ng	96
24) Pentachlorophenol	16.649	266	645	0.502	ng	93
25) Phenanthrene	17.021	178	6261	0.364	ng	100
26) Anthracene	17.120	178	5447	0.347	ng	99
28) Fluoranthene	19.054	202	6641	0.350	ng	99
30) Pyrene	19.416	202	6586	0.384	ng	100
32) Benzo(a)anthracene	21.171	228	4507	0.354	ng	98
33) Chrysene	21.224	228	5182	0.366	ng	99
34) Bis(2-ethylhexyl)phtha...	21.108	149	2880	0.359	ng	# 98
36) Indeno(1,2,3-cd)pyrene	25.573	276	4552	0.359	ng	98
37) Benzo(b)fluoranthene	22.722	252	4604	0.358	ng	99
38) Benzo(k)fluoranthene	22.766	252	4587	0.349	ng	98
39) Benzo(a)pyrene	23.284	252	3801	0.353	ng	97
40) Dibenzo(a,h)anthracene	25.591	278	3442	0.352	ng	98
41) Benzo(g,h,i)perylene	26.243	276	4160	0.371	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037189.D
Acq On : 09 Jun 2025 10:54
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Jun 09 11:21:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:52:03 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.589 min Scan# 61
Delta R.T. -0.001 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 2093
Ion Ratio Lower Upper
152 100
150 148.6 123.2 184.8
115 67.6 56.6 85.0




#2
1,4-Dioxane
Concen: 0.389 ng
RT: 3.119 min Scan# 33
Delta R.T. -0.007 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion: 88 Resp: 1086
Ion Ratio Lower Upper
88 100
43 48.8 43.5 65.3
58 82.5 67.7 101.5



Abundance Scan 77 (3.437 min): BN037145.D\data.ms (-73)



#3

n-Nitrosodimethylamine

Concen: 0.396 ng

RT: 3.437 min Scan# 77

Delta R.T. -0.000 min

Lab File: BN037189.D

Acq: 09 Jun 2025 10:54

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Tgt Ion: 42 Resp: 2219


Ion Ratio Lower Upper

42 100


74 65.1 53.0 79.4

44 6.7 5.9 8.9


Abundance Scan 77 (3.437 min): BN037189.D\data.ms




Abundance



Abundance Scan 77 (3.437 min): BN037189.D\data.ms (-56)



Abundance Scan 319 (5.184 min): BN037145.D\data.ms (-31)



#4

2-Fluorophenol

Concen: 0.373 ng

RT: 5.184 min Scan# 319

Delta R.T. -0.000 min

Lab File: BN037189.D

Acq: 09 Jun 2025 10:54

Tgt Ion: 112 Resp: 1933


Ion Ratio Lower Upper

112 100


64 70.3 56.3 84.5

63 45.2 36.2 54.4


Abundance Scan 319 (5.184 min): BN037189.D\data.ms




Abundance



Abundance Scan 319 (5.184 min): BN037189.D\data.ms (-29)









#5
Phenol-d6
Concen: 0.375 ng
RT: 6.766 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	99	Resp:	2353
Ion	Ratio	Lower	Upper
99	100		
42	38.1	31.3	46.9
71	47.1	38.2	57.2



#6
bis(2-Chloroethyl)ether
Concen: 0.393 ng
RT: 7.019 min Scan# 573
Delta R.T. -0.007 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:	93	Resp:	2354
Ion	Ratio	Lower	Upper
93	100		
63	87.8	68.6	103.0
95	31.6	24.3	36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.362 min Scan# 945
Delta R.T. -0.011 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	5342
Ion	Ratio	Lower	Upper
136	100		
137	12.7	9.7	14.5
54	15.6	9.7	14.5#
68	8.2	5.4	8.2#



#8
Nitrobenzene-d5
Concen: 0.400 ng
RT: 8.739 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:	82	Resp:	2254
Ion	Ratio	Lower	Upper
82	100		
128	36.9	26.9	40.3
54	70.1	61.4	92.2









#9
Naphthalene
Concen: 0.389 ng
RT: 10.415 min Scan# 949
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 6000
Ion Ratio Lower Upper
128 100
129 12.0 9.8 14.8
127 15.2 12.3 18.5



#10
Hexachlorobutadiene
Concen: 0.402 ng
RT: 10.703 min Scan# 976
Delta R.T. -0.011 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:225 Resp: 1350
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.2 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.401 ng
RT: 11.966 min Scan# 1142
Delta R.T. -0.005 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 2980
Ion Ratio Lower Upper
152 100
151 22.1 17.1 25.7



#12
2-Methylnaphthalene
Concen: 0.373 ng
RT: 12.042 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:142 Resp: 3683
Ion Ratio Lower Upper
142 100
141 90.6 74.6 111.8
115 49.3 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.235 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 2894
Ion Ratio Lower Upper
164 100
162 107.7 85.5 128.3
160 55.2 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 0.352 ng
RT: 15.730 min Scan# 1599
Delta R.T. -0.012 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:330 Resp: 410
Ion Ratio Lower Upper
330 100
332 97.8 77.1 115.7
141 57.6 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 0.397 ng
RT: 12.858 min Scan# 1318
Delta R.T. -0.005 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.6	29.6	44.4
170	25.2	20.3	30.5



#16
Acenaphthylene
Concen: 0.356 ng
RT: 13.957 min Scan# 1436
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.2	16.3	24.5
153	12.9	10.6	15.8









#17
Acenaphthene
Concen: 0.366 ng
RT: 14.299 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 3370
Ion Ratio Lower Upper
154 100
153 117.2 93.8 140.8
152 62.5 50.5 75.7



#18
Fluorene
Concen: 0.359 ng
RT: 15.293 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:166 Resp: 4348
Ion Ratio Lower Upper
166 100
165 99.8 81.1 121.7
167 13.2 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.984 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	13.1	11.3	16.9



#20
4,6-Dinitro-2-methylphenol
Concen: 0.487 ng
RT: 15.378 min Scan# 1569
Delta R.T. -0.011 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion	Ratio	Lower	Upper
198	100		
51	144.9	125.2	187.8
105	71.9	57.1	85.7









#21
4-Bromophenyl-phenylether
Concen: 0.369 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 1282
Ion Ratio Lower Upper
248 100
250 102.6 76.1 114.1
141 59.8 60.1 90.1#



#22
Hexachlorobenzene
Concen: 0.396 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:284 Resp: 1488
Ion Ratio Lower Upper
284 100
142 55.6 44.0 66.0
249 35.3 29.7 44.5









#23
Atrazine
Concen: 0.355 ng
RT: 16.463 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
200	100		
173	30.6	28.1	42.1
215	48.7	39.3	58.9



#24
Pentachlorophenol
Concen: 0.502 ng
RT: 16.649 min Scan# 1673
Delta R.T. -0.012 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion	Ratio	Lower	Upper
266	100		
264	64.5	49.3	73.9
268	68.5	49.0	73.4









#25
Phenanthrene
Concen: 0.364 ng
RT: 17.021 min Scan# 1703
Delta R.T. -0.012 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 6261
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.2 12.3 18.5



#26
Anthracene
Concen: 0.347 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:178 Resp: 5447
Ion Ratio Lower Upper
178 100
176 19.3 15.2 22.8
179 15.1 12.9 19.3









#27
Fluoranthene-d10
Concen: 0.378 ng
RT: 19.026 min Scan# 1934
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
212	100		
106	13.2	10.6	15.8
104	8.2	6.6	9.8



#28
Fluoranthene
Concen: 0.350 ng
RT: 19.054 min Scan# 1940
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion	Ratio	Lower	Upper
202	100		
101	10.2	8.7	13.1
203	17.0	13.5	20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.180 min Scan# 2118
Delta R.T. -0.009 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 3516
Ion Ratio Lower Upper
240 100
120 18.2 9.0 13.4#
236 30.7 23.0 34.4



#30
Pyrene
Concen: 0.384 ng
RT: 19.416 min Scan# 2018
Delta R.T. -0.005 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54


Tgt Ion:202 Resp: 6586
Ion Ratio Lower Upper
202 100
200 21.2 17.0 25.6
203 17.8 14.2 21.4








#31
Terphenyl-d14
Concen: 0.386 ng
RT: 19.630 min Scan# 2064
Delta R.T. -0.005 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:244 Resp: 3196
Ion Ratio Lower Upper
244 100
212 12.4 10.0 15.0
122 18.3 13.2 19.8



#32
Benzo(a)anthracene
Concen: 0.354 ng
RT: 21.171 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:228 Resp: 4507
Ion Ratio Lower Upper
228 100
226 28.9 22.6 33.8
229 21.0 16.2 24.2





#33

Chrysene

Concen: 0.366 ng

RT: 21.224 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037189.D

Acq: 09 Jun 2025 10:54

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:228 Resp: 5182

Ion Ratio Lower Upper

228 100

226 31.0 25.2 37.8

229 20.8 16.8 25.2



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.359 ng

RT: 21.108 min Scan# 2268

Delta R.T. -0.009 min

Lab File: BN037189.D

Acq: 09 Jun 2025 10:54




Tgt Ion:149 Resp: 2880


Ion Ratio Lower Upper

149 100

167 27.2 21.0 31.4

279 4.4 2.9 4.3#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.377 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	264	260	265
Resp:	3185		
Ratio	100	27.8	72.4
Lower		22.1	55.8
Upper		33.1	83.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.359 ng
RT: 25.573 min Scan# 3496
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion	276	138	277
Resp:	4552		
Ratio	100	24.2	24.5
Lower		21.0	19.4
Upper		31.6	29.2









#37
Benzo(b)fluoranthene
Concen: 0.358 ng
RT: 22.722 min Scan# 2521
Delta R.T. -0.000 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 4604
Ion Ratio Lower Upper
252 100
253 28.2 22.3 33.5
125 17.4 13.2 19.8



#38
Benzo(k)fluoranthene
Concen: 0.349 ng
RT: 22.766 min Scan# 2536
Delta R.T. 0.003 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Tgt Ion:252 Resp: 4587
Ion Ratio Lower Upper
252 100
253 29.3 22.2 33.4
125 16.6 13.2 19.8





#39

Benzo(a)pyrene

Concen: 0.353 ng

RT: 23.284 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN037189.D

Acq: 09 Jun 2025 10:54

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 3801

Ion Ratio Lower Upper

252 100

253 31.9 25.0 37.4

125 23.8 17.0 25.6



#40

Dibenzo(a,h)anthracene

Concen: 0.352 ng

RT: 25.591 min Scan# 3502

Delta R.T. 0.003 min

Lab File: BN037189.D

Acq: 09 Jun 2025 10:54




Tgt Ion:278 Resp: 3442


Ion Ratio Lower Upper

278 100

139 21.1 17.6 26.4


279 33.7 26.0 39.0







#41
Benzo(g,h,i)perylene
Concen: 0.371 ng
RT: 26.243 min Scan# 31
Delta R.T. 0.003 min
Lab File: BN037189.D
Acq: 09 Jun 2025 10:54

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion: 276 Resp: 4160

Ion	Ratio	Lower	Upper
276	100		
277	27.6	20.9	31.3
138	25.5	20.8	31.2



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037189.D
 Acq On : 09 Jun 2025 10:54
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 09 11:21:12 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.533	0.519	2.6	101	0.00
3	n-Nitrosodimethylamine	1.071	1.060	1.0	100	0.00
4 S	2-Fluorophenol	0.989	0.924	6.6	98	0.00
5 S	Phenol-d6	1.199	1.124	6.3	99	0.00
6	bis(2-Chloroethyl)ether	1.144	1.125	1.7	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	-0.01
8 S	Nitrobenzene-d5	0.422	0.422	0.0	103	0.00
9	Naphthalene	1.154	1.123	2.7	103	0.00
10	Hexachlorobutadiene	0.251	0.253	-0.8	99	-0.01
11 SURR	2-Methylnaphthalene-d10	0.557	0.558	-0.2	101	0.00
12	2-Methylnaphthalene	0.740	0.689	6.9	102	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.161	0.142	11.8	97	-0.01
15 S	2-Fluorobiphenyl	1.705	1.691	0.8	104	0.00
16	Acenaphthylene	1.961	1.747	10.9	99	0.00
17	Acenaphthene	1.273	1.164	8.6	101	0.00
18	Fluorene	1.674	1.502	10.3	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.077	0.062	19.5	122	-0.01
21	4-Bromophenyl-phenylether	0.262	0.242	7.6	96	0.00
22	Hexachlorobenzene	0.283	0.280	1.1	102	0.00
23	Atrazine	0.216	0.192	11.1	100	0.00
24	Pentachlorophenol	0.124	0.122	1.6	128	-0.01
25	Phenanthrene	1.296	1.180	9.0	96	-0.01
26	Anthracene	1.183	1.026	13.3	97	0.00
27 SURR	Fluoranthene-d10	1.016	0.960	5.5	96	0.00
28	Fluoranthene	1.432	1.251	12.6	96	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
30	Pyrene	1.953	1.873	4.1	96	0.00
31 S	Terphenyl-d14	0.942	0.909	3.5	95	0.00
32	Benzo(a)anthracene	1.448	1.282	11.5	93	0.00
33	Chrysene	1.612	1.474	8.6	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.914	0.819	10.4	99	0.00
35 I	Perylene-d12	1.000	1.000	0.0	94	0.00
36	Indeno(1,2,3-cd)pyrene	1.591	1.429	10.2	90	0.00
37	Benzo(b)fluoranthene	1.615	1.446	10.5	96	0.00
38	Benzo(k)fluoranthene	1.648	1.440	12.6	93	0.00
39 C	Benzo(a)pyrene	1.352	1.193	11.8	92	0.00
40	Dibenzo(a,h)anthracene	1.227	1.081	11.9	88	0.00
41	Benzo(g,h,i)perylene	1.410	1.306	7.4	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037189.D
 Acq On : 09 Jun 2025 10:54
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 09 11:21:12 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.389	2.8	101	0.00
3	n-Nitrosodimethylamine	0.400	0.396	1.0	100	0.00
4 S	2-Fluorophenol	0.400	0.373	6.8	98	0.00
5 S	Phenol-d6	0.400	0.375	6.3	99	0.00
6	bis(2-Chloroethyl)ether	0.400	0.393	1.8	99	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	102	-0.01
8 S	Nitrobenzene-d5	0.400	0.400	0.0	103	0.00
9	Naphthalene	0.400	0.389	2.8	103	0.00
10	Hexachlorobutadiene	0.400	0.402	-0.5	99	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.401	-0.3	101	0.00
12	2-Methylnaphthalene	0.400	0.373	6.8	102	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	100	0.00
14 S	2,4,6-Tribromophenol	0.400	0.352	12.0	97	-0.01
15 S	2-Fluorobiphenyl	0.400	0.397	0.8	104	0.00
16	Acenaphthylene	0.400	0.356	11.0	99	0.00
17	Acenaphthene	0.400	0.366	8.5	101	0.00
18	Fluorene	0.400	0.359	10.3	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	97	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.487	-21.7	122	-0.01
21	4-Bromophenyl-phenylether	0.400	0.369	7.8	96	0.00
22	Hexachlorobenzene	0.400	0.396	1.0	102	0.00
23	Atrazine	0.400	0.355	11.3	100	0.00
24	Pentachlorophenol	0.400	0.502	-25.5#	128	-0.01
25	Phenanthrene	0.400	0.364	9.0	96	-0.01
26	Anthracene	0.400	0.347	13.3	97	0.00
27 SURR	Fluoranthene-d10	0.400	0.378	5.5	96	0.00
28	Fluoranthene	0.400	0.350	12.5	96	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	93	0.00
30	Pyrene	0.400	0.384	4.0	96	0.00
31 S	Terphenyl-d14	0.400	0.386	3.5	95	0.00
32	Benzo(a)anthracene	0.400	0.354	11.5	93	0.00
33	Chrysene	0.400	0.366	8.5	93	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.359	10.3	99	0.00
35 I	Perylene-d12	0.400	0.400	0.0	94	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.359	10.3	90	0.00
37	Benzo(b)fluoranthene	0.400	0.358	10.5	96	0.00
38	Benzo(k)fluoranthene	0.400	0.349	12.8	93	0.00
39 C	Benzo(a)pyrene	0.400	0.353	11.8	92	0.00
40	Dibenzo(a,h)anthracene	0.400	0.352	12.0	88	0.00
41	Benzo(g,h,i)perylene	0.400	0.371	7.3	91	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG No.: Q2209

Instrument ID: BNA_N Calibration Date/Time: 06/09/2025 23:11

Lab File ID: BN037204.D Init. Calib. Date(s): 06/03/2025 06/03/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:39 15:14

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.557	0.554		-0.5	20.0
Fluoranthene-d10	1.016	0.919		-9.5	20.0
2-Fluorophenol	0.989	0.935		-5.5	20.0
Phenol-d6	1.199	1.161		-3.2	20.0
Nitrobenzene-d5	0.422	0.423		0.2	20.0
2-Fluorobiphenyl	1.705	1.685		-1.2	20.0
2,4,6-Tribromophenol	0.161	0.151		-6.2	20.0
Terphenyl-d14	0.942	0.919		-2.4	20.0
Benzo(a)anthracene	1.448	1.292		-10.8	20.0
Benzo(b)fluoranthene	1.615	1.383		-14.4	20.0
Benzo(k)fluoranthene	1.648	1.487		-9.8	20.0
Benzo(a)pyrene	1.352	1.233		-8.8	20.0
Benzo(g,h,i)perylene	1.410	1.404		-0.4	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037204.D
 Acq On : 09 Jun 2025 23:11
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Jun 10 04:04:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	1688	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	4447	0.400	ng	0.00
13) Acenaphthene-d10	14.234	164	2457	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	4471	0.400	ng	0.00
29) Chrysene-d12	21.180	240	2829	0.400	ng	# 0.00
35) Perylene-d12	23.377	264	2684	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	1579	0.378	ng	0.00
5) Phenol-d6	6.766	99	1959	0.387	ng	0.00
8) Nitrobenzene-d5	8.739	82	1880	0.401	ng	0.00
11) 2-Methylnaphthalene-d10	11.965	152	2463	0.398	ng	0.00
14) 2,4,6-Tribromophenol	15.730	330	370	0.374	ng	-0.01
15) 2-Fluorobiphenyl	12.858	172	4140	0.395	ng	0.00
27) Fluoranthene-d10	19.026	212	4111	0.362	ng	0.00
31) Terphenyl-d14	19.635	244	2600	0.390	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.119	88	836	0.372	ng	96
3) n-Nitrosodimethylamine	3.429	42	1892	0.419	ng	93
6) bis(2-Chloroethyl)ether	7.019	93	1846	0.382	ng	95
9) Naphthalene	10.415	128	4918	0.383	ng	99
10) Hexachlorobutadiene	10.703	225	1140	0.408	ng	# 100
12) 2-Methylnaphthalene	12.041	142	3069	0.373	ng	98
16) Acenaphthylene	13.956	152	4318	0.358	ng	100
17) Acenaphthene	14.299	154	2830	0.362	ng	99
18) Fluorene	15.293	166	3684	0.358	ng	99
20) 4,6-Dinitro-2-methylph...	15.378	198	360	0.549	ng	# 76
21) 4-Bromophenyl-phenylether	16.189	248	1090	0.372	ng	92
22) Hexachlorobenzene	16.301	284	1245	0.394	ng	99
23) Atrazine	16.462	200	893	0.369	ng	96
24) Pentachlorophenol	16.649	266	562	0.512	ng	99
25) Phenanthrene	17.021	178	5255	0.363	ng	100
26) Anthracene	17.120	178	4594	0.348	ng	99
28) Fluoranthene	19.054	202	5351	0.334	ng	99
30) Pyrene	19.416	202	5310	0.385	ng	99
32) Benzo(a)anthracene	21.171	228	3655	0.357	ng	98
33) Chrysene	21.215	228	4111	0.361	ng	98
34) Bis(2-ethylhexyl)phtha...	21.117	149	2430	0.376	ng	98
36) Indeno(1,2,3-cd)pyrene	25.576	276	4264	0.399	ng	97
37) Benzo(b)fluoranthene	22.722	252	3713	0.343	ng	96
38) Benzo(k)fluoranthene	22.766	252	3991	0.361	ng	98
39) Benzo(a)pyrene	23.284	252	3310	0.365	ng	# 93
40) Dibenzo(a,h)anthracene	25.587	278	3252	0.395	ng	98
41) Benzo(g,h,i)perylene	26.242	276	3769	0.398	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037204.D
Acq On : 09 Jun 2025 23:11
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Jun 10 04:04:24 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:52:03 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.589 min Scan# 61
Delta R.T. -0.001 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:152 Resp: 1688
Ion Ratio Lower Upper
152 100
150 154.9 123.2 184.8
115 69.5 56.6 85.0




#2
1,4-Dioxane
Concen: 0.372 ng
RT: 3.119 min Scan# 33
Delta R.T. -0.007 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion: 88 Resp: 836
Ion Ratio Lower Upper
88 100
43 59.2 43.5 65.3
58 86.4 67.7 101.5



Abundance Scan 77 (3.437 min): BN037145.D\data.ms (-73)



#3

n-Nitrosodimethylamine

Concen: 0.419 ng

RT: 3.429 min Scan# 7

Delta R.T. -0.007 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11


Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Abundance Scan 76 (3.429 min): BN037204.D\data.ms



Tgt Ion: 42 Resp: 1892


Ion Ratio Lower Upper

42 100


74 60.5 53.0 79.4

44 7.2 5.9 8.9


Abundance



Abundance Scan 76 (3.429 min): BN037204.D\data.ms (-56)



Abundance Scan 319 (5.184 min): BN037145.D\data.ms (-31)



#4

2-Fluorophenol

Concen: 0.378 ng

RT: 5.184 min Scan# 319

Delta R.T. -0.000 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11

Tgt Ion: 112 Resp: 1579


Ion Ratio Lower Upper

112 100


64 68.8 56.3 84.5

63 46.4 36.2 54.4


Abundance Scan 319 (5.184 min): BN037204.D\data.ms




Abundance



Abundance Scan 319 (5.184 min): BN037204.D\data.ms (-29)









#5
Phenol-d6
Concen: 0.387 ng
RT: 6.766 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
99	100		
42	41.3	31.3	46.9
71	48.6	38.2	57.2



#6
bis(2-Chloroethyl)ether
Concen: 0.382 ng
RT: 7.019 min Scan# 573
Delta R.T. -0.007 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion	Ratio	Lower	Upper
93	100		
63	90.4	68.6	103.0
95	32.3	24.3	36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	4447
Ion	Ratio	Lower	Upper
136	100		
137	12.2	9.7	14.5
54	10.2	9.7	14.5
68	5.8	5.4	8.2



#8
Nitrobenzene-d5
Concen: 0.401 ng
RT: 8.739 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion:	82	Resp:	1880
Ion	Ratio	Lower	Upper
82	100		
128	37.5	26.9	40.3
54	73.2	61.4	92.2









#9
Naphthalene
Concen: 0.383 ng
RT: 10.415 min Scan# 949
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 4918
Ion Ratio Lower Upper
128 100
129 12.4 9.8 14.8
127 14.6 12.3 18.5



#10
Hexachlorobutadiene
Concen: 0.408 ng
RT: 10.703 min Scan# 976
Delta R.T. -0.011 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion:225 Resp: 1140
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.8 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.398 ng
RT: 11.965 min Scan# 1143
Delta R.T. -0.005 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 2463
Ion Ratio Lower Upper
152 100
151 22.1 17.1 25.7



#12
2-Methylnaphthalene
Concen: 0.373 ng
RT: 12.041 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion:142 Resp: 3069
Ion Ratio Lower Upper
142 100
141 90.6 74.6 111.8
115 50.7 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.234 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 2457
Ion Ratio Lower Upper
164 100
162 106.1 85.5 128.3
160 56.4 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 0.374 ng
RT: 15.730 min Scan# 1599
Delta R.T. -0.013 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion:330 Resp: 370
Ion Ratio Lower Upper
330 100
332 96.5 77.1 115.7
141 57.3 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 0.395 ng
RT: 12.858 min Scan# 1318
Delta R.T. -0.005 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.5	29.6	44.4
170	24.5	20.3	30.5



#16
Acenaphthylene
Concen: 0.358 ng
RT: 13.956 min Scan# 1436
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.1	16.3	24.5
153	13.3	10.6	15.8









#17
Acenaphthene
Concen: 0.362 ng
RT: 14.299 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 2830
Ion Ratio Lower Upper
154 100
153 119.2 93.8 140.8
152 63.6 50.5 75.7



#18
Fluorene
Concen: 0.358 ng
RT: 15.293 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion:166 Resp: 3684
Ion Ratio Lower Upper
166 100
165 100.3 81.1 121.7
167 13.4 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:188 Resp: 4471

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 13.0 11.3 16.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.549 ng

RT: 15.378 min Scan# 1569

Delta R.T. -0.011 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11




Tgt Ion:198 Resp: 360


Ion Ratio Lower Upper

198 100

51 114.1 125.2 187.8#

105 66.5 57.1 85.7









#21
4-Bromophenyl-phenylether
Concen: 0.372 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
248	100		
250	101.3	76.1	114.1
141	66.1	60.1	90.1



#22
Hexachlorobenzene
Concen: 0.394 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion	Ratio	Lower	Upper
284	100		
142	55.7	44.0	66.0
249	36.2	29.7	44.5









#23
Atrazine
Concen: 0.369 ng
RT: 16.462 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
200	100		
173	31.4	28.1	42.1
215	47.7	39.3	58.9



#24
Pentachlorophenol
Concen: 0.512 ng
RT: 16.649 min Scan# 1673
Delta R.T. -0.012 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.6	49.3	73.9
268	61.7	49.0	73.4









#25
Phenanthrene
Concen: 0.363 ng
RT: 17.021 min Scan# 1703
Delta R.T. -0.013 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 5255
Ion Ratio Lower Upper
178 100
176 19.3 15.7 23.5
179 15.5 12.3 18.5



#26
Anthracene
Concen: 0.348 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion:178 Resp: 4594
Ion Ratio Lower Upper
178 100
176 18.9 15.2 22.8
179 15.5 12.9 19.3





#27

Fluoranthene-d10

Concen: 0.362 ng

RT: 19.026 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:212 Resp: 4111

Ion Ratio Lower Upper

212 100

106 12.7 10.6 15.8

104 7.8 6.6 9.8



#28

Fluoranthene

Concen: 0.334 ng

RT: 19.054 min Scan# 1940

Delta R.T. -0.000 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11




Tgt Ion:202 Resp: 5351


Ion Ratio Lower Upper

202 100

101 10.1 8.7 13.1

203 17.3 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.180 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 2829
Ion Ratio Lower Upper
240 100
120 18.0 9.0 13.4#
236 31.2 23.0 34.4



#30
Pyrene
Concen: 0.385 ng
RT: 19.416 min Scan# 2018
Delta R.T. -0.005 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion:202 Resp: 5310
Ion Ratio Lower Upper
202 100
200 20.9 17.0 25.6
203 17.7 14.2 21.4









#31
Terphenyl-d14
Concen: 0.390 ng
RT: 19.635 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	12.6	10.0	15.0
122	16.0	13.2	19.8



#32
Benzo(a)anthracene
Concen: 0.357 ng
RT: 21.171 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.3	22.6	33.8
229	21.5	16.2	24.2









#33
Chrysene
Concen: 0.361 ng
RT: 21.215 min Scan# 2115
Delta R.T. -0.009 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
228	100		
226	32.9	25.2	37.8
229	20.3	16.8	25.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.376 ng
RT: 21.117 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion	Ratio	Lower	Upper
149	100		
167	27.4	21.0	31.4
279	3.7	2.9	4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.377 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	264	Resp:	2684
Ion Ratio	Lower	Upper	
264	100		
260	28.5	22.1	33.1
265	76.2	55.8	83.8



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.399 ng
RT: 25.576 min Scan# 3497
Delta R.T. 0.003 min
Lab File: BN037204.D
Acq: 09 Jun 2025 23:11

Tgt Ion:	276	Resp:	4264
Ion Ratio	Lower	Upper	
276	100		
138	23.5	21.0	31.6
277	24.8	19.4	29.2





#37

Benzo(b)fluoranthene

Concen: 0.343 ng

RT: 22.722 min Scan# 2521

Delta R.T. -0.000 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 3713

Ion Ratio Lower Upper

252 100

253 29.4 22.3 33.5

125 18.6 13.2 19.8



#38

Benzo(k)fluoranthene

Concen: 0.361 ng

RT: 22.766 min Scan# 2536

Delta R.T. 0.003 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11




Tgt Ion:252 Resp: 3991


Ion Ratio Lower Upper

252 100

253 29.5 22.2 33.4

125 16.3 13.2 19.8





#39

Benzo(a)pyrene

Concen: 0.365 ng

RT: 23.284 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 3310

Ion Ratio Lower Upper

252 100

253 33.7 25.0 37.4

125 26.8 17.0 25.6#



#40

Dibenzo(a,h)anthracene

Concen: 0.395 ng

RT: 25.587 min Scan# 3501

Delta R.T. -0.000 min

Lab File: BN037204.D

Acq: 09 Jun 2025 23:11




Tgt Ion:278 Resp: 3252


Ion Ratio Lower Upper

278 100

139 21.2 17.6 26.4

279 34.0 26.0 39.0





#41

Benzo(g,h,i)perylene

Concen: 0.398 ng

RT: 26.242 min Scan# 31

Delta R.T. 0.003 min

Lab File: BN037204.D


Acq: 09 Jun 2025 23:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:276 Resp: 3769

Ion Ratio Lower Upper

276 100

277 27.5 20.9 31.3

138 26.2 20.8 31.2



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037204.D
 Acq On : 09 Jun 2025 23:11
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 10 04:04:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.533	0.495	7.1	78	0.00
3	n-Nitrosodimethylamine	1.071	1.121	-4.7	85	0.00
4 S	2-Fluorophenol	0.989	0.935	5.5	80	0.00
5 S	Phenol-d6	1.199	1.161	3.2	83	0.00
6	bis(2-Chloroethyl)ether	1.144	1.094	4.4	78	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.422	0.423	-0.2	86	0.00
9	Naphthalene	1.154	1.106	4.2	84	0.00
10	Hexachlorobutadiene	0.251	0.256	-2.0	84	-0.01
11 SURR	2-Methylnaphthalene-d10	0.557	0.554	0.5	84	0.00
12	2-Methylnaphthalene	0.740	0.690	6.8	85	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
14 S	2,4,6-Tribromophenol	0.161	0.151	6.2	88	-0.01
15 S	2-Fluorobiphenyl	1.705	1.685	1.2	88	0.00
16	Acenaphthylene	1.961	1.757	10.4	85	0.00
17	Acenaphthene	1.273	1.152	9.5	85	0.00
18	Fluorene	1.674	1.499	10.5	84	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
20	4,6-Dinitro-2-methylphenol	0.077	0.081	-5.2	132	-0.01
21	4-Bromophenyl-phenylether	0.262	0.244	6.9	82	0.00
22	Hexachlorobenzene	0.283	0.278	1.8	85	0.00
23	Atrazine	0.216	0.200	7.4	88	0.00
24	Pentachlorophenol	0.124	0.126	-1.6	112	-0.01
25	Phenanthrene	1.296	1.175	9.3	81	-0.01
26	Anthracene	1.183	1.028	13.1	81	0.00
27 SURR	Fluoranthene-d10	1.016	0.919	9.5	77	0.00
28	Fluoranthene	1.432	1.197	16.4	77	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
30	Pyrene	1.953	1.877	3.9	77	0.00
31 S	Terphenyl-d14	0.942	0.919	2.4	77	0.00
32	Benzo(a)anthracene	1.448	1.292	10.8	75	0.00
33	Chrysene	1.612	1.453	9.9	74	0.00
34	Bis(2-ethylhexyl)phthalate	0.914	0.859	6.0	83	0.00
35 I	Perylene-d12	1.000	1.000	0.0	79	0.00
36	Indeno(1,2,3-cd)pyrene	1.591	1.589	0.1	84	0.00
37	Benzo(b)fluoranthene	1.615	1.383	14.4	77	0.00
38	Benzo(k)fluoranthene	1.648	1.487	9.8	81	0.00
39 C	Benzo(a)pyrene	1.352	1.233	8.8	80	0.00
40	Dibenzo(a,h)anthracene	1.227	1.212	1.2	83	0.00
41	Benzo(g,h,i)perylene	1.410	1.404	0.4	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037204.D
 Acq On : 09 Jun 2025 23:11
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 10 04:04:24 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	80	0.00
2	1,4-Dioxane	0.400	0.372	7.0	78	0.00
3	n-Nitrosodimethylamine	0.400	0.419	-4.7	85	0.00
4 S	2-Fluorophenol	0.400	0.378	5.5	80	0.00
5 S	Phenol-d6	0.400	0.387	3.3	83	0.00
6	bis(2-Chloroethyl)ether	0.400	0.382	4.5	78	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	0.00
8 S	Nitrobenzene-d5	0.400	0.401	-0.3	86	0.00
9	Naphthalene	0.400	0.383	4.3	84	0.00
10	Hexachlorobutadiene	0.400	0.408	-2.0	84	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.398	0.5	84	0.00
12	2-Methylnaphthalene	0.400	0.373	6.8	85	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	85	0.00
14 S	2,4,6-Tribromophenol	0.400	0.374	6.5	88	-0.01
15 S	2-Fluorobiphenyl	0.400	0.395	1.3	88	0.00
16	Acenaphthylene	0.400	0.358	10.5	85	0.00
17	Acenaphthene	0.400	0.362	9.5	85	0.00
18	Fluorene	0.400	0.358	10.5	84	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	82	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.549	-37.3#	132	-0.01
21	4-Bromophenyl-phenylether	0.400	0.372	7.0	82	0.00
22	Hexachlorobenzene	0.400	0.394	1.5	85	0.00
23	Atrazine	0.400	0.369	7.8	88	0.00
24	Pentachlorophenol	0.400	0.512	-28.0#	112	-0.01
25	Phenanthrene	0.400	0.363	9.3	81	-0.01
26	Anthracene	0.400	0.348	13.0	81	0.00
27 SURR	Fluoranthene-d10	0.400	0.362	9.5	77	0.00
28	Fluoranthene	0.400	0.334	16.5	77	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	75	0.00
30	Pyrene	0.400	0.385	3.8	77	0.00
31 S	Terphenyl-d14	0.400	0.390	2.5	77	0.00
32	Benzo(a)anthracene	0.400	0.357	10.8	75	0.00
33	Chrysene	0.400	0.361	9.8	74	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.376	6.0	83	0.00
35 I	Perylene-d12	0.400	0.400	0.0	79	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.399	0.3	84	0.00
37	Benzo(b)fluoranthene	0.400	0.343	14.2	77	0.00
38	Benzo(k)fluoranthene	0.400	0.361	9.8	81	0.00
39 C	Benzo(a)pyrene	0.400	0.365	8.8	80	0.00
40	Dibenzo(a,h)anthracene	0.400	0.395	1.3	83	0.00
41	Benzo(g,h,i)perylene	0.400	0.398	0.5	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4





QC SAMPLE DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037142.D
Acq On : 03 Jun 2025 10:21
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Jun 04 01:52:03 2025






AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2469

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	69.8	142869	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	58.7	120176	PASS
70	69	0.00	2	0.5	575	PASS
127	198	10	80	53.9	110283	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	204619	PASS
199	198	5	9	6.8	13950	PASS
275	198	10	60	24.4	49829	PASS
365	198	1	100	4.5	9270	PASS
441	198	0.01	100	10.3	21027	PASS
442	442	50	100	100.0	124989	PASS
443	442	15	24	19.8	24771	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037142.D
Acq On : 03 Jun 2025 10:21
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jun 03 13:49:49 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 27 07:40:28 2025
Response via : Initial Calibration



TIC: BN037142.D\data.ms

(70) Pentachlorophenol (C)

16.651min (-0.005) 16176.78 ng




response 200710

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	67.97
264.00	61.60	66.57
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060325\
Data File : BN037142.D
Acq On : 03 Jun 2025 10:21
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jun 03 13:56:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue May 27 07:40:28 2025
Response via : Initial Calibration



TIC: BN037142.D\data.ms

(77) Benzidine

19.257min (-0.002) 0.00 ng

response 1234162

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.54
183.00	13.20	13.40
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/3/2025	BNA_N	<u>BN037142.D</u>
Compound Name	Response	Retention Time
DDT	724224	20.492
DDD	9498	20.104
DDE	658	19.545
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
10156	734380	1.38



Instrument :
BNA_N
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037188.D
Acq On : 09 Jun 2025 10:15
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Jun 04 01:52:03 2025






AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	74.0	99844	PASS
68	69	0.00	2	0.7	559	PASS
69	198	0.00	100	59.6	80408	PASS
70	69	0.00	2	0.6	489	PASS
127	198	10	80	53.0	71472	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	134965	PASS
199	198	5	9	6.9	9301	PASS
275	198	10	60	24.1	32565	PASS
365	198	1	100	4.4	5933	PASS
441	198	0.01	100	8.6	11593	PASS
442	442	50	100	100.0	76653	PASS
443	442	15	24	18.5	14197	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037188.D
Acq On : 09 Jun 2025 10:15
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jun 09 14:37:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 02:36:29 2025
Response via : Initial Calibration



TIC: BN037188.D\data.ms

(70) Pentachlorophenol (C)

16.639min (-0.007) 18937.43 ng




response 119636

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.72
264.00	61.60	66.86
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037188.D
Acq On : 09 Jun 2025 10:15
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jun 09 14:33:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 02:36:29 2025
Response via : Initial Calibration



TIC: BN037188.D\data.ms

(77) Benzidine

19.245min (-0.009) 0.00 ng

response 528346

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.53
183.00	13.20	12.40
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/9/2025	BNA_N	<u>BN037188.D</u>
Compound Name	Response	Retention Time
DDT	398680	20.486
DDD	7508	20.045
DDE	1010	19.539
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8518	407198	2.09



Instrument :
BNA_N
ClientSampleId :
DFTPP

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037203.D
Acq On : 09 Jun 2025 22:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Jun 04 01:52:03 2025






AutoFind: Scans 2474, 2475, 2476; Background Corrected with Scan 2468

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	79.8	88176	PASS
68	69	0.00	2	1.5	1111	PASS
69	198	0.00	100	65.5	72469	PASS
70	69	0.00	2	0.6	429	PASS
127	198	10	80	56.9	62893	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	110557	PASS
199	198	5	9	6.8	7484	PASS
275	198	10	60	24.4	27021	PASS
365	198	1	100	4.3	4720	PASS
441	198	0.01	100	8.8	9692	PASS
442	442	50	100	100.0	57680	PASS
443	442	15	24	20.4	11739	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037203.D
Acq On : 09 Jun 2025 22:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jun 10 06:02:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 10 06:01:37 2025
Response via : Initial Calibration



TIC: BN037203.D\data.ms

(70) Pentachlorophenol (C)

16.639min (0.000) 275.85 ng




response 102519

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.86
264.00	61.60	57.89
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037203.D
Acq On : 09 Jun 2025 22:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jun 10 06:02:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 10 06:01:37 2025
Response via : Initial Calibration



TIC: BN037203.D\data.ms

(77) Benzidine

19.245min (0.000) 0.00 ng

response 432254

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.09
183.00	13.20	12.58
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
6/9/2025	BNA_N	<u>BN037203.D</u>
Compound Name	Response	Retention Time
DDT	300238	20.486
DDD	4412	20.051
DDE	217	19.539
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
4629	304867	1.52

Instrument :
BNA_N
ClientSampleId :
DFTPP

Report of Analysis

Client:	G Environmental	Date Collected:	
Project:	Power	Date Received:	
Client Sample ID:	PB168336BL	SDG No.:	Q2209
Lab Sample ID:	PB168336BL	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037190.D	1	06/06/25 11:54	06/09/25 11:30	PB168336

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
56-55-3	Benzo(a)anthracene	0.040	U	0.040	0.10	ug/L
205-99-2	Benzo(b)fluoranthene	0.040	U	0.040	0.10	ug/L
207-08-9	Benzo(k)fluoranthene	0.050	U	0.050	0.10	ug/L
50-32-8	Benzo(a)pyrene	0.040	U	0.040	0.10	ug/L
191-24-2	Benzo(g,h,i)perylene	0.040	U	0.040	0.10	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.36		30 (20) - 150 (139)	91%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.40		30 (54) - 150 (157)	99%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		30 (27) - 130 (154)	92%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		30 (30) - 130 (155)	101%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		30 (54) - 130 (175)	105%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	1820		7.589		
1146-65-2	Naphthalene-d8	4230		10.372		
15067-26-2	Acenaphthene-d10	2100		14.245		
1517-22-2	Phenanthrene-d10	3500		16.996		
1719-03-5	Chrysene-d12	2450		21.189		
1520-96-3	Perylene-d12	2290		23.386		

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037190.D
 Acq On : 09 Jun 2025 11:30
 Operator : RC/JU
 Sample : PB168336BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB168336BL

Quant Time: Jun 09 12:24:55 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	1816	0.400	ng	0.00
7) Naphthalene-d8	10.372	136	4227	0.400	ng	# 0.00
13) Acenaphthene-d10	14.245	164	2101	0.400	ng	0.01
19) Phenanthrene-d10	16.996	188	3500	0.400	ng	0.01
29) Chrysene-d12	21.189	240	2446	0.400	ng	# 0.00
35) Perylene-d12	23.386	264	2291	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.192	112	1866	0.416	ng	0.00
5) Phenol-d6	6.773	99	1994	0.366	ng	0.00
8) Nitrobenzene-d5	8.749	82	1646	0.369	ng	0.01
11) 2-Methylnaphthalene-d10	11.970	152	2143	0.364	ng	0.00
14) 2,4,6-Tribromophenol	15.755	330	207	0.245	ng	0.01
15) 2-Fluorobiphenyl	12.863	172	3619	0.404	ng	0.00
27) Fluoranthene-d10	19.026	212	3511	0.395	ng	0.00
31) Terphenyl-d14	19.635	244	2421	0.420	ng	0.00


Target Compounds	Qvalue
-----	-----


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037190.D
Acq On : 09 Jun 2025 11:30
Operator : RC/JU
Sample : PB168336BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB168336BL


Quant Time: Jun 09 12:24:55 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:52:03 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.589 min Scan# 61
Delta R.T. -0.001 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30

Instrument :
BNA_N
ClientSampleId :
PB168336BL







Tgt Ion:152 Resp: 1816
Ion Ratio Lower Upper
152 100
150 153.4 123.2 184.8
115 69.5 56.6 85.0



#4
2-Fluorophenol
Concen: 0.416 ng
RT: 5.192 min Scan# 320
Delta R.T. 0.007 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30

Tgt Ion:112 Resp: 1866
Ion Ratio Lower Upper
112 100
64 69.8 56.3 84.5
63 47.4 36.2 54.4









#5
Phenol-d6
Concen: 0.366 ng
RT: 6.773 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30




Instrument :
BNA_N
ClientSampleId :
PB168336BL


Tgt Ion:	99	Resp:	1994
Ion	Ratio	Lower	Upper
99	100		
42	34.5	31.3	46.9
71	48.8	38.2	57.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.372 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30

Tgt Ion:	136	Resp:	4227
Ion	Ratio	Lower	Upper
136	100		
137	13.6	9.7	14.5
54	13.8	9.7	14.5
68	8.2	5.4	8.2









#8
Nitrobenzene-d5
Concen: 0.369 ng
RT: 8.749 min Scan# 792
Delta R.T. 0.011 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30




Instrument :
BNA_N
ClientSampleId :
PB168336BL


Tgt Ion: 82 Resp: 1646
Ion Ratio Lower Upper
82 100
128 38.7 26.9 40.3
54 71.6 61.4 92.2



#11
2-Methylnaphthalene-d10
Concen: 0.364 ng
RT: 11.970 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30

Tgt Ion: 152 Resp: 2143
Ion Ratio Lower Upper
152 100
151 21.8 17.1 25.7









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.245 min Scan# 1463
Delta R.T. 0.011 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30




Instrument :
BNA_N
ClientSampleId :
PB168336BL


Tgt Ion	164	Resp:	2101
Ion Ratio	Lower	Upper	
164	100		
162	105.1	85.5	128.3
160	55.6	44.6	67.0



#14
2,4,6-Tribromophenol
Concen: 0.245 ng
RT: 15.755 min Scan# 1601
Delta R.T. 0.012 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30

Tgt Ion	330	Resp:	207
Ion Ratio	Lower	Upper	
330	100		
332	98.1	77.1	115.7
141	66.7	46.4	69.6









#15
2-Fluorobiphenyl
Concen: 0.404 ng
RT: 12.863 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30




Instrument :
BNA_N
ClientSampleId :
PB168336BL


Tgt Ion	Ratio	Lower	Upper
172	100		
171	38.8	29.6	44.4
170	26.7	20.3	30.5



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.996 min Scan# 1701
Delta R.T. 0.012 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	12.5	11.3	16.9





#27

Fluoranthene-d10

Concen: 0.395 ng

RT: 19.026 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BN037190.D

Acq: 09 Jun 2025 11:30

Instrument :

BNA_N

ClientSampleId :

PB168336BL





Tgt Ion:212 Resp: 3511

Ion Ratio Lower Upper

212 100

106 12.0 10.6 15.8

104 7.4 6.6 9.8



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.189 min Scan# 2277

Delta R.T. -0.000 min

Lab File: BN037190.D

Acq: 09 Jun 2025 11:30




Tgt Ion:240 Resp: 2446


Ion Ratio Lower Upper

240 100

120 16.8 9.0 13.4

236 30.4 23.0 34.4









#31
Terphenyl-d14
Concen: 0.420 ng
RT: 19.635 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30




Instrument :
BNA_N
ClientSampleId :
PB168336BL

Tgt Ion	Ratio	Lower	Upper
244	100		
212	13.3	10.0	15.0
122	17.0	13.2	19.8



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.386 min Scan# 2748
Delta R.T. 0.009 min
Lab File: BN037190.D
Acq: 09 Jun 2025 11:30

Tgt Ion	Ratio	Lower	Upper
264	100		
260	27.2	22.1	33.1
265	87.7	55.8	83.8



Report of Analysis

Client:	G Environmental	Date Collected:	
Project:	Power	Date Received:	
Client Sample ID:	PB168336BS	SDG No.:	Q2209
Lab Sample ID:	PB168336BS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037201.D	1	06/06/25 11:54	06/09/25 20:40	PB168336

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
56-55-3	Benzo(a)anthracene	0.36		0.040	0.10	ug/L
205-99-2	Benzo(b)fluoranthene	0.35		0.040	0.10	ug/L
207-08-9	Benzo(k)fluoranthene	0.36		0.050	0.10	ug/L
50-32-8	Benzo(a)pyrene	0.39		0.040	0.10	ug/L
191-24-2	Benzo(g,h,i)perylene	0.42		0.040	0.10	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.36		30 (20) - 150 (139)	90%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.30		30 (54) - 150 (157)	76%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.36		30 (27) - 130 (154)	90%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.38		30 (30) - 130 (155)	95%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		30 (54) - 130 (175)	95%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	2230	7.589			
1146-65-2	Naphthalene-d8	5470	10.361			
15067-26-2	Acenaphthene-d10	2610	14.234			
1517-22-2	Phenanthrene-d10	4250	16.984			
1719-03-5	Chrysene-d12	2470	21.188			
1520-96-3	Perylene-d12	2370	23.377			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037201.D
 Acq On : 09 Jun 2025 20:40
 Operator : RC/JU
 Sample : PB168336BS
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB168336BS

Quant Time: Jun 10 04:03:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	2227	0.400	ng	0.00
7) Naphthalene-d8	10.361	136	5466	0.400	ng	#-0.01
13) Acenaphthene-d10	14.234	164	2607	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	4253	0.400	ng	0.00
29) Chrysene-d12	21.188	240	2468	0.400	ng	# 0.00
35) Perylene-d12	23.377	264	2373	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.191	112	2001	0.363	ng	0.00
5) Phenol-d6	6.766	99	2345	0.351	ng	0.00
8) Nitrobenzene-d5	8.738	82	2065	0.358	ng	0.00
11) 2-Methylnaphthalene-d10	11.965	152	2755	0.362	ng	0.00
14) 2,4,6-Tribromophenol	15.742	330	307	0.292	ng	0.00
15) 2-Fluorobiphenyl	12.858	172	4234	0.381	ng	0.00
27) Fluoranthene-d10	19.026	212	3278	0.303	ng	0.00
31) Terphenyl-d14	19.634	244	2215	0.381	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.119	88	1196	0.403	ng	# 44
3) n-Nitrosodimethylamine	3.429	42	2277	0.382	ng	# 91
6) bis(2-Chloroethyl)ether	7.019	93	2167	0.340	ng	95
9) Naphthalene	10.415	128	5395	0.342	ng	100
10) Hexachlorobutadiene	10.703	225	1252	0.364	ng	# 97
12) 2-Methylnaphthalene	12.041	142	3113	0.308	ng	99
16) Acenaphthylene	13.956	152	4871	0.381	ng	100
17) Acenaphthene	14.298	154	2908	0.350	ng	100
18) Fluorene	15.293	166	3677	0.337	ng	100
20) 4,6-Dinitro-2-methylph...	15.389	198	337	0.545	ng	# 68
21) 4-Bromophenyl-phenylether	16.189	248	1073	0.385	ng	91
22) Hexachlorobenzene	16.301	284	1184	0.394	ng	98
23) Atrazine	16.462	200	810	0.352	ng	98
24) Pentachlorophenol	16.648	266	393	0.433	ng	98
25) Phenanthrene	17.033	178	4954	0.360	ng	100
26) Anthracene	17.120	178	4498	0.358	ng	98
28) Fluoranthene	19.054	202	4519	0.297	ng	100
30) Pyrene	19.416	202	4443	0.369	ng	99
32) Benzo(a)anthracene	21.170	228	3255	0.364	ng	98
33) Chrysene	21.224	228	3659	0.368	ng	99
34) Bis(2-ethylhexyl)phtha...	21.108	149	1974	0.350	ng	100
36) Indeno(1,2,3-cd)pyrene	25.573	276	4127	0.437	ng	98
37) Benzo(b)fluoranthene	22.722	252	3317	0.346	ng	92
38) Benzo(k)fluoranthene	22.766	252	3519	0.360	ng	94
39) Benzo(a)pyrene	23.283	252	3154	0.393	ng	94
40) Dibenzo(a,h)anthracene	25.590	278	3219	0.442	ng	96
41) Benzo(g,h,i)perylene	26.248	276	3529	0.422	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037201.D
Acq On : 09 Jun 2025 20:40
Operator : RC/JU
Sample : PB168336BS
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB168336BS


Quant Time: Jun 10 04:03:40 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:52:03 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.589 min Scan# 61
Delta R.T. -0.001 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Instrument :
BNA_N
ClientSampleId :
PB168336BS






Tgt Ion:152 Resp: 2227
Ion Ratio Lower Upper
152 100
150 154.5 123.2 184.8
115 69.7 56.6 85.0




#2
1,4-Dioxane
Concen: 0.403 ng
RT: 3.119 min Scan# 33
Delta R.T. -0.008 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion: 88 Resp: 1196
Ion Ratio Lower Upper
88 100
43 127.3 43.5 65.3#
58 108.4 67.7 101.5#



Abundance Scan 77 (3.437 min): BN037145.D\data.ms (-73)



#3

n-Nitrosodimethylamine

Concen: 0.382 ng

RT: 3.429 min Scan# 70

Delta R.T. -0.008 min

Lab File: BN037201.D

Acq: 09 Jun 2025 20:40

Instrument :

BNA_N


ClientSampleId :

PB168336BS


Tgt Ion: 42 Resp: 2277

Ion	Ratio	Lower	Upper
42	100		
74	58.8	53.0	79.4
44	5.2	5.9	8.9#


Abundance Scan 76 (3.429 min): BN037201.D\data.ms




Abundance



Abundance Scan 76 (3.429 min): BN037201.D\data.ms (-56)



Abundance Scan 319 (5.184 min): BN037145.D\data.ms (-31)



#4

2-Fluorophenol

Concen: 0.363 ng

RT: 5.191 min Scan# 320

Delta R.T. 0.007 min


Lab File: BN037201.D

Acq: 09 Jun 2025 20:40


Tgt Ion: 112 Resp: 2001

Ion	Ratio	Lower	Upper
112	100		
64	69.6	56.3	84.5
63	43.6	36.2	54.4


Abundance Scan 320 (5.191 min): BN037201.D\data.ms




Abundance



Abundance Scan 320 (5.191 min): BN037201.D\data.ms (-29)









#5
Phenol-d6
Concen: 0.351 ng
RT: 6.766 min Scan# 51
Delta R.T. -0.008 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion: 99 Resp: 2345
Ion Ratio Lower Upper
99 100
42 38.7 31.3 46.9
71 48.7 38.2 57.2



#6
bis(2-Chloroethyl)ether
Concen: 0.340 ng
RT: 7.019 min Scan# 573
Delta R.T. -0.008 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion: 93 Resp: 2167
Ion Ratio Lower Upper
93 100
63 90.4 68.6 103.0
95 33.0 24.3 36.5









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.361 min Scan# 945
Delta R.T. -0.011 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion:	136	Resp:	5466
Ion	Ratio	Lower	Upper
136	100		
137	12.3	9.7	14.5
54	15.3	9.7	14.5
68	8.1	5.4	8.2



#8
Nitrobenzene-d5
Concen: 0.358 ng
RT: 8.738 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion:	82	Resp:	2065
Ion	Ratio	Lower	Upper
82	100		
128	39.6	26.9	40.3
54	68.4	61.4	92.2









#9
Naphthalene
Concen: 0.342 ng
RT: 10.415 min Scan# 949
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion	128	129	127
Resp:	5395		
Ion Ratio	100	12.0	15.3
Lower		9.8	12.3
Upper		14.8	18.5



#10
Hexachlorobutadiene
Concen: 0.364 ng
RT: 10.703 min Scan# 976
Delta R.T. -0.011 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion	225	223	227
Resp:	1252		
Ion Ratio	100	0.0	64.9
Lower		0.0	50.3
Upper		0.0	75.5









#11
2-Methylnaphthalene-d10
Concen: 0.362 ng
RT: 11.965 min Scan# 1142
Delta R.T. -0.005 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion:152 Resp: 2755
Ion Ratio Lower Upper
152 100
151 20.6 17.1 25.7



#12
2-Methylnaphthalene
Concen: 0.308 ng
RT: 12.041 min Scan# 1157
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion:142 Resp: 3113
Ion Ratio Lower Upper
142 100
141 91.8 74.6 111.8
115 51.0 41.0 61.4









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.234 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion:164 Resp: 2607
Ion Ratio Lower Upper
164 100
162 109.3 85.5 128.3
160 56.6 44.6 67.0



#14
2,4,6-Tribromophenol
Concen: 0.292 ng
RT: 15.742 min Scan# 1600
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion:330 Resp: 307
Ion Ratio Lower Upper
330 100
332 89.9 77.1 115.7
141 58.3 46.4 69.6









#15
2-Fluorobiphenyl
Concen: 0.381 ng
RT: 12.858 min Scan# 1318
Delta R.T. -0.005 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.8	29.6	44.4
170	24.8	20.3	30.5



#16
Acenaphthylene
Concen: 0.381 ng
RT: 13.956 min Scan# 1436
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.4	16.3	24.5
153	13.1	10.6	15.8









#17
Acenaphthene
Concen: 0.350 ng
RT: 14.298 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion:154 Resp: 2908
Ion Ratio Lower Upper
154 100
153 117.8 93.8 140.8
152 63.4 50.5 75.7



#18
Fluorene
Concen: 0.337 ng
RT: 15.293 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion:166 Resp: 3677
Ion Ratio Lower Upper
166 100
165 101.0 81.1 121.7
167 13.8 10.8 16.2





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 16.984 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN037201.D

Acq: 09 Jun 2025 20:40

Instrument :

BNA_N

ClientSampleId :

PB168336BS





Tgt Ion:188 Resp: 4253

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.3 11.3 16.9



#20

4,6-Dinitro-2-methylphenol

Concen: 0.545 ng

RT: 15.389 min Scan# 1570

Delta R.T. -0.000 min

Lab File: BN037201.D

Acq: 09 Jun 2025 20:40




Tgt Ion:198 Resp: 337


Ion Ratio Lower Upper

198 100

51 103.3 125.2 187.8#

105 59.7 57.1 85.7









#21
4-Bromophenyl-phenylether
Concen: 0.385 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion	Ratio	Lower	Upper
248	100		
250	99.9	76.1	114.1
141	62.3	60.1	90.1



#22
Hexachlorobenzene
Concen: 0.394 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion	Ratio	Lower	Upper
284	100		
142	53.0	44.0	66.0
249	36.7	29.7	44.5









#23
Atrazine
Concen: 0.352 ng
RT: 16.462 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion:200 Resp: 810
Ion Ratio Lower Upper
200 100
173 32.9 28.1 42.1
215 48.7 39.3 58.9



#24
Pentachlorophenol
Concen: 0.433 ng
RT: 16.648 min Scan# 1673
Delta R.T. -0.013 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion:266 Resp: 393
Ion Ratio Lower Upper
266 100
264 60.3 49.3 73.9
268 63.4 49.0 73.4









#25
Phenanthrene
Concen: 0.360 ng
RT: 17.033 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion:	178	Resp:	4954
Ion	Ratio	Lower	Upper
178	100		
176	19.8	15.7	23.5
179	15.5	12.3	18.5



#26
Anthracene
Concen: 0.358 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion:	178	Resp:	4498
Ion	Ratio	Lower	Upper
178	100		
176	18.7	15.2	22.8
179	14.7	12.9	19.3









#27
Fluoranthene-d10
Concen: 0.303 ng
RT: 19.026 min Scan# 1934
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion	Ratio	Lower	Upper
212	100		
106	13.2	10.6	15.8
104	8.4	6.6	9.8



#28
Fluoranthene
Concen: 0.297 ng
RT: 19.054 min Scan# 1940
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion	Ratio	Lower	Upper
202	100		
101	10.8	8.7	13.1
203	16.8	13.5	20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.188 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion:240 Resp: 2468
Ion Ratio Lower Upper
240 100
120 13.6 9.0 13.4#
236 29.6 23.0 34.4



#30
Pyrene
Concen: 0.369 ng
RT: 19.416 min Scan# 2018
Delta R.T. -0.005 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion:202 Resp: 4443
Ion Ratio Lower Upper
202 100
200 21.8 17.0 25.6
203 17.9 14.2 21.4









#31
Terphenyl-d14
Concen: 0.381 ng
RT: 19.634 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion	Ratio	Lower	Upper
244	100		
212	13.2	10.0	15.0
122	16.6	13.2	19.8



#32
Benzo(a)anthracene
Concen: 0.364 ng
RT: 21.170 min Scan# 2275
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.7	22.6	33.8
229	20.5	16.2	24.2









#33
Chrysene
Concen: 0.368 ng
RT: 21.224 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion	Ratio	Lower	Upper
228	100		
226	32.0	25.2	37.8
229	21.4	16.8	25.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.350 ng
RT: 21.108 min Scan# 2268
Delta R.T. -0.009 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.2	21.0	31.4
279	3.6	2.9	4.3









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.377 min Scan# 21
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion:264 Resp: 2373
Ion Ratio Lower Upper
264 100
260 27.6 22.1 33.1
265 90.6 55.8 83.8#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.437 ng
RT: 25.573 min Scan# 3496
Delta R.T. -0.000 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion:276 Resp: 4127
Ion Ratio Lower Upper
276 100
138 24.5 21.0 31.6
277 24.4 19.4 29.2





#37

Benzo(b)fluoranthene

Concen: 0.346 ng

RT: 22.722 min Scan# 2521

Delta R.T. -0.000 min

Lab File: BN037201.D

Acq: 09 Jun 2025 20:40

Instrument :

BNA_N

ClientSampleId :

PB168336BS





Tgt Ion:252 Resp: 3317

Ion Ratio Lower Upper

252 100

253 32.4 22.3 33.5

125 19.7 13.2 19.8



#38

Benzo(k)fluoranthene

Concen: 0.360 ng

RT: 22.766 min Scan# 2536

Delta R.T. 0.003 min

Lab File: BN037201.D

Acq: 09 Jun 2025 20:40




Tgt Ion:252 Resp: 3519


Ion Ratio Lower Upper

252 100

253 31.6 22.2 33.4

125 18.4 13.2 19.8









#39
Benzo(a)pyrene
Concen: 0.393 ng
RT: 23.283 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40




Instrument :
BNA_N
ClientSampleId :
PB168336BS


Tgt Ion:252 Resp: 3154
Ion Ratio Lower Upper
252 100
253 35.3 25.0 37.4
125 23.1 17.0 25.6



#40
Dibenzo(a,h)anthracene
Concen: 0.442 ng
RT: 25.590 min Scan# 3502
Delta R.T. 0.003 min
Lab File: BN037201.D
Acq: 09 Jun 2025 20:40

Tgt Ion:278 Resp: 3219
Ion Ratio Lower Upper
278 100
139 22.2 17.6 26.4
279 35.8 26.0 39.0





#41

Benzo(g,h,i)perylene

Concen: 0.422 ng

RT: 26.248 min Scan# 31

Delta R.T. 0.008 min

Lab File: BN037201.D


Acq: 09 Jun 2025 20:40

Instrument :

BNA_N

ClientSampleId :

PB168336BS





Tgt Ion:276 Resp: 3529

Ion Ratio Lower Upper

276 100

277 26.6 20.9 31.3

138 25.8 20.8 31.2



Report of Analysis

Client:	G Environmental	Date Collected:	06/05/25
Project:	Power	Date Received:	06/05/25
Client Sample ID:	MW-11A-13.5-060525MS	SDG No.:	Q2209
Lab Sample ID:	Q2250-02MS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	960 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037192.D	1	06/06/25 11:54	06/09/25 14:33	PB168336

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
56-55-3	Benzo(a)anthracene	0.45		0.040	0.10	ug/L
205-99-2	Benzo(b)fluoranthene	0.40		0.040	0.10	ug/L
207-08-9	Benzo(k)fluoranthene	0.38		0.050	0.10	ug/L
50-32-8	Benzo(a)pyrene	0.40		0.040	0.10	ug/L
191-24-2	Benzo(g,h,i)perylene	0.39		0.040	0.10	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.30		30 (20) - 150 (139)	75%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 (54) - 150 (157)	92%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		30 (27) - 130 (154)	79%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		30 (30) - 130 (155)	86%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.47		30 (54) - 130 (175)	118%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	2140	7.589			
1146-65-2	Naphthalene-d8	5670	10.361			
15067-26-2	Acenaphthene-d10	2990	14.234			
1517-22-2	Phenanthrene-d10	5390	16.984			
1719-03-5	Chrysene-d12	3450	21.188			
1520-96-3	Perylene-d12	3180	23.38			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037192.D
 Acq On : 09 Jun 2025 14:33
 Operator : RC/JU
 Sample : Q2250-02MS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-11A-13.5-060525MS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 06/10/2025
 Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 09 15:40:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	2144	0.400	ng	0.00
7) Naphthalene-d8	10.361	136	5670	0.400	ng	#-0.01
13) Acenaphthene-d10	14.234	164	2991	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	5389	0.400	ng	0.00
29) Chrysene-d12	21.188	240	3448	0.400	ng	0.00
35) Perylene-d12	23.380	264	3177	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.184	112	842	0.159	ng	0.00
5) Phenol-d6	6.766	99	649	0.101	ng	0.00
8) Nitrobenzene-d5	8.739	82	1888	0.316	ng	0.00
11) 2-Methylnaphthalene-d10	11.965	152	2380m	0.302	ng	0.00
14) 2,4,6-Tribromophenol	15.730	330	466	0.387	ng	-0.01
15) 2-Fluorobiphenyl	12.858	172	4371	0.343	ng	0.00
27) Fluoranthene-d10	19.026	212	5042	0.368	ng	0.00
31) Terphenyl-d14	19.630	244	3837	0.473	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.111	88	8692	3.041	ng	94
3) n-Nitrosodimethylamine	3.429	42	693	0.121	ng	# 80
6) bis(2-Chloroethyl)ether	7.019	93	2028	0.331	ng	96
9) Naphthalene	10.415	128	5151	0.315	ng	99
10) Hexachlorobutadiene	10.703	225	918	0.258	ng	# 100
12) 2-Methylnaphthalene	12.036	142	3206	0.306	ng	98
16) Acenaphthylene	13.956	152	5430	0.370	ng	100
17) Acenaphthene	14.299	154	3253	0.342	ng	99
18) Fluorene	15.293	166	4617	0.369	ng	98
20) 4,6-Dinitro-2-methylph...	15.378	198	513	0.599	ng	# 58
21) 4-Bromophenyl-phenylether	16.189	248	1414	0.400	ng	# 83
22) Hexachlorobenzene	16.301	284	1382	0.363	ng	99
23) Atrazine	16.462	200	1269	0.435	ng	# 91
24) Pentachlorophenol	16.636	266	1565	0.901	ng	98
25) Phenanthrene	17.021	178	7297	0.418	ng	99
26) Anthracene	17.120	178	6246	0.392	ng	98
28) Fluoranthene	19.054	202	7085	0.367	ng	# 97
30) Pyrene	19.416	202	7143	0.424	ng	99
32) Benzo(a)anthracene	21.171	228	5416	0.434	ng	99
33) Chrysene	21.224	228	5649	0.407	ng	100
34) Bis(2-ethylhexyl)phtha...	21.117	149	3531	0.448	ng	99
36) Indeno(1,2,3-cd)pyrene	25.576	276	5130	0.406	ng	99
37) Benzo(b)fluoranthene	22.725	252	4923m	0.384	ng	
38) Benzo(k)fluoranthene	22.766	252	4787	0.366	ng	94
39) Benzo(a)pyrene	23.284	252	4119	0.383	ng	93
40) Dibenzo(a,h)anthracene	25.590	278	4003	0.411	ng	100
41) Benzo(g,h,i)perylene	26.245	276	4218	0.377	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037192.D
 Acq On : 09 Jun 2025 14:33
 Operator : RC/JU
 Sample : Q2250-02MS
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 MW-11A-13.5-060525MS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 06/10/2025
 Supervised By :Jagrut Upadhyay 06/10/2025


Quant Time: Jun 09 15:40:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.589 min Scan# 61
Delta R.T. -0.001 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:152 Resp: 214
Ion Ratio Lower Upper
152 100
150 153.5 123.2 184.8
115 69.4 56.6 85.0




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#2
1,4-Dioxane
Concen: 3.041 ng
RT: 3.111 min Scan# 32
Delta R.T. -0.015 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33

Tgt Ion: 88 Resp: 8692
Ion Ratio Lower Upper
88 100
43 62.0 43.5 65.3
58 87.9 67.7 101.5



Abundance Scan 77 (3.437 min): BN037145.D\data.ms (-73)



#3

n-Nitrosodimethylamine

Concen: 0.121 ng

RT: 3.429 min Scan# 70

Delta R.T. -0.007 min

Lab File: BN037192.D

Acq: 09 Jun 2025 14:33


Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MS

Abundance Scan 76 (3.429 min): BN037192.D\data.ms



Tgt Ion: 42 Resp: 69

Ion Ratio Lower Upper

42 100

74 81.1 53.0 79.4

44 19.9 5.9 8.9



Manual Integrations

APPROVED


Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025

Abundance Scan 76 (3.429 min): BN037192.D\data.ms (-56)



Abundance Scan 319 (5.184 min): BN037145.D\data.ms (-31)



#4

2-Fluorophenol

Concen: 0.159 ng

RT: 5.184 min Scan# 319

Delta R.T. -0.000 min

Lab File: BN037192.D

Acq: 09 Jun 2025 14:33

Tgt Ion:112 Resp: 842



Ion Ratio Lower Upper

112 100


64 71.0 56.3 84.5


63 44.7 36.2 54.4

Abundance Scan 319 (5.184 min): BN037192.D\data.ms



Abundance Scan 319 (5.184 min): BN037192.D\data.ms (-29)









#5
Phenol-d6
Concen: 0.101 ng
RT: 6.766 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33

Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS

Tgt Ion: 99 Resp: 649
Ion Ratio Lower Upper
99 100
42 53.5 31.3 46.9
71 57.5 38.2 57.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.331 ng
RT: 7.019 min Scan# 573
Delta R.T. -0.007 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion: 93 Resp: 2028
Ion Ratio Lower Upper
93 100
63 89.6 68.6 103.0
95 32.3 24.3 36.5





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.361 min Scan# 945
Delta R.T. -0.011 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:136 Resp: 5670
Ion Ratio Lower Upper
136 100
137 12.1 9.7 14.5
54 15.1 9.7 14.5
68 8.2 5.4 8.2#




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#8
Nitrobenzene-d5
Concen: 0.316 ng
RT: 8.739 min Scan# 792
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion: 82 Resp: 1888
Ion Ratio Lower Upper
82 100
128 39.2 26.9 40.3
54 66.9 61.4 92.2





#9
Naphthalene
Concen: 0.315 ng
RT: 10.415 min Scan# 949
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:128 Resp: 515
Ion Ratio Lower Upper
128 100
129 12.6 9.8 14.8
127 15.1 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#10
Hexachlorobutadiene
Concen: 0.258 ng
RT: 10.703 min Scan# 976
Delta R.T. -0.011 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion:225 Resp: 918
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.7 50.3 75.5





#11
2-Methylnaphthalene-d10
Concen: 0.302 ng m
RT: 11.965 min Scan# 1143
Delta R.T. -0.005 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:152 Resp: 2380
Ion Ratio Lower Upper
152 100
151 21.0 17.1 25.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#12
2-Methylnaphthalene
Concen: 0.306 ng
RT: 12.036 min Scan# 1156
Delta R.T. -0.005 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion:142 Resp: 3206
Ion Ratio Lower Upper
142 100
141 91.9 74.6 111.8
115 49.2 41.0 61.4





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.234 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS




Tgt Ion:164 Resp: 299
Ion Ratio Lower Upper
164 100
162 105.0 85.5 128.3
160 52.6 44.6 67.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#14
2,4,6-Tribromophenol
Concen: 0.387 ng
RT: 15.730 min Scan# 1599
Delta R.T. -0.013 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Tgt Ion:330 Resp: 466
Ion Ratio Lower Upper
330 100
332 92.7 77.1 115.7
141 66.7 46.4 69.6





#15
2-Fluorobiphenyl
Concen: 0.343 ng
RT: 12.858 min Scan# 1318
Delta R.T. -0.005 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:172 Resp: 437
Ion Ratio Lower Upper
172 100
171 35.9 29.6 44.4
170 24.3 20.3 30.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#16
Acenaphthylene
Concen: 0.370 ng
RT: 13.956 min Scan# 1436
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion:152 Resp: 5430
Ion Ratio Lower Upper
152 100
151 20.3 16.3 24.5
153 13.2 10.6 15.8





#17
Acenaphthene
Concen: 0.342 ng
RT: 14.299 min Scan# 1468
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:154 Resp: 325
Ion Ratio Lower Upper
154 100
153 116.9 93.8 140.8
152 61.8 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#18
Fluorene
Concen: 0.369 ng
RT: 15.293 min Scan# 1561
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion:166 Resp: 4617
Ion Ratio Lower Upper
166 100
165 98.9 81.1 121.7
167 13.4 10.8 16.2





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.984 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:188 Resp: 5389
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.6 11.3 16.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#20
4,6-Dinitro-2-methylphenol
Concen: 0.599 ng
RT: 15.378 min Scan# 1569
Delta R.T. -0.011 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion:198 Resp: 513
Ion Ratio Lower Upper
198 100
51 90.7 125.2 187.8#
105 50.9 57.1 85.7#





#21
4-Bromophenyl-phenylether
Concen: 0.400 ng
RT: 16.189 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:248 Resp: 1414
Ion Ratio Lower Upper
248 100
250 101.9 76.1 114.1
141 50.6 60.1 90.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#22
Hexachlorobenzene
Concen: 0.363 ng
RT: 16.301 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion:284 Resp: 1382
Ion Ratio Lower Upper
284 100
142 54.3 44.0 66.0
249 35.8 29.7 44.5





#23
Atrazine
Concen: 0.435 ng
RT: 16.462 min Scan# 1658
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:200 Resp: 1269
Ion Ratio Lower Upper
200 100
173 26.3 28.1 42.1
215 45.9 39.3 58.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#24
Pentachlorophenol
Concen: 0.901 ng
RT: 16.636 min Scan# 1672
Delta R.T. -0.025 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion:266 Resp: 1565
Ion Ratio Lower Upper
266 100
264 61.0 49.3 73.9
268 63.3 49.0 73.4





#25
Phenanthrene
Concen: 0.418 ng
RT: 17.021 min Scan# 1703
Delta R.T. -0.013 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:178 Resp: 729
Ion Ratio Lower Upper
178 100
176 20.3 15.7 23.5
179 15.2 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#26
Anthracene
Concen: 0.392 ng
RT: 17.120 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33

Tgt Ion:178 Resp: 6246
Ion Ratio Lower Upper
178 100
176 19.1 15.2 22.8
179 14.7 12.9 19.3





#27

Fluoranthene-d10

Concen: 0.368 ng

RT: 19.026 min Scan# 1934

Delta R.T. -0.000 min

Lab File: BN037192.D

Acq: 09 Jun 2025 14:33

Instrument :

BNA_N

ClientSampleId :





MW-11A-13.5-060525MS

Tgt Ion:212 Resp: 504

Ion	Ratio	Lower	Upper
212	100		
106	13.6	10.6	15.8
104	8.3	6.6	9.8

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#28

Fluoranthene

Concen: 0.367 ng

RT: 19.054 min Scan# 1940




Delta R.T. -0.000 min


Lab File: BN037192.D

Acq: 09 Jun 2025 14:33

Tgt Ion:202 Resp: 7085

Ion	Ratio	Lower	Upper
202	100		
101	13.6	8.7	13.1#
203	17.1	13.5	20.3





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.188 min Scan# 2118

Delta R.T. -0.000 min

Lab File: BN037192.D


Acq: 09 Jun 2025 14:33

Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MS






Tgt Ion	Ratio	Lower	Upper
240	100		
120	12.6	9.0	13.4
236	30.6	23.0	34.4

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#30

Pyrene


Concen: 0.424 ng

RT: 19.416 min Scan# 2018



Delta R.T. -0.005 min


Lab File: BN037192.D

Acq: 09 Jun 2025 14:33




Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.6	17.0	25.6
203	17.5	14.2	21.4





#31
 Terphenyl-d14
 Concen: 0.473 ng
 RT: 19.630 min Scan# 2065
 Delta R.T. -0.005 min
 Lab File: BN037192.D
 Acq: 09 Jun 2025 14:33




Instrument :
 BNA_N
 ClientSampleId :
 MW-11A-13.5-060525MS



Tgt Ion:244 Resp: 383
 Ion Ratio Lower Upper
 244 100
 212 12.1 10.0 15.0
 122 18.5 13.2 19.8




Manual Integrations
 APPROVED


Reviewed By :Rahul Chavli 06/10/2025
 Supervised By :Jagrut Upadhyay 06/10/2025



#32
 Benzo(a)anthracene
 Concen: 0.434 ng
 RT: 21.171 min Scan# 2275
 Delta R.T. -0.000 min
 Lab File: BN037192.D
 Acq: 09 Jun 2025 14:33


Tgt Ion:228 Resp: 5416
 Ion Ratio Lower Upper
 228 100
 226 28.5 22.6 33.8
 229 20.7 16.2 24.2





#33
Chrysene
Concen: 0.407 ng
RT: 21.224 min Scan# 2119
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MS



Tgt Ion:228 Resp: 5649
Ion Ratio Lower Upper
228 100
226 31.3 25.2 37.8
229 21.0 16.8 25.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.448 ng
RT: 21.117 min Scan# 2269
Delta R.T. -0.000 min
Lab File: BN037192.D
Acq: 09 Jun 2025 14:33


Tgt Ion:149 Resp: 3531
Ion Ratio Lower Upper
149 100
167 25.6 21.0 31.4
279 3.5 2.9 4.3





#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.380 min Scan# 21
 Delta R.T. 0.003 min
 Lab File: BN037192.D
 Acq: 09 Jun 2025 14:33




Instrument :
 BNA_N
 ClientSampleId :
 MW-11A-13.5-060525MS



Tgt Ion:264 Resp: 317
 Ion Ratio Lower Upper
 264 100
 260 27.8 22.1 33.1
 265 81.8 55.8 83.8




Manual Integrations
 APPROVED


Reviewed By :Rahul Chavli 06/10/2025
 Supervised By :Jagrut Upadhyay 06/10/2025



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.406 ng
 RT: 25.576 min Scan# 3497
 Delta R.T. 0.003 min
 Lab File: BN037192.D
 Acq: 09 Jun 2025 14:33

Tgt Ion:276 Resp: 5130
 Ion Ratio Lower Upper
 276 100
 138 26.9 21.0 31.6
 277 24.2 19.4 29.2





#37

Benzo(b)fluoranthene

Concen: 0.384 ng m

RT: 22.725 min Scan# 2521

Delta R.T. 0.003 min

Lab File: BN037192.D

Acq: 09 Jun 2025 14:33

Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MS





Tgt Ion:252	Resp:	4923
Ion Ratio	Lower	Upper
252	100	
253	30.6	22.3 33.5
125	19.9	13.2 19.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#38

Benzo(k)fluoranthene

Concen: 0.366 ng




RT: 22.766 min Scan# 2536


Delta R.T. 0.003 min

Lab File: BN037192.D

Acq: 09 Jun 2025 14:33

Tgt Ion:252	Resp:	4787
Ion Ratio	Lower	Upper
252	100	
253	30.3	22.2 33.4
125	19.7	13.2 19.8





#39

Benzo(a)pyrene

Concen: 0.383 ng

RT: 23.284 min Scan# 2712

Delta R.T. 0.003 min

Lab File: BN037192.D

Acq: 09 Jun 2025 14:33

Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MS





Tgt Ion:	252	Resp:	4119
Ion Ratio	Lower	Upper	
252	100		
253	34.7	25.0	37.4
125	25.3	17.0	25.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.411 ng




RT: 25.590 min Scan# 3502


Delta R.T. 0.003 min

Lab File: BN037192.D

Acq: 09 Jun 2025 14:33

Tgt Ion:	278	Resp:	4003
Ion Ratio	Lower	Upper	
278	100		
139	22.2	17.6	26.4
279	32.5	26.0	39.0





#41

Benzo(g,h,i)perylene

Concen: 0.377 ng

RT: 26.245 min Scan# 31

Delta R.T. 0.006 min

Lab File: BN037192.D


Acq: 09 Jun 2025 14:33

Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MS



Tgt Ion:276 Resp: 4218

Ion Ratio Lower Upper

276 100

277 26.5 20.9 31.3



138 27.0 20.8 31.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



Report of Analysis

Client:	G Environmental	Date Collected:	06/05/25
Project:	Power	Date Received:	06/05/25
Client Sample ID:	MW-11A-13.5-060525MSD	SDG No.:	Q2209
Lab Sample ID:	Q2250-03MSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN037193.D	1	06/06/25 11:54	06/09/25 15:47	PB168336

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
56-55-3	Benzo(a)anthracene	0.44		0.040	0.10	ug/L
205-99-2	Benzo(b)fluoranthene	0.38		0.040	0.10	ug/L
207-08-9	Benzo(k)fluoranthene	0.39		0.050	0.10	ug/L
50-32-8	Benzo(a)pyrene	0.38		0.040	0.10	ug/L
191-24-2	Benzo(g,h,i)perylene	0.39		0.040	0.10	ug/L
SURROGATES						
7297-45-2	2-Methylnaphthalene-d10	0.30		30 (20) - 150 (139)	75%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 (54) - 150 (157)	91%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		30 (27) - 130 (154)	79%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		30 (30) - 130 (155)	86%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		30 (54) - 130 (175)	113%	SPK: 0.4
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	2170	7.59			
1146-65-2	Naphthalene-d8	5650	10.362			
15067-26-2	Acenaphthene-d10	2930	14.235			
1517-22-2	Phenanthrene-d10	5140	16.984			
1719-03-5	Chrysene-d12	3420	21.18			
1520-96-3	Perylene-d12	3340	23.374			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
 Data File : BN037193.D
 Acq On : 09 Jun 2025 15:47
 Operator : RC/JU
 Sample : Q2250-03MSD
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-11A-13.5-060525MSD

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 06/10/2025
 Supervised By :Jagrut Upadhyay 06/10/2025

Quant Time: Jun 09 16:53:21 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 04 01:52:03 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.590	152	2169	0.400	ng	0.00
7) Naphthalene-d8	10.362	136	5646	0.400	ng	#-0.01
13) Acenaphthene-d10	14.235	164	2926	0.400	ng	0.00
19) Phenanthrene-d10	16.984	188	5139	0.400	ng	0.00
29) Chrysene-d12	21.180	240	3419	0.400	ng	# 0.00
35) Perylene-d12	23.374	264	3336	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.185	112	842	0.157	ng	0.00
5) Phenol-d6	6.773	99	692	0.106	ng	0.00
8) Nitrobenzene-d5	8.739	82	1894	0.318	ng	0.00
11) 2-Methylnaphthalene-d10	11.966	152	2373m	0.302	ng	0.00
14) 2,4,6-Tribromophenol	15.730	330	451	0.383	ng	-0.01
15) 2-Fluorobiphenyl	12.858	172	4311	0.346	ng	0.00
27) Fluoranthene-d10	19.022	212	4775	0.366	ng	0.00
31) Terphenyl-d14	19.630	244	3637	0.452	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.112	88	9354	3.235	ng	95
3) n-Nitrosodimethylamine	3.430	42	740	0.127	ng	# 83
6) bis(2-Chloroethyl)ether	7.019	93	2082	0.336	ng	97
9) Naphthalene	10.415	128	5110	0.314	ng	98
10) Hexachlorobutadiene	10.703	225	906	0.255	ng	# 98
12) 2-Methylnaphthalene	12.037	142	3210	0.307	ng	98
16) Acenaphthylene	13.957	152	5333	0.372	ng	100
17) Acenaphthene	14.299	154	3131	0.336	ng	99
18) Fluorene	15.293	166	4472	0.365	ng	98
20) 4,6-Dinitro-2-methylph...	15.379	198	471	0.587	ng	# 63
21) 4-Bromophenyl-phenylether	16.190	248	1330	0.395	ng	# 79
22) Hexachlorobenzene	16.301	284	1342	0.369	ng	98
23) Atrazine	16.463	200	1184	0.426	ng	# 93
24) Pentachlorophenol	16.636	266	1464	0.888	ng	98
25) Phenanthrene	17.021	178	6930	0.416	ng	99
26) Anthracene	17.120	178	5874	0.387	ng	98
28) Fluoranthene	19.054	202	6813	0.370	ng	# 97
30) Pyrene	19.417	202	6824	0.409	ng	100
32) Benzo(a)anthracene	21.171	228	5337	0.431	ng	100
33) Chrysene	21.216	228	5681	0.412	ng	98
34) Bis(2-ethylhexyl)phtha...	21.108	149	3448	0.442	ng	100
36) Indeno(1,2,3-cd)pyrene	25.573	276	5422	0.409	ng	99
37) Benzo(b)fluoranthene	22.720	252	5075m	0.377	ng	
38) Benzo(k)fluoranthene	22.761	252	5262	0.383	ng	95
39) Benzo(a)pyrene	23.281	252	4286	0.380	ng	# 88
40) Dibenzo(a,h)anthracene	25.585	278	4235	0.414	ng	99
41) Benzo(g,h,i)perylene	26.240	276	4499	0.383	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN060925\
Data File : BN037193.D
Acq On : 09 Jun 2025 15:47
Operator : RC/JU
Sample : Q2250-03MSD
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD

Quant Time: Jun 09 16:53:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN060325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jun 04 01:52:03 2025
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.590 min Scan# 61
Delta R.T. -0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion: 152 Resp: 2169
Ion Ratio Lower Upper
152 100
150 150.2 123.2 184.8
115 67.7 56.6 85.0




Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#2
1,4-Dioxane
Concen: 3.235 ng
RT: 3.112 min Scan# 32
Delta R.T. -0.014 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47

Tgt Ion: 88 Resp: 9354
Ion Ratio Lower Upper
88 100
43 60.0 43.5 65.3
58 87.0 67.7 101.5



Abundance Scan 77 (3.437 min): BN037145.D\data.ms (-73)



#3

n-Nitrosodimethylamine

Concen: 0.127 ng

RT: 3.430 min Scan# 70

Delta R.T. -0.007 min

Lab File: BN037193.D

Acq: 09 Jun 2025 15:47


Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MSD

Abundance Scan 76 (3.430 min): BN037193.D\data.ms



Tgt Ion: 42 Resp: 740

Ion Ratio Lower Upper

42 100

74 80.4 53.0 79.4

44 12.2 5.9 8.9



Manual Integrations

APPROVED


Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025

Abundance Scan 76 (3.430 min): BN037193.D\data.ms (-56)



Abundance Scan 319 (5.184 min): BN037145.D\data.ms (-31)



#4

2-Fluorophenol

Concen: 0.157 ng

RT: 5.185 min Scan# 319

Delta R.T. 0.000 min

Lab File: BN037193.D

Acq: 09 Jun 2025 15:47

Tgt Ion:112 Resp: 842



Ion Ratio Lower Upper

112 100


64 71.9 56.3 84.5


63 48.2 36.2 54.4

Abundance Scan 319 (5.185 min): BN037193.D\data.ms




Abundance Scan 319 (5.185 min): BN037193.D\data.ms (-29)





#5
Phenol-d6
Concen: 0.106 ng
RT: 6.773 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion: 99 Resp: 692
Ion Ratio Lower Upper
99 100
42 46.7 31.3 46.9
71 51.4 38.2 57.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.336 ng
RT: 7.019 min Scan# 573
Delta R.T. -0.007 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion: 93 Resp: 2082
Ion Ratio Lower Upper
93 100
63 87.8 68.6 103.0
95 32.5 24.3 36.5





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.362 min Scan# 945
Delta R.T. -0.011 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:136 Resp: 5640
Ion Ratio Lower Upper
136 100
137 12.6 9.7 14.5
54 15.8 9.7 14.5
68 8.5 5.4 8.2#




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#8
Nitrobenzene-d5
Concen: 0.318 ng
RT: 8.739 min Scan# 792
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion: 82 Resp: 1894
Ion Ratio Lower Upper
82 100
128 40.0 26.9 40.3
54 70.1 61.4 92.2





#9
Naphthalene
Concen: 0.314 ng
RT: 10.415 min Scan# 949
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:128 Resp: 5110
Ion Ratio Lower Upper
128 100
129 12.8 9.8 14.8
127 14.5 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#10
Hexachlorobutadiene
Concen: 0.255 ng
RT: 10.703 min Scan# 976
Delta R.T. -0.011 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:225 Resp: 906
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.5 50.3 75.5





#11
2-Methylnaphthalene-d10
Concen: 0.302 ng m
RT: 11.966 min Scan# 1142
Delta R.T. -0.005 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:152 Resp: 237
Ion Ratio Lower Upper
152 100
151 21.5 17.1 25.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#12
2-Methylnaphthalene
Concen: 0.307 ng
RT: 12.037 min Scan# 1156
Delta R.T. -0.005 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:142 Resp: 3210
Ion Ratio Lower Upper
142 100
141 91.6 74.6 111.8
115 48.9 41.0 61.4





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.235 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD




Tgt Ion:164 Resp: 2920
Ion Ratio Lower Upper
164 100
162 105.1 85.5 128.3
160 54.2 44.6 67.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025




#14
2,4,6-Tribromophenol
Concen: 0.383 ng
RT: 15.730 min Scan# 1599
Delta R.T. -0.012 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Tgt Ion:330 Resp: 451
Ion Ratio Lower Upper
330 100
332 95.6 77.1 115.7
141 63.6 46.4 69.6





#15
2-Fluorobiphenyl
Concen: 0.346 ng
RT: 12.858 min Scan# 1318
Delta R.T. -0.005 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:172 Resp: 431
Ion Ratio Lower Upper
172 100
171 37.0 29.6 44.4
170 25.4 20.3 30.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#16
Acenaphthylene
Concen: 0.372 ng
RT: 13.957 min Scan# 1436
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:152 Resp: 5333
Ion Ratio Lower Upper
152 100
151 20.3 16.3 24.5
153 13.4 10.6 15.8





#17
Acenaphthene
Concen: 0.336 ng
RT: 14.299 min Scan# 1468
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:154 Resp: 313
Ion Ratio Lower Upper
154 100
153 118.6 93.8 140.8
152 62.3 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#18
Fluorene
Concen: 0.365 ng
RT: 15.293 min Scan# 1561
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:166 Resp: 4472
Ion Ratio Lower Upper
166 100
165 99.3 81.1 121.7
167 13.3 10.8 16.2





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.984 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:188 Resp: 5139
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.7 11.3 16.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#20
4,6-Dinitro-2-methylphenol
Concen: 0.587 ng
RT: 15.379 min Scan# 1569
Delta R.T. -0.010 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:198 Resp: 471
Ion Ratio Lower Upper
198 100
51 97.7 125.2 187.8#
105 54.7 57.1 85.7#





#21
4-Bromophenyl-phenylether
Concen: 0.395 ng
RT: 16.190 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:248 Resp: 1330
Ion Ratio Lower Upper
248 100
250 106.3 76.1 114.1
141 47.8 60.1 90.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#22
Hexachlorobenzene
Concen: 0.369 ng
RT: 16.301 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:284 Resp: 1342
Ion Ratio Lower Upper
284 100
142 53.1 44.0 66.0
249 36.2 29.7 44.5





#23
Atrazine
Concen: 0.426 ng
RT: 16.463 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:200 Resp: 1184
Ion Ratio Lower Upper
200 100
173 26.9 28.1 42.1
215 47.4 39.3 58.9




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#24
Pentachlorophenol
Concen: 0.888 ng
RT: 16.636 min Scan# 1672
Delta R.T. -0.025 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:266 Resp: 1464
Ion Ratio Lower Upper
266 100
264 63.6 49.3 73.9
268 62.1 49.0 73.4





#25
Phenanthrene
Concen: 0.416 ng
RT: 17.021 min Scan# 1703
Delta R.T. -0.012 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:178 Resp: 6930
Ion Ratio Lower Upper
178 100
176 19.9 15.7 23.5
179 15.1 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#26
Anthracene
Concen: 0.387 ng
RT: 17.120 min Scan# 1711
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:178 Resp: 5874
Ion Ratio Lower Upper
178 100
176 19.4 15.2 22.8
179 14.7 12.9 19.3





#27
Fluoranthene-d10
Concen: 0.366 ng
RT: 19.022 min Scan# 1934
Delta R.T. -0.005 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:212 Resp: 4779
Ion Ratio Lower Upper
212 100
106 13.6 10.6 15.8
104 8.3 6.6 9.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#28
Fluoranthene
Concen: 0.370 ng
RT: 19.054 min Scan# 1940
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:202 Resp: 6813
Ion Ratio Lower Upper
202 100
101 13.6 8.7 13.1#
203 17.2 13.5 20.3





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.180 min Scan# 2118
Delta R.T. -0.009 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:240 Resp: 3419
Ion Ratio Lower Upper
240 100
120 17.9 9.0 13.4
236 30.6 23.0 34.4




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#30
Pyrene
Concen: 0.409 ng
RT: 19.417 min Scan# 2018
Delta R.T. -0.005 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:202 Resp: 6824
Ion Ratio Lower Upper
202 100
200 21.4 17.0 25.6
203 17.6 14.2 21.4





#31
Terphenyl-d14
Concen: 0.452 ng
RT: 19.630 min Scan# 2065
Delta R.T. -0.005 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:244 Resp: 3637
Ion Ratio Lower Upper
244 100
212 11.6 10.0 15.0
122 16.8 13.2 19.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#32
Benzo(a)anthracene
Concen: 0.431 ng
RT: 21.171 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47


Tgt Ion:228 Resp: 5337
Ion Ratio Lower Upper
228 100
226 28.2 22.6 33.8
229 20.5 16.2 24.2





#33
Chrysene
Concen: 0.412 ng
RT: 21.216 min Scan# 2116
Delta R.T. -0.009 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47




Instrument :
BNA_N
ClientSampleId :
MW-11A-13.5-060525MSD



Tgt Ion:228 Resp: 568
Ion Ratio Lower Upper
228 100
226 32.2 25.2 37.8
229 20.0 16.8 25.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 06/10/2025
Supervised By :Jagrut Upadhyay 06/10/2025



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.442 ng
RT: 21.108 min Scan# 2268
Delta R.T. -0.009 min
Lab File: BN037193.D
Acq: 09 Jun 2025 15:47

Tgt Ion:149 Resp: 3448
Ion Ratio Lower Upper
149 100
167 26.0 21.0 31.4
279 3.7 2.9 4.3





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.374 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN037193.D


Acq: 09 Jun 2025 15:47

Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MSD



Tgt Ion:264 Resp: 3330

Ion Ratio Lower Upper

264 100

260 26.8 22.1 33.1




265 79.8 55.8 83.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.409 ng

RT: 25.573 min Scan# 3496

Delta R.T. 0.000 min

Lab File: BN037193.D

Acq: 09 Jun 2025 15:47




Tgt Ion:276 Resp: 5422


Ion Ratio Lower Upper

276 100

138 26.2 21.0 31.6

277 24.7 19.4 29.2





#37

Benzo(b)fluoranthene

Concen: 0.377 ng m

RT: 22.720 min Scan# 2520

Delta R.T. -0.003 min

Lab File: BN037193.D

Acq: 09 Jun 2025 15:47

Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MSD





Tgt Ion:252	Resp:	507
Ion Ratio	Lower	Upper
252	100	
253	30.9	22.3 33.5
125	20.1	13.2 19.8

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#38

Benzo(k)fluoranthene

Concen: 0.383 ng




RT: 22.761 min Scan# 2534


Delta R.T. -0.003 min

Lab File: BN037193.D

Acq: 09 Jun 2025 15:47

Tgt Ion:252	Resp:	5262
Ion Ratio	Lower	Upper
252	100	
253	30.1	22.2 33.4
125	18.6	13.2 19.8





#39

Benzo(a)pyrene

Concen: 0.380 ng

RT: 23.281 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037193.D


Acq: 09 Jun 2025 15:47

Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MSD






Tgt Ion:	252	Resp:	4280
Ion Ratio	Lower	Upper	
252	100		
253	35.8	25.0	37.4
125	29.7	17.0	25.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025



#40

Dibenzo(a,h)anthracene


Concen: 0.414 ng

RT: 25.585 min Scan# 3500



Delta R.T. -0.003 min


Lab File: BN037193.D

Acq: 09 Jun 2025 15:47



Tgt Ion:	278	Resp:	4235
Ion Ratio	Lower	Upper	
278	100		
139	23.1	17.6	26.4
279	32.0	26.0	39.0





#41

Benzo(g,h,i)perylene

Concen: 0.383 ng

RT: 26.240 min Scan# 31

Delta R.T. 0.000 min

Lab File: BN037193.D


Acq: 09 Jun 2025 15:47

Instrument :

BNA_N

ClientSampleId :

MW-11A-13.5-060525MSD



Tgt Ion:276 Resp: 4499

Ion Ratio Lower Upper

276 100

277 26.5 20.9 31.3



138 28.6 20.8 31.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 06/10/2025

Supervised By :Jagrut Upadhyay 06/10/2025





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

BN060325

Instrument

BNA_n

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

BN060925

Instrument

BNA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q2250-02MS	BN037192.D	2-Methylnaphthalene-d10	Rahul	6/10/2025 11:44:39 AM	Jagrut	6/10/2025 1:37:49 PM	Peak Integrated by Software
Q2250-02MS	BN037192.D	Benzo(b)fluoranthene	Rahul	6/10/2025 11:44:39 AM	Jagrut	6/10/2025 1:37:49 PM	Peak Integrated by Software
Q2250-03MSD	BN037193.D	2-Methylnaphthalene-d10	Rahul	6/10/2025 11:44:42 AM	Jagrut	6/10/2025 1:37:52 PM	Peak Integrated by Software
Q2250-03MSD	BN037193.D	Benzo(b)fluoranthene	Rahul	6/10/2025 11:44:42 AM	Jagrut	6/10/2025 1:37:52 PM	Peak Integrated by Software

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN060325

Review By	Rahul	Review On	6/4/2025 11:44:25 AM
Supervise By	Jagrut	Supervise On	6/5/2025 10:56:16 AM
SubDirectory	BN060325	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn060325
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC	SP6779		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN037142.D	03 Jun 2025 10:21	RC/JU	Ok
2	SSTDIC0.1	BN037143.D	03 Jun 2025 11:39	RC/JU	Ok
3	SSTDIC0.2	BN037144.D	03 Jun 2025 12:15	RC/JU	Ok
4	SSTDIC0.4	BN037145.D	03 Jun 2025 12:51	RC/JU	Ok
5	SSTDIC0.8	BN037146.D	03 Jun 2025 13:26	RC/JU	Ok
6	SSTDIC1.6	BN037147.D	03 Jun 2025 14:02	RC/JU	Ok
7	SSTDIC3.2	BN037148.D	03 Jun 2025 14:38	RC/JU	Ok
8	SSTDIC5.0	BN037149.D	03 Jun 2025 15:14	RC/JU	Ok
9	SSTDICV0.4	BN037150.D	03 Jun 2025 15:53	RC/JU	Ok
10	PB168238BL	BN037151.D	03 Jun 2025 17:05	RC/JU	Not Ok
11	Q2181-01	BN037152.D	03 Jun 2025 17:41	RC/JU	Dilution
12	Q2181-01DL	BN037153.D	03 Jun 2025 18:18	RC/JU	Ok
13	SSTDIC0.4	BN037154.D	03 Jun 2025 18:54	RC/JU	Ok
14	DFTPP	BN037155.D	03 Jun 2025 20:10	RC/JU	Ok
15	SSTDIC0.4	BN037156.D	03 Jun 2025 20:49	RC/JU	Ok
16	PB168238BL	BN037157.D	03 Jun 2025 21:25	RC/JU	Not Ok
17	Q2162-03	BN037158.D	03 Jun 2025 22:01	RC/JU	Ok
18	Q2162-07	BN037159.D	03 Jun 2025 22:37	RC/JU	Ok
19	Q2162-09	BN037160.D	03 Jun 2025 23:13	RC/JU	Ok
20	Q2162-10	BN037161.D	03 Jun 2025 23:49	RC/JU	Ok
21	PB168238BS	BN037162.D	04 Jun 2025 00:25	RC/JU	Not Ok

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN060325

Review By	Rahul	Review On	6/4/2025 11:44:25 AM
Supervise By	Jagrut	Supervise On	6/5/2025 10:56:16 AM
SubDirectory	BN060325	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn060325
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC	SP6779		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PB168238BSD	BN037163.D	04 Jun 2025 01:01	RC/JU	Not Ok
23	SSTDCCC0.4	BN037164.D	04 Jun 2025 02:13	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN060925

Review By	Rahul	Review On	6/10/2025 11:45:37 AM
Supervise By	Jagrut	Supervise On	6/10/2025 1:38:10 PM
SubDirectory	BN060925	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn060325
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC	SP6779		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN037188.D	09 Jun 2025 10:15	RC/JU	Ok
2	SSTDCCC0.4	BN037189.D	09 Jun 2025 10:54	RC/JU	Ok
3	PB168336BL	BN037190.D	09 Jun 2025 11:30	RC/JU	Ok
4	Q2250-01	BN037191.D	09 Jun 2025 12:06	RC/JU	Ok
5	Q2250-02MS	BN037192.D	09 Jun 2025 14:33	RC/JU	Ok,M
6	Q2250-03MSD	BN037193.D	09 Jun 2025 15:47	RC/JU	Ok,M
7	Q2251-03	BN037194.D	09 Jun 2025 16:26	RC/JU	Ok
8	Q2251-05	BN037195.D	09 Jun 2025 17:02	RC/JU	Ok
9	Q2251-06	BN037196.D	09 Jun 2025 17:39	RC/JU	Ok
10	Q2253-01	BN037197.D	09 Jun 2025 18:15	RC/JU	Ok
11	Q2253-02	BN037198.D	09 Jun 2025 18:51	RC/JU	Ok
12	Q2250-05	BN037199.D	09 Jun 2025 19:27	RC/JU	Ok
13	Q2254-01	BN037200.D	09 Jun 2025 20:04	RC/JU	Ok
14	PB168336BS	BN037201.D	09 Jun 2025 20:40	RC/JU	Ok
15	SSTDCCC0.4	BN037202.D	09 Jun 2025 21:16	RC/JU	Ok
16	DFTPP	BN037203.D	09 Jun 2025 22:32	RC/JU	Ok
17	SSTDCCC0.4	BN037204.D	09 Jun 2025 23:11	RC/JU	Ok
18	PB168336BL	BN037205.D	09 Jun 2025 23:48	RC/JU	Not Ok
19	Q2234-01	BN037206.D	10 Jun 2025 00:24	RC/JU	Ok
20	Q2234-05	BN037207.D	10 Jun 2025 01:00	RC/JU	Ok
21	Q2234-06	BN037208.D	10 Jun 2025 01:36	RC/JU	Ok

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN060925

Review By	Rahul	Review On	6/10/2025 11:45:37 AM
Supervise By	Jagrut	Supervise On	6/10/2025 1:38:10 PM
SubDirectory	BN060925	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn060325
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC	SP6779		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q2234-07	BN037209.D	10 Jun 2025 02:12	RC/JU	Dilution
23	Q2250-04	BN037210.D	10 Jun 2025 02:49	RC/JU	Ok
24	Q2209-01	BN037211.D	10 Jun 2025 03:25	RC/JU	Ok
25	Q2210-01	BN037212.D	10 Jun 2025 04:01	RC/JU	Ok
26	Q2234-07DL	BN037213.D	10 Jun 2025 09:49	RC/JU	Ok
27	SSTDCCC0.4	BN037214.D	10 Jun 2025 10:25	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN060325

Review By	Rahul	Review On	6/4/2025 11:44:25 AM
Supervise By	Jagrut	Supervise On	6/5/2025 10:56:16 AM
SubDirectory	BN060325	HP Acquire Method	BNA_N, 8270_HP Processing Method bn060325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6779 SP6740,1ul/100ul sample SP6768

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037142.D	03 Jun 2025 10:21		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN037143.D	03 Jun 2025 11:39	Compound #20,24 removed from 0.1 PPM	RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN037144.D	03 Jun 2025 12:15		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN037145.D	03 Jun 2025 12:51	Compound #20,24 kept on LR.	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN037146.D	03 Jun 2025 13:26		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN037147.D	03 Jun 2025 14:02		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN037148.D	03 Jun 2025 14:38	Method is good for DOD and NONDOD.	RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN037149.D	03 Jun 2025 15:14		RC/JU	Ok
9	SSTDICV0.4	ICVBN060325	BN037150.D	03 Jun 2025 15:53		RC/JU	Ok
10	PB168238BL	PB168238BL	BN037151.D	03 Jun 2025 17:05	Not Used	RC/JU	Not Ok
11	Q2181-01	38072-062624	BN037152.D	03 Jun 2025 17:41	Need 50X Dilution	RC/JU	Dilution
12	Q2181-01DL	38072-062624DL	BN037153.D	03 Jun 2025 18:18		RC/JU	Ok
13	SSTDCCC0.4	SSTDCCC0.4EC	BN037154.D	03 Jun 2025 18:54		RC/JU	Ok
14	DFTPP	DFTPP	BN037155.D	03 Jun 2025 20:10		RC/JU	Ok
15	SSTDCCC0.4	SSTDCCC0.4	BN037156.D	03 Jun 2025 20:49		RC/JU	Ok
16	PB168238BL	PB168238BL	BN037157.D	03 Jun 2025 21:25	Not Used	RC/JU	Not Ok
17	Q2162-03	BP-VPB-182-GW-580-5	BN037158.D	03 Jun 2025 22:01		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN060325

Review By	Rahul	Review On	6/4/2025 11:44:25 AM
Supervise By	Jagrut	Supervise On	6/5/2025 10:56:16 AM
SubDirectory	BN060325	HP Acquire Method	BNA_N, 8270_HP Processing Method bn060325
STD. NAME	STD REF.#		
Tune/Reschk	SP6757		
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775		
CCC	SP6779		
Internal Standard/PEM	SP6740,1ul/100ul sample		
ICV/I.BLK	SP6768		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	Q2162-07	BP-VPB-182-GW-620-6	BN037159.D	03 Jun 2025 22:37		RC/JU	Ok
19	Q2162-09	BP-VPB-182-DUP-2025	BN037160.D	03 Jun 2025 23:13		RC/JU	Ok
20	Q2162-10	BP-VPB-182-EB-20250	BN037161.D	03 Jun 2025 23:49		RC/JU	Ok
21	PB168238BS	PB168238BS	BN037162.D	04 Jun 2025 00:25	Recovery Fail for 1,4 Dioxane from low side	RC/JU	Not Ok
22	PB168238BSD	PB168238BSD	BN037163.D	04 Jun 2025 01:01	Recovery Fail for 1,4 Dioxane from low side	RC/JU	Not Ok
23	SSTDCCC0.4	SSTDCCC0.4EC	BN037164.D	04 Jun 2025 02:13		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN060925

Review By	Rahul	Review On	6/10/2025 11:45:37 AM
Supervise By	Jagrut	Supervise On	6/10/2025 1:38:10 PM
SubDirectory	BN060925	HP Acquire Method	BNA_N, 8270_HP Processing Method bn060325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775
CCC	SP6779
Internal Standard/PEM	SP6740,1ul/100ul sample
ICV/I.BLK	SP6768
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037188.D	09 Jun 2025 10:15		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN037189.D	09 Jun 2025 10:54		RC/JU	Ok
3	PB168336BL	PB168336BL	BN037190.D	09 Jun 2025 11:30		RC/JU	Ok
4	Q2250-01	MW-11A-13.5-060525	BN037191.D	09 Jun 2025 12:06		RC/JU	Ok
5	Q2250-02MS	MW-11A-13.5-060525M	BN037192.D	09 Jun 2025 14:33		RC/JU	Ok,M
6	Q2250-03MSD	MW-11A-13.5-060525M	BN037193.D	09 Jun 2025 15:47		RC/JU	Ok,M
7	Q2251-03	BP-VPB-182-GW-760-7	BN037194.D	09 Jun 2025 16:26		RC/JU	Ok
8	Q2251-05	BP-VPB-182-EB-20250	BN037195.D	09 Jun 2025 17:02		RC/JU	Ok
9	Q2251-06	VPB182-HYD-2025060	BN037196.D	09 Jun 2025 17:39		RC/JU	Ok
10	Q2253-01	RW8-SP100-20250605	BN037197.D	09 Jun 2025 18:15		RC/JU	Ok
11	Q2253-02	RW8-SP303-20250605	BN037198.D	09 Jun 2025 18:51		RC/JU	Ok
12	Q2250-05	EB02-060525	BN037199.D	09 Jun 2025 19:27		RC/JU	Ok
13	Q2254-01	BP-VPB-182-GW-810-8	BN037200.D	09 Jun 2025 20:04		RC/JU	Ok
14	PB168336BS	PB168336BS	BN037201.D	09 Jun 2025 20:40		RC/JU	Ok
15	SSTDCCC0.4	SSTDCCC0.4EC	BN037202.D	09 Jun 2025 21:16		RC/JU	Ok
16	DFTPP	DFTPP	BN037203.D	09 Jun 2025 22:32		RC/JU	Ok
17	SSTDCCC0.4	SSTDCCC0.4	BN037204.D	09 Jun 2025 23:11		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN060925

Review By	Rahul	Review On	6/10/2025 11:45:37 AM
Supervise By	Jagrut	Supervise On	6/10/2025 1:38:10 PM
SubDirectory	BN060925	HP Acquire Method	BNA_N, 8270_HP Processing Method bn060325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6757 SP6781,SP6780,SP6779,SP6778,SP6777,SP6776,SP6775
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6779 SP6740,1ul/100ul sample SP6768

18	PB168336BL	PB168336BL	BN037205.D	09 Jun 2025 23:48	analyzed to check contamination.	RC/JU	Not Ok
19	Q2234-01	MW-17B-55-060425	BN037206.D	10 Jun 2025 00:24		RC/JU	Ok
20	Q2234-05	MW-18B-56-060425	BN037207.D	10 Jun 2025 01:00		RC/JU	Ok
21	Q2234-06	MW-18B-56-060425-FD	BN037208.D	10 Jun 2025 01:36		RC/JU	Ok
22	Q2234-07	MW-19B-72-060425	BN037209.D	10 Jun 2025 02:12	Need 2X dilutiion	RC/JU	Dilution
23	Q2250-04	MW-06-6.5-060525	BN037210.D	10 Jun 2025 02:49		RC/JU	Ok
24	Q2209-01	P01W	BN037211.D	10 Jun 2025 03:25		RC/JU	Ok
25	Q2210-01	TW1	BN037212.D	10 Jun 2025 04:01		RC/JU	Ok
26	Q2234-07DL	MW-19B-72-060425DL	BN037213.D	10 Jun 2025 09:49		RC/JU	Ok
27	SSTDCCC0.4	SSTDCCC0.4EC	BN037214.D	10 Jun 2025 10:25		RC/JU	Ok

M : Manual Integration

SOP ID: M3510C,3580A-Extraction SVOC-20

Clean Up SOP #: N/A

Matrix : Water

Welgh By: N/A **Extraction By:** RS

Balance check: N/A **Filter By:** RJ

Balance ID: N/A **pH Meter ID:** N/A

pH Strip Lot#: E3880 **Hood ID:** 4,5,6,7

Extraction Method: ☒ Separatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Extraction Start Date : 06/06/2025

Extraction Start Time : 11:54

Extraction End Date : 06/06/2025

Extraction End Time : 17:10

Concentration By: EH

Supervisor By : RUPESH

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	0.4 PPM	SP6756
Surrogate	1.0ML	0.4 PPM	SP6758
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3939
Baked Na2SO4	N/A	EP2620
10N NaoH	N/A	EP2609
H2SO4 1:1	N/A	EP2610
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210443. pH Adjusted<2 with 1:1 H2SO4 &>11 with 10 N NaOH.

KD Bath ID: WATER BATH-1,2 **Envap ID:** NEVAP-02

KD Bath Temperature: 60 °C **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Recelved By/Location
6/6/25	RS (EX4-Lab)	JH / SVOC
17:15	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-20

Concentration Date: 06/06/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB168336BL	SBLK336	SVOC-SIMGrou p1	1000	6	RUPEsh	ritesh	1			SEP-1
PB168336BS	SLCS336	SVOC-SIMGrou p1	1000	6	RUPEsh	ritesh	1			2
Q2209-01	P01W	SVOC-SIMGrou p1	990	6	RUPEsh	ritesh	1	C		3
Q2210-01	TW1	SVOCMS Group2	1000	6	RUPEsh	ritesh	1	E		4
Q2234-01	MW-17B-55-060425	SVOC-SIMGrou p1	980	6	RUPEsh	ritesh	1	H		5
Q2234-05	MW-18B-56-060425	SVOC-SIMGrou p1	970	11	RUPEsh	ritesh	1	E		6
Q2234-06	MW-18B-56-060425-FD	SVOC-SIMGrou p1	1000	11	RUPEsh	ritesh	1	E		7
Q2234-07	MW-19B-72-060425	SVOC-SIMGrou p1	980	6	RUPEsh	ritesh	1	E		8
Q2250-01	MW-11A-13.5-060525	SVOC-SIMGrou p1	930	6	RUPEsh	ritesh	1	E		9
Q2250-02	Q2250-01MS	SVOC-SIMGrou p1	960	6	RUPEsh	ritesh	1	E		10
Q2250-03	Q2250-01MSD	SVOC-SIMGrou p1	990	6	RUPEsh	ritesh	1	E		11
Q2250-04	MW-06-6.5-060525	SVOC-SIMGrou p1	970	6	RUPEsh	ritesh	1	A		12
Q2250-05	EB02-060525	SVOC-SIMGrou p1	990	6	RUPEsh	ritesh	1	J		13
Q2251-03	BP-VPB-182-GW-760-762	SVOC-SIMGrou p1	850	6	RUPEsh	ritesh	1	C		14
Q2251-05	BP-VPB-182-EB-2025060 4	SVOC-SIMGrou p1	870	6	RUPEsh	ritesh	1	C		15
Q2251-06	VPB182-HYD-20250605	SVOC-SIMGrou p1	890	6	RUPEsh	ritesh	1	C		16
Q2253-01	RW8-SP100-20250605	SVOC-SIMGrou p1	1000	6	RUPEsh	ritesh	1	B		SEP-1
Q2253-02	RW8-SO303-20250605	SVOC-SIMGrou p1	1000	6	RUPEsh	ritesh	1	D		2
Q2254-01	BP-VPB-182-GW-810-812	SVOC-SIMGrou p1	890	6	RUPEsh	ritesh	1			3

* Extracts relinquished on the same date as received.

RS
6/6

16820891
11:47
6/6/25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q2250

WorkList ID : 190013

Department : Extraction

Date : 06-06-2025 11:47:44

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q2209-01	P01W	Water	SVOC-SIMGroup1	Cool 4 deg C	GENV01	N31	06/04/2025	8270-Modified
Q2210-01	TW1	Water	SVOCMS Group2	Cool 4 deg C	GENV01	L31	06/03/2025	8270-Modified
Q2234-01	MW-17B-55-060425	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	N31	06/04/2025	8270-Modified
Q2234-05	MW-18B-56-060425	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	N31	06/04/2025	8270-Modified
Q2234-06	MW-18B-56-060425-FD	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	N31	06/04/2025	8270-Modified
Q2234-07	MW-19B-72-060425	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	N31	06/04/2025	8270-Modified
Q2250-01	MW-11A-13.5-060525	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	D22	06/05/2025	8270-Modified
Q2250-02	Q2250-01MS	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	D22	06/05/2025	8270-Modified
Q2250-03	Q2250-01MSD	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	D22	06/05/2025	8270-Modified
Q2250-04	MW-06-6.5-060525	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	D22	06/05/2025	8270-Modified
Q2250-05	EB02-060525	Water	SVOC-SIMGroup1	Cool 4 deg C	JACO05	D22	06/05/2025	8270-Modified
Q2251-03	BP-VPB-182-GW-760-762	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	L31	06/03/2025	8270-Modified
Q2251-05	BP-VPB-182-EB-20250604	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	L31	06/04/2025	8270-Modified
Q2251-06	VPB182-HYD-20250605	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	L31	06/05/2025	8270-Modified
Q2253-01	RW8-SP100-20250605	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D21	06/05/2025	8270-Modified
Q2253-02	RW8-SO303-20250605	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D21	06/05/2025	8270-Modified
Q2254-01	BP-VPB-182-GW-810-812	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D21	06/05/2025	8270-Modified

Date/Time 6/6/25 11:47
Raw Sample Received by: RS (BGA-606)
Raw Sample Relinquished by: JD GSN

Date/Time 6/6/25 12:45
Raw Sample Received by: JD GSN
Raw Sample Relinquished by: RS (BGA-606)



SHIPPING DOCUMENTS

CLIENT INFORMATION

COMPANY: Gen Environmental
ADDRESS: 8 CARRAGE lane
CITY: Shelton STATE: NJ ZIP: 07876
ATTENTION:
PHONE: FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Power
PROJECT NO.: LOCATION: NJ
PROJECT MANAGER: GV
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Gen Environmental PO#:
ADDRESS: 8 CARRAGE
CITY: Shelton STATE: NJ ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) Standard DAYS*
HARDCOPY (DATA PACKAGE): DAYS*
EDD: DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other excel
☒ EDD FORMAT NJ DEP spec

1 2 3 4 5 6 7 8 9

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES										COMMENTS ← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	
1.	<u>P01W</u>	<u>GW</u>			<u>6/4/25</u>	<u>1100</u>	<u>4</u>	<u>X</u>	<u>X</u>	<u>X</u>							
2.																	
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME:	RECEIVED BY: 1. <u>[Signature]</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>3.4</u> °C
RELINQUISHED BY SAMPLER: 2.	DATE/TIME:	RECEIVED BY: 2. <u>[Signature]</u>	Comments:
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2209	GENV01	Order Date : 6/4/2025 1:52:00 PM	Project Mgr :
Client Name : G Environmental		Project Name : Power	Report Type : Level 1 NJ Reduce
Client Contact : Gary Landis		Receive DateTime : 6/4/2025 1:10:00 PM	EDD Type : HAZ/EXCEL
Invoice Name : G Environmental		Purchase Order :	Hard Copy Date :
Invoice Contact : Gary Landis			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2209-01	P01W	Water	06/04/2025	11:00	VOCMS Group1		8260-Low		10 Bus. Days

Relinquished By : CL

Date / Time : 6/4/25 1435

Received By : Samy

Date / Time : 06/04/25 14:35

Storage Area : VOA Refridgerator Room