

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG NO.: Q2209

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 06/09/2025

Lab File ID: BN037189.D Time Analyzed: 10:54

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2093	7.589	5342	10.36	2894	14.24
UPPER LIMIT	4186	8.089	10684	10.862	5788	14.735
LOWER LIMIT	1046.5	7.089	2671	9.862	1447	13.735
EPA SAMPLE NO.						
01 PB168336BL	1816	7.59	4227	10.37	2101	14.25
02 MW-11A-13.5-060525MS	2144	7.59	5670	10.36	2991	14.23
03 MW-11A-13.5-060525MSD	2169	7.59	5646	10.36	2926	14.24
04 PB168336BS	2227	7.59	5466	10.36	2607	14.23

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.