

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG No.: Q2209
Instrument ID: BNA_P Calibration Date(s): 06/06/2025 06/06/2025
Calibration Time(s): 10:30 15:18

LAB FILE ID:		RRF2.5 = BP024860.D			RRF005 = BP024861.D		RRF010 = BP024862.D	
		RRF020 = BP024863.D			RRF040 = BP024864.D		RRF050 = BP024865.D	
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2-Fluorophenol		1.127	1.139	1.207	1.163	1.283	1.198	4.9
Benzaldehyde			1.038	1.071	0.873	0.985	0.914	17.0
Phenol-d6		1.507	1.528	1.588	1.545	1.676	1.585	4.5
bis(2-Chloroethyl)ether		1.222	1.277	1.334	1.234	1.363	1.285	4.3
2,2-oxybis(1-Chloropropane)		1.748	1.732	1.722	1.583	1.751	1.682	4.5
Acetophenone		0.506	0.521	0.511	0.491	0.535	0.505	4.6
n-Nitroso-di-n-propylamine	0.984	1.101	1.105	1.107	1.043	1.141	1.078	4.9
Nitrobenzene-d5		0.407	0.397	0.423	0.404	0.444	0.412	4.9
Hexachloroethane		0.591	0.565	0.581	0.545	0.611	0.576	3.7
Nitrobenzene		0.366	0.351	0.375	0.360	0.392	0.366	4.8
Isophorone		0.704	0.678	0.724	0.694	0.764	0.713	3.9
bis(2-Chloroethoxy)methane		0.414	0.408	0.438	0.414	0.465	0.426	4.7
Naphthalene		1.071	1.022	1.044	0.989	1.079	1.025	4.9
4-Chloroaniline		0.397	0.401	0.435	0.426	0.471	0.429	6.7
Hexachlorobutadiene		0.203	0.199	0.208	0.194	0.218	0.201	4.8
Caprolactam			0.098	0.109	0.110	0.118	0.109	6.9
2-Methylnaphthalene		0.659	0.638	0.659	0.633	0.696	0.650	4.5
Hexachlorocyclopentadiene			0.259	0.315	0.337	0.404	0.346	15.7
2-Fluorobiphenyl		1.542	1.507	1.517	1.390	1.563	1.485	4.4
1,1-Biphenyl		1.477	1.456	1.485	1.386	1.509	1.453	3.1
2-Chloronaphthalene		1.123	1.104	1.135	1.069	1.171	1.117	3.2
2-Nitroaniline		0.289	0.324	0.344	0.346	0.371	0.343	8.6
Dimethylphthalate		1.515	1.450	1.501	1.400	1.550	1.475	3.4
Acenaphthylene		1.880	1.851	1.892	1.768	1.939	1.863	3.2
2,6-Dinitrotoluene		0.299	0.301	0.326	0.312	0.339	0.318	4.9
3-Nitroaniline		0.263	0.292	0.337	0.338	0.367	0.330	11.7
Acenaphthene		1.106	1.064	1.090	1.020	1.087	1.067	2.9
Dibenzofuran		1.815	1.721	1.756	1.627	1.757	1.713	4.2
2,4-Dinitrotoluene		0.390	0.416	0.457	0.437	0.487	0.445	7.4
Diethylphthalate		1.501	1.474	1.487	1.393	1.545	1.470	3.3
4-Chlorophenyl-phenylether		0.711	0.668	0.689	0.637	0.709	0.677	4.1
Fluorene		1.437	1.394	1.420	1.304	1.434	1.384	3.8
4-Nitroaniline		0.239	0.235	0.307	0.311	0.336	0.299	14.7
n-Nitrosodiphenylamine		0.627	0.608	0.633	0.597	0.659	0.620	3.7

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2209 SAS No.: Q2209 SDG No.: Q2209
Instrument ID: BNA_P Calibration Date(s): 06/06/2025 06/06/2025
Calibration Time(s): 10:30 15:18

LAB FILE ID:		RRF2.5 = BP024860.D		RRF005 = BP024861.D		RRF010 = BP024862.D		
		RRF020 = BP024863.D		RRF040 = BP024864.D		RRF050 = BP024865.D		
COMPOUND	RRF2.5	RRF005	RRF010	RRF020	RRF040	RRF050	RRF	% RSD
2,4,6-Tribromophenol		0.256	0.264	0.279	0.267	0.298	0.277	5.3
4-Bromophenyl-phenylether		0.224	0.215	0.226	0.213	0.246	0.226	4.8
Hexachlorobenzene		0.278	0.268	0.272	0.260	0.290	0.274	3.3
Atrazine		0.213	0.212	0.231	0.217	0.244	0.225	5.2
Phenanthrene		1.158	1.108	1.110	1.056	1.161	1.105	4.1
Anthracene		1.129	1.093	1.137	1.072	1.188	1.119	3.6
Carbazole		1.023	1.013	1.057	0.998	1.112	1.038	3.8
Di-n-butylphthalate		1.178	1.245	1.326	1.272	1.421	1.285	5.8
Fluoranthene		1.300	1.287	1.307	1.223	1.344	1.281	3.3
Pyrene		1.307	1.195	1.261	1.184	1.322	1.249	4.4
Terphenyl-d14		1.178	1.073	1.146	1.089	1.164	1.116	4.6
Butylbenzylphthalate		0.508	0.529	0.581	0.562	0.641	0.572	7.7
3,3-Dichlorobenzidine			0.468	0.513	0.493	0.542	0.508	5.3
Benzo (a) anthracene		1.312	1.234	1.288	1.219	1.347	1.279	3.7
Chrysene		1.252	1.174	1.229	1.144	1.279	1.212	4.1
Bis(2-ethylhexyl)phthalate		0.715	0.780	0.846	0.797	0.921	0.820	7.9
Di-n-octyl phthalate			1.320	1.438	1.384	1.587	1.445	6.2
Benzo (b) fluoranthene		1.103	1.104	1.133	1.127	1.232	1.145	4.1
Benzo (k) fluoranthene		1.165	1.144	1.181	1.106	1.259	1.165	4.1
Benzo (a) pyrene		1.096	1.069	1.127	1.070	1.214	1.118	4.5
Indeno (1,2,3-cd) pyrene		1.427	1.402	1.469	1.412	1.559	1.459	3.9
Dibenzo (a,h) anthracene		1.151	1.143	1.202	1.143	1.279	1.188	4.3
Benzo (g,h,i) perylene		1.172	1.127	1.183	1.136	1.261	1.179	3.9
1,2,4,5-Tetrachlorobenzene		0.568	0.561	0.568	0.538	0.601	0.568	3.3
1,4-Dioxane		0.564	0.529	0.524	0.495	0.546	0.527	4.2

All other compounds must meet a minimum RRF of 0.010.

Form VI SV-1