

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: GENV01

Lab Code: CHEM

SAS No.: Q2209

SDG NO.: Q2209

Lab File ID: BP024870.D

DFTPP Injection Date: 06/09/2025

Instrument ID: BNA_P

DFTPP Injection Time: 10:03

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.2
68	Less than 2.0% of mass 69	0.6 (1.7) 1
69	Mass 69 relative abundance	37.7
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	48.1
197	Less than 2.0% of mass 198	0.1
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	30.7
365	Greater than 1% of mass 198	4.3
441	Present, but less than mass 443	11.6
442	Greater than 50% of mass 198	75.5
443	15.0 - 24.0% of mass 442	14.5 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024871.D	06/09/2025	10:44
PB168323BL	PB168323BL	BP024872.D	06/09/2025	11:24
PB168323BS	PB168323BS	BP024873.D	06/09/2025	12:05
P01W	Q2209-01	BP024875.D	06/09/2025	13:31
GW-MW01-060425MS	Q2230-03MS	BP024879.D	06/09/2025	16:14
GW-MW01-060425MSD	Q2230-04MSD	BP024880.D	06/09/2025	16:55