

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN060625

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	105	0.00
2 T	Dichlorodifluoromethane	0.499	0.551	-10.4	116	0.00
3 P	Chloromethane	0.644	0.637	1.1	112	0.00
4 C	Vinyl Chloride	0.664	0.697	-5.0#	114	0.00
5 T	Bromomethane	0.372	0.411	-10.5	121	0.00
6 T	Chloroethane	0.429	0.439	-2.3	113	0.00
7 T	Trichlorofluoromethane	0.868	0.899	-3.6	113	0.00
8 T	Diethyl Ether	0.378	0.401	-6.1	116	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.585	-7.3	118	0.00
10 T	Methyl Iodide	0.707	0.752	-6.4	116	0.00
11 T	Tert butyl alcohol	0.182	0.191	-4.9	114	0.00
12 CM	1,1-Dichloroethene	0.557	0.575	-3.2#	113	0.00
13 T	Acrolein	0.057	0.055	3.5	128	0.00
14 T	Allyl chloride	0.923	0.932	-1.0	113	0.00
15 T	Acrylonitrile	0.425	0.440	-3.5	114	0.00
16 T	Acetone	0.355	0.350	1.4	114	0.00
17 T	Carbon Disulfide	1.539	1.556	-1.1	115	0.00
18 T	Methyl Acetate	1.035	1.074	-3.8	115	0.00
19 T	Methyl tert-butyl Ether	2.015	2.084	-3.4	113	0.00
20 T	Methylene Chloride	0.665	0.660	0.8	115	0.00
21 T	trans-1,2-Dichloroethene	0.619	0.612	1.1	113	0.00
22 T	Diisopropyl ether	1.945	2.018	-3.8	114	0.00
23 T	Vinyl Acetate	1.644	1.677	-2.0	111	0.00
24 P	1,1-Dichloroethane	1.120	1.153	-2.9	114	0.00
25 T	2-Butanone	0.577	0.591	-2.4	113	0.00
26 T	2,2-Dichloropropane	0.871	0.993	-14.0	116	0.00
27 T	cis-1,2-Dichloroethene	0.740	0.756	-2.2	114	0.00
28 T	Bromochloromethane	0.550	0.499	9.3	113	0.00
29 T	Tetrahydrofuran	0.376	0.386	-2.7	113	0.00
30 C	Chloroform	1.118	1.139	-1.9#	113	0.00
31 T	Cyclohexane	1.086	1.090	-0.4	114	0.00
32 T	1,1,1-Trichloroethane	0.951	0.962	-1.2	113	0.00
33 S	1,2-Dichloroethane-d4	0.669	0.698	-4.3	147	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	104	0.00
35 S	Dibromofluoromethane	0.296	0.322	-8.8	153#	0.00
36 T	1,1-Dichloropropene	0.442	0.460	-4.1	115	0.00
37 T	Ethyl Acetate	0.562	0.563	-0.2	109	0.00
38 T	Carbon Tetrachloride	0.433	0.455	-5.1	116	0.00
39 T	Methylcyclohexane	0.605	0.637	-5.3	116	0.00
40 TM	Benzene	1.444	1.471	-1.9	114	0.00
41 T	Methacrylonitrile	0.316	0.332	-5.1	114	0.00
42 TM	1,2-Dichloroethane	0.438	0.451	-3.0	114	0.00
43 T	Isopropyl Acetate	0.902	0.912	-1.1	111	0.00
44 TM	Trichloroethene	0.342	0.355	-3.8	113	0.00
45 C	1,2-Dichloropropane	0.351	0.361	-2.8#	113	0.00
46 T	Dibromomethane	0.233	0.247	-6.0	116	0.00
47 T	Bromodichloromethane	0.480	0.498	-3.8	113	0.00
48 T	Methyl methacrylate	0.415	0.426	-2.7	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
 Data File : VN086869.D
 Acq On : 06 Jun 2025 15:54
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN060625

Quant Time: Jun 07 02:16:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 07 02:12:50 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	112	0.00
50 S	Toluene-d8	1.173	1.246	-6.2	151#	0.00
51 T	4-Methyl-2-Pentanone	0.543	0.582	-7.2	113	0.00
52 CM	Toluene	0.882	0.918	-4.1#	114	0.00
53 T	t-1,3-Dichloropropene	0.537	0.570	-6.1	114	0.00
54 T	cis-1,3-Dichloropropene	0.574	0.603	-5.1	114	0.00
55 T	1,1,2-Trichloroethane	0.340	0.350	-2.9	113	0.00
56 T	Ethyl methacrylate	0.541	0.613	-13.3	115	0.00
57 T	1,3-Dichloropropane	0.589	0.605	-2.7	112	0.00
58 T	2-Chloroethyl Vinyl ether	0.323	0.317	1.9	120	0.00
59 T	2-Hexanone	0.350	0.397	-13.4	112	0.00
60 T	Dibromochloromethane	0.354	0.372	-5.1	114	0.00
61 T	1,2-Dibromoethane	0.348	0.361	-3.7	114	0.00
62 S	4-Bromofluorobenzene	0.436	0.471	-8.0	151#	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	105	0.00
64 T	Tetrachloroethene	0.316	0.319	-0.9	114	0.00
65 PM	Chlorobenzene	1.103	1.122	-1.7	115	0.00
66 T	1,1,1,2-Tetrachloroethane	0.354	0.370	-4.5	115	0.00
67 C	Ethyl Benzene	1.899	1.960	-3.2#	114	0.00
68 T	m/p-Xylenes	0.727	0.756	-4.0	113	0.00
69 T	o-Xylene	0.696	0.730	-4.9	114	0.00
70 T	Styrene	1.192	1.257	-5.5	113	0.00
71 P	Bromoform	0.263	0.288	-9.5	114	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.643	3.775	-3.6	114	0.00
74 T	N-amyl acetate	1.273	1.380	-8.4	112	0.00
75 P	1,1,2,2-Tetrachloroethane	1.234	1.278	-3.6	112	0.00
76 T	1,2,3-Trichloropropane	1.189	1.279	-7.6	112	0.00
77 T	Bromobenzene	0.835	0.874	-4.7	114	0.00
78 T	n-propylbenzene	4.427	4.639	-4.8	114	0.00
79 T	2-Chlorotoluene	2.655	2.742	-3.3	113	0.00
80 T	1,3,5-Trimethylbenzene	3.007	3.175	-5.6	113	0.00
81 T	trans-1,4-Dichloro-2-butene	0.516	0.575	-11.4	125	0.00
82 T	4-Chlorotoluene	2.686	2.787	-3.8	114	0.00
83 T	tert-Butylbenzene	2.753	2.892	-5.0	114	0.00
84 T	1,2,4-Trimethylbenzene	3.016	3.198	-6.0	113	0.00
85 T	sec-Butylbenzene	3.998	4.192	-4.9	113	0.00
86 T	p-Isopropyltoluene	3.305	3.514	-6.3	114	0.00
87 T	1,3-Dichlorobenzene	1.644	1.684	-2.4	112	0.00
88 T	1,4-Dichlorobenzene	1.676	1.728	-3.1	113	0.00
89 T	n-Butylbenzene	3.201	3.356	-4.8	113	0.00
90 T	Hexachloroethane	0.560	0.592	-5.7	114	0.00
91 T	1,2-Dichlorobenzene	1.579	1.621	-2.7	111	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.295	0.290	1.7	110	0.00
93 T	1,2,4-Trichlorobenzene	1.008	1.048	-4.0	112	0.00
94 T	Hexachlorobutadiene	0.376	0.397	-5.6	113	0.00
95 T	Naphthalene	3.753	3.904	-4.0	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN060625\
Data File : VN086869.D
Acq On : 06 Jun 2025 15:54
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN060625

Quant Time: Jun 07 02:16:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N060625W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 07 02:12:50 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	1.013	-1.1	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6