

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV29	Mercury	3.69	4.0	92	90 - 110	CV	06/06/2025	10:03	LB136036

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Mercury	4.74	5.0	95	90 - 110	CV	06/06/2025	10:07	LB136036
CCV09	Mercury	4.88	5.0	98	90 - 110	CV	06/06/2025	10:37	LB136036
CCV10	Mercury	4.71	5.0	94	90 - 110	CV	06/06/2025	11:04	LB136036
CCV11	Mercury	4.55	5.0	91	90 - 110	CV	06/06/2025	11:41	LB136036
CCV12	Mercury	4.77	5.0	96	90 - 110	CV	06/06/2025	12:06	LB136036
CCV13	Mercury	5.10	5.0	102	90 - 110	CV	06/06/2025	12:33	LB136036

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2550	2500	102	90 - 110	P	06/06/2025	13:16	LB136052
	Antimony	1020	1000	102	90 - 110	P	06/06/2025	13:16	LB136052
	Arsenic	1010	1000	101	90 - 110	P	06/06/2025	13:16	LB136052
	Barium	507	520	98	90 - 110	P	06/06/2025	13:16	LB136052
	Beryllium	485	510	95	90 - 110	P	06/06/2025	13:16	LB136052
	Cadmium	502	510	98	90 - 110	P	06/06/2025	13:16	LB136052
	Calcium	9890	10000	99	90 - 110	P	06/06/2025	13:16	LB136052
	Chromium	522	520	100	90 - 110	P	06/06/2025	13:16	LB136052
	Cobalt	523	520	101	90 - 110	P	06/06/2025	13:16	LB136052
	Copper	527	510	103	90 - 110	P	06/06/2025	13:16	LB136052
	Iron	10300	10000	103	90 - 110	P	06/06/2025	13:16	LB136052
	Lead	980	1000	98	90 - 110	P	06/06/2025	13:16	LB136052
	Magnesium	6090	6000	102	90 - 110	P	06/06/2025	13:16	LB136052
	Manganese	520	520	100	90 - 110	P	06/06/2025	13:16	LB136052
	Nickel	533	530	100	90 - 110	P	06/06/2025	13:16	LB136052
	Potassium	9630	9900	97	90 - 110	P	06/06/2025	13:16	LB136052
	Selenium	998	1000	100	90 - 110	P	06/06/2025	13:16	LB136052
	Silver	243	250	97	90 - 110	P	06/06/2025	13:16	LB136052
	Sodium	10100	10000	101	90 - 110	P	06/06/2025	13:16	LB136052
	Thallium	985	1000	98	90 - 110	P	06/06/2025	13:16	LB136052
	Vanadium	489	500	98	90 - 110	P	06/06/2025	13:16	LB136052
	Zinc	991	1000	99	90 - 110	P	06/06/2025	13:16	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental SDG No.: Q2210
Contract: GENV01 Lab Code: CHEM Case No.: Q2210 SAS No.: Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	105	100	105	80 - 120	P	06/06/2025	13:29	LB136052
	Antimony	47.2	50.0	94	80 - 120	P	06/06/2025	13:29	LB136052
	Arsenic	21.6	20.0	108	80 - 120	P	06/06/2025	13:29	LB136052
	Barium	92.0	100	92	80 - 120	P	06/06/2025	13:29	LB136052
	Beryllium	5.86	6.0	98	80 - 120	P	06/06/2025	13:29	LB136052
	Cadmium	6.02	6.0	100	80 - 120	P	06/06/2025	13:29	LB136052
	Calcium	1940	2000	97	80 - 120	P	06/06/2025	13:29	LB136052
	Chromium	9.45	10.0	94	80 - 120	P	06/06/2025	13:29	LB136052
	Cobalt	29.2	30.0	97	80 - 120	P	06/06/2025	13:29	LB136052
	Copper	21.0	20.0	105	80 - 120	P	06/06/2025	13:29	LB136052
	Iron	106	100	106	80 - 120	P	06/06/2025	13:29	LB136052
	Lead	13.9	12.0	116	80 - 120	P	06/06/2025	13:29	LB136052
	Magnesium	1960	2000	98	80 - 120	P	06/06/2025	13:29	LB136052
	Manganese	19.2	20.0	96	80 - 120	P	06/06/2025	13:29	LB136052
	Nickel	39.7	40.0	99	80 - 120	P	06/06/2025	13:29	LB136052
	Potassium	2290	2000	115	80 - 120	P	06/06/2025	13:29	LB136052
	Selenium	22.5	20.0	112	80 - 120	P	06/06/2025	13:29	LB136052
	Silver	9.79	10.0	98	80 - 120	P	06/06/2025	13:29	LB136052
	Sodium	1630	2000	82	80 - 120	P	06/06/2025	13:29	LB136052
	Thallium	41.0	40.0	102	80 - 120	P	06/06/2025	13:29	LB136052
	Vanadium	38.8	40.0	97	80 - 120	P	06/06/2025	13:29	LB136052
	Zinc	43.4	40.0	108	80 - 120	P	06/06/2025	13:29	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9970	10000	100	90 - 110	P	06/06/2025	14:05	LB136052
	Antimony	5060	5000	101	90 - 110	P	06/06/2025	14:05	LB136052
	Arsenic	5060	5000	101	90 - 110	P	06/06/2025	14:05	LB136052
	Barium	9720	10000	97	90 - 110	P	06/06/2025	14:05	LB136052
	Beryllium	244	250	98	90 - 110	P	06/06/2025	14:05	LB136052
	Cadmium	2460	2500	98	90 - 110	P	06/06/2025	14:05	LB136052
	Calcium	24000	25000	96	90 - 110	P	06/06/2025	14:05	LB136052
	Chromium	995	1000	100	90 - 110	P	06/06/2025	14:05	LB136052
	Cobalt	2460	2500	98	90 - 110	P	06/06/2025	14:05	LB136052
	Copper	1250	1250	100	90 - 110	P	06/06/2025	14:05	LB136052
	Iron	4810	5000	96	90 - 110	P	06/06/2025	14:05	LB136052
	Lead	4910	5000	98	90 - 110	P	06/06/2025	14:05	LB136052
	Magnesium	23900	25000	95	90 - 110	P	06/06/2025	14:05	LB136052
	Manganese	2360	2500	95	90 - 110	P	06/06/2025	14:05	LB136052
	Nickel	2460	2500	98	90 - 110	P	06/06/2025	14:05	LB136052
	Potassium	24400	25000	98	90 - 110	P	06/06/2025	14:05	LB136052
	Selenium	5120	5000	102	90 - 110	P	06/06/2025	14:05	LB136052
	Silver	1240	1250	99	90 - 110	P	06/06/2025	14:05	LB136052
	Sodium	24400	25000	98	90 - 110	P	06/06/2025	14:05	LB136052
	Thallium	5010	5000	100	90 - 110	P	06/06/2025	14:05	LB136052
	Vanadium	2420	2500	97	90 - 110	P	06/06/2025	14:05	LB136052
	Zinc	2480	2500	99	90 - 110	P	06/06/2025	14:05	LB136052
CCV02	Aluminum	9950	10000	100	90 - 110	P	06/06/2025	14:56	LB136052
	Antimony	5150	5000	103	90 - 110	P	06/06/2025	14:56	LB136052
	Arsenic	5110	5000	102	90 - 110	P	06/06/2025	14:56	LB136052
	Barium	9740	10000	97	90 - 110	P	06/06/2025	14:56	LB136052
	Beryllium	235	250	94	90 - 110	P	06/06/2025	14:56	LB136052
	Cadmium	2470	2500	99	90 - 110	P	06/06/2025	14:56	LB136052
	Calcium	23900	25000	96	90 - 110	P	06/06/2025	14:56	LB136052
	Chromium	993	1000	99	90 - 110	P	06/06/2025	14:56	LB136052
	Cobalt	2480	2500	99	90 - 110	P	06/06/2025	14:56	LB136052
	Copper	1280	1250	102	90 - 110	P	06/06/2025	14:56	LB136052
	Iron	4920	5000	98	90 - 110	P	06/06/2025	14:56	LB136052
	Lead	4960	5000	99	90 - 110	P	06/06/2025	14:56	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23700	25000	95	90 - 110	P	06/06/2025	14:56	LB136052
	Manganese	2370	2500	95	90 - 110	P	06/06/2025	14:56	LB136052
	Nickel	2480	2500	99	90 - 110	P	06/06/2025	14:56	LB136052
	Potassium	25000	25000	100	90 - 110	P	06/06/2025	14:56	LB136052
	Selenium	5170	5000	103	90 - 110	P	06/06/2025	14:56	LB136052
	Silver	1250	1250	100	90 - 110	P	06/06/2025	14:56	LB136052
	Sodium	25100	25000	100	90 - 110	P	06/06/2025	14:56	LB136052
	Thallium	5070	5000	101	90 - 110	P	06/06/2025	14:56	LB136052
	Vanadium	2430	2500	97	90 - 110	P	06/06/2025	14:56	LB136052
	Zinc	2490	2500	100	90 - 110	P	06/06/2025	14:56	LB136052
CCV03	Aluminum	10200	10000	102	90 - 110	P	06/06/2025	15:50	LB136052
	Antimony	5160	5000	103	90 - 110	P	06/06/2025	15:50	LB136052
	Arsenic	5150	5000	103	90 - 110	P	06/06/2025	15:50	LB136052
	Barium	9690	10000	97	90 - 110	P	06/06/2025	15:50	LB136052
	Beryllium	253	250	101	90 - 110	P	06/06/2025	15:50	LB136052
	Cadmium	2510	2500	100	90 - 110	P	06/06/2025	15:50	LB136052
	Calcium	24500	25000	98	90 - 110	P	06/06/2025	15:50	LB136052
	Chromium	998	1000	100	90 - 110	P	06/06/2025	15:50	LB136052
	Cobalt	2510	2500	100	90 - 110	P	06/06/2025	15:50	LB136052
	Copper	1280	1250	103	90 - 110	P	06/06/2025	15:50	LB136052
	Iron	4640	5000	93	90 - 110	P	06/06/2025	15:50	LB136052
	Lead	5000	5000	100	90 - 110	P	06/06/2025	15:50	LB136052
	Magnesium	24600	25000	98	90 - 110	P	06/06/2025	15:50	LB136052
	Manganese	2420	2500	97	90 - 110	P	06/06/2025	15:50	LB136052
	Nickel	2510	2500	100	90 - 110	P	06/06/2025	15:50	LB136052
	Potassium	23600	25000	94	90 - 110	P	06/06/2025	15:50	LB136052
	Selenium	5210	5000	104	90 - 110	P	06/06/2025	15:50	LB136052
	Silver	1240	1250	99	90 - 110	P	06/06/2025	15:50	LB136052
	Sodium	23900	25000	95	90 - 110	P	06/06/2025	15:50	LB136052
	Thallium	5080	5000	102	90 - 110	P	06/06/2025	15:50	LB136052
	Vanadium	2460	2500	98	90 - 110	P	06/06/2025	15:50	LB136052
	Zinc	2470	2500	99	90 - 110	P	06/06/2025	15:50	LB136052
CCV04	Aluminum	9810	10000	98	90 - 110	P	06/06/2025	16:46	LB136052
	Antimony	4960	5000	99	90 - 110	P	06/06/2025	16:46	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	4920	5000	98	90 - 110	P	06/06/2025	16:46	LB136052
	Barium	9750	10000	98	90 - 110	P	06/06/2025	16:46	LB136052
	Beryllium	237	250	95	90 - 110	P	06/06/2025	16:46	LB136052
	Cadmium	2450	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
	Calcium	24200	25000	97	90 - 110	P	06/06/2025	16:46	LB136052
	Chromium	993	1000	99	90 - 110	P	06/06/2025	16:46	LB136052
	Cobalt	2450	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
	Copper	1240	1250	99	90 - 110	P	06/06/2025	16:46	LB136052
	Iron	4980	5000	100	90 - 110	P	06/06/2025	16:46	LB136052
	Lead	4890	5000	98	90 - 110	P	06/06/2025	16:46	LB136052
	Magnesium	23900	25000	96	90 - 110	P	06/06/2025	16:46	LB136052
	Manganese	2400	2500	96	90 - 110	P	06/06/2025	16:46	LB136052
	Nickel	2450	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
	Potassium	24600	25000	98	90 - 110	P	06/06/2025	16:46	LB136052
	Selenium	4920	5000	98	90 - 110	P	06/06/2025	16:46	LB136052
	Silver	1230	1250	98	90 - 110	P	06/06/2025	16:46	LB136052
	Sodium	25000	25000	100	90 - 110	P	06/06/2025	16:46	LB136052
	Thallium	4710	5000	94	90 - 110	P	06/06/2025	16:46	LB136052
	Vanadium	2430	2500	97	90 - 110	P	06/06/2025	16:46	LB136052
	Zinc	2460	2500	98	90 - 110	P	06/06/2025	16:46	LB136052
CCV05	Aluminum	9840	10000	98	90 - 110	P	06/06/2025	17:40	LB136052
	Antimony	4960	5000	99	90 - 110	P	06/06/2025	17:40	LB136052
	Arsenic	4930	5000	99	90 - 110	P	06/06/2025	17:40	LB136052
	Barium	9480	10000	95	90 - 110	P	06/06/2025	17:40	LB136052
	Beryllium	243	250	97	90 - 110	P	06/06/2025	17:40	LB136052
	Cadmium	2390	2500	96	90 - 110	P	06/06/2025	17:40	LB136052
	Calcium	23700	25000	95	90 - 110	P	06/06/2025	17:40	LB136052
	Chromium	977	1000	98	90 - 110	P	06/06/2025	17:40	LB136052
	Cobalt	2410	2500	96	90 - 110	P	06/06/2025	17:40	LB136052
	Copper	1250	1250	100	90 - 110	P	06/06/2025	17:40	LB136052
	Iron	4770	5000	95	90 - 110	P	06/06/2025	17:40	LB136052
	Lead	4810	5000	96	90 - 110	P	06/06/2025	17:40	LB136052
	Magnesium	23600	25000	94	90 - 110	P	06/06/2025	17:40	LB136052
	Manganese	2370	2500	95	90 - 110	P	06/06/2025	17:40	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2410	2500	96	90 - 110	P	06/06/2025	17:40	LB136052
	Potassium	23900	25000	96	90 - 110	P	06/06/2025	17:40	LB136052
	Selenium	5010	5000	100	90 - 110	P	06/06/2025	17:40	LB136052
	Silver	1230	1250	99	90 - 110	P	06/06/2025	17:40	LB136052
	Sodium	24000	25000	96	90 - 110	P	06/06/2025	17:40	LB136052
	Thallium	4840	5000	97	90 - 110	P	06/06/2025	17:40	LB136052
	Vanadium	2400	2500	96	90 - 110	P	06/06/2025	17:40	LB136052
	Zinc	2450	2500	98	90 - 110	P	06/06/2025	17:40	LB136052
CCV06	Aluminum	9230	10000	92	90 - 110	P	06/07/2025	03:03	LB136052
	Antimony	4480	5000	90	90 - 110	P	06/07/2025	03:03	LB136052
	Arsenic	4390	5000	88	90 - 110	P	06/07/2025	03:03	LB136052
	Barium	9840	10000	98	90 - 110	P	06/07/2025	03:03	LB136052
	Beryllium	257	250	103	90 - 110	P	06/07/2025	03:03	LB136052
	Cadmium	2300	2500	92	90 - 110	P	06/07/2025	03:03	LB136052
	Calcium	23900	25000	96	90 - 110	P	06/07/2025	03:03	LB136052
	Chromium	907	1000	91	90 - 110	P	06/07/2025	03:03	LB136052
	Cobalt	2290	2500	92	90 - 110	P	06/07/2025	03:03	LB136052
	Copper	1170	1250	94	90 - 110	P	06/07/2025	03:03	LB136052
	Iron	4160	5000	83	90 - 110	P	06/07/2025	03:03	LB136052
	Lead	4580	5000	92	90 - 110	P	06/07/2025	03:03	LB136052
	Magnesium	24100	25000	96	90 - 110	P	06/07/2025	03:03	LB136052
	Manganese	2470	2500	99	90 - 110	P	06/07/2025	03:03	LB136052
	Nickel	2300	2500	92	90 - 110	P	06/07/2025	03:03	LB136052
	Potassium	20100	25000	81	90 - 110	P	06/07/2025	03:03	LB136052
	Selenium	4340	5000	87	90 - 110	P	06/07/2025	03:03	LB136052
	Silver	1120	1250	90	90 - 110	P	06/07/2025	03:03	LB136052
	Sodium	20700	25000	83	90 - 110	P	06/07/2025	03:03	LB136052
	Thallium	4470	5000	89	90 - 110	P	06/07/2025	03:03	LB136052
	Vanadium	2390	2500	96	90 - 110	P	06/07/2025	03:03	LB136052
	Zinc	2190	2500	88	90 - 110	P	06/07/2025	03:03	LB136052
CCV07	Aluminum	9310	10000	93	90 - 110	P	06/07/2025	03:40	LB136052
	Antimony	4500	5000	90	90 - 110	P	06/07/2025	03:40	LB136052
	Arsenic	4400	5000	88	90 - 110	P	06/07/2025	03:40	LB136052
	Barium	10300	10000	103	90 - 110	P	06/07/2025	03:40	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	249	250	100	90 - 110	P	06/07/2025	03:40	LB136052
	Cadmium	2330	2500	93	90 - 110	P	06/07/2025	03:40	LB136052
	Calcium	24600	25000	98	90 - 110	P	06/07/2025	03:40	LB136052
	Chromium	951	1000	95	90 - 110	P	06/07/2025	03:40	LB136052
	Cobalt	2320	2500	93	90 - 110	P	06/07/2025	03:40	LB136052
	Copper	1180	1250	94	90 - 110	P	06/07/2025	03:40	LB136052
	Iron	4870	5000	98	90 - 110	P	06/07/2025	03:40	LB136052
	Lead	4630	5000	92	90 - 110	P	06/07/2025	03:40	LB136052
	Magnesium	24500	25000	98	90 - 110	P	06/07/2025	03:40	LB136052
	Manganese	2530	2500	101	90 - 110	P	06/07/2025	03:40	LB136052
	Nickel	2330	2500	93	90 - 110	P	06/07/2025	03:40	LB136052
	Potassium	22700	25000	91	90 - 110	P	06/07/2025	03:40	LB136052
	Selenium	4350	5000	87	90 - 110	P	06/07/2025	03:40	LB136052
	Silver	1180	1250	95	90 - 110	P	06/07/2025	03:40	LB136052
	Sodium	23300	25000	93	90 - 110	P	06/07/2025	03:40	LB136052
	Thallium	4510	5000	90	90 - 110	P	06/07/2025	03:40	LB136052
	Vanadium	2470	2500	99	90 - 110	P	06/07/2025	03:40	LB136052
	Zinc	2300	2500	92	90 - 110	P	06/07/2025	03:40	LB136052

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	7520	8000	94	90 - 110	P	06/10/2025	14:37	LB136097
	Antimony	4020	4000	100	90 - 110	P	06/10/2025	14:37	LB136097
	Arsenic	3850	4000	96	90 - 110	P	06/10/2025	14:37	LB136097
	Barium	7450	8000	93	90 - 110	P	06/10/2025	14:37	LB136097
	Beryllium	198	200	99	90 - 110	P	06/10/2025	14:37	LB136097
	Cadmium	1920	2000	96	90 - 110	P	06/10/2025	14:37	LB136097
	Calcium	18500	20000	92	90 - 110	P	06/10/2025	14:37	LB136097
	Chromium	794	800	99	90 - 110	P	06/10/2025	14:37	LB136097
	Cobalt	1940	2000	97	90 - 110	P	06/10/2025	14:37	LB136097
	Copper	994	1000	99	90 - 110	P	06/10/2025	14:37	LB136097
	Iron	3770	4000	94	90 - 110	P	06/10/2025	14:37	LB136097
	Lead	3800	4000	95	90 - 110	P	06/10/2025	14:37	LB136097
	Magnesium	18700	20000	93	90 - 110	P	06/10/2025	14:37	LB136097
	Manganese	1860	2000	93	90 - 110	P	06/10/2025	14:37	LB136097
	Nickel	1930	2000	97	90 - 110	P	06/10/2025	14:37	LB136097
	Potassium	21800	20000	109	90 - 110	P	06/10/2025	14:37	LB136097
	Selenium	3880	4000	97	90 - 110	P	06/10/2025	14:37	LB136097
	Silver	978	1000	98	90 - 110	P	06/10/2025	14:37	LB136097
	Sodium	21900	20000	109	90 - 110	P	06/10/2025	14:37	LB136097
	Thallium	3740	4000	94	90 - 110	P	06/10/2025	14:37	LB136097
	Vanadium	1900	2000	95	90 - 110	P	06/10/2025	14:37	LB136097
	Zinc	1980	2000	99	90 - 110	P	06/10/2025	14:37	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	110	100	110	80 - 120	P	06/10/2025	14:41	LB136097
	Antimony	55.0	50.0	110	80 - 120	P	06/10/2025	14:41	LB136097
	Arsenic	21.2	20.0	106	80 - 120	P	06/10/2025	14:41	LB136097
	Barium	101	100	101	80 - 120	P	06/10/2025	14:41	LB136097
	Beryllium	6.23	6.0	104	80 - 120	P	06/10/2025	14:41	LB136097
	Cadmium	6.31	6.0	105	80 - 120	P	06/10/2025	14:41	LB136097
	Calcium	2040	2000	102	80 - 120	P	06/10/2025	14:41	LB136097
	Chromium	10.8	10.0	108	80 - 120	P	06/10/2025	14:41	LB136097
	Cobalt	30.5	30.0	102	80 - 120	P	06/10/2025	14:41	LB136097
	Copper	22.6	20.0	113	80 - 120	P	06/10/2025	14:41	LB136097
	Iron	108	100	108	80 - 120	P	06/10/2025	14:41	LB136097
	Lead	12.8	12.0	107	80 - 120	P	06/10/2025	14:41	LB136097
	Magnesium	2030	2000	102	80 - 120	P	06/10/2025	14:41	LB136097
	Manganese	21.1	20.0	106	80 - 120	P	06/10/2025	14:41	LB136097
	Nickel	41.6	40.0	104	80 - 120	P	06/10/2025	14:41	LB136097
	Potassium	1960	2000	98	80 - 120	P	06/10/2025	14:41	LB136097
	Selenium	19.7	20.0	98	80 - 120	P	06/10/2025	14:41	LB136097
	Silver	10.5	10.0	105	80 - 120	P	06/10/2025	14:41	LB136097
	Sodium	1920	2000	96	80 - 120	P	06/10/2025	14:41	LB136097
	Thallium	41.3	40.0	103	80 - 120	P	06/10/2025	14:41	LB136097
	Vanadium	38.7	40.0	97	80 - 120	P	06/10/2025	14:41	LB136097
	Zinc	43.7	40.0	109	80 - 120	P	06/10/2025	14:41	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	10100	10000	101	90 - 110	P	06/10/2025	15:30	LB136097
	Antimony	5060	5000	101	90 - 110	P	06/10/2025	15:30	LB136097
	Arsenic	5050	5000	101	90 - 110	P	06/10/2025	15:30	LB136097
	Barium	9840	10000	98	90 - 110	P	06/10/2025	15:30	LB136097
	Beryllium	262	250	105	90 - 110	P	06/10/2025	15:30	LB136097
	Cadmium	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Calcium	24900	25000	100	90 - 110	P	06/10/2025	15:30	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	15:30	LB136097
	Cobalt	2500	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Copper	1270	1250	101	90 - 110	P	06/10/2025	15:30	LB136097
	Iron	4860	5000	97	90 - 110	P	06/10/2025	15:30	LB136097
	Lead	5030	5000	100	90 - 110	P	06/10/2025	15:30	LB136097
	Magnesium	25000	25000	100	90 - 110	P	06/10/2025	15:30	LB136097
	Manganese	2480	2500	99	90 - 110	P	06/10/2025	15:30	LB136097
	Nickel	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Potassium	24200	25000	97	90 - 110	P	06/10/2025	15:30	LB136097
	Selenium	5080	5000	102	90 - 110	P	06/10/2025	15:30	LB136097
	Silver	1270	1250	101	90 - 110	P	06/10/2025	15:30	LB136097
	Sodium	24100	25000	96	90 - 110	P	06/10/2025	15:30	LB136097
	Thallium	5100	5000	102	90 - 110	P	06/10/2025	15:30	LB136097
	Vanadium	2510	2500	100	90 - 110	P	06/10/2025	15:30	LB136097
	Zinc	2560	2500	102	90 - 110	P	06/10/2025	15:30	LB136097
CCV02	Aluminum	10000	10000	100	90 - 110	P	06/10/2025	16:28	LB136097
	Antimony	5040	5000	101	90 - 110	P	06/10/2025	16:28	LB136097
	Arsenic	5070	5000	101	90 - 110	P	06/10/2025	16:28	LB136097
	Barium	9830	10000	98	90 - 110	P	06/10/2025	16:28	LB136097
	Beryllium	261	250	104	90 - 110	P	06/10/2025	16:28	LB136097
	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	16:28	LB136097
	Calcium	24900	25000	100	90 - 110	P	06/10/2025	16:28	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	16:28	LB136097
	Cobalt	2530	2500	101	90 - 110	P	06/10/2025	16:28	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	16:28	LB136097
	Iron	4930	5000	99	90 - 110	P	06/10/2025	16:28	LB136097
	Lead	5070	5000	101	90 - 110	P	06/10/2025	16:28	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	25000	25000	100	90 - 110	P	06/10/2025	16:28	LB136097
	Manganese	2460	2500	98	90 - 110	P	06/10/2025	16:28	LB136097
	Nickel	2540	2500	101	90 - 110	P	06/10/2025	16:28	LB136097
	Potassium	24500	25000	98	90 - 110	P	06/10/2025	16:28	LB136097
	Selenium	5100	5000	102	90 - 110	P	06/10/2025	16:28	LB136097
	Silver	1270	1250	102	90 - 110	P	06/10/2025	16:28	LB136097
	Sodium	24400	25000	97	90 - 110	P	06/10/2025	16:28	LB136097
	Thallium	5070	5000	101	90 - 110	P	06/10/2025	16:28	LB136097
	Vanadium	2510	2500	100	90 - 110	P	06/10/2025	16:28	LB136097
	Zinc	2570	2500	103	90 - 110	P	06/10/2025	16:28	LB136097
CCV03	Aluminum	9930	10000	99	90 - 110	P	06/10/2025	17:36	LB136097
	Antimony	5110	5000	102	90 - 110	P	06/10/2025	17:36	LB136097
	Arsenic	5130	5000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Barium	10200	10000	102	90 - 110	P	06/10/2025	17:36	LB136097
	Beryllium	243	250	97	90 - 110	P	06/10/2025	17:36	LB136097
	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	17:36	LB136097
	Calcium	25000	25000	100	90 - 110	P	06/10/2025	17:36	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Cobalt	2530	2500	101	90 - 110	P	06/10/2025	17:36	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	17:36	LB136097
	Iron	5390	5000	108	90 - 110	P	06/10/2025	17:36	LB136097
	Lead	5080	5000	102	90 - 110	P	06/10/2025	17:36	LB136097
	Magnesium	24700	25000	99	90 - 110	P	06/10/2025	17:36	LB136097
	Manganese	2500	2500	100	90 - 110	P	06/10/2025	17:36	LB136097
	Nickel	2540	2500	102	90 - 110	P	06/10/2025	17:36	LB136097
	Potassium	26400	25000	105	90 - 110	P	06/10/2025	17:36	LB136097
	Selenium	5140	5000	103	90 - 110	P	06/10/2025	17:36	LB136097
	Silver	1280	1250	103	90 - 110	P	06/10/2025	17:36	LB136097
	Sodium	26700	25000	107	90 - 110	P	06/10/2025	17:36	LB136097
	Thallium	4990	5000	100	90 - 110	P	06/10/2025	17:36	LB136097
	Vanadium	2540	2500	101	90 - 110	P	06/10/2025	17:36	LB136097
	Zinc	2580	2500	103	90 - 110	P	06/10/2025	17:36	LB136097
CCV04	Aluminum	9910	10000	99	90 - 110	P	06/10/2025	18:28	LB136097
	Antimony	5130	5000	103	90 - 110	P	06/10/2025	18:28	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5200	5000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Barium	10300	10000	103	90 - 110	P	06/10/2025	18:28	LB136097
	Beryllium	246	250	99	90 - 110	P	06/10/2025	18:28	LB136097
	Cadmium	2580	2500	103	90 - 110	P	06/10/2025	18:28	LB136097
	Calcium	25200	25000	101	90 - 110	P	06/10/2025	18:28	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	18:28	LB136097
	Cobalt	2560	2500	102	90 - 110	P	06/10/2025	18:28	LB136097
	Copper	1290	1250	103	90 - 110	P	06/10/2025	18:28	LB136097
	Iron	5420	5000	108	90 - 110	P	06/10/2025	18:28	LB136097
	Lead	5160	5000	103	90 - 110	P	06/10/2025	18:28	LB136097
	Magnesium	24800	25000	99	90 - 110	P	06/10/2025	18:28	LB136097
	Manganese	2510	2500	100	90 - 110	P	06/10/2025	18:28	LB136097
	Nickel	2570	2500	103	90 - 110	P	06/10/2025	18:28	LB136097
	Potassium	25900	25000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Selenium	5220	5000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	18:28	LB136097
	Sodium	26000	25000	104	90 - 110	P	06/10/2025	18:28	LB136097
	Thallium	5110	5000	102	90 - 110	P	06/10/2025	18:28	LB136097
	Vanadium	2550	2500	102	90 - 110	P	06/10/2025	18:28	LB136097
	Zinc	2540	2500	101	90 - 110	P	06/10/2025	18:28	LB136097
CCV05	Aluminum	9840	10000	98	90 - 110	P	06/10/2025	19:18	LB136097
	Antimony	5120	5000	102	90 - 110	P	06/10/2025	19:18	LB136097
	Arsenic	5170	5000	103	90 - 110	P	06/10/2025	19:18	LB136097
	Barium	10200	10000	102	90 - 110	P	06/10/2025	19:18	LB136097
	Beryllium	243	250	97	90 - 110	P	06/10/2025	19:18	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Calcium	24900	25000	100	90 - 110	P	06/10/2025	19:18	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/10/2025	19:18	LB136097
	Cobalt	2530	2500	101	90 - 110	P	06/10/2025	19:18	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	19:18	LB136097
	Iron	5340	5000	107	90 - 110	P	06/10/2025	19:18	LB136097
	Lead	5090	5000	102	90 - 110	P	06/10/2025	19:18	LB136097
	Magnesium	24400	25000	98	90 - 110	P	06/10/2025	19:18	LB136097
	Manganese	2490	2500	100	90 - 110	P	06/10/2025	19:18	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2540	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Potassium	25200	25000	101	90 - 110	P	06/10/2025	19:18	LB136097
	Selenium	5190	5000	104	90 - 110	P	06/10/2025	19:18	LB136097
	Silver	1260	1250	101	90 - 110	P	06/10/2025	19:18	LB136097
	Sodium	25200	25000	101	90 - 110	P	06/10/2025	19:18	LB136097
	Thallium	5000	5000	100	90 - 110	P	06/10/2025	19:18	LB136097
	Vanadium	2540	2500	102	90 - 110	P	06/10/2025	19:18	LB136097
	Zinc	2510	2500	100	90 - 110	P	06/10/2025	19:18	LB136097
CCV06	Aluminum	10100	10000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Antimony	5120	5000	102	90 - 110	P	06/10/2025	20:09	LB136097
	Arsenic	5160	5000	103	90 - 110	P	06/10/2025	20:09	LB136097
	Barium	9960	10000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Beryllium	262	250	105	90 - 110	P	06/10/2025	20:09	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Calcium	25100	25000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	20:09	LB136097
	Cobalt	2540	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	20:09	LB136097
	Iron	5150	5000	103	90 - 110	P	06/10/2025	20:09	LB136097
	Lead	5110	5000	102	90 - 110	P	06/10/2025	20:09	LB136097
	Magnesium	25000	25000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Manganese	2500	2500	100	90 - 110	P	06/10/2025	20:09	LB136097
	Nickel	2550	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Potassium	24100	25000	96	90 - 110	P	06/10/2025	20:09	LB136097
	Selenium	5190	5000	104	90 - 110	P	06/10/2025	20:09	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	20:09	LB136097
	Sodium	23800	25000	95	90 - 110	P	06/10/2025	20:09	LB136097
	Thallium	5010	5000	100	90 - 110	P	06/10/2025	20:09	LB136097
	Vanadium	2560	2500	102	90 - 110	P	06/10/2025	20:09	LB136097
	Zinc	2570	2500	103	90 - 110	P	06/10/2025	20:09	LB136097
CCV07	Aluminum	10100	10000	101	90 - 110	P	06/10/2025	21:04	LB136097
	Antimony	5160	5000	103	90 - 110	P	06/10/2025	21:04	LB136097
	Arsenic	5230	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
	Barium	10100	10000	101	90 - 110	P	06/10/2025	21:04	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	261	250	104	90 - 110	P	06/10/2025	21:04	LB136097
	Cadmium	2600	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Calcium	25600	25000	102	90 - 110	P	06/10/2025	21:04	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	21:04	LB136097
	Cobalt	2590	2500	103	90 - 110	P	06/10/2025	21:04	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	21:04	LB136097
	Iron	5190	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
	Lead	5210	5000	104	90 - 110	P	06/10/2025	21:04	LB136097
	Magnesium	25400	25000	102	90 - 110	P	06/10/2025	21:04	LB136097
	Manganese	2550	2500	102	90 - 110	P	06/10/2025	21:04	LB136097
	Nickel	2590	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Potassium	23600	25000	94	90 - 110	P	06/10/2025	21:04	LB136097
	Selenium	5230	5000	105	90 - 110	P	06/10/2025	21:04	LB136097
	Silver	1280	1250	102	90 - 110	P	06/10/2025	21:04	LB136097
	Sodium	23500	25000	94	90 - 110	P	06/10/2025	21:04	LB136097
	Thallium	5090	5000	102	90 - 110	P	06/10/2025	21:04	LB136097
	Vanadium	2590	2500	104	90 - 110	P	06/10/2025	21:04	LB136097
	Zinc	2540	2500	102	90 - 110	P	06/10/2025	21:04	LB136097
CCV08	Aluminum	10100	10000	101	90 - 110	P	06/10/2025	22:04	LB136097
	Antimony	5200	5000	104	90 - 110	P	06/10/2025	22:04	LB136097
	Arsenic	5240	5000	105	90 - 110	P	06/10/2025	22:04	LB136097
	Barium	10100	10000	101	90 - 110	P	06/10/2025	22:04	LB136097
	Beryllium	262	250	105	90 - 110	P	06/10/2025	22:04	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
	Calcium	25400	25000	101	90 - 110	P	06/10/2025	22:04	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	22:04	LB136097
	Cobalt	2560	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
	Copper	1280	1250	103	90 - 110	P	06/10/2025	22:04	LB136097
	Iron	5140	5000	103	90 - 110	P	06/10/2025	22:04	LB136097
	Lead	5130	5000	102	90 - 110	P	06/10/2025	22:04	LB136097
	Magnesium	25300	25000	101	90 - 110	P	06/10/2025	22:04	LB136097
	Manganese	2540	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
	Nickel	2560	2500	103	90 - 110	P	06/10/2025	22:04	LB136097
	Potassium	23500	25000	94	90 - 110	P	06/10/2025	22:04	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5280	5000	106	90 - 110	P	06/10/2025	22:04	LB136097
	Silver	1280	1250	102	90 - 110	P	06/10/2025	22:04	LB136097
	Sodium	23400	25000	94	90 - 110	P	06/10/2025	22:04	LB136097
	Thallium	5010	5000	100	90 - 110	P	06/10/2025	22:04	LB136097
	Vanadium	2580	2500	103	90 - 110	P	06/10/2025	22:04	LB136097
	Zinc	2550	2500	102	90 - 110	P	06/10/2025	22:04	LB136097
CCV09	Aluminum	10000	10000	100	90 - 110	P	06/10/2025	22:56	LB136097
	Antimony	5110	5000	102	90 - 110	P	06/10/2025	22:56	LB136097
	Arsenic	5150	5000	103	90 - 110	P	06/10/2025	22:56	LB136097
	Barium	10100	10000	101	90 - 110	P	06/10/2025	22:56	LB136097
	Beryllium	260	250	104	90 - 110	P	06/10/2025	22:56	LB136097
	Cadmium	2550	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Calcium	25100	25000	101	90 - 110	P	06/10/2025	22:56	LB136097
	Chromium	1050	1000	105	90 - 110	P	06/10/2025	22:56	LB136097
	Cobalt	2550	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Copper	1270	1250	102	90 - 110	P	06/10/2025	22:56	LB136097
	Iron	5160	5000	103	90 - 110	P	06/10/2025	22:56	LB136097
	Lead	5100	5000	102	90 - 110	P	06/10/2025	22:56	LB136097
	Magnesium	25000	25000	100	90 - 110	P	06/10/2025	22:56	LB136097
	Manganese	2510	2500	100	90 - 110	P	06/10/2025	22:56	LB136097
	Nickel	2560	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Potassium	24100	25000	96	90 - 110	P	06/10/2025	22:56	LB136097
	Selenium	5180	5000	104	90 - 110	P	06/10/2025	22:56	LB136097
	Silver	1290	1250	103	90 - 110	P	06/10/2025	22:56	LB136097
	Sodium	23900	25000	96	90 - 110	P	06/10/2025	22:56	LB136097
	Thallium	5020	5000	100	90 - 110	P	06/10/2025	22:56	LB136097
	Vanadium	2540	2500	102	90 - 110	P	06/10/2025	22:56	LB136097
	Zinc	2580	2500	103	90 - 110	P	06/10/2025	22:56	LB136097
CCV10	Aluminum	10200	10000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Antimony	5170	5000	103	90 - 110	P	06/10/2025	23:57	LB136097
	Arsenic	5180	5000	104	90 - 110	P	06/10/2025	23:57	LB136097
	Barium	10000	10000	100	90 - 110	P	06/10/2025	23:57	LB136097
	Beryllium	264	250	106	90 - 110	P	06/10/2025	23:57	LB136097
	Cadmium	2540	2500	102	90 - 110	P	06/10/2025	23:57	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210

Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV10	Calcium	25400	25000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Chromium	1060	1000	106	90 - 110	P	06/10/2025	23:57	LB136097
	Cobalt	2550	2500	102	90 - 110	P	06/10/2025	23:57	LB136097
	Copper	1280	1250	102	90 - 110	P	06/10/2025	23:57	LB136097
	Iron	5200	5000	104	90 - 110	P	06/10/2025	23:57	LB136097
	Lead	5100	5000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Magnesium	25400	25000	102	90 - 110	P	06/10/2025	23:57	LB136097
	Manganese	2530	2500	101	90 - 110	P	06/10/2025	23:57	LB136097
	Nickel	2560	2500	102	90 - 110	P	06/10/2025	23:57	LB136097
	Potassium	24000	25000	96	90 - 110	P	06/10/2025	23:57	LB136097
	Selenium	5230	5000	105	90 - 110	P	06/10/2025	23:57	LB136097
	Silver	1310	1250	104	90 - 110	P	06/10/2025	23:57	LB136097
	Sodium	23600	25000	94	90 - 110	P	06/10/2025	23:57	LB136097
	Thallium	4940	5000	99	90 - 110	P	06/10/2025	23:57	LB136097
	Vanadium	2570	2500	103	90 - 110	P	06/10/2025	23:57	LB136097
	Zinc	2630	2500	105	90 - 110	P	06/10/2025	23:57	LB136097
CCV11	Aluminum	10200	10000	102	90 - 110	P	06/11/2025	01:35	LB136097
	Antimony	5240	5000	105	90 - 110	P	06/11/2025	01:35	LB136097
	Arsenic	5280	5000	106	90 - 110	P	06/11/2025	01:35	LB136097
	Barium	10100	10000	101	90 - 110	P	06/11/2025	01:35	LB136097
	Beryllium	259	250	104	90 - 110	P	06/11/2025	01:35	LB136097
	Cadmium	2570	2500	103	90 - 110	P	06/11/2025	01:35	LB136097
	Calcium	25100	25000	100	90 - 110	P	06/11/2025	01:35	LB136097
	Chromium	1040	1000	104	90 - 110	P	06/11/2025	01:35	LB136097
	Cobalt	2580	2500	103	90 - 110	P	06/11/2025	01:35	LB136097
	Copper	1290	1250	103	90 - 110	P	06/11/2025	01:35	LB136097
	Iron	5060	5000	101	90 - 110	P	06/11/2025	01:35	LB136097
	Lead	5160	5000	103	90 - 110	P	06/11/2025	01:35	LB136097
	Magnesium	24900	25000	100	90 - 110	P	06/11/2025	01:35	LB136097
	Manganese	2520	2500	101	90 - 110	P	06/11/2025	01:35	LB136097
	Nickel	2590	2500	104	90 - 110	P	06/11/2025	01:35	LB136097
	Potassium	23800	25000	95	90 - 110	P	06/11/2025	01:35	LB136097
	Selenium	5320	5000	106	90 - 110	P	06/11/2025	01:35	LB136097
	Silver	1280	1250	103	90 - 110	P	06/11/2025	01:35	LB136097

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Sodium	23700	25000	95	90 - 110	P	06/11/2025	01:35	LB136097
	Thallium	4950	5000	99	90 - 110	P	06/11/2025	01:35	LB136097
	Vanadium	2560	2500	102	90 - 110	P	06/11/2025	01:35	LB136097
	Zinc	2600	2500	104	90 - 110	P	06/11/2025	01:35	LB136097
CCV12	Aluminum	10200	10000	102	90 - 110	P	06/11/2025	02:21	LB136097
	Antimony	4840	5000	97	90 - 110	P	06/11/2025	02:21	LB136097
	Arsenic	4850	5000	97	90 - 110	P	06/11/2025	02:21	LB136097
	Barium	10000	10000	100	90 - 110	P	06/11/2025	02:21	LB136097
	Beryllium	264	250	105	90 - 110	P	06/11/2025	02:21	LB136097
	Cadmium	2400	2500	96	90 - 110	P	06/11/2025	02:21	LB136097
	Calcium	24900	25000	99	90 - 110	P	06/11/2025	02:21	LB136097
	Chromium	1030	1000	103	90 - 110	P	06/11/2025	02:21	LB136097
	Cobalt	2420	2500	97	90 - 110	P	06/11/2025	02:21	LB136097
	Copper	1200	1250	96	90 - 110	P	06/11/2025	02:21	LB136097
	Iron	4930	5000	99	90 - 110	P	06/11/2025	02:21	LB136097
	Lead	4810	5000	96	90 - 110	P	06/11/2025	02:21	LB136097
	Magnesium	24900	25000	100	90 - 110	P	06/11/2025	02:21	LB136097
	Manganese	2490	2500	100	90 - 110	P	06/11/2025	02:21	LB136097
	Nickel	2420	2500	97	90 - 110	P	06/11/2025	02:21	LB136097
	Potassium	23100	25000	92	90 - 110	P	06/11/2025	02:21	LB136097
	Selenium	4880	5000	98	90 - 110	P	06/11/2025	02:21	LB136097
	Silver	1280	1250	102	90 - 110	P	06/11/2025	02:21	LB136097
	Sodium	22800	25000	91	90 - 110	P	06/11/2025	02:21	LB136097
	Thallium	4660	5000	93	90 - 110	P	06/11/2025	02:21	LB136097
	Vanadium	2530	2500	101	90 - 110	P	06/11/2025	02:21	LB136097
	Zinc	2590	2500	104	90 - 110	P	06/11/2025	02:21	LB136097