



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental SDG No.: Q2210
Contract: GENV01 Lab Code: CHEM Case No.: Q2210 SAS No.: Q2210
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	107	70 - 130	CV	06/06/2025	10:12	LB136036
CRI01	Aluminum	104	100	104	65 - 135	P	06/06/2025	13:37	LB136052
	Antimony	48.7	50.0	97	65 - 135	P	06/06/2025	13:37	LB136052
	Arsenic	21.3	20.0	107	65 - 135	P	06/06/2025	13:37	LB136052
	Barium	91.8	100	92	65 - 135	P	06/06/2025	13:37	LB136052
	Beryllium	5.67	6.0	94	65 - 135	P	06/06/2025	13:37	LB136052
	Cadmium	5.92	6.0	99	65 - 135	P	06/06/2025	13:37	LB136052
	Calcium	1900	2000	95	65 - 135	P	06/06/2025	13:37	LB136052
	Chromium	9.75	10.0	98	65 - 135	P	06/06/2025	13:37	LB136052
	Cobalt	29.5	30.0	98	65 - 135	P	06/06/2025	13:37	LB136052
	Copper	20.8	20.0	104	65 - 135	P	06/06/2025	13:37	LB136052
	Iron	112	100	112	65 - 135	P	06/06/2025	13:37	LB136052
	Lead	13.8	12.0	115	65 - 135	P	06/06/2025	13:37	LB136052
	Magnesium	1930	2000	96	65 - 135	P	06/06/2025	13:37	LB136052
	Manganese	18.7	20.0	93	65 - 135	P	06/06/2025	13:37	LB136052
	Nickel	39.5	40.0	99	65 - 135	P	06/06/2025	13:37	LB136052
	Potassium	2490	2000	124	65 - 135	P	06/06/2025	13:37	LB136052
	Selenium	19.7	20.0	98	65 - 135	P	06/06/2025	13:37	LB136052
	Silver	10.1	10.0	101	65 - 135	P	06/06/2025	13:37	LB136052
	Sodium	1800	2000	90	65 - 135	P	06/06/2025	13:37	LB136052
	Thallium	39.1	40.0	98	65 - 135	P	06/06/2025	13:37	LB136052
	Vanadium	36.5	40.0	91	65 - 135	P	06/06/2025	13:37	LB136052
	Zinc	45.4	40.0	113	65 - 135	P	06/06/2025	13:37	LB136052
CRI01	Aluminum	105	100	105	65 - 135	P	06/10/2025	14:50	LB136097
	Antimony	53.3	50.0	107	65 - 135	P	06/10/2025	14:50	LB136097
	Arsenic	22.9	20.0	114	65 - 135	P	06/10/2025	14:50	LB136097
	Barium	98.7	100	99	65 - 135	P	06/10/2025	14:50	LB136097
	Beryllium	6.22	6.0	104	65 - 135	P	06/10/2025	14:50	LB136097
	Cadmium	6.07	6.0	101	65 - 135	P	06/10/2025	14:50	LB136097
	Calcium	2030	2000	102	65 - 135	P	06/10/2025	14:50	LB136097
	Chromium	10.7	10.0	107	65 - 135	P	06/10/2025	14:50	LB136097
	Cobalt	30.4	30.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Copper	22.7	20.0	114	65 - 135	P	06/10/2025	14:50	LB136097
	Iron	111	100	111	65 - 135	P	06/10/2025	14:50	LB136097
	Lead	13.4	12.0	112	65 - 135	P	06/10/2025	14:50	LB136097

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Magnesium	2040	2000	102	65 - 135	P	06/10/2025	14:50	LB136097
	Manganese	21.0	20.0	105	65 - 135	P	06/10/2025	14:50	LB136097
	Nickel	40.8	40.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Potassium	1990	2000	99	65 - 135	P	06/10/2025	14:50	LB136097
	Selenium	19.5	20.0	98	65 - 135	P	06/10/2025	14:50	LB136097
	Silver	10.6	10.0	106	65 - 135	P	06/10/2025	14:50	LB136097
	Sodium	1890	2000	95	65 - 135	P	06/10/2025	14:50	LB136097
	Thallium	40.8	40.0	102	65 - 135	P	06/10/2025	14:50	LB136097
	Vanadium	39.1	40.0	98	65 - 135	P	06/10/2025	14:50	LB136097
	Zinc	44.1	40.0	110	65 - 135	P	06/10/2025	14:50	LB136097