



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	G Environmental				SDG No.:	Q2210				
Contract:	GENV01	Lab Code:	CHEM		Case No.:	Q2210	SAS No.:	Q2210		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB29	Mercury	0.076	+/-0.2	U	0.20	CV	06/06/2025	10:05	LB136036	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q2210

Contract: GENV01

Lab Code: CHEM

Case No.: Q2210

SAS No.: Q2210

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Mercury	0.076	+/-0.2	U	0.20	CV	06/06/2025	10:10	LB136036
CCB09	Mercury	0.076	+/-0.2	U	0.20	CV	06/06/2025	10:40	LB136036
CCB10	Mercury	0.076	+/-0.2	U	0.20	CV	06/06/2025	11:07	LB136036
CCB11	Mercury	0.076	+/-0.2	U	0.20	CV	06/06/2025	11:43	LB136036
CCB12	Mercury	0.076	+/-0.2	U	0.20	CV	06/06/2025	12:08	LB136036
CCB13	Mercury	0.076	+/-0.2	U	0.20	CV	06/06/2025	12:35	LB136036

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2210</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2210</u>	SAS No.: <u>Q2210</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	06/06/2025	13:33	LB136052
	Antimony	6.76	+/-25	U	50.0	P	06/06/2025	13:33	LB136052
	Arsenic	5.12	+/-10	U	20.0	P	06/06/2025	13:33	LB136052
	Barium	14.6	+/-50	U	100	P	06/06/2025	13:33	LB136052
	Beryllium	0.56	+/-3	U	6.00	P	06/06/2025	13:33	LB136052
	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	13:33	LB136052
	Calcium	234	+/-1000	U	2000	P	06/06/2025	13:33	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	13:33	LB136052
	Cobalt	2.26	+/-15	U	30.0	P	06/06/2025	13:33	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	13:33	LB136052
	Iron	23.4	+/-50	U	100	P	06/06/2025	13:33	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	13:33	LB136052
	Magnesium	244	+/-1000	U	2000	P	06/06/2025	13:33	LB136052
	Manganese	5.94	+/-10	U	20.0	P	06/06/2025	13:33	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	13:33	LB136052
	Potassium	918	+/-1000	U	2000	P	06/06/2025	13:33	LB136052
	Selenium	9.64	+/-10	U	20.0	P	06/06/2025	13:33	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	13:33	LB136052
	Sodium	868	+/-1000	U	2000	P	06/06/2025	13:33	LB136052
	Thallium	4.38	+/-20	U	40.0	P	06/06/2025	13:33	LB136052
	Vanadium	6.26	+/-20	U	40.0	P	06/06/2025	13:33	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	13:33	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2210</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2210</u>	SAS No.: <u>Q2210</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	06/06/2025	14:09	LB136052
	Antimony	6.76	+/-25	U	50.0	P	06/06/2025	14:09	LB136052
	Arsenic	5.12	+/-10	U	20.0	P	06/06/2025	14:09	LB136052
	Barium	14.6	+/-50	U	100	P	06/06/2025	14:09	LB136052
	Beryllium	0.56	+/-3	U	6.00	P	06/06/2025	14:09	LB136052
	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	14:09	LB136052
	Calcium	234	+/-1000	U	2000	P	06/06/2025	14:09	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	14:09	LB136052
	Cobalt	2.26	+/-15	U	30.0	P	06/06/2025	14:09	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	14:09	LB136052
	Iron	23.4	+/-50	U	100	P	06/06/2025	14:09	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	14:09	LB136052
	Magnesium	244	+/-1000	U	2000	P	06/06/2025	14:09	LB136052
	Manganese	5.94	+/-10	U	20.0	P	06/06/2025	14:09	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	14:09	LB136052
	Potassium	918	+/-1000	U	2000	P	06/06/2025	14:09	LB136052
	Selenium	9.64	+/-10	U	20.0	P	06/06/2025	14:09	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	14:09	LB136052
	Sodium	868	+/-1000	U	2000	P	06/06/2025	14:09	LB136052
	Thallium	4.38	+/-20	U	40.0	P	06/06/2025	14:09	LB136052
	Vanadium	6.26	+/-20	U	40.0	P	06/06/2025	14:09	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	14:09	LB136052
CCB02	Aluminum	11.3	+/-50	U	100	P	06/06/2025	15:00	LB136052
	Antimony	6.76	+/-25	U	50.0	P	06/06/2025	15:00	LB136052
	Arsenic	5.12	+/-10	U	20.0	P	06/06/2025	15:00	LB136052
	Barium	14.6	+/-50	U	100	P	06/06/2025	15:00	LB136052
	Beryllium	0.56	+/-3	U	6.00	P	06/06/2025	15:00	LB136052
	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	15:00	LB136052
	Calcium	234	+/-1000	U	2000	P	06/06/2025	15:00	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	15:00	LB136052
	Cobalt	2.26	+/-15	U	30.0	P	06/06/2025	15:00	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	15:00	LB136052
	Iron	23.4	+/-50	U	100	P	06/06/2025	15:00	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	15:00	LB136052
	Magnesium	244	+/-1000	U	2000	P	06/06/2025	15:00	LB136052
	Manganese	5.94	+/-10	U	20.0	P	06/06/2025	15:00	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	15:00	LB136052
	Potassium	918	+/-1000	U	2000	P	06/06/2025	15:00	LB136052
	Selenium	9.64	+/-10	U	20.0	P	06/06/2025	15:00	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2210</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2210</u>	SAS No.: <u>Q2210</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	06/06/2025	15:00	LB136052
	Sodium	868	+/-1000	U	2000	P	06/06/2025	15:00	LB136052
	Thallium	4.38	+/-20	U	40.0	P	06/06/2025	15:00	LB136052
	Vanadium	6.26	+/-20	U	40.0	P	06/06/2025	15:00	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	15:00	LB136052
CCB03	Aluminum	11.3	+/-50	U	100	P	06/06/2025	15:57	LB136052
	Antimony	6.76	+/-25	U	50.0	P	06/06/2025	15:57	LB136052
	Arsenic	5.12	+/-10	U	20.0	P	06/06/2025	15:57	LB136052
	Barium	14.6	+/-50	U	100	P	06/06/2025	15:57	LB136052
	Beryllium	0.56	+/-3	U	6.00	P	06/06/2025	15:57	LB136052
	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	15:57	LB136052
	Calcium	234	+/-1000	U	2000	P	06/06/2025	15:57	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	15:57	LB136052
	Cobalt	2.26	+/-15	U	30.0	P	06/06/2025	15:57	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	15:57	LB136052
	Iron	23.4	+/-50	U	100	P	06/06/2025	15:57	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	15:57	LB136052
	Magnesium	244	+/-1000	U	2000	P	06/06/2025	15:57	LB136052
	Manganese	5.94	+/-10	U	20.0	P	06/06/2025	15:57	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	15:57	LB136052
	Potassium	918	+/-1000	U	2000	P	06/06/2025	15:57	LB136052
	Selenium	9.64	+/-10	U	20.0	P	06/06/2025	15:57	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	15:57	LB136052
	Sodium	868	+/-1000	U	2000	P	06/06/2025	15:57	LB136052
	Thallium	4.38	+/-20	U	40.0	P	06/06/2025	15:57	LB136052
	Vanadium	6.26	+/-20	U	40.0	P	06/06/2025	15:57	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	15:57	LB136052
CCB04	Aluminum	11.3	+/-50	U	100	P	06/06/2025	16:51	LB136052
	Antimony	6.76	+/-25	U	50.0	P	06/06/2025	16:51	LB136052
	Arsenic	5.12	+/-10	U	20.0	P	06/06/2025	16:51	LB136052
	Barium	14.6	+/-50	U	100	P	06/06/2025	16:51	LB136052
	Beryllium	0.56	+/-3	U	6.00	P	06/06/2025	16:51	LB136052
	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	16:51	LB136052
	Calcium	234	+/-1000	U	2000	P	06/06/2025	16:51	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	16:51	LB136052
	Cobalt	2.26	+/-15	U	30.0	P	06/06/2025	16:51	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	16:51	LB136052
	Iron	23.4	+/-50	U	100	P	06/06/2025	16:51	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	16:51	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2210</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2210</u>	SAS No.: <u>Q2210</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	06/06/2025	16:51	LB136052
	Manganese	5.94	+/-10	U	20.0	P	06/06/2025	16:51	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	16:51	LB136052
	Potassium	918	+/-1000	U	2000	P	06/06/2025	16:51	LB136052
	Selenium	9.64	+/-10	U	20.0	P	06/06/2025	16:51	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	16:51	LB136052
	Sodium	868	+/-1000	U	2000	P	06/06/2025	16:51	LB136052
	Thallium	4.38	+/-20	U	40.0	P	06/06/2025	16:51	LB136052
	Vanadium	6.26	+/-20	U	40.0	P	06/06/2025	16:51	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	16:51	LB136052
CCB05	Aluminum	11.3	+/-50	U	100	P	06/06/2025	17:46	LB136052
	Antimony	6.76	+/-25	U	50.0	P	06/06/2025	17:46	LB136052
	Arsenic	5.12	+/-10	U	20.0	P	06/06/2025	17:46	LB136052
	Barium	14.6	+/-50	U	100	P	06/06/2025	17:46	LB136052
	Beryllium	0.56	+/-3	U	6.00	P	06/06/2025	17:46	LB136052
	Cadmium	0.50	+/-3	U	6.00	P	06/06/2025	17:46	LB136052
	Calcium	234	+/-1000	U	2000	P	06/06/2025	17:46	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/06/2025	17:46	LB136052
	Cobalt	2.26	+/-15	U	30.0	P	06/06/2025	17:46	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/06/2025	17:46	LB136052
	Iron	23.4	+/-50	U	100	P	06/06/2025	17:46	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/06/2025	17:46	LB136052
	Magnesium	244	+/-1000	U	2000	P	06/06/2025	17:46	LB136052
	Manganese	5.94	+/-10	U	20.0	P	06/06/2025	17:46	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/06/2025	17:46	LB136052
	Potassium	918	+/-1000	U	2000	P	06/06/2025	17:46	LB136052
	Selenium	9.64	+/-10	U	20.0	P	06/06/2025	17:46	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/06/2025	17:46	LB136052
	Sodium	868	+/-1000	U	2000	P	06/06/2025	17:46	LB136052
	Thallium	4.38	+/-20	U	40.0	P	06/06/2025	17:46	LB136052
	Vanadium	6.26	+/-20	U	40.0	P	06/06/2025	17:46	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/06/2025	17:46	LB136052
CCB06	Aluminum	11.3	+/-50	U	100	P	06/07/2025	03:07	LB136052
	Antimony	6.76	+/-25	U	50.0	P	06/07/2025	03:07	LB136052
	Arsenic	5.12	+/-10	U	20.0	P	06/07/2025	03:07	LB136052
	Barium	14.6	+/-50	U	100	P	06/07/2025	03:07	LB136052
	Beryllium	0.56	+/-3	U	6.00	P	06/07/2025	03:07	LB136052
	Cadmium	0.50	+/-3	U	6.00	P	06/07/2025	03:07	LB136052
	Calcium	234	+/-1000	U	2000	P	06/07/2025	03:07	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q2210							
Contract: GENV01	Lab Code: CHEM	Case No.: Q2210	SAS No.: Q2210						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	06/07/2025	03:07	LB136052
	Cobalt	2.26	+/-15	U	30.0	P	06/07/2025	03:07	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/07/2025	03:07	LB136052
	Iron	23.4	+/-50	U	100	P	06/07/2025	03:07	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/07/2025	03:07	LB136052
	Magnesium	244	+/-1000	U	2000	P	06/07/2025	03:07	LB136052
	Manganese	5.94	+/-10	U	20.0	P	06/07/2025	03:07	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/07/2025	03:07	LB136052
	Potassium	918	+/-1000	U	2000	P	06/07/2025	03:07	LB136052
	Selenium	9.64	+/-10	U	20.0	P	06/07/2025	03:07	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/07/2025	03:07	LB136052
	Sodium	868	+/-1000	U	2000	P	06/07/2025	03:07	LB136052
	Thallium	5.29	+/-20	J	40.0	P	06/07/2025	03:07	LB136052
	Vanadium	6.26	+/-20	U	40.0	P	06/07/2025	03:07	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/07/2025	03:07	LB136052
CCB07	Aluminum	13.6	+/-50	J	100	P	06/07/2025	03:44	LB136052
	Antimony	6.76	+/-25	U	50.0	P	06/07/2025	03:44	LB136052
	Arsenic	5.12	+/-10	U	20.0	P	06/07/2025	03:44	LB136052
	Barium	14.6	+/-50	U	100	P	06/07/2025	03:44	LB136052
	Beryllium	0.56	+/-3	U	6.00	P	06/07/2025	03:44	LB136052
	Cadmium	0.50	+/-3	U	6.00	P	06/07/2025	03:44	LB136052
	Calcium	234	+/-1000	U	2000	P	06/07/2025	03:44	LB136052
	Chromium	2.12	+/-5	U	10.0	P	06/07/2025	03:44	LB136052
	Cobalt	2.26	+/-15	U	30.0	P	06/07/2025	03:44	LB136052
	Copper	4.60	+/-10	U	20.0	P	06/07/2025	03:44	LB136052
	Iron	23.4	+/-50	U	100	P	06/07/2025	03:44	LB136052
	Lead	2.30	+/-6	U	12.0	P	06/07/2025	03:44	LB136052
	Magnesium	244	+/-1000	U	2000	P	06/07/2025	03:44	LB136052
	Manganese	5.94	+/-10	U	20.0	P	06/07/2025	03:44	LB136052
	Nickel	3.06	+/-20	U	40.0	P	06/07/2025	03:44	LB136052
	Potassium	918	+/-1000	U	2000	P	06/07/2025	03:44	LB136052
	Selenium	9.64	+/-10	U	20.0	P	06/07/2025	03:44	LB136052
	Silver	1.62	+/-5	U	10.0	P	06/07/2025	03:44	LB136052
	Sodium	868	+/-1000	U	2000	P	06/07/2025	03:44	LB136052
	Thallium	6.19	+/-20	J	40.0	P	06/07/2025	03:44	LB136052
	Vanadium	6.26	+/-20	U	40.0	P	06/07/2025	03:44	LB136052
	Zinc	16.7	+/-20	U	40.0	P	06/07/2025	03:44	LB136052

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2210</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2210</u>	SAS No.: <u>Q2210</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	06/10/2025	14:45	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	14:45	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	14:45	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	14:45	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	14:45	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	14:45	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	14:45	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	14:45	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	14:45	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	14:45	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	14:45	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	14:45	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	14:45	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	14:45	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	14:45	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	14:45	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	14:45	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	14:45	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	14:45	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2210</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2210</u>	SAS No.: <u>Q2210</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	06/10/2025	15:35	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	15:35	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	15:35	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	15:35	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	15:35	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	15:35	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	15:35	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	15:35	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	15:35	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	15:35	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	15:35	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	15:35	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	15:35	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	15:35	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	15:35	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	15:35	LB136097
CCB02	Aluminum	11.3	+/-50	U	100	P	06/10/2025	16:32	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	16:32	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	16:32	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	16:32	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	16:32	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	16:32	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	16:32	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	16:32	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	16:32	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	16:32	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	16:32	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	16:32	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	16:32	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	16:32	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	16:32	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	16:32	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q2210							
Contract: GENV01	Lab Code: CHEM	Case No.: Q2210	SAS No.: Q2210						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	06/10/2025	16:32	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	16:32	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	16:32	LB136097
CCB03	Aluminum	11.3	+/-50	U	100	P	06/10/2025	17:40	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	17:40	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	17:40	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	17:40	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	17:40	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	17:40	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	17:40	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	17:40	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	17:40	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	17:40	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	17:40	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	17:40	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	17:40	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	17:40	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	17:40	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	17:40	LB136097
CCB04	Aluminum	11.3	+/-50	U	100	P	06/10/2025	18:32	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	18:32	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	18:32	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	18:32	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	18:32	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	18:32	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	18:32	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	18:32	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	18:32	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	18:32	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q2210							
Contract: GENV01	Lab Code: CHEM	Case No.: Q2210	SAS No.: Q2210						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	06/10/2025	18:32	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	18:32	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	18:32	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	18:32	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	18:32	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	18:32	LB136097
CCB05	Aluminum	11.3	+/-50	U	100	P	06/10/2025	19:22	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	19:22	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	19:22	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	19:22	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	19:22	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	19:22	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	19:22	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	19:22	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	19:22	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	19:22	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	19:22	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	19:22	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	19:22	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	19:22	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	19:22	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	19:22	LB136097
CCB06	Aluminum	11.3	+/-50	U	100	P	06/10/2025	20:14	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	20:14	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	20:14	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	20:14	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	20:14	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	20:14	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q2210							
Contract: GENV01	Lab Code: CHEM	Case No.: Q2210	SAS No.: Q2210						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	20:14	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	20:14	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	20:14	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	20:14	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	20:14	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	20:14	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	20:14	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	20:14	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	20:14	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	20:14	LB136097
CCB07	Aluminum	11.3	+/-50	U	100	P	06/10/2025	21:08	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	21:08	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	21:08	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	21:08	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	21:08	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	21:08	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	21:08	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	21:08	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	21:08	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	21:08	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	21:08	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	21:08	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	21:08	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	21:08	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	21:08	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	21:08	LB136097
CCB08	Aluminum	11.3	+/-50	U	100	P	06/10/2025	22:08	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	22:08	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2210</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2210</u>	SAS No.: <u>Q2210</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	22:08	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	22:08	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	22:08	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	22:08	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	22:08	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	22:08	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	22:08	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	22:08	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	22:08	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	22:08	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	22:08	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	22:08	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	22:08	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	22:08	LB136097
CCB09	Aluminum	11.3	+/-50	U	100	P	06/10/2025	23:01	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/10/2025	23:01	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Barium	14.6	+/-50	U	100	P	06/10/2025	23:01	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/10/2025	23:01	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/10/2025	23:01	LB136097
	Calcium	234	+/-1000	U	2000	P	06/10/2025	23:01	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/10/2025	23:01	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/10/2025	23:01	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Iron	23.4	+/-50	U	100	P	06/10/2025	23:01	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/10/2025	23:01	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/10/2025	23:01	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
	Potassium	918	+/-1000	U	2000	P	06/10/2025	23:01	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/10/2025	23:01	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/10/2025	23:01	LB136097
	Sodium	868	+/-1000	U	2000	P	06/10/2025	23:01	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2210</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2210</u>	SAS No.: <u>Q2210</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	4.38	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/10/2025	23:01	LB136097
CCB10	Aluminum	11.3	+/-50	U	100	P	06/11/2025	00:02	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/11/2025	00:02	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
	Barium	14.6	+/-50	U	100	P	06/11/2025	00:02	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/11/2025	00:02	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	00:02	LB136097
	Calcium	234	+/-1000	U	2000	P	06/11/2025	00:02	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/11/2025	00:02	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/11/2025	00:02	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
	Iron	23.4	+/-50	U	100	P	06/11/2025	00:02	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	00:02	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/11/2025	00:02	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Potassium	918	+/-1000	U	2000	P	06/11/2025	00:02	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/11/2025	00:02	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/11/2025	00:02	LB136097
	Sodium	868	+/-1000	U	2000	P	06/11/2025	00:02	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	00:02	LB136097
	Aluminum	11.3	+/-50	U	100	P	06/11/2025	01:39	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/11/2025	01:39	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	01:39	LB136097
	Barium	14.6	+/-50	U	100	P	06/11/2025	01:39	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/11/2025	01:39	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	01:39	LB136097
	Calcium	234	+/-1000	U	2000	P	06/11/2025	01:39	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/11/2025	01:39	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/11/2025	01:39	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	01:39	LB136097
	Iron	23.4	+/-50	U	100	P	06/11/2025	01:39	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	01:39	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/11/2025	01:39	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/11/2025	01:39	LB136097

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q2210</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q2210</u>	SAS No.: <u>Q2210</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB11	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
	Potassium	918	+/-1000	U	2000	P	06/11/2025	01:39	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/11/2025	01:39	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/11/2025	01:39	LB136097
	Sodium	868	+/-1000	U	2000	P	06/11/2025	01:39	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	01:39	LB136097
CCB12	Aluminum	11.3	+/-50	U	100	P	06/11/2025	02:25	LB136097
	Antimony	6.76	+/-25	U	50.0	P	06/11/2025	02:25	LB136097
	Arsenic	5.12	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Barium	14.6	+/-50	U	100	P	06/11/2025	02:25	LB136097
	Beryllium	0.56	+/-3	U	6.00	P	06/11/2025	02:25	LB136097
	Cadmium	0.50	+/-3	U	6.00	P	06/11/2025	02:25	LB136097
	Calcium	234	+/-1000	U	2000	P	06/11/2025	02:25	LB136097
	Chromium	2.12	+/-5	U	10.0	P	06/11/2025	02:25	LB136097
	Cobalt	2.26	+/-15	U	30.0	P	06/11/2025	02:25	LB136097
	Copper	4.60	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Iron	23.4	+/-50	U	100	P	06/11/2025	02:25	LB136097
	Lead	2.30	+/-6	U	12.0	P	06/11/2025	02:25	LB136097
	Magnesium	244	+/-1000	U	2000	P	06/11/2025	02:25	LB136097
	Manganese	5.94	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Nickel	3.06	+/-20	U	40.0	P	06/11/2025	02:25	LB136097
	Potassium	918	+/-1000	U	2000	P	06/11/2025	02:25	LB136097
	Selenium	9.64	+/-10	U	20.0	P	06/11/2025	02:25	LB136097
	Silver	1.62	+/-5	U	10.0	P	06/11/2025	02:25	LB136097
	Sodium	868	+/-1000	U	2000	P	06/11/2025	02:25	LB136097
	Thallium	4.38	+/-20	U	40.0	P	06/11/2025	02:25	LB136097
	Vanadium	6.26	+/-20	U	40.0	P	06/11/2025	02:25	LB136097
	Zinc	16.7	+/-20	U	40.0	P	06/11/2025	02:25	LB136097