

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental **SDG No.:** Q2210
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q2210 **SAS No.:** Q2210

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168284BS							
Aluminum	ug/L	1000	1010		101	80 - 120	P
Antimony	ug/L	400	418		104	80 - 120	P
Arsenic	ug/L	400	402		100	80 - 120	P
Barium	ug/L	100	94.5		94	80 - 120	P
Beryllium	ug/L	100	102		102	80 - 120	P
Cadmium	ug/L	100	100		100	80 - 120	P
Calcium	ug/L	500	505	J	101	80 - 120	P
Chromium	ug/L	200	204		102	80 - 120	P
Cobalt	ug/L	100	102		102	80 - 120	P
Copper	ug/L	150	157		105	80 - 120	P
Iron	ug/L	1500	1460		97	80 - 120	P
Lead	ug/L	500	490		98	80 - 120	P
Magnesium	ug/L	1000	978	J	98	80 - 120	P
Manganese	ug/L	100	99.5		100	80 - 120	P
Nickel	ug/L	250	253		101	80 - 120	P
Potassium	ug/L	5000	4670		93	80 - 120	P
Selenium	ug/L	1000	1020		102	80 - 120	P
Silver	ug/L	37.5	36.6		98	80 - 120	P
Sodium	ug/L	1500	1360		91	80 - 120	P
Thallium	ug/L	1000	969		97	80 - 120	P
Vanadium	ug/L	150	148		99	80 - 120	P
Zinc	ug/L	100	103		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q2210

Contract: GENV01

Lab Code: CHEM

Case No.: Q2210

SAS No.: Q2210

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168317BS Mercury	ug/L	4.0	3.53		88	80 - 120	CV