

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2210

Client: G Environmental

Analytical Method: 8270-Modified DataFile: BN037201.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168336BS	Benzo(a)anthracene	0.4	0.36	ug/L	90				70 (54)	130 (130)	
	Benzo(b)fluoranthene	0.4	0.35	ug/L	88				70 (65)	130 (121)	
	Benzo(k)fluoranthene	0.4	0.36	ug/L	90				70 (72)	130 (119)	
	Benzo(a)pyrene	0.4	0.39	ug/L	98				70 (68)	130 (120)	
	Benzo(g,h,i)perylene	0.4	0.42	ug/L	105				70 (76)	130 (117)	