

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: GENV01

Lab Code: CHEM

SAS No.: Q2210

SDG NO.: Q2210

Lab File ID: BN037203.D

DFTPP Injection Date: 06/09/2025

Instrument ID: BNA_N

DFTPP Injection Time: 22:32

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	79.8
68	Less than 2.0% of mass 69	1 (1.5) 1
69	Mass 69 relative abundance	65.5
70	Less than 2.0% of mass 69	0.4 (0.6) 1
127	10.0 - 80.0% of mass 198	56.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	24.4
365	Greater than 1% of mass 198	4.3
441	Present, but less than mass 443	8.8
442	Greater than 50% of mass 198	52.2
443	15.0 - 24.0% of mass 442	10.6 (20.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN037204.D	06/09/2025	23:11
TW1	Q2210-01	BN037212.D	06/10/2025	04:01