

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2229

Project: Monthly 2025

Run No.: LB136028

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136028BS							
BOD5	mg/L	198	210		106	1	84.6-115.4	06/05/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2229

Project: Monthly 2025

Run No.: LB136054

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136054BS							
TSS	mg/L	550	533		97	1	90-110	06/09/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2229

Project: Monthly 2025

Run No.: LB136055

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136055BS							
TPH	mg/L	20.0	16.9		84	1	78-114	06/09/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q2229

Project: Monthly 2025

Run No.: LB136055

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB136055BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	06/09/2025