

Surrogate Summary

SW-846

SDG No.: Q2230

Client: CDM Smith

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168323BL	PB168323BL	Nitrobenzene-d5	100	85.2	85		67	132
		2-Fluorobiphenyl	100	84.3	84		52	132
		Terphenyl-d14	100	90.1	90		42	152
PB168323BS	PB168323BS	Nitrobenzene-d5	100	79.4	79		67	132
		2-Fluorobiphenyl	100	77.9	78		52	132
		Terphenyl-d14	100	79.2	79		42	152
Q2230-01	FB-060425	Nitrobenzene-d5	100	91.2	91		67	132
		2-Fluorobiphenyl	100	85.1	85		52	132
		Terphenyl-d14	100	86.2	86		42	152
Q2230-02	GW-MW01-060425	Nitrobenzene-d5	100	93.0	93		67	132
		2-Fluorobiphenyl	100	85.9	86		52	132
		Terphenyl-d14	100	78.5	79		42	152
Q2230-03MS	GW-MW01-060425MS	Nitrobenzene-d5	100	90.5	91		67	132
		2-Fluorobiphenyl	100	85.6	86		52	132
		Terphenyl-d14	100	80.4	80		42	152
Q2230-04MSD	GW-MW01-060425MSD	Nitrobenzene-d5	100	88.0	88		67	132
		2-Fluorobiphenyl	100	85.9	86		52	132
		Terphenyl-d14	100	89.3	89		42	152
Q2230-05	GW-MW901-060425	Nitrobenzene-d5	100	85.9	86		67	132
		2-Fluorobiphenyl	100	81.4	81		52	132
		Terphenyl-d14	100	82.1	82		42	152