

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2230

Analytical Method: SW8270E

Client: CDM Smith

DataFile: BP024879.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2230-03MS		Client Sample ID: GW-MW01-060425MS									
Naphthalene	52.1	2.20	42.7	ug/L	78				17	157	
Acenaphthylene	52.1	0	50.7	ug/L	97				40	141	
Acenaphthene	52.1	0	52.1	ug/L	100				37	146	
Fluorene	52.1	0	53.8	ug/L	103				39	144	
Phenanthrene	52.1	0	53.7	ug/L	103				40	147	
Anthracene	52.1	0	53.8	ug/L	103				41	146	
Fluoranthene	52.1	0	57.7	ug/L	111				42	146	
Pyrene	52.1	0	51.4	ug/L	99				41	149	
Benzo(a)anthracene	52.1	0	54.6	ug/L	105				41	147	
Chrysene	52.1	0	53.4	ug/L	102				44	144	
Benzo(b)fluoranthene	52.1	0	56.2	ug/L	108				40	150	
Benzo(k)fluoranthene	52.1	0	53.1	ug/L	102				40	147	
Benzo(a)pyrene	52.1	0	54.1	ug/L	104				42	147	
Indeno(1,2,3-cd)pyrene	52.1	0	56.3	ug/L	108				30	166	
Dibenz(a,h)anthracene	52.1	0	57.5	ug/L	110				23	172	
Benzo(g,h,i)perylene	52.1	0	56.4	ug/L	108				27	167	