

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2230

Analytical Method: SW8270E

Client: CDM Smith

DataFile: BP024880.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID: Q2230-04MSD		Client Sample ID: GW-MW01-060425MSD									
Naphthalene	53.2	2.20	42.1	ug/L	75		4		17	157	20
Acenaphthylene	53.2	0	51.1	ug/L	96		1		40	141	20
Acenaphthene	53.2	0	52.7	ug/L	99		1		37	146	20
Fluorene	53.2	0	53.9	ug/L	101		2		39	144	20
Phenanthrene	53.2	0	53.6	ug/L	101		2		40	147	20
Anthracene	53.2	0	53.0	ug/L	100		3		41	146	20
Fluoranthene	53.2	0	54.9	ug/L	103		7		42	146	20
Pyrene	53.2	0	56.8	ug/L	107		8		41	149	20
Benzo(a)anthracene	53.2	0	55.1	ug/L	104		1		41	147	20
Chrysene	53.2	0	54.7	ug/L	103		1		44	144	20
Benzo(b)fluoranthene	53.2	0	56.0	ug/L	105		3		40	150	20
Benzo(k)fluoranthene	53.2	0	55.1	ug/L	104		2		40	147	20
Benzo(a)pyrene	53.2	0	54.8	ug/L	103		1		42	147	20
Indeno(1,2,3-cd)pyrene	53.2	0	55.3	ug/L	104		4		30	166	20
Dibenz(a,h)anthracene	53.2	0	55.8	ug/L	105		5		23	172	20
Benzo(g,h,i)perylene	53.2	0	54.4	ug/L	102		6		27	167	20