

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2230 SAS No.: Q2230 SDG No.: Q2230

Instrument ID: MSVOA_X Calibration Date/Time: 06/10/2025 09:07

Lab File ID: VX046586.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:42 12:57

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.479	1.615		9.19	20
Toluene	0.908	0.992		9.25	20
Ethyl Benzene	2.053	2.179		6.14	20
m/p-Xylenes	0.745	0.802		7.65	20
o-Xylene	0.720	0.775		7.64	20
Isopropylbenzene	3.930	4.253		8.22	20
n-propylbenzene	4.757	5.071		6.6	20
1,3,5-Trimethylbenzene	3.275	3.526		7.66	20
tert-Butylbenzene	3.334	3.617		8.49	20
1,2,4-Trimethylbenzene	3.313	3.499		5.61	20
sec-Butylbenzene	4.284	4.548		6.16	20
p-Isopropyltoluene	3.572	3.774		5.66	20
n-Butylbenzene	3.507	3.673		4.73	20
1,2-Dichloroethane-d4	0.890	0.757		-14.94	20
Dibromofluoromethane	0.368	0.360		-2.17	20
Toluene-d8	1.212	1.152		-4.95	20
4-Bromofluorobenzene	0.502	0.468		-6.77	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.