

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2230 SAS No.: Q2230 SDG No.: Q2230

Instrument ID: MSVOA_X Calibration Date/Time: 06/11/2025 14:46

Lab File ID: VX046636.D Init. Calib. Date(s): 06/06/2025 06/06/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:42 12:57

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.479	1.534		3.72	20
Toluene	0.908	0.958		5.51	20
Ethyl Benzene	2.053	2.124		3.46	20
m/p-Xylenes	0.745	0.792		6.31	20
o-Xylene	0.720	0.786		9.17	20
Isopropylbenzene	3.930	4.163		5.93	20
n-propylbenzene	4.757	4.937		3.78	20
1,3,5-Trimethylbenzene	3.275	3.443		5.13	20
tert-Butylbenzene	3.334	3.554		6.6	20
1,2,4-Trimethylbenzene	3.313	3.476		4.92	20
sec-Butylbenzene	4.284	4.505		5.16	20
p-Isopropyltoluene	3.572	3.722		4.2	20
n-Butylbenzene	3.507	3.559		1.48	20
1,2-Dichloroethane-d4	0.890	0.830		-6.74	20
Dibromofluoromethane	0.368	0.382		3.8	20
Toluene-d8	1.212	1.219		0.58	20
4-Bromofluorobenzene	0.502	0.503		0.2	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.