



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061125\
Data File : VX046635.D
Acq On : 11 Jun 2025 14:01
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X060625W.M
Title : SW846 8260
Last Update : Fri Jun 06 16:56:12 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.0	6619	PASS
75	95	30	60	52.7	16610	PASS
95	95	100	100	100.0	31491	PASS
96	95	5	9	6.8	2149	PASS
173	174	0.00	2	1.6	335	PASS
174	95	50	100	66.1	20804	PASS
175	174	5	9	7.5	1565	PASS
176	174	95	101	96.3	20030	PASS
177	176	5	9	7.0	1411	PASS