

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060625

Review By	Mahesh Dadoda	Review On	6/9/2025 1:15:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/9/2025 1:19:58 PM
SubDirectory	VX060625	HP Acquire Method	HP Processing Method 82X060625W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134153
Initial Calibration Stds	VP134235,VP134236,VP134237,VP134238,VP134239,VP134240
CCC	VP134154
Internal Standard/PEM	
ICV/I.BLK	VP134241
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046516.D	06 Jun 2025 08:47		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046517.D	06 Jun 2025 09:13		JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046518.D	06 Jun 2025 09:42	for TCLP	JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX046519.D	06 Jun 2025 10:18	Comp #05 fail	JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046520.D	06 Jun 2025 10:40	LR-13,16,17	JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046521.D	06 Jun 2025 11:02		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046522.D	06 Jun 2025 11:25		JC/MD	Ok,M
8	IBLK	IBLK	VX046523.D	06 Jun 2025 11:47		JC/MD	Ok
9	VSTDIC001	VSTDIC001	VX046524.D	06 Jun 2025 12:57		JC/MD	Ok,M
10	VSTDICV050	ICV VX060625	VX046525.D	06 Jun 2025 14:11		JC/MD	Ok,M
11	VX0606MBL01	VX0606MBL01	VX046526.D	06 Jun 2025 14:39		JC/MD	Ok
12	VX0606WBL01	VX0606WBL01	VX046527.D	06 Jun 2025 15:02		JC/MD	Ok
13	VX0606WBS01	VX0606WBS01	VX046528.D	06 Jun 2025 15:25		JC/MD	Ok,M
14	VX0606MBS01	VX0606MBS01	VX046529.D	06 Jun 2025 15:51		JC/MD	Ok,M
15	Q2168-11MEDL	C2MEDL	VX046530.D	06 Jun 2025 16:13		JC/MD	Ok
16	VX0606WBSD01	VX0606WBSD01	VX046531.D	06 Jun 2025 16:36		JC/MD	Ok,M
17	Q2194-02	COMP-12	VX046532.D	06 Jun 2025 16:58	vial A pH#5.0	JC/MD	Ok
18	Q2194-04	COMP-13	VX046533.D	06 Jun 2025 17:21	vial A pH#5.0	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060625

Review By	Maresh Dadoda	Review On	6/9/2025 1:15:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/9/2025 1:19:58 PM
SubDirectory	VX060625	HP Acquire Method	HP Processing Method 82X060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134153 VP134235,VP134236,VP134237,VP134238,VP134239,VP134240		
CCC Internal Standard/PEM	VP134154		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134241		

19	Q2207-09	BU-703-COMP-01	VX046534.D	06 Jun 2025 17:43	vial A pH#5.0	JC/MD	Ok,M
20	Q2207-18	BU-703-COMP-02	VX046535.D	06 Jun 2025 18:06	vial A pH#5.0	JC/MD	Ok,M
21	Q2207-27	BU-703-COMP-03	VX046536.D	06 Jun 2025 18:28	vial A pH#5.0	JC/MD	Ok,M
22	Q2207-36	BU-703-COMP-04	VX046537.D	06 Jun 2025 18:51	vial A pH#5.0	JC/MD	Ok,M
23	Q2207-45	BU-703-COMP-05	VX046538.D	06 Jun 2025 19:13	vial A pH#5.0	JC/MD	Ok,M
24	Q2208-09	BU-703-COMP-06	VX046539.D	06 Jun 2025 19:35	vial A pH#5.0	JC/MD	Ok,M
25	Q2208-18	BU-703-COMP-07	VX046540.D	06 Jun 2025 19:58	vial A pH#5.0	JC/MD	Ok,M
26	Q2208-27	BU-703-COMP-08	VX046541.D	06 Jun 2025 20:20	vial A pH#5.0	JC/MD	Ok,M
27	Q2208-36	BU-703-COMP-09	VX046542.D	06 Jun 2025 20:42	vial A pH#5.0	JC/MD	Ok,M
28	Q2236-01	WC-A4-05A-G	VX046543.D	06 Jun 2025 21:04	vial A pH#5.0 Out of tune	JC/MD	Not Ok
29	VSTDCCC050	VSTDCCC050EC	VX046544.D	06 Jun 2025 21:26	Out of tune	JC/MD	Not Ok
30	BFB	BFB	VX046545.D	06 Jun 2025 23:59		JC/MD	Ok
31	VSTDCCC050	VSTDCCC050	VX046546.D	07 Jun 2025 00:34		JC/MD	Ok,M
32	VX0606WBL02	VX0606WBL02	VX046547.D	07 Jun 2025 01:17		JC/MD	Ok
33	VX0606WBS02	VX0606WBS02	VX046548.D	07 Jun 2025 02:00		JC/MD	Ok,M
34	VX0606WBSD02	VX0606WBSD02	VX046549.D	07 Jun 2025 02:22		JC/MD	Ok,M
35	PB168312TB	PB168312TB	VX046550.D	07 Jun 2025 02:43		JC/MD	Ok,M
36	PB168272TB	PB168272TB	VX046551.D	07 Jun 2025 03:05		JC/MD	Ok,M
37	Q2236-05	WC-A2-04-G	VX046552.D	07 Jun 2025 03:26	Surrogate Fail	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX060625

Review By	Maresh Dadoda	Review On	6/9/2025 1:15:21 PM
Supervise By	Semsettin Yesilyurt	Supervise On	6/9/2025 1:19:58 PM
SubDirectory	VX060625	HP Acquire Method	HP Processing Method 82X060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134153 VP134235,VP134236,VP134237,VP134238,VP134239,VP134240		
CCC Internal Standard/PEM	VP134154		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134241		

38	Q2236-09	WC-A2-05-G	VX046553.D	07 Jun 2025 03:48	Internal Standard Fail	JC/MD	ReRun
39	Q2236-13	WC-A2-06-G	VX046554.D	07 Jun 2025 04:09	vial A pH#5.0	JC/MD	Ok
40	Q2236-17	WC-A2-07-G	VX046555.D	07 Jun 2025 04:31	Internal Standard Fail; Surrogate fail	JC/MD	ReRun
41	Q2227-04	TP07-MHH-WC	VX046556.D	07 Jun 2025 04:52		JC/MD	Ok,M
42	Q2228-04	TP08-MHI-WC	VX046557.D	07 Jun 2025 05:14	Internal Standard Fail	JC/MD	ReRun
43	Q2235-01	WC-A2-08-G	VX046558.D	07 Jun 2025 05:36	Internal Standard Fail	JC/MD	ReRun
44	Q2240-04	TP-3	VX046559.D	07 Jun 2025 05:57	Internal Standard Fail	JC/MD	ReRun
45	Q2240-08	TP-2	VX046560.D	07 Jun 2025 06:18	Internal Standard Fail	JC/MD	ReRun
46	Q2240-12	TP-1	VX046561.D	07 Jun 2025 06:40	Internal Standard Fail	JC/MD	ReRun
47	Q2241-04	TP-N	VX046562.D	07 Jun 2025 07:01	Internal Standard Fail	JC/MD	ReRun
48	Q2241-08	TP-S	VX046563.D	07 Jun 2025 07:23	Internal Standard Fail	JC/MD	ReRun
49	Q2226-04	TP06-MHI-WC	VX046564.D	07 Jun 2025 07:44	Internal Standard Fail	JC/MD	ReRun
50	VSTDCCC050	VSTDCCC050EC	VX046565.D	07 Jun 2025 08:06		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061025

Review By	John Carlone	Review On	6/11/2025 10:17:47 AM
Supervise By	Maresh Dadoda	Supervise On	6/11/2025 11:14:33 AM
SubDirectory	VX061025	HP Acquire Method	HP Processing Method 82X060625W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134200
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134201,VP134202

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046585.D	10 Jun 2025 08:36		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046586.D	10 Jun 2025 09:07	pH#Lot#V12668	JC/MD	Ok,M
3	VX0610MBL01	VX0610MBL01	VX046587.D	10 Jun 2025 09:36		JC/MD	Ok
4	VX0610WBL01	VX0610WBL01	VX046588.D	10 Jun 2025 09:57		JC/MD	Ok
5	VX0610WBS01	VX0610WBS01	VX046589.D	10 Jun 2025 10:18		JC/MD	Ok,M
6	VX0610WBSD01	VX0610WBSD01	VX046590.D	10 Jun 2025 10:44		JC/MD	Ok,M
7	Q2262-02	ARS20-0032	VX046591.D	10 Jun 2025 11:05	vial B pH#5.0	JC/MD	Ok
8	Q2262-04	ARS20-0001	VX046592.D	10 Jun 2025 11:27	vial B pH#5.0	JC/MD	Ok
9	IBLK	IBLK	VX046593.D	10 Jun 2025 11:48		JC/MD	Ok
10	Q2230-01	FB-060425	VX046594.D	10 Jun 2025 12:09	vial A pH<2 FB	JC/MD	Ok
11	Q2230-06	TB060425	VX046595.D	10 Jun 2025 12:30	vial A pH<2 TB	JC/MD	Ok
12	Q2233-01DL	MW-18B-56-060425DL	VX046596.D	10 Jun 2025 12:51	vial B pH<2	JC/MD	Ok
13	Q2233-03DL	MW-18B-56-060425-FD	VX046597.D	10 Jun 2025 13:13	vial B pH<2	JC/MD	Ok
14	Q2233-04DL	MW-19B-72-060425DL	VX046598.D	10 Jun 2025 13:34	vial B pH<2	JC/MD	Ok
15	Q2234-01DL	MW-17B-55-060425DL	VX046599.D	10 Jun 2025 13:55	vial B pH<2	JC/MD	Ok
16	IBLK	IBLK	VX046600.D	10 Jun 2025 14:16		JC/MD	Ok
17	Q2230-05	GW-MW901-060425	VX046601.D	10 Jun 2025 14:38	vial A pH<2	JC/MD	Ok,M
18	Q2230-02	GW-MW01-060425	VX046602.D	10 Jun 2025 14:59	vial A pH<2	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061025

Review By	John Carlone	Review On	6/11/2025 10:17:47 AM
Supervise By	Maresh Dadoda	Supervise On	6/11/2025 11:14:33 AM
SubDirectory	VX061025	HP Acquire Method	HP Processing Method 82X060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134200		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134201,VP134202		

19	Q2230-03MS	GW-MW01-060425MS	VX046603.D	10 Jun 2025 15:20	Internal Standard Fail; Surrogate Fail; Spike not added	JC/MD	Not Ok
20	Q2230-04MSD	GW-MW01-060425MS	VX046604.D	10 Jun 2025 15:42	Internal Standard Fail; Surrogate Fail; Spike not added	JC/MD	Not Ok
21	IBLK	IBLK	VX046605.D	10 Jun 2025 16:03		JC/MD	Ok
22	Q2249-01	MW-06-6.5-060525	VX046606.D	10 Jun 2025 16:24	Run @ lower dilution	JC/MD	Not Ok
23	IBLK	IBLK	VX046607.D	10 Jun 2025 16:46		JC/MD	Ok
24	Q2233-01	MW-18B-56-060425	VX046608.D	10 Jun 2025 17:07	vial A pH<2 Need 10X	JC/MD	Dilution
25	Q2233-03	MW-18B-56-060425-FD	VX046609.D	10 Jun 2025 17:28	vial A pH<2 Need 10X	JC/MD	Dilution
26	Q2233-04	MW-19B-72-060425	VX046610.D	10 Jun 2025 17:50	vial A pH<2 Need 100X	JC/MD	Dilution
27	Q2234-01	MW-17B-55-060425	VX046611.D	10 Jun 2025 18:11	vial A pH<2 Need 100X	JC/MD	Dilution
28	VSTDCCC050	VSTDCCC050EC	VX046612.D	10 Jun 2025 18:53		JC/MD	Ok,M
29	BFB	BFB	VX046613.D	10 Jun 2025 19:57		JC/MD	Ok
30	VSTDICC005	VSTDICC005	VX046614.D	10 Jun 2025 20:19		JC/MD	Ok,M
31	VSTDICCC020	VSTDICCC020	VX046615.D	10 Jun 2025 20:40		JC/MD	Ok,M
32	VSTDICC050	VSTDICC050	VX046616.D	10 Jun 2025 21:01		JC/MD	Ok,M
33	VSTDICC100	VSTDICC100	VX046617.D	10 Jun 2025 21:22		JC/MD	Ok,M
34	VSTDICC150	VSTDICC150	VX046618.D	10 Jun 2025 21:44		JC/MD	Ok,M
35	IBLK	IBLK	VX046619.D	10 Jun 2025 22:05		JC/MD	Ok
36	VSTDICV020	ICVVX061025	VX046620.D	10 Jun 2025 22:26		JC/MD	Ok,M
37	VX0610WBS02	VX0610WBS02	VX046621.D	10 Jun 2025 23:09		JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061025

Review By	John Carlone	Review On	6/11/2025 10:17:47 AM
Supervise By	Maresh Dadoda	Supervise On	6/11/2025 11:14:33 AM
SubDirectory	VX061025	HP Acquire Method	HP Processing Method 82X060625W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP134200		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134201,VP134202		

38	VX0610WBSD02	VX0610WBSD02	VX046622.D	10 Jun 2025 23:30		JC/MD	Ok,M
39	VX0610WBL02	VX0610WBL02	VX046623.D	11 Jun 2025 00:12		JC/MD	Ok
40	Q2203-01	001-WILLETS-PT-BLVD	VX046624.D	11 Jun 2025 00:33	vial A pH<2 Need 5X	JC/MD	Dilution
41	Q2203-04	002-35TH-AVE(JUNE)	VX046625.D	11 Jun 2025 00:54	vial A pH<2 Need 5X	JC/MD	Dilution
42	VSTDCCC020	VSTDCCC020EC	VX046626.D	11 Jun 2025 01:16		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061125

Review By	John Carlone	Review On	6/12/2025 10:06:05 AM
Supervise By	Maresh Dadoda	Supervise On	6/12/2025 3:53:50 PM
SubDirectory	VX061125	HP Acquire Method	HP Processing Method 624X061025W.M

STD. NAME	STD REF.#
Tune/Reschk	VP134259,VP134275
Initial Calibration Stds	VP134316,VP134317,VP134318,VP134319,VP134320
CCC	VP134260,VP134261,VP134276,VP134277
Internal Standard/PEM	
ICV/I.BLK	VP134321
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046627.D	11 Jun 2025 10:45		JC/MD	Ok
2	VSTDCCC020	VSTDCCC020	VX046628.D	11 Jun 2025 11:12	pH#Lot#V12668	JC/MD	Ok,M
3	VX0611WBS01	VX0611WBS01	VX046629.D	11 Jun 2025 11:43		JC/MD	Ok,M
4	VX0611WBSD01	VX0611WBSD01	VX046630.D	11 Jun 2025 12:09		JC/MD	Ok,M
5	VX0611WBL01	VX0611WBL01	VX046631.D	11 Jun 2025 12:31		JC/MD	Ok
6	Q2203-01DL	001-WILLETTS-PT-BLVD	VX046632.D	11 Jun 2025 12:57	vial B pH<2	JC/MD	Ok,M
7	Q2203-04DL	002-35TH-AVE(JUNE)	VX046633.D	11 Jun 2025 13:18	vial B pH<2	JC/MD	Ok,M
8	Q2264-01	EFF-WW	VX046634.D	11 Jun 2025 13:40	vial A pH<2 foamy sample	JC/MD	Ok,M
9	BFB	BFB	VX046635.D	11 Jun 2025 14:01		JC/MD	Ok
10	VSTDCCC050	VSTDCCC050	VX046636.D	11 Jun 2025 14:46		JC/MD	Ok,M
11	VX0611WBL02	VX0611WBL02	VX046637.D	11 Jun 2025 15:14		JC/MD	Ok
12	VX0611MBL01	VX0611MBL01	VX046638.D	11 Jun 2025 15:36		JC/MD	Ok
13	VX0611WBS02	VX0611WBS02	VX046639.D	11 Jun 2025 15:58		JC/MD	Ok,M
14	VX0611WBSD02	VX0611WBSD02	VX046640.D	11 Jun 2025 16:25		JC/MD	Ok,M
15	Q2230-03MS	GW-MW01-060425MS	VX046641.D	11 Jun 2025 16:47		JC/MD	Ok,M
16	Q2230-04MSD	GW-MW01-060425MSI	VX046642.D	11 Jun 2025 17:09		JC/MD	Ok,M
17	IBLK	IBLK	VX046643.D	11 Jun 2025 17:31		JC/MD	Ok
18	Q2249-01	MW-06-6.5-060525	VX046644.D	11 Jun 2025 17:53	Internal Standard Fail	JC/MD	ReRun

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX061125

Review By	John Carlone	Review On	6/12/2025 10:06:05 AM
Supervise By	Maresh Dadoda	Supervise On	6/12/2025 3:53:50 PM
SubDirectory	VX061125	HP Acquire Method	HP Processing Method 624X061025W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP134259,VP134275		
Initial Calibration Stds	VP134316,VP134317,VP134318,VP134319,VP134320		
CCC	VP134260,VP134261,VP134276,VP134277		
Internal Standard/PEM			
ICV/I.BLK	VP134321		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2271-04	TP12-MHK-WC	VX046645.D	11 Jun 2025 18:15	Internal Standard Fail	JC/MD	ReRun
20	Q2272-04	TP-6	VX046646.D	11 Jun 2025 18:37	Internal Standard Fail ;Surrogate fail	JC/MD	ReRun
21	Q2273-04	WC-4	VX046647.D	11 Jun 2025 18:59	Internal Standard Fail	JC/MD	ReRun
22	Q2273-08	WC-6	VX046648.D	11 Jun 2025 19:21	Internal Standard Fail ;Surrogate fail	JC/MD	ReRun
23	Q2274-04	TP-13-MHP-WC	VX046649.D	11 Jun 2025 19:43	Internal Standard Fail; Surrogate fail	JC/MD	ReRun
24	Q2278-04	TP-2	VX046650.D	11 Jun 2025 20:04	Surrogate Fail	JC/MD	ReRun
25	Q2280-02	VNJ-210	VX046651.D	11 Jun 2025 20:26	Internal Standard Fail	JC/MD	ReRun
26	Q2280-04	RT-4643	VX046652.D	11 Jun 2025 20:48	Internal Standard Fail	JC/MD	ReRun
27	Q2285-05	HAM-CONCRETE	VX046653.D	11 Jun 2025 21:10	Internal Standard Fail	JC/MD	ReRun
28	PB168370TB	PB168370TB	VX046654.D	11 Jun 2025 21:32		JC/MD	Ok,M
29	VSTDCCC050	VSTDCCC050EC	VX046655.D	11 Jun 2025 21:53		JC/MD	Ok,M

M : Manual Integration