

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2230

Client: CDM Smith

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec			RPD		Limits		RPD
		Result			Rec	Qual	RPD	Qual	Low	High		
Lab Sample ID :	Q2230-03MS	Client Sample ID :	GW-MW01-060425MS					Datafile :	VX046641.D			
Benzene	50	0.36	58.1	ug/L	115				81	128		
Toluene	50	1.60	62.2	ug/L	121	*			85	115		
Ethyl Benzene	50	3.80	65.8	ug/L	124				81	128		
m/p-Xylenes	100	16.1	150	ug/L	134	*			69	129		
o-Xylene	50	20.0	94.6	ug/L	149	*			75	127		
Isopropylbenzene	50	1.30	62.9	ug/L	123	*			76	121		
N-propylbenzene	50	2.50	63.6	ug/L	122	*			79	116		
1,3,5-Trimethylbenzene	50	7.40	72.5	ug/L	130				40	169		
tert-Butylbenzene	50	0	61.8	ug/L	124	*			80	116		
1,2,4-Trimethylbenzene	50	26.0	100	ug/L	148	*			81	116		
Sec-butylbenzene	50	1.00	60.8	ug/L	120				70	122		
p-Isopropyltoluene	50	0	60.8	ug/L	122				64	124		
n-Butylbenzene	50	1.40	60.9	ug/L	119				53	140		