

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2230

Client: CDM Smith

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec		RPD	Limits			RPD
		Result			Qual	RPD		Qual	Low	High	
Lab Sample ID :	Q2230-04MSD	Client Sample ID :	GW-MW01-060425MSD					Datafile :	VX046642.D		
Benzene	50	0.36	50.7	ug/L	101		13		81	128	20
Toluene	50	1.60	54.6	ug/L	106		13		85	115	20
Ethyl Benzene	50	3.80	58.5	ug/L	109		13		81	128	20
m/p-Xylenes	100	16.1	140	ug/L	124		8		69	129	20
o-Xylene	50	20.0	95.5	ug/L	151	*	1		75	127	20
Isopropylbenzene	50	1.30	53.5	ug/L	104		17		76	121	20
N-propylbenzene	50	2.50	55.2	ug/L	105		15		79	116	20
1,3,5-Trimethylbenzene	50	7.40	66.8	ug/L	119		9		40	169	20
tert-Butylbenzene	50	0	52.1	ug/L	104		17		80	116	20
1,2,4-Trimethylbenzene	50	26.0	110	ug/L	168	*	13		81	116	20
Sec-butylbenzene	50	1.00	52.0	ug/L	102		16		70	122	20
p-Isopropyltoluene	50	0	51.8	ug/L	104		16		64	124	20
n-Butylbenzene	50	1.40	53.5	ug/L	104		13		53	140	20