

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2230

Client: CDM Smith

Analytical Method: SW8260-Low

Datafile : VX046589.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0610WBS01	Benzene	20	21.0	ug/L	105			82	109	
	Toluene	20	21.2	ug/L	106			82	110	
	Ethyl Benzene	20	20.3	ug/L	102			83	109	
	m/p-Xylenes	40	41.6	ug/L	104			82	110	
	o-Xylene	20	21.1	ug/L	106			83	109	
	Isopropylbenzene	20	20.2	ug/L	101			83	112	
	N-propylbenzene	20	20.0	ug/L	100			83	112	
	1,3,5-Trimethylbenzene	20	20.5	ug/L	103			85	112	
	tert-Butylbenzene	20	20.4	ug/L	102			83	112	
	1,2,4-Trimethylbenzene	20	20.8	ug/L	104			85	111	
	Sec-butylbenzene	20	20.5	ug/L	103			81	114	
	p-Isopropyltoluene	20	20.2	ug/L	101			78	116	
	n-Butylbenzene	20	19.9	ug/L	100			75	115	