

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2230 SAS No.: Q2230 SDG NO.: Q2230
 Lab File ID: VX046586.D Date Analyzed: 06/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:07
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	90758	5.56	150623	6.76	134767	10.06
UPPER LIMIT	181516	6.056	301246	7.263	269534	10.555
LOWER LIMIT	45379	5.056	75311.5	6.263	67383.5	9.555
EPA SAMPLE NO.						
FB-060425	109731	5.56	215111	6.77	220551	10.06
GW-MW01-060425	112673	5.56	225413	6.77	232357	10.06
GW-MW901-060425	112231	5.56	225068	6.77	232456	10.06
TB060425	114604	5.57	224360	6.77	234275	10.06
VX0610WBL01	106110	5.56	211553	6.77	217338	10.06
VX0610WBS01	84052	5.56	143506	6.77	131426	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.