

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2230 SAS No.: Q2230 SDG NO.: Q2230
 Lab File ID: VX046636.D Date Analyzed: 06/11/2025
 Instrument ID: MSVOA_X Time Analyzed: 14:46
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	77468	5.56	133266	6.76	119186	10.06
UPPER LIMIT	154936	6.062	266532	7.263	238372	10.555
LOWER LIMIT	38734	5.062	66633	6.263	59593	9.555
EPA SAMPLE NO.						
GW-MW01-060425MS	80829	5.57	141022	6.77	125449	10.06
GW-MW01-060425MSD	63422	5.57	110741	6.78	101042	10.06
VX0611WBL02	73470	5.56	147517	6.77	153084	10.06
VX0611WBS02	79398	5.57	136732	6.78	125565	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.